

# CRM Product Performance Report 2014

Q1 Edition



CRM Quality Pledge

I improve  
the quality  
of patient care  
and all things  
Boston Scientific

## Advancing Science for Life.

For almost forty years, meaningful innovation at Boston Scientific Rhythm Management has helped patients live healthier, longer lives. We are committed to providing performance data which are accurate, transparent and responsive to topics of contemporary clinical interest. This Q1 2014 report includes data through January 17, 2014.

This report meets or exceeds the AdvaMed Industry Guidance for Uniform Reporting of Clinical Performance, and address recommendations from the Heart Rhythm Society Task Force on Lead Performance. With increased interest in lead performance, our **Product Performance Report** provides the most comprehensive presentation of lead performance data available, including:

- ✓ U.S. Lead survival probability
- ✓ Worldwide malfunction counts and patterns
- ✓ Longitudinal Surveillance Registry lead survival probability
- ✓ Acute (first month) lead observations
- ✓ Chronic (after first month) lead complications
- ✓ Malfunctions reported before and during an implant procedure

Your feedback is always welcome, and plays a vital role in our effort to continuously improve our products and services, advancing science to transform the lives of our patients.

Sincerely,

Renold J. Russie  
Vice President, Quality Assurance

## **Medical Review Board**

**Ronald D. Berger, M.D., PhD**  
Professor of Medicine  
Johns Hopkins University

**Stephen R. Shorofsky, M.D., PhD**  
Professor of Medicine  
University of Maryland, School of Medicine

## **Boston Scientific Reviewers**

**Renold Russie**  
Vice President, Quality Assurance

**Anjun Sharma, M.D.**  
Vice President of Medical Safety

**Mark Gayle**  
Global Vice President, Post-Market Quality Assurance

**Jack Litzau, M.S.**  
Principal Statistician

# Statistical Methodology

## What Is Device Survival Probability?

Medical journals have traditionally used patient survival probability to display information on treatment option effectiveness. In the report, **pulse generator and lead** survival probabilities convey information about long-term performance of implantable cardiac rhythm management products.

Survival probability shows the percentage of implanted devices that remain implanted and in service at various points in a product's service life, in the absence of competing risks, such as natural mortality or voluntary explants. Conceptually, a pulse generator of high reliability and large battery capacity or low current drain remains near 100% survival until eventually, normal battery depletion begins to cause significant numbers of devices to be removed, and the device survival probability drops rapidly. For example, a device survival probability of 99% indicates that within the stated implant duration, the pulse generator had a 1% risk of removal for battery depletion or for incurring a malfunction that required replacement. Survival probabilities are provided with and without normal battery depletions, depicted as "Battery Depletions and Malfunctions" and "Malfunctions Only," respectively.

Boston Scientific estimates survival probability in compliance with international standard ISO 5841-2: 2000 (E). Survival probability is calculated at a given time by separately estimating the probability of surviving each interval and multiplying the survival probabilities of all intervals through which a device has passed. To estimate the probability of surviving any interval, the number of units that successfully functioned during the interval is divided by the number of units exposed to malfunction/depletion during the interval. The number of units exposed is calculated using the actuarial method, where device suspensions in an interval are distributed uniformly across the interval. Reasons for device suspension from survival probability statistics are detailed in the report section entitled U.S. Reason for Out of Service.

Survival probabilities are statistical estimates subject to uncertainty. To quantify this uncertainty, 95% confidence intervals of survival probabilities are computed. Greenwood's formula is used to estimate the standard error of the calculated survival probabilities, and confidence intervals are constructed using a logit transformation, assuming the transformed values are normally distributed. For example, 99.36% (-0.3/+0.2) represents an interval of 99.06% to 99.56%, and these intervals are constructed such that 95% of the time they will contain the true survival probability. These confidence intervals are not symmetric due to the transformation method described previously.

## Inclusion Criteria for Pulse Generator and Lead Survival Probability Datasets

Pulse generator survival probability is reported for U.S. implanted devices in product families which meet inclusion criteria described below. Lead survival probability is reported for both the U.S. Registered Implant population and for leads enrolled in the Longitude Surveillance Registry, for product families which meet inclusion criteria described below.

To be included in survival probability statistics, a device must first be successfully implanted (defined in this report as occurring upon pocket closure). Prophylactic device removals are tracked as part of the active population up until the time the device is removed from service; devices removed prophylactically which are not identified as malfunctions at the time of explant do not contribute to a reduction in survival probability. Reasons for device explant or out of service, if known, are provided in this report for each pulse generator product/product grouping.

Survival probabilities are based on devices registered as implanted in the United States. Privacy laws in many other geographies preclude manufacturers from obtaining specific patient implant and explant information, thus device survival probabilities cannot be constructed from these data. Boston Scientific believes, however, that U.S. experience is generally representative of worldwide performance. The Malfunction Details for leads and pulse generators reflect worldwide malfunctions, inclusive of U.S. data.

Criteria for inclusion of product families in this report are in compliance with the *AdvaMed Industry Guidance for Uniform Reporting of Clinical Performance of Pulse Generators and Leads*. Survival estimates are provided for product families once they have at least 10,000 cumulative U.S. implant months. The minimum interval sample size is 200 U.S. implanted units. Pulse generator product families with less than 200 total remaining estimated active U.S. devices are not included in this report. Lead product families that received original U.S. market release approval twenty or more years ago are not included in this report. Survival estimates for leads enrolled in the Longitude Surveillance Registry are provided for product families once they have at least 200 enrolled leads. The minimum interval sample size is 50 leads which have been followed for at least 6 months.

Estimated Longevity information is provided for pulse generator products in the U.S. Survival Probability - Battery Depletions and Malfunctions graphs, depicted as a blue bar on the x axis for Years Implanted. The estimated longevity values from the Instructions for Use for each product family are used to construct the blue longevity bars on their U.S. Survival Probability graph. They represent the range of estimated longevity based on a variety of programmed settings and therapy usage.

Survival probability data are presented in tabular and graphical formats online at [www.bostonscientific.com/ppr](http://www.bostonscientific.com/ppr). Performance data for Intermedics products may also be found on [www.bostonscientific.com/ppr](http://www.bostonscientific.com/ppr). Specific inclusion criteria for pulse generator and lead survival probability datasets are described here. Not all products may be approved for use in all geographies, as product approval is geography specific.

Worldwide distribution, U.S. registered implant and U.S. estimated active implant numbers have been rounded to provide population size context.

To convey implant experience for a product family, average device age and U.S. approval date are provided. The U.S. approval date listed is the earliest date Boston Scientific received approval for one or more of the models in the family. For Longitude Surveillance Registry data, the number of enrolled leads and their cumulative followup months are also provided for context.

### **Survival Probability – Battery Depletions and Malfunctions (Pulse Generators)**

Reduction in survival probability for **pulse generators** is due to:

- Devices removed for normal battery depletion
- Device malfunctions occurring while implanted, as confirmed by returned product analysis
- Devices removed from service and reported to have exhibited premature battery depletion, but not confirmed by laboratory analysis, whether returned or not—also known as “unconfirmed reports of premature battery depletions.”

### **Survival Probability – Malfunctions Only (Pulse Generators)**

Reduction in survival probability for **pulse generators** is due only to:

- Device malfunctions occurring while implanted, as confirmed by returned product analysis; premature battery depletions are considered device malfunctions.

In this case, normal battery depletions do not contribute to the reduction in survival probability; rather, reduction in survival probability is due only to confirmed pulse generator malfunctions. Furthermore, unconfirmed reports of premature battery depletions do not reduce “Malfunctions Only” survival probability. Put another way, this information depicts the percentage of confirmed malfunction-free devices remaining in service at various intervals in the product's service life, based on returned product analysis.

### **Survival Probability — Complications and Malfunctions (Leads)**

The AdvaMed *Industry Guidance for Uniform Reporting of Clinical Performance of Cardiac Rhythm Management Pulse Generators and Leads*, published in May 2009, outlines a methodology for lead survival probability inclusion. Boston Scientific has applied this methodology to all lead families being implanted as of May 2009, and will apply it to all future lead families as they are included in the Product Performance Report. Worldwide malfunctions are not included for older lead families. .

Reduction in survival probability is due to:

- Leads and lead segments returned for analysis and determined to be non-compliant in form, fit, or function at any time while implanted
- Leads and lead segments returned for analysis with reported observations 30 days or more post-implant, but for which analysis was inconclusive or a reported complication was unconfirmed
- Leads removed from service but not returned for laboratory analysis, with reported complications 30 days or more post-implant

### **Further Adjustments for Device and Lead Survival**

Because underreporting of patient deaths unrelated to device function would result in overestimation of pulse generator or lead survival by overstating the number of devices in service, Boston Scientific addresses this underreporting in two ways . First, regular updates are obtained from the Social Security Administration about deceased persons and compared to Boston Scientific patient data to learn about patients who have died but whose deaths had not been reported to Boston Scientific. Second, Boston Scientific uses 10% annual patient mortality as a baseline and adjusts reported patient deaths in any interval for which reports are less than the baseline rate. No adjustment is applied to account for underreporting of malfunctions, as the rate of underreporting is unknown.

Boston Scientific does not make statistical adjustments to account for underreporting of battery depletion. However, as mentioned earlier, Boston Scientific includes non-returned devices removed from service for battery depletion with no associated complaint as normal battery depletions.

### **Categorization of Malfunctions for Survival Probability Reporting**

Malfunctions represent pulse generators and leads removed from service and confirmed through laboratory analysis to have operated outside the specified performance limits established by Boston Scientific while implanted and in service. In addition, leads utilize AdvaMed methodology which includes Extrinsic Factor malfunctions occurring 30 days or more post-implant, where laboratory analysis is inconclusive or unconfirmed. Device damage occurring during or after explant, or caused by external factors including those warned against in product labeling (such as ionizing therapeutic radiation), are not reported as device malfunctions in survival data. Damage to a pulse generator caused by a lead malfunction is reported as a lead malfunction. Malfunctions are further classified according to their impact on therapy, as follows:

#### **Malfunction With Compromised Therapy —**

The condition when a device is confirmed through laboratory analysis to have malfunctioned in a manner that compromised pacing or defibrillation therapy (including complete loss or partial degradation) while implanted and in service.

Examples include (but are not limited to): sudden loss of battery voltage; accelerated current drain such that low battery was not detected before loss of therapy; sudden malfunction during defibrillation therapy resulting in aborted therapy delivery; intermittent malfunction in which therapy is compromised while in the malfunction state.

#### **Malfunction Without Compromised Therapy —**

The condition when a device is confirmed through laboratory analysis to have malfunctioned in a manner that did not compromise pacing or defibrillation therapy while implanted and in service. Malfunctions in which critical patient-protective pacing and defibrillation therapies remain available are included here.

Examples include (but are not limited to): error affecting diagnostic functions, telemetry function, data storage; malfunction of a component that causes the battery to lose power quickly enough to result in premature battery depletion, but slowly enough that the condition is detected through normal follow-up before therapy is lost; mechanical problems with connector header that do not affect therapy.

### **Categorization of Normal Battery Depletion for Survival Probability Reporting**

Per the AdvaMed *Industry Guidance for Uniform Reporting of Clinical Performance of Pulse Generators and Leads*, **Normal Battery Depletion** is defined as the condition when

- a) A device is returned with no associated complaint and the device has reached its elective replacement indicator(s) with implant time that meets or exceeds the nominal (50 percentile) predicted longevity at default (labeled) settings, or
- b) A device is returned and the device has reached its elective replacement indicator(s) with implant time exceeding 75% of the expected longevity using actual device settings and therapeutic use.

Boston Scientific includes within this count both returned *and non-returned* devices removed from service for battery depletion with no associated complaint. In conformance with the AdvaMed guidance document, Boston Scientific performs battery usage analysis, including battery status verification, on all devices returned without a complaint. We continue to include non-returned devices reported by our customers as being removed from service due to normal battery depletion within this count.

### **Boston Scientific CRM's Corrective and Preventive Actions (CAPA) System**

Boston Scientific strives to provide implantable devices of high quality and reliability. However, these devices are not perfect and may exhibit malfunctions at a low rate of occurrence. Device performance information is received from many sources through various channels. Boston Scientific monitors information from many sources including suppliers, testing, manufacturing and field performance to identify opportunities for improvement.

When a device is returned to Boston Scientific, laboratory technicians and engineers assess overall device function and perform analysis using specific tests related to the clinical observation(s). Test results are compared to original manufacturing records and design intent. Clinical observations are added to laboratory findings to help determine root cause of the clinical observation(s). Each discrete event is then compared to other similar-appearing events. If a pattern is detected, actions are taken to identify a common root cause, and improvements intended to improve product reliability and/or performance may be implemented. Observations from supplier data and internal manufacturing operations also lead to opportunities for improvement. Improvements, when made, may include design changes in existing or subsequent generations, manufacturing and supplier process modifications, software updates, educational communications, and/or labeling changes as examples. Improvement implementation may vary by geography due to various factors including regulatory review timing. They may not be applied to every product susceptible to the malfunction pattern, and may not mitigate or eliminate the potential for additional malfunctions. In cases where an improvement is made to an approved product line, devices made without the improvement may continue to be distributed where such products meet our high reliability and performance standards, particularly when changes are incremental and in accordance with our overall philosophy of continuous product improvement.

Improvements are closely monitored for effectiveness. Boston Scientific informs regulatory bodies of each significant event that poses potential risk to patient health to meet regulatory obligations, and shares returned product investigation findings with physicians. The malfunction details section for pulse generators and leads includes a summary of these findings.

In summary, thorough investigation of internal and external data coupled with low trigger levels for improvements creates a continuous product improvement system that is very responsive to patient and physician needs. Boston Scientific is committed to characterizing and presenting to our customers an accurate picture of product performance and addressing identified issues in a timely fashion.



# Malfunction Details: Overview

Boston Scientific CRM pursues product quality and reliability with a passion. We therefore continuously monitor product performance to make improvements whenever possible. Worldwide Malfunction tables provide a count and description of malfunctions associated with the majority of actively in-service Boston Scientific products. Intermedics co-branded product data are included in corresponding pacemaker and pacing lead malfunction counts and details. Information presented is based on malfunctions reported to and analyzed by Boston Scientific. Each table contains malfunction counts listed by category, pattern and therapy availability.

## Category

Malfunctions are categorized by the nature of their root cause. For example, a malfunction due to the software within a pulse generator is listed in the Software category. There are four pulse generator malfunction categories and five malfunction categories for leads (described below).

## Patterns

Patients and physicians have asked for more access to Quality System details; therefore, we provide information on patterns of product performance. Patterns listed are informational and do not represent actions that need to be taken. Boston Scientific is committed to direct communication when predicted product performance fails to achieve design or performance expectations or when actions may be taken to improve patient outcomes. Malfunctions associated with product advisories are denoted. Refer to the Product Advisories section for more information.

Each pattern description includes:

- **Clinical Manifestation and Root Cause** – Malfunctions for each product are characterized according to root cause. Descriptions provide clinical observations and/or analysis findings associated with each malfunction pattern listed in this report. Malfunctions listed within “Other” either do not yet have an identified root cause, or are related to a proprietary product feature, such as connectors or seal rings.
- **Improvement Implementation** – All of the patterns listed are thoroughly investigated and analyzed. As part of Boston Scientific's process of continuous improvement, when possible, improvements have been or will be implemented in response to identified malfunction patterns. Improvements may include product design changes in existing or subsequent generations, manufacturing process modifications, software updates, educational communications or labeling changes. Improvement implementation may vary by geography due to various factors, including regulatory review timing. They may not be applied to every product susceptible to the malfunction pattern, and may not completely mitigate or eliminate the potential for additional malfunctions.

Pattern information in this report is dynamic. Pattern names, superscript number assignments and descriptions may all change from quarter to quarter; as Boston Scientific's investigations progress and improvements are implemented, updated information is provided.

## Therapy Availability

Malfunctions are further classified according to their impact on therapy, as follows:

- **Malfunction With Compromised Therapy** – The condition when a device is confirmed through laboratory analysis to have malfunctioned in a manner that compromised pacing or defibrillation therapy (including complete loss or partial degradation) while implanted and in service. Examples include (but are not limited to): sudden loss of battery voltage; accelerated current drain such that low battery was not detected before loss of therapy; sudden malfunction during defibrillation therapy resulting in aborted therapy delivery; intermittent malfunction in which therapy is compromised while in the malfunction state.
- **Malfunction Without Compromised Therapy** – The condition when a device is confirmed through laboratory analysis to have malfunctioned in a manner that did not compromise pacing or defibrillation therapy while implanted and in service. Malfunctions in which critical patient-protective pacing and defibrillation therapies remain available are included here. Examples include (but are not limited to): error affecting diagnostic functions, telemetry function, data storage; malfunction of a component that causes the battery to lose power quickly enough to result in premature battery depletion, but slowly enough that the condition is detected through normal follow-up before therapy is lost; mechanical problems with connector header that do not affect therapy.

For lead malfunctions listed in the Extrinsic Factors category, therapy availability may be known, not reported or unable to be determined. When known, these malfunctions are reported in the appropriate therapy availability column. When unknown, because the lead was taken out of service and returned, it is assumed that therapy may have been compromised, and will be reported in the With Compromised Therapy column.

## Pulse Generator Confirmed Malfunctions

Pulse generator confirmed malfunctions represent devices removed from service and confirmed through laboratory analysis to have operated outside the performance limits established by Boston Scientific while implanted and in service. Device damage occurring during or after explant, or caused by external factors including those warned against in product labeling (e.g. therapeutic radiation), are not considered device malfunctions. Damage to a pulse generator caused by a lead malfunction is reported as a lead malfunction.

## Lead Malfunctions

The Boston Scientific Product Performance Report is in compliance with the May 2009 AdvaMed Industry Guidance for Uniform Reporting of Clinical Performance of Cardiac Rhythm Management Pulse Generators and Leads. This methodology also addresses the Recommendations from the Heart Rhythm Society Task Force on Lead Performance Policies and Guidelines.

## Malfunction Categories for Leads

Lead malfunction categories include Conductor, Insulation, Crimps/Welds/Bonds, Other and Extrinsic Factors, and include the following:

- **Conductor:** Conductor break with complete or intermittent loss of continuity that could interrupt current flow (e.g. fractured conductors), including those associated with clavicle flex fatigue or crush damage.
- **Insulation:** Any lead insulation breach. Examples include: 1) proximal abrasions associated with lead-on-lead or lead-on-PG contact in the pocket, 2) mid-lead insulation damage caused by clavicle flex fatigue or crush, suture or suture sleeve, insulation wear in the region of vein insertion, and 3) distal region wear due to lead-on-lead (intracardiac), lead-on-heart valve or lead-on-other anatomy contact.
- **Crimps/Welds/Bonds:** Any interruption in the conductor or lead body associated with a point of connection. Typically demonstrated by high or low shocking/pacing impedance, undersensing or oversensing.
- **Other:** Includes specific proprietary lead mechanical attributes, such as lead-incorporated sensors, connectors, seal rings or the 4–Site connector, or any malfunction modes not included in the three categories above.
- **Extrinsic Factors:** Lead complication where the identified lead was removed from service and returned for analysis, where analysis was either inconclusive or the complication was not confirmed. Inconclusive includes leads where only portions of the lead were available for return, or the returned lead was damaged by the explantation process. Unconfirmed includes when lab analysis could not determine an out of specification condition (including complaints of malfunction or complications such as dislodgements, perforations or failure to capture).

The categories of Conductor, Insulation, Crimps/Welds/Bonds and Other represent malfunctions for leads removed from service and confirmed through laboratory analysis to have operated outside the performance limits established by Boston Scientific while implanted and in service. The Extrinsic Factors category represents leads with reported complications for which the leads were removed from service and returned, but for which laboratory analysis was inconclusive or the complication was unconfirmed. For the Extrinsic Factors category only, malfunctions are included for leads implanted greater than 30 days.

## Supporting Greater Return of Explanted Devices

The Heart Rhythm Society (HRS) Task Force on Device Performance Policies and Guidelines stated that knowledge, confidence, and trust in cardiac rhythm management devices can be strengthened through enhancing systems that increase the return of devices to the manufacturer. Boston Scientific CRM shares in this belief and supports the HRS-specified actions geared toward achieving the goal of greater device return to the manufacturer, including post-mortem device interrogation, explantation and return to the manufacturer.<sup>1</sup>

In this section of the report, Boston Scientific provides device return rate data with the goal of raising awareness and improving current device explant and return rates. Figure 1 on the following page depicts the percentage of devices reported to have been explanted and then returned to Boston Scientific for various product therapy types.

### Help Us Provide You With More Complete Product Performance Data

#### ***Reporting Adverse Events***

The data in this report reflect Boston Scientific's understanding of product performance. We acknowledge that there is underreporting. If you have product performance observations to report, please contact your local Boston Scientific sales representative or Boston Scientific's Technical Services department at:

United States: Phone 1.800.CARDIAC (1.800.227.3422) or 1.651.582.2698.

International: Please refer to the Country Offices List for local contact information.


E-mail: [crmevent@bsci.com](mailto:crmevent@bsci.com)

#### ***Returning Products to Boston Scientific***

Boston Scientific provides a Returned Products Kit (Model 6499) that includes proper forms, shipping/packaging (biohazard bags), and a prepaid shipping label. It can be ordered at no charge through Boston Scientific's Customer Service department at 1.800.CARDIAC (1.800.227.3422) or 1.651.582.2698, or you can order a Returned Products Kit online at [www.bostonscientific.com/ppr](http://www.bostonscientific.com/ppr).



<sup>1</sup>Carlson et al. Recommendations from the Heart Rhythm Society Task Force. Heart Rhythm. October 2006; 3(special issue):1251 — 1252.



	2004	2005	2006	2007	2008	2009	2010	2011	2012	2013 (through 9/30)
Explants	1328	4512	4382	4688	5288	8146	9001	7464	5884	2864
Returns	600	3088	2430	2951	3255	4811	5666	4750	3700	1827
<b>% Returned</b>	<b>45%</b>	<b>68%</b>	<b>55%</b>	<b>63%</b>	<b>62%</b>	<b>59%</b>	<b>63%</b>	<b>64%</b>	<b>63%</b>	<b>64%</b>
Explants	15299	16487	10213	11532	15731	20196	20798	17615	13216	6815
Returns	5144	8687	5253	6945	9487	11970	13479	11651	8619	4495
<b>% Returned</b>	<b>34%</b>	<b>53%</b>	<b>51%</b>	<b>60%</b>	<b>60%</b>	<b>59%</b>	<b>65%</b>	<b>66%</b>	<b>65%</b>	<b>66%</b>
Explants	14477	21674	17746	19082	20921	21545	21469	20411	19010	9331
Returns	4115	10192	7848	9101	10346	10907	11643	11763	10780	5306
<b>% Returned</b>	<b>28%</b>	<b>47%</b>	<b>44%</b>	<b>48%</b>	<b>49%</b>	<b>51%</b>	<b>54%</b>	<b>58%</b>	<b>57%</b>	<b>57%</b>

**Figure 1.** Percentage of U.S. explanted devices as reported and returned to Boston Scientific CRM.

**INCEPTA CRT-D 4-Site**


Models N160/N162/P162

U.S. Survival Probability	Worldwide Malfunction Details	Product Advisories
---------------------------	-------------------------------	--------------------


**U.S. Summary**

U.S. Registered Implants: 5,000	U.S. Normal Battery Depletions: 1
U.S. Approval Date: November 2011	U.S. Unconfirmed Reports of Premature Battery Depletion : 0
U.S. Estimated Active Implants: 5,000	U.S. Malfunctions:0
	Without Compromised Therapy:0
	With Compromised Therapy:0

**Battery Depletions and Malfunctions**



**Malfunctions Only**



**U.S. Survival Probability**

	Year	1	2	3	4	5	6	7	8	9	10
Non Advisory Population Registered Implants: 5000	Depletions and Malfunctions (%) (Confidence Interval)	99.98 (-0.1/+0.0)	99.98 @ 23 mo. (-0.1/+0.0)	-	-	-	-	-	-	-	-
	Malfunctions Only (%) (Confidence Interval)	100.00 (-0.0/+0.0)	100.00 @ 23 mo. (-0.0/+0.0)	-	-	-	-	-	-	-	-
	Effective Sample Size	2270	211	-	-	-	-	-	-	-	-

**INCEPTA CRT-D 4-Site**

Models N160/N162/P162

<b>U.S. Survival Probability</b>	<b>Worldwide Malfunction Details</b>	<b>Product Advisories</b>
----------------------------------	--------------------------------------	---------------------------

**INCEPTA CRT-D 4-Site**  
**Models N160/N162/P162**

**Worldwide Distribution:** 11,000  
**Worldwide Confirmed Malfunctions:** 1

	Without Compromised Therapy	With Compromised Therapy	Total
<b>Electrical</b>	-	-	<b>0</b>
<b>Mechanical</b>	-	<b>1</b>	<b>1</b>
<sup>103</sup> Transformer	-	1	
<b>Software</b>	-	-	<b>0</b>
<b>Other</b>	-	-	<b>0</b>
Non-patterned	-	-	
<b>WW Confirmed Malfunctions</b>	<b>0</b>	<b>1</b>	<b>1</b>

[More details](#) about malfunctions

[References](#) cited in table above

**INCEPTA CRT-D**


Models N161/N163/N164/N165/P163/  
P165

U.S. Survival Probability	Worldwide Malfunction Details	Product Advisories
---------------------------	-------------------------------	--------------------


**U.S. Summary**

U.S. Registered Implants: 8,000	U.S. Normal Battery Depletions: 0
U.S. Approval Date: November 2011	U.S. Unconfirmed Reports of Premature Battery Depletion : 1
U.S. Estimated Active Implants: 7,000	U.S. Malfunctions:2
	Without Compromised Therapy:0
	With Compromised Therapy:2

**Battery Depletions and Malfunctions**



**Malfunctions Only**



**U.S. Survival Probability**


	Year	1	2	3	4	5	6	7	8	9	10
Non Advisory Population	Depletions and Malfunctions(%) (Confidence Interval)	99.94 (-0.1/+0.0)	99.94 @ 22 mo. (-0.1/+0.0)	-	-	-	-	-	-	-	-
Registered Implants: 8000	Malfunctions Only(%) (Confidence Interval)	99.97 (-0.1/+0.0)	99.97 @ 22 mo. (-0.1/+0.0)	-	-	-	-	-	-	-	-
	Effective Sample Size	3112	289	-	-	-	-	-	-	-	-



**INCEPTA CRT-D**

Models N161/N163/N164/N165/P163/  
P165

U.S. Survival Probability	Worldwide Malfunction Details	Product Advisories
---------------------------	-------------------------------	--------------------

**INCEPTA CRT-D**  
Models N161/N163/N164/N165/P163/  
P165 

**Worldwide Distribution:** 11,000  
**Worldwide Confirmed Malfunctions:** 2

	Without Compromised Therapy	With Compromised Therapy	Total
<b>Electrical</b>	-	1	1
<sup>112</sup> High-voltage capacitor	-	1	
<b>Mechanical</b>	-	1	1
<sup>103</sup> Transformer	-	1	
<b>Software</b>	-	-	0
<b>Other</b>	-	-	0
Non-patterned	-	-	
<b>WW Confirmed Malfunctions</b>	<b>0</b>	<b>2</b>	<b>2</b>

[More details](#) about malfunctions

[References](#) cited in table above

**ENERGEN CRT-D 4-Site**


Models N140/N142/P142

U.S. Survival Probability	Worldwide Malfunction Details	Product Advisories
---------------------------	-------------------------------	--------------------


**U.S. Summary**

U.S. Registered Implants: 9,000	U.S. Normal Battery Depletions: 0
U.S. Approval Date: November 2011	U.S. Unconfirmed Reports of Premature Battery Depletion : 1
U.S. Estimated Active Implants: 9,000	U.S. Malfunctions:0
	Without Compromised Therapy:0
	With Compromised Therapy:0

**Battery Depletions and Malfunctions**



**Malfunctions Only**



**U.S. Survival Probability**


	Year	1	2	3	4	5	6	7	8	9	10
Non Advisory Population Registered Implants: 9000	Depletions and Malfunctions(%) (Confidence Interval)	99.98 (-0.1/+0.0)	99.98 @ 23 mo. (-0.1/+0.0)	-	-	-	-	-	-	-	-
	Malfunctions Only(%) (Confidence Interval)	100.00 (-0.0/+0.0)	100.00 @ 23 mo. (-0.0/+0.0)	-	-	-	-	-	-	-	-
	Effective Sample Size	4084	238	-	-	-	-	-	-	-	-

**ENERGEN CRT-D 4-Site**

Models N140/N142/P142

<b>U.S. Survival Probability</b>	<b>Worldwide Malfunction Details</b>	<b>Product Advisories</b>
----------------------------------	--------------------------------------	---------------------------

**ENERGEN CRT-D 4-Site**  
**Models N140/N142/P142**



**Worldwide Distribution:** 14,000  
**Worldwide Confirmed Malfunctions:** 2

	Without Compromised Therapy	With Compromised Therapy	Total
<b>Electrical</b>	<b>1</b>	<b>-</b>	<b>1</b>
<sup>121</sup> Integrated circuit	1	-	
<b>Mechanical</b>	<b>-</b>	<b>-</b>	<b>0</b>
<b>Software</b>	<b>-</b>	<b>-</b>	<b>0</b>
<b>Other</b>	<b>1</b>	<b>-</b>	<b>1</b>
Non-patterned	1	-	
<b>WW Confirmed Malfunctions</b>	<b>2</b>	<b>0</b>	<b>2</b>

[More details](#) about malfunctions

[References](#) cited in table above

## ENERGEN CRT-D


Models N141/N143/P143

U.S. Survival Probability	Worldwide Malfunction Details	Product Advisories
---------------------------	-------------------------------	--------------------


### U.S. Summary

<p>U.S. Registered Implants: 9,000                  U.S. Approval Date: November 2011                  U.S. Estimated Active Implants: 8,000</p>	<p>U.S. Normal Battery Depletions: 0                  U.S. Unconfirmed Reports of Premature Battery Depletion : 0                  U.S. Malfunctions:7                  Without Compromised Therapy:4                  With Compromised Therapy:3</p>
--	---

### Battery Depletions and Malfunctions



### Malfunctions Only




### U.S. Survival Probability

	Year	1	2	3	4	5	6	7	8	9	10
Non Advisory Population Registered Implants: 9000	Depletions and Malfunctions(%) (Confidence Interval)	99.88 (-0.1/+0.1)	99.88 @ 22 mo. (-0.1/+0.1)	-	-	-	-	-	-	-	-
	Malfunctions Only(%) (Confidence Interval)	99.88 (-0.1/+0.1)	99.88 @ 22 mo. (-0.1/+0.1)	-	-	-	-	-	-	-	-
	Effective Sample Size	3817	359	-	-	-	-	-	-	-	-

## ENERGEN CRT-D

Models N141/N143/P143

<b>U.S. Survival Probability</b>	<b>Worldwide Malfunction Details</b>	<b>Product Advisories</b>
----------------------------------	--------------------------------------	---------------------------

<b>ENERGEN CRT-D</b> <b>Models N141/N143/P143</b> 			
<b>Worldwide Distribution:</b> 12,000			
<b>Worldwide Confirmed Malfunctions:</b> 8			
	Without Compromised Therapy	With Compromised Therapy	Total
<b>Electrical</b>	<b>2</b>	<b>1</b>	<b>3</b>
<sup>110</sup> Safety Core-electrocautery	1	-	
<sup>117</sup> Low-voltage capacitors	1	-	
<sup>121</sup> Integrated circuit	-	1	
<b>Mechanical</b>	<b>-</b>	<b>3</b>	<b>3</b>
<sup>103</sup> Transformer	-	3	
<b>Software</b>	<b>1</b>	<b>-</b>	<b>1</b>
<sup>122</sup> Memory errors	1	-	
<b>Other</b>	<b>1</b>	<b>-</b>	<b>1</b>
Non-patterned	1	-	
<b>WW Confirmed Malfunctions</b>	<b>4</b>	<b>4</b>	<b>8</b>

[More details](#) about malfunctions


[References](#) cited in table above

**PUNCTUA CRT-D 4-Site**

Models N050/N052/P052

<b>U.S. Survival Probability</b>	<b>Worldwide Malfunction Details</b>	<b>Product Advisories</b>
----------------------------------	--------------------------------------	---------------------------

**PUNCTUA CRT-D 4-Site**  
**Models N050/N052/P052**



**Worldwide Distribution: 2,000**  
**Worldwide Confirmed Malfunctions: 0**

	Without Compromised Therapy	With Compromised Therapy	Total
<b>Electrical</b>	-	-	<b>0</b>
<b>Mechanical</b>	-	-	<b>0</b>
<b>Software</b>	-	-	<b>0</b>
<b>Other</b>	-	-	<b>0</b>
Non-patterned	-	-	
<b>WW Confirmed Malfunctions</b>	<b>0</b>	<b>0</b>	<b>0</b>

[More details](#) about malfunctions


[References](#) cited in table above

**PUNCTUA CRT-D**

Models N051/N053/P053

<b>U.S. Survival Probability</b>	<b>Worldwide Malfunction Details</b>	<b>Product Advisories</b>
----------------------------------	--------------------------------------	---------------------------

**PUNCTUA CRT-D**  
**Models N051/N053/P053**



**Worldwide Distribution: 3,000**  
**Worldwide Confirmed Malfunctions: 0**

	Without Compromised Therapy	With Compromised Therapy	Total
<b>Electrical</b>	-	-	<b>0</b>
<b>Mechanical</b>	-	-	<b>0</b>
<b>Software</b>	-	-	<b>0</b>
<b>Other</b>	-	-	<b>0</b>
Non-patterned	-	-	
<b>WW Confirmed Malfunctions</b>	<b>0</b>	<b>0</b>	<b>0</b>

[More details](#) about malfunctions

[References](#) cited in table above

CRM PRODUCT PERFORMANCE REPORT Q1 2014

COGNIS


Models N106/N107/N108/N118/N119/  
N120/P106/P107/P108

U.S. Survival Probability	Worldwide Malfunction Details	Product Advisories
---------------------------	-------------------------------	--------------------


U.S. Summary

U.S. Registered Implants: 75,000	U.S. Normal Battery Depletions: 290
U.S. Approval Date: May 2008	U.S. Unconfirmed Reports of Premature Battery Depletion : 20
U.S. Estimated Active Implants: 52,000	U.S. Malfunctions:317
	Without Compromised Therapy:215
	With Compromised Therapy:102

Battery Depletions and Malfunctions



Malfunctions Only



U.S. Survival Probability

	Year	1	2	3	4	5	6	7	8	9	10
Non Advisory Population Registered Implants: 40000	Depletions and Malfunctions(%) (Confidence Interval)	99.92 (-0.0/+0.0)	99.84 (-0.0/+0.0)	99.68 (-0.1/+0.1)	99.44 @ 47 mo. (-0.2/+0.1)	-	-	-	-	-	-
	Malfunctions Only(%) (Confidence Interval)	99.93 (-0.0/+0.0)	99.89 (-0.0/+0.0)	99.85 (-0.1/+0.0)	99.74 @ 47 mo. (-0.2/+0.1)	-	-	-	-	-	-
	Effective Sample Size	35512	29399	12165	327	-	-	-	-	-	-
Advisory: 01-Dec-09 (Subpectoral Implant 2009)* Registered Implants: 32,000	Depletions and Malfunctions(%) (Confidence Interval)	99.76 (-0.1/+0.1)	99.63 (-0.1/+0.1)	99.37 (-0.2/+0.1)	98.60 (-0.3/+0.3)	97.03 (-0.6/+0.5)	96.33 @ 65 (-0.6/+0.6)	-	-	-	-
	Malfunctions Only(%) (Confidence Interval)	99.82 (-0.1/+0.1)	99.71 (-0.1/+0.1)	99.61 (-0.1/+0.1)	99.23 (-0.1/+0.1)	98.74 (-0.3/+0.2)	98.66 @ 65 (-0.3/+0.2)	-	-	-	-
	Effective Sample Size	27518	24405	21708	18655	4251	497	-	-	-	-
Advisory: 29-Aug-13 (Low Voltage Capacitor 2013)* Registered Implants: 12,000	Depletions and Malfunctions(%) (Confidence Interval)	99.81 (-0.1/+0.1)	99.70 (-0.1/+0.1)	99.37 (-0.2/+0.1)	97.83 (-0.4/+0.3)	95.18 (-1.2/+1.0)	93.67 @ 65 (-1.4/+1.2)	-	-	-	-
	Malfunctions Only(%) (Confidence Interval)	99.81 (-0.1/+0.1)	99.70 (-0.1/+0.1)	99.37 (-0.2/+0.1)	97.83 (-0.4/+0.3)	95.18 (-1.2/+1.0)	93.67 @ 65 (-1.4/+1.2)	-	-	-	-
	Effective Sample Size	27518	24405	21708	18655	4251	497	-	-	-	-



Malfunctions Only(%) (Confidence Interval)	99.86 (-0.1/+0.1)	99.76 (-0.1/+0.1)	99.59 (-0.2/+0.1)	98.48 (-0.3/+0.3)	97.22 (-0.6/+0.4)	97.11 @ 65 (-0.8/+0.7)	-	-	-	-
Effective Sample Size	10396	9179	8167	5495	796	241	-	-	-	-

\*Devices subject to an advisory. Refer to the Advisories for more details. Devices may be part of more than one advisory.

**COGNIS**

Models N106/N107/N108/N118/N119/  
N120/P106/P107/P108

U.S. Survival Probability	Worldwide Malfunction Details	Product Advisories
---------------------------	-------------------------------	--------------------

**COGNIS**  
Models N106/N107/N108/N118/N119/  
N120/P106/P107/P108 

Worldwide Distribution: 108,000  
Worldwide Confirmed Malfunctions: 415

	Without Compromised Therapy	With Compromised Therapy	Total
<b>Electrical</b>	<b>210</b>	<b>56</b>	<b>266</b>
<sup>1</sup> Low Voltage Capacitor (Advisory issued)	114	10	
<sup>110</sup> Safety Core-electrocautery	40	16	
<sup>112</sup> High-voltage capacitor	1	4	
<sup>117</sup> Low-voltage capacitors	7	-	
<sup>121</sup> Integrated circuit	7	19	
<sup>123</sup> High voltage circuit	-	1	
<sup>124</sup> Battery	15	2	
<sup>125</sup> Low-voltage capacitor	26	4	
<b>Mechanical</b>	<b>29</b>	<b>75</b>	<b>104</b>
<sup>5</sup> Subpectoral implant 2009 (Advisory issued)	12	35	
<sup>103</sup> Transformer	-	9	
<sup>108</sup> Difficulty securing lead	9	9	
<sup>115</sup> Header contacts	4	7	
<sup>129</sup> Header	4	15	
<b>Software</b>	<b>11</b>	<b>-</b>	<b>11</b>
<sup>116</sup> Safety Core-programming	1	-	
<sup>119</sup> Alert messages not displayed post-EOL	2	-	
<sup>122</sup> Memory errors	8	-	
<b>Other</b>	<b>26</b>	<b>8</b>	<b>34</b>
Non-patterned	26	8	
<b>WW Confirmed Malfunctions</b>	<b>276</b>	<b>139</b>	<b>415</b>

[More details](#) about malfunctions

[References](#) cited in table above

CRM PRODUCT PERFORMANCE REPORT Q1 2014

LIVIAN


Models H220/H225/H240/H245

U.S. Survival Probability	Worldwide Malfunction Details	Product Advisories
---------------------------	-------------------------------	--------------------


**U.S. Summary**

U.S. Registered Implants: 5,000	U.S. Normal Battery Depletions: 398
U.S. Approval Date: February 2008	U.S. Unconfirmed Reports of Premature Battery Depletion : 0
U.S. Estimated Active Implants: 2,000	U.S. Malfunctions:8
	Without Compromised Therapy:5
	With Compromised Therapy:3

**Battery Depletions and Malfunctions**



**Malfunctions Only**



**U.S. Survival Probability**

	Year	1	2	3	4	5	6	7	8	9	10
Non Advisory Population	Depletions and Malfunctions(%)	99.95	99.68	99.05	96.44	88.24	75.83	-	-	-	-
	(Confidence Interval)	(-0.1/+0.0)	(-0.2/+0.1)	(-0.4/+0.3)	(-0.8/+0.6)	(-1.4/+1.3)	@ 69 mo. (-2.7/+2.5)				
Registered Implants: 5000	Malfunctions Only(%)	99.98	99.87	99.80	99.77	99.77	99.77	-	-	-	-
	(Confidence Interval)	(-0.1/+0.0)	(-0.2/+0.1)	(-0.2/+0.1)	(-0.2/+0.1)	(-0.2/+0.1)	@ 69 mo. (-0.2/+0.1)				
	Effective Sample Size	3997	3482	2974	2384	1516	227	-	-	-	-

**LIVIAN**

Models H220/H225/H240/H245

<b>U.S. Survival Probability</b>	<b>Worldwide Malfunction Details</b>	<b>Product Advisories</b>
----------------------------------	--------------------------------------	---------------------------

**LIVIAN**  
Models H220/H225/H240/H245

**Worldwide Distribution:** 6,000  
**Worldwide Confirmed Malfunctions:** 9

	Without Compromised Therapy	With Compromised Therapy	Total
<b>Electrical</b>	<b>1</b>	<b>2</b>	<b>3</b>
<sup>42</sup> Integrated circuit	1	2	
<b>Mechanical</b>	<b>1</b>	-	<b>1</b>
<sup>47</sup> Seal plug	1	-	
<b>Software</b>	<b>1</b>	-	<b>1</b>
<sup>60</sup> Memory error	1	-	
<b>Other</b>	<b>2</b>	<b>2</b>	<b>4</b>
Non-patterned	-	2	
<sup>57</sup> Battery depletion	2	-	
<b>WW Confirmed Malfunctions</b>	<b>5</b>	<b>4</b>	<b>9</b>

[More details](#) about malfunctions

[References](#) cited in table above

CRM PRODUCT PERFORMANCE REPORT Q1 2014

LIVIAN HE


Models H227/H229/H247/H249

U.S. Survival Probability	Worldwide Malfunction Details	Product Advisories
---------------------------	-------------------------------	--------------------


**U.S. Summary**

U.S. Registered Implants: 6,000	U.S. Normal Battery Depletions: 528
U.S. Approval Date: February 2008	U.S. Unconfirmed Reports of Premature Battery Depletion : 4
U.S. Estimated Active Implants: 3,000	U.S. Malfunctions:4
	Without Compromised Therapy:2
	With Compromised Therapy:2

**Battery Depletions and Malfunctions**



**Malfunctions Only**




**U.S. Survival Probability**

	Year	1	2	3	4	5	6	7	8	9	10
Non Advisory Population	Depletions and Malfunctions(%)	99.98	99.74	98.89	96.40	87.50	71.48	-	-	-	-
	(Confidence Interval)	(-0.1/+0.0)	(-0.2/+0.1)	(-0.4/+0.3)	(-0.7/+0.6)	(-1.4/+1.2)	@ 68 mo. (-2.5/+2.4)				
Registered Implants: 6000	Malfunctions Only(%)	99.98	99.92	99.92	99.92	99.92	99.92	-	-	-	-
	(Confidence Interval)	(-0.1/+0.0)	(-0.1/+0.1)	(-0.1/+0.1)	(-0.1/+0.1)	(-0.1/+0.1)	@ 68 mo. (-0.1/+0.1)				
	Effective Sample Size	4942	4308	3593	2804	1597	372	-	-	-	-

**LIVIAN HE**

Models H227/H229/H247/H249

<b>U.S. Survival Probability</b>	<b>Worldwide Malfunction Details</b>	<b>Product Advisories</b>
----------------------------------	--------------------------------------	---------------------------

**LIVIAN HE**  
Models H227/H229/H247/H249 

**Worldwide Distribution:** 7,000  
**Worldwide Confirmed Malfunctions:** 6

	Without Compromised Therapy	With Compromised Therapy	Total
<b>Electrical</b>	<b>1</b>	<b>1</b>	<b>2</b>
<sup>42</sup> Integrated circuit	1	1	
<b>Mechanical</b>	-	<b>2</b>	<b>2</b>
<sup>108</sup> Difficulty securing lead	-	2	
<b>Software</b>	-	-	<b>0</b>
<b>Other</b>	<b>1</b>	<b>1</b>	<b>2</b>
Non-patterned	1	-	
<sup>57</sup> Battery depletion	-	1	
<b>WW Confirmed Malfunctions</b>	<b>2</b>	<b>4</b>	<b>6</b>

[More details](#) about malfunctions

[References](#) cited in table above

CRM PRODUCT PERFORMANCE REPORT Q1 2014


CONTAK RENEWAL 3 RF

Models H210/H215


U.S. Survival Probability	Worldwide Malfunction Details	Product Advisories
---------------------------	-------------------------------	--------------------

<b>U.S. Summary</b>  U.S. Registered Implants: 21,000 U.S. Approval Date: February 2005 U.S. Estimated Active Implants: 3,000	U.S. Normal Battery Depletions: 6,309 U.S. Unconfirmed Reports of Premature Battery Depletion : 27 U.S. Malfunctions:175 Without Compromised Therapy:157 With Compromised Therapy:18
---	--

Battery Depletions and Malfunctions



Malfunctions Only



U.S. Survival Probability

	Year	1	2	3	4	5	6	7	8	9	10
Non Advisory Population	Depletions and Malfunctions(%) (Confidence Interval)	99.94 (-0.3/+0.0)	99.82 (-0.4/+0.1)	99.17 (-0.6/+0.4)	96.26 (-1.2/+0.9)	88.48 (-2.0/+1.8)	65.03 @ 71 mo. (-3.6/+3.5)	-	-	-	-
	Registered Implants: 2000										
	Malfunctions Only(%) (Confidence Interval)	99.94 (-0.3/+0.0)	99.94 (-0.3/+0.0)	99.94 (-0.3/+0.0)	99.78 (-0.5/+0.2)	99.78 (-0.5/+0.2)	99.78 @ 71 mo. (-0.5/+0.2)	-	-	-	-
	Effective Sample Size	1735	1524	1321	1126	872	215	-	-	-	-
21-Jul-10 Magnetic Reed Switch 2010*	Depletions and Malfunctions(%) (Confidence Interval)	99.91 (-0.1/+0.0)	99.75 (-0.1/+0.1)	98.81 (-0.2/+0.2)	96.00 (-0.4/+0.4)	86.62 (-0.7/+0.7)	58.53 (-1.2/+1.1)	20.92 (-1.5/+1.5)	20.33 @ 85 mo. (-1.5/+1.6)	-	-
	Registered Implants: 15000										
	Malfunctions Only(%) (Confidence Interval)	99.96 (-0.1/+0.0)	99.88 (-0.1/+0.0)	99.77 (-0.1/+0.1)	99.53 (-0.2/+0.1)	99.28 (-0.2/+0.2)	99.19 (-0.2/+0.2)	98.97 (-0.3/+0.2)	98.97 @ 85 mo. (-0.3/+0.2)	-	-
	Effective Sample Size	12967	11434	9925	8445	6670	3716	285	210	-	-
05-Apr-07 and 04-Mar-09 Shortened Replacement Window*	Depletions and Malfunctions(%) (Confidence Interval)	99.83 (-0.2/+0.1)	99.41 (-0.3/+0.2)	97.57 (-0.6/+0.5)	91.86 (-1.2/+1.0)	72.82 (-2.0/+1.9)	30.99 (-2.2/+2.2)	14.82 @ 78 mo. (-1.7/+1.9)	-	-	-
	Registered Implants:										
	Effective Sample Size										

4000	Malfunions Only (%) (Confidence Interval)	99.89 (-0.2/+0.1)	99.66 (-0.3/+0.2)	98.82 (-0.5/+0.3)	97.09 (-0.7/+0.6)	94.91 (-1.0/+0.9)	93.55 (-1.3/+1.1)	92.74 @ 78 mo. (-1.6/+1.4)	-	-	-
	Effective Sample Size	3377	2941	2484	2036	1398	501	206	-	-	-
	11-Mar-06 Low Level Current*	Depletions and Malfunions (%) (Confidence Interval)	99.90 (-0.1/+0.0)	99.67 (-0.1/+0.1)	98.62 (-0.2/+0.2)	95.47 (-0.4/+0.4)	84.74 (-0.7/+0.7)	54.17 (-1.0/+1.0)	19.70 (-1.1/+1.2)	17.59 @ 88 mo. (-1.2/+1.3)	-
Registered Implants: 19000	Malfunions Only (%) (Confidence Interval)	99.94 (-0.0/+0.0)	99.84 (-0.1/+0.1)	99.64 (-0.1/+0.1)	99.23 (-0.2/+0.1)	98.73 (-0.2/+0.2)	98.48 (-0.3/+0.2)	98.20 (-0.3/+0.3)	98.20 @ 88 mo. (-0.3/+0.3)	-	-
	Effective Sample Size	16380	14429	12470	10564	8197	4387	462	200	-	-

\*Devices subject to an advisory. Refer to the Advisories for more details. Devices may be part of more than one advisory.



**CONTAK RENEWAL 3 RF**

Models H210/H215

<b>U.S. Survival Probability</b>	<b>Worldwide Malfunction Details</b>	<b>Product Advisories</b>
----------------------------------	--------------------------------------	---------------------------

**CONTAK RENEWAL 3 RF**  
Models H210/H215 

**Worldwide Distribution:** 21,000  
**Worldwide Confirmed Malfunctions:** 177

	Without Compromised Therapy	With Compromised Therapy	Total
<b>Electrical</b>	<b>143</b>	<b>5</b>	<b>148</b>
<sup>7</sup> Shortened replacement window (Advisory issued)	84	2	
<sup>21</sup> Extended charge time post-mid-life	1	-	
<sup>32</sup> Capacitor	2	-	
<sup>42</sup> Integrated circuit	8	3	
<sup>64</sup> Capacitor	1	-	
<sup>69</sup> Capacitor	3	-	
<sup>81</sup> Mid-life display of replacement indicators	13	-	
<sup>82</sup> High-voltage capacitor	2	-	
<sup>109</sup> Low-voltage capacitor	29	-	
<b>Mechanical</b>	<b>8</b>	<b>10</b>	<b>18</b>
<sup>4</sup> Magnetic reed switch 2010 (Advisory issued)	5	5	
<sup>15</sup> Magnetic switch (Advisory issued)	-	1	
<sup>47</sup> Seal plug	2	-	
<sup>93</sup> Setscrew	1	-	
<sup>95</sup> Seal plug	-	1	
<sup>118</sup> Bent flex circuit	-	3	
<b>Software</b>	<b>3</b>	<b>-</b>	<b>3</b>
<sup>25</sup> Parameter errors	1	-	
<sup>80</sup> Memory location	1	-	
<sup>106</sup> Misaligned markers	1	-	
<b>Other</b>	<b>5</b>	<b>3</b>	<b>8</b>
Non-patterned	-	2	
<sup>57</sup> Battery depletion	5	1	
<b>WW Confirmed Malfunctions</b>	<b>159</b>	<b>18</b>	<b>177</b>

[More details](#) about malfunctions


[References](#) cited in table above

**CONTAK RENEWAL 3 RF HE**

Models H217/H219

<b>U.S. Survival Probability</b>	<b>Worldwide Malfunction Details</b>	<b>Product Advisories</b>
----------------------------------	--------------------------------------	---------------------------

<b>U.S. Summary</b>	
U.S. Registered Implants: 18,000	U.S. Normal Battery Depletions: 6,110
U.S. Approval Date: February 2005	U.S. Unconfirmed Reports of Premature Battery Depletion : 24
U.S. Estimated Active Implants: 2,000	U.S. Malfunctions:141
	Without Compromised Therapy:121
	With Compromised Therapy:20



<b>U.S. Survival Probability</b>		Year	1	2	3	4	5	6	7	8	9	10
Non Advisory Population	Depletions and Malfunctions(%) (Confidence Interval)	99.93 (-0.4/+0.1)	99.63 (-0.5/+0.2)	98.85 (-0.8/+0.5)	95.13 (-1.5/+1.1)	84.05 (-2.5/+2.2)	52.71 @ 70 mo. (-3.9/+3.8)	-	-	-	-	-
	Malfunctions Only(%) (Confidence Interval)	99.93 (-0.4/+0.1)	99.93 (-0.4/+0.1)	99.93 (-0.4/+0.1)	99.93 (-0.4/+0.1)	99.93 (-0.4/+0.1)	99.93 @ 70 mo. (-0.4/+0.1)	-	-	-	-	-
	Effective Sample Size	1457	1269	1097	926	708	211	-	-	-	-	-
21-Jul-10 Magnetic Reed Switch 2010*	Depletions and Malfunctions(%) (Confidence Interval)	99.90 (-0.1/+0.0)	99.71 (-0.1/+0.1)	98.41 (-0.3/+0.2)	93.88 (-0.6/+0.5)	77.77 (-1.0/+1.0)	35.12 (-1.3/+1.3)	20.19 @ 83 mo. (-1.3/+1.3)	-	-	-	-
	Malfunctions Only(%) (Confidence Interval)	99.94 (-0.1/+0.0)	99.89 (-0.1/+0.0)	99.75 (-0.1/+0.1)	99.43 (-0.2/+0.1)	99.31 (-0.2/+0.2)	99.16 (-0.3/+0.2)	99.16 @ 83 mo. (-0.3/+0.2)	-	-	-	-
	Effective Sample Size	10724	9409	8074	6756	4919	1722	241	-	-	-	-
05-Apr-07 and 04-Mar-09 Shortened Replacement Window*	Depletions and Malfunctions(%) (Confidence Interval)	99.88 (-0.2/+0.1)	99.36 (-0.4/+0.2)	96.77 (-0.8/+0.6)	87.27 (-1.5/+1.3)	55.67 (-2.3/+2.3)	18.06 (-1.9/+2.1)	16.57 @ 73 mo. (-1.9/+2.0)	-	-	-	-
	Malfunctions Only(%) (Confidence Interval)	99.88 (-0.2/+0.1)	99.36 (-0.4/+0.2)	96.77 (-0.8/+0.6)	87.27 (-1.5/+1.3)	55.67 (-2.3/+2.3)	18.06 (-1.9/+2.1)	16.57 @ 73 mo. (-1.9/+2.0)	-	-	-	-
	Effective Sample Size	10724	9409	8074	6756	4919	1722	241	-	-	-	-


4000	Malfunctions Only (%) (Confidence Interval)	99.91 (-0.2/+0.1)	99.65 (-0.3/+0.2)	98.59 (-0.5/+0.4)	96.01 (-0.9/+0.8)	95.18 (-1.0/+0.9)	94.55 (-1.4/+1.2)	93.65 @ 73 mo. (-2.0/+1.6)	-	-	-
	Effective Sample Size	3013	2622	2211	1722	931	231	210	-	-	-
	11-Mar-06 Low Level Current*	Depletions and Malfunctions (%) (Confidence Interval)	99.89 (-0.1/+0.0)	99.64 (-0.1/+0.1)	98.10 (-0.3/+0.2)	92.69 (-0.5/+0.5)	74.01 (-0.9/+0.9)	32.19 (-1.1/+1.1)	19.03 (-1.1/+1.1)	18.54 @ 87 mo. (-1.1/+1.2)	-
Registered Implants: 16000	Malfunctions Only (%) (Confidence Interval)	99.93 (-0.1/+0.0)	99.84 (-0.1/+0.1)	99.56 (-0.1/+0.1)	98.94 (-0.2/+0.2)	98.71 (-0.2/+0.2)	98.49 (-0.3/+0.3)	98.02 (-0.8/+0.6)	98.02 @ 87 mo. (-0.8/+0.6)	-	-
	Effective Sample Size	13773	12074	10336	8548	5981	2037	347	227	-	-

\*Devices subject to an advisory. Refer to the Advisories for more details. Devices may be part of more than one advisory.

**CONTAK RENEWAL 3 RF HE**

Models H217/H219

U.S. Survival Probability	Worldwide Malfunction Details	Product Advisories
---------------------------	-------------------------------	--------------------

<b>CONTAK RENEWAL 3 RF HE</b> <b>Models H217/H219</b> 			
<b>Worldwide Distribution: 18,000</b>			
<b>Worldwide Confirmed Malfunctions: 141</b>			
	Without Compromised Therapy	With Compromised Therapy	Total
<b>Electrical</b>	<b>111</b>	<b>7</b>	<b>118</b>
<sup>7</sup> Shortened replacement window (Advisory issued)	65	5	
<sup>21</sup> Extended charge time post-mid-life	12	-	
<sup>32</sup> Capacitor	1	-	
<sup>42</sup> Integrated circuit	5	1	
<sup>64</sup> Capacitor	1	-	
<sup>81</sup> Mid-life display of replacement indicators	8	-	
<sup>89</sup> Integrated circuit	1	-	
<sup>109</sup> Low-voltage capacitor	18	1	
<b>Mechanical</b>	<b>3</b>	<b>8</b>	<b>11</b>
<sup>4</sup> Magnetic reed switch 2010 (Advisory issued)	3	7	
<sup>93</sup> Setscrew	-	1	
<b>Software</b>	<b>-</b>	<b>-</b>	<b>0</b>
<b>Other</b>	<b>7</b>	<b>5</b>	<b>12</b>
Non-patterned	4	4	
<sup>57</sup> Battery depletion	3	1	
<b>WW Confirmed Malfunctions</b>	<b>121</b>	<b>20</b>	<b>141</b>

[More details](#) about malfunctions

[References](#) cited in table above

## CONTAK RENEWAL 3

Models H170/H175


<b>U.S. Survival Probability</b>	<b>Worldwide Malfunction Details</b>	<b>Product Advisories</b>
----------------------------------	--------------------------------------	---------------------------

### U.S. Summary


U.S. Registered Implants: 34,000  
 U.S. Approval Date: June 2003  
 U.S. Estimated Active Implants: 2,000

U.S. Normal Battery Depletions: 11,863  
 U.S. Unconfirmed Reports of Premature Battery Depletion : 72  
 U.S. Malfunctions:973  
 Without Compromised Therapy:926  
 With Compromised Therapy:47

### Battery Depletions and Malfunctions



### Malfunctions Only



— Non Advisory Population  
- - - 05-Apr-2007 and 04-Mar-2009: Shortened Replacement Window  
- - - 23-Jun-2005: Magnetic Switch  
- - - 21-Jul-2010: Magnetic Reed Switch 2010  
- - - 10-Mar-2007: Product Update - Mid-life Display of Replacement Indicators  
— Estimated Longevity

### U.S. Survival Probability

	Year	1	2	3	4	5	6	7	8	9	10
Non Advisory Population	Depletions and Malfunctions(%) (Confidence Interval)	99.94 (-0.3/+0.0)	99.72 (-0.5/+0.2)	98.57 (-0.8/+0.5)	95.41 (-1.4/+1.1)	85.15 (-2.5/+2.2)	45.38 @ 71 mo. (-4.3/+4.3)	-	-	-	-
	Registered Implants: 2000										
	Malfunctions Only(%) (Confidence Interval)	99.94 (-0.3/+0.0)	99.94 (-0.3/+0.0)	99.85 (-0.5/+0.1)	99.28 (-0.7/+0.4)	99.00 (-0.9/+0.5)	98.22 @ 71 mo. (-1.9/+0.9)	-	-	-	-
	Effective Sample Size	1504	1320	1133	914	638	235	-	-	-	-
21-Jul-10 Magnetic Reed Switch 2010*	Depletions and Malfunctions(%) (Confidence Interval)	99.91 (-0.3/+0.1)	99.43 (-0.5/+0.3)	98.14 (-0.8/+0.5)	94.13 (-1.3/+1.1)	82.29 (-2.2/+2.0)	49.28 (-3.0/+3.0)	29.35 @ 77 mo. (-2.9/+3.1)	-	-	-
	Registered Implants: 2000										
	Malfunctions Only(%) (Confidence Interval)	99.91 (-0.3/+0.1)	99.54 (-0.4/+0.2)	99.04 (-0.6/+0.4)	98.77 (-0.7/+0.4)	98.13 (-0.9/+0.6)	97.78 (-1.0/+0.7)	97.78 @ 77 mo. (-1.0/+0.7)	-	-	-
	Effective Sample Size	2060	1778	1530	1264	951	486	211	-	-	-
05-Apr-07 and 04-Mar-09 Shortened Replacement Window*	Depletions and Malfunctions(%) (Confidence Interval)	99.76 (-0.1/+0.1)	98.67 (-0.3/+0.2)	95.23 (-0.5/+0.5)	86.00 (-0.9/+0.8)	64.36 (-1.3/+1.2)	26.45 (-1.3/+1.3)	14.12 (-1.0/+1.1)	12.87 (-1.0/+1.1)	12.87 @ 98 mo. (-1.0/+1.1)	-
	Registered Implants:										
	Shortened Replacement Window*										


10000	Malfunctions Only (%) (Confidence Interval)	99.78 (-0.1/+0.1)	98.95 (-0.2/+0.2)	97.13 (-0.4/+0.4)	94.05 (-0.6/+0.5)	90.05 (-0.8/+0.8)	88.13 (-1.0/+0.9)	87.78 (-1.1/+1.0)	87.58 (-1.2/+1.1)	87.58 @ 98 mo. (-1.2/+1.1)	—
	Effective Sample Size	8903	7754	6531	5105	3285	1129	484	255	210	—
10-Mar-07 Product Update - Mid-life Display of Replacement Indicators*	Depletions and Malfunctions (%) (Confidence Interval)	99.84 (-0.1/+0.0)	99.62 (-0.1/+0.1)	97.83 (-0.3/+0.2)	91.12 (-0.5/+0.5)	73.94 (-0.8/+0.8)	33.27 (-1.0/+1.0)	16.38 (-0.8/+0.8)	15.15 (-0.8/+0.8)	15.00 (-0.8/+0.8)	14.94 @ 116 mo. (-0.8/+0.8)
Registered Implants: 21000	Malfunctions Only (%) (Confidence Interval)	99.87 (-0.1/+0.0)	99.78 (-0.1/+0.1)	99.50 (-0.1/+0.1)	98.11 (-0.3/+0.2)	97.56 (-0.3/+0.3)	96.86 (-0.4/+0.4)	96.36 (-0.5/+0.5)	96.16 (-0.6/+0.5)	96.16 (-0.6/+0.5)	96.16 @ 116 mo. (-0.8/+0.5)
	Effective Sample Size	17334	15154	13017	10427	7260	2731	1029	828	553	219
12-May-06 Premature Battery Depletion*	Survival probability data not provided because this population does not meet report inclusion criteria (see Statistical Methodology for more details). Refer to Product Advisories for more information.										
23-Jun-05 Magnetic Switch*	Depletions and Malfunctions (%) (Confidence Interval)	99.82 (-0.1/+0.0)	99.51 (-0.1/+0.1)	97.65 (-0.2/+0.2)	91.01 (-0.5/+0.5)	74.13 (-0.8/+0.8)	33.81 (-0.9/+0.9)	16.31 (-0.8/+0.8)	14.96 (-0.8/+0.8)	14.80 (-0.8/+0.8)	14.74 @ 116 mo. (-0.8/+0.8)
Registered Implants: 23000	Malfunctions Only (%) (Confidence Interval)	99.85 (-0.1/+0.0)	99.70 (-0.1/+0.1)	99.30 (-0.1/+0.1)	97.88 (-0.3/+0.2)	97.21 (-0.3/+0.3)	96.54 (-0.4/+0.3)	96.07 (-0.5/+0.4)	95.89 (-0.6/+0.5)	95.89 (-0.6/+0.5)	95.89 @ 116 mo. (-0.6/+0.5)
	Effective Sample Size	19269	16832	14470	11631	8156	3110	1163	931	552	218

\*Devices subject to an advisory. Refer to the Advisories for more details. Devices may be part of more than one advisory.

## CONTAK RENEWAL 3

Models H170/H175

U.S. Survival Probability	Worldwide Malfunction Details	Product Advisories
---------------------------	-------------------------------	--------------------

<b>CONTAK RENEWAL 3</b> <b>Models H170/H175</b> 			
<b>Worldwide Distribution:</b> 34,000			
<b>Worldwide Confirmed Malfunctions:</b> 975			
	Without Compromised Therapy	With Compromised Therapy	Total
<b>Electrical</b>	<b>859</b>	<b>26</b>	<b>885</b>
<sup>7</sup> Shortened replacement window (Advisory issued)	320	13	
<sup>9</sup> Premature battery depletion (Advisory issued)	18	-	
<sup>21</sup> Extended charge time post-mid-life	46	-	
<sup>27</sup> Integrated circuit	1	1	
<sup>32</sup> Capacitor	3	1	
<sup>42</sup> Integrated circuit	2	5	
<sup>64</sup> Capacitor	9	3	
<sup>69</sup> Capacitor	12	-	
<sup>71</sup> Device tones	1	-	
<sup>81</sup> Mid-life display of replacement indicators	203	-	
<sup>89</sup> Integrated circuit	1	1	
<sup>109</sup> Low-voltage capacitor	243	2	
<b>Mechanical</b>	<b>37</b>	<b>16</b>	<b>53</b>
<sup>4</sup> Magnetic reed switch 2010 (Advisory issued)	-	1	
<sup>10</sup> Subpectoral implant (Advisory issued)	-	5	
<sup>15</sup> Magnetic switch (Advisory issued)	-	2	
<sup>33</sup> Header	5	2	
<sup>47</sup> Seal plug	26	4	
<sup>61</sup> Adhesive consistency	-	1	
<sup>93</sup> Setscrew	4	1	
<sup>95</sup> Seal plug	1	-	
<sup>102</sup> Cracked solder joint	1	-	
<b>Software</b>	<b>3</b>	<b>-</b>	<b>3</b>
<sup>80</sup> Memory location	1	-	
<sup>106</sup> Misaligned markers	2	-	
<b>Other</b>	<b>29</b>	<b>5</b>	<b>34</b>
Non-patterned	18	4	
<sup>26</sup> Firmware error	1	-	
<sup>57</sup> Battery depletion	10	1	
<b>WW Confirmed Malfunctions</b>	<b>928</b>	<b>47</b>	<b>975</b>

[More details](#) about malfunctions

[References](#) cited in table above


CONTAK RENEWAL 3 HE

Models H177/H179


U.S. Survival Probability	Worldwide Malfunction Details	Product Advisories
---------------------------	-------------------------------	--------------------

<b>U.S. Summary</b>	
U.S. Registered Implants: 24,000	U.S. Normal Battery Depletions: 8,088
U.S. Approval Date: June 2003	U.S. Unconfirmed Reports of Premature Battery Depletion : 92
U.S. Estimated Active Implants: 1,000	U.S. Malfunctions:875
	Without Compromised Therapy:833
	With Compromised Therapy:42

Battery Depletions and Malfunctions



Malfunctions Only



U.S. Survival Probability

	Year	1	2	3	4	5	6	7	8	9	10
Non Advisory Population Registered Implants: 1000	Depletions and Malfunctions(%) (Confidence Interval)	99.90 (-0.6/+0.1)	99.78 (-0.7/+0.2)	98.73 (-1.1/+0.6)	94.57 (-2.0/+1.5)	75.02 (-4.0/+3.7)	59.55 @ 64 mo. (-4.8/+4.6)	-	-	-	-
	Malfunctions Only(%) (Confidence Interval)	100.00 (-0.0/+0.0)	100.00 (-0.0/+0.0)	100.00 (-0.0/+0.0)	99.67 (-1.0/+0.2)	99.67 (-1.0/+0.2)	99.27 @ 65 mo. (-1.8/+0.5)	-	-	-	-
	Effective Sample Size	976	850	697	544	325	212	-	-	-	-
21-Jul-10 Magnetic Reed Switch 2010* Registered Implants: 1000	Depletions and Malfunctions(%) (Confidence Interval)	99.87 (-0.8/+0.1)	99.38 (-1.0/+0.4)	98.04 (-1.6/+0.9)	94.29 (-2.5/+1.8)	74.18 (-4.6/+4.1)	59.82 @ 65 mo. (-5.1/+4.9)	-	-	-	-
	Malfunctions Only(%) (Confidence Interval)	99.87 (-0.8/+0.1)	99.72 (-0.8/+0.2)	99.52 (-1.0/+0.3)	99.27 (-1.2/+0.5)	99.27 (-1.2/+0.5)	99.27 @ 65 mo. (-1.2/+0.5)	-	-	-	-
	Effective Sample Size	689	585	494	405	267	201	-	-	-	-
05-Apr-07 and 04-Mar-09 Shortened Replacement Window* Registered Implants: 400	Depletions and Malfunctions(%) (Confidence Interval)	99.84 (-0.1/+0.1)	98.53 (-0.3/+0.3)	93.45 (-0.7/+0.6)	78.72 (-1.1/+1.1)	46.21 (-1.5/+1.5)	18.91 (-1.2/+1.3)	14.41 (-1.1/+1.2)	13.55 (-1.1/+1.2)	-	-
	Malfunctions Only(%) (Confidence Interval)	99.84 (-0.1/+0.1)	98.53 (-0.3/+0.3)	93.45 (-0.7/+0.6)	78.72 (-1.1/+1.1)	46.21 (-1.5/+1.5)	18.91 (-1.2/+1.3)	14.41 (-1.1/+1.2)	13.55 (-1.1/+1.2)	-	-
	Effective Sample Size	400	400	400	400	400	400	400	400	400	400



9000	Malfunctions Only (%) (Confidence Interval)	99.89 (-0.1/+0.1)	99.07 (-0.3/+0.2)	96.98 (-0.5/+0.4)	92.57 (-0.8/+0.7)	88.46 (-1.0/+0.9)	85.14 (-1.5/+1.3)	84.05 (-1.7/+1.6)	83.11 (-1.9/+1.8)	--	--
	Effective Sample Size	7303	6340	5238	3822	1876	625	398	211	--	--
10-Mar-07 Product Update - Mid-life Display of Replacement Indicators*	Depletions and Malfunctions (%) (Confidence Interval)	99.84 (-0.1/+0.1)	99.47 (-0.2/+0.1)	96.81 (-0.4/+0.3)	87.31 (-0.7/+0.7)	58.66 (-1.2/+1.2)	24.17 (-1.1/+1.2)	17.77 (-1.0/+1.1)	16.14 (-1.0/+1.1)	15.87 (-1.0/+1.1)	15.87 @ 112 mo. (-1.0/+1.1)
Registered Implants: 14000	Malfunctions Only (%) (Confidence Interval)	99.90 (-0.1/+0.0)	99.78 (-0.1/+0.1)	99.04 (-0.2/+0.2)	97.57 (-0.4/+0.3)	96.38 (-0.5/+0.4)	94.25 (-0.9/+0.8)	92.76 (-1.2/+1.0)	91.76 (-1.4/+1.2)	91.76 (-1.4/+1.2)	91.76 @ 112 mo. (-1.4/+1.2)
	Effective Sample Size	11233	9719	8240	6373	3621	1178	721	570	358	203
12-May-06 Premature Battery Depletion*	Survival probability data not provided because this population does not meet report inclusion criteria (see Statistical Methodology for more details). Refer to Product Advisories for more information.										
23-Jun-05 Magnetic Switch*	Depletions and Malfunctions (%) (Confidence Interval)	99.83 (-0.1/+0.1)	99.38 (-0.2/+0.1)	96.73 (-0.4/+0.3)	87.63 (-0.7/+0.7)	59.40 (-1.1/+1.1)	24.80 (-1.1/+1.1)	18.09 (-1.0/+1.0)	16.33 (-1.0/+1.0)	16.00 (-0.9/+1.0)	16.00 @ 112 mo. (-0.9/+1.0)
Registered Implants: 15000	Malfunctions Only (%) (Confidence Interval)	99.88 (-0.1/+0.0)	99.65 (-0.1/+0.1)	98.89 (-0.2/+0.2)	97.44 (-0.3/+0.3)	96.09 (-0.5/+0.4)	93.89 (-0.8/+0.7)	92.15 (-1.1/+1.0)	91.04 (-1.4/+1.2)	91.04 (-1.4/+1.2)	91.04 @ 112 mo. (-1.4/+1.2)
	Effective Sample Size	12701	11005	9336	7261	4165	1393	841	662	355	201

\*Devices subject to an advisory. Refer to the Advisories for more details. Devices may be part of more than one advisory.

## CONTAK RENEWAL 3 HE

Models H177/H179

U.S. Survival Probability	Worldwide Malfunction Details	Product Advisories
---------------------------	-------------------------------	--------------------

CONTAK RENEWAL 3 HE Models H177/H179			
Worldwide Distribution: 23,000			
Worldwide Confirmed Malfunctions: 876			
	Without Compromised Therapy	With Compromised Therapy	Total
<b>Electrical</b>	<b>785</b>	<b>25</b>	<b>810</b>
<sup>7</sup> Shortened replacement window (Advisory issued)	290	12	
<sup>9</sup> Premature battery depletion (Advisory issued)	10	-	
<sup>21</sup> Extended charge time post-mid-life	98	-	
<sup>32</sup> Capacitor	2	-	
<sup>42</sup> Integrated circuit	3	2	
<sup>64</sup> Capacitor	10	4	
<sup>69</sup> Capacitor	7	-	
<sup>81</sup> Mid-life display of replacement indicators	165	-	
<sup>82</sup> High-voltage capacitor	-	1	
<sup>109</sup> Low-voltage capacitor	200	3	
<sup>111</sup> Resistor	-	3	
<b>Mechanical</b>	<b>26</b>	<b>11</b>	<b>37</b>
<sup>4</sup> Magnetic reed switch 2010 (Advisory issued)	-	2	
<sup>10</sup> Subpectoral implant (Advisory issued)	-	3	
<sup>15</sup> Magnetic switch (Advisory issued)	2	-	
<sup>33</sup> Header	3	1	
<sup>47</sup> Seal plug	14	3	
<sup>93</sup> Setscrew	7	2	
<b>Software</b>	<b>1</b>	<b>-</b>	<b>1</b>
<sup>106</sup> Misaligned markers	1	-	
<b>Other</b>	<b>22</b>	<b>6</b>	<b>28</b>
Non-patterned	15	5	
<sup>57</sup> Battery depletion	7	1	
<b>WW Confirmed Malfunctions</b>	<b>834</b>	<b>42</b>	<b>876</b>

[More details](#) about malfunctions

[References](#) cited in table above

**CONTAK RENEWAL 4**

Models H190/H195

<b>U.S. Survival Probability</b>	<b>Worldwide Malfunction Details</b>	<b>Product Advisories</b>
----------------------------------	--------------------------------------	---------------------------

**CONTAK RENEWAL 4  
Models H190/H195****Worldwide Distribution: 18,000****Worldwide Confirmed Malfunctions: 352**

	Without Compromised Therapy	With Compromised Therapy	Total
<b>Electrical</b>	<b>308</b>	<b>11</b>	<b>319</b>
<sup>7</sup> Shortened replacement window (Advisory issued)	159	5	
<sup>9</sup> Premature battery depletion (Advisory issued)	14	-	
<sup>21</sup> Extended charge time post-mid-life	9	-	
<sup>27</sup> Integrated circuit	2	-	
<sup>32</sup> Capacitor	-	1	
<sup>42</sup> Integrated circuit	2	3	
<sup>64</sup> Capacitor	-	1	
<sup>69</sup> Capacitor	3	-	
<sup>81</sup> Mid-life display of replacement indicators	63	-	
<sup>89</sup> Integrated circuit	-	1	
<sup>109</sup> Low-voltage capacitor	56	-	
<b>Mechanical</b>	<b>7</b>	<b>14</b>	<b>21</b>
<sup>4</sup> Magnetic reed switch 2010 (Advisory issued)	-	3	
<sup>10</sup> Subpectoral implant (Advisory issued)	-	7	
<sup>15</sup> Magnetic switch (Advisory issued)	-	1	
<sup>33</sup> Header	2	-	
<sup>47</sup> Seal plug	3	-	
<sup>66</sup> Circuit connection	-	1	
<sup>93</sup> Setscrew	-	1	
<sup>101</sup> Reed switch	1	1	
<sup>102</sup> Cracked solder joint	1	-	
<b>Software</b>	<b>-</b>	<b>-</b>	<b>0</b>
<b>Other</b>	<b>6</b>	<b>6</b>	<b>12</b>
Non-patterned	2	3	
<sup>57</sup> Battery depletion	4	3	
<b>WW Confirmed Malfunctions</b>	<b>321</b>	<b>31</b>	<b>352</b>

[More details](#) about malfunctions[References](#) cited in table above

**CONTAK RENEWAL 4 AVT**

Models M170/M175

U.S. Survival Probability	Worldwide Malfunction Details	Product Advisories
---------------------------	-------------------------------	--------------------

CONTAK RENEWAL 4 AVT Models M170/M175			
Worldwide Distribution: 2,000			
Worldwide Confirmed Malfunctions: 24			
	Without Compromised Therapy	With Compromised Therapy	Total
<b>Electrical</b>	<b>15</b>	<b>-</b>	<b>15</b>
<sup>7</sup> Shortened replacement window (Advisory issued)	8	-	
<sup>21</sup> Extended charge time post-mid-life	1	-	
<sup>32</sup> Capacitor	1	-	
<sup>42</sup> Integrated circuit	1	-	
<sup>69</sup> Capacitor	1	-	
<sup>81</sup> Mid-life display of replacement indicators	1	-	
<sup>109</sup> Low-voltage capacitor	2	-	
<b>Mechanical</b>	<b>2</b>	<b>-</b>	<b>2</b>
<sup>47</sup> Seal plug	1	-	
<sup>93</sup> Setscrew	1	-	
<b>Software</b>	<b>-</b>	<b>-</b>	<b>0</b>
<b>Other</b>	<b>6</b>	<b>1</b>	<b>7</b>
Non-patterned	2	-	
<sup>57</sup> Battery depletion	4	1	
<b>WW Confirmed Malfunctions</b>	<b>23</b>	<b>1</b>	<b>24</b>


[More details](#) about malfunctions

[References](#) cited in table above

**CONTAK RENEWAL 4 AVT HE**

Models M177/M179

<b>U.S. Survival Probability</b>	<b>Worldwide Malfunction Details</b>	<b>Product Advisories</b>
----------------------------------	--------------------------------------	---------------------------

<b>CONTAK RENEWAL 4 AVT HE</b> <b>Models M177/M179</b> 			
<b>Worldwide Distribution:</b> 1,000			
<b>Worldwide Confirmed Malfunctions:</b> 32			
	Without Compromised Therapy	With Compromised Therapy	Total
<b>Electrical</b>	<b>26</b>	<b>-</b>	<b>26</b>
<sup>7</sup> Shortened replacement window (Advisory issued)	17	-	
<sup>9</sup> Premature battery depletion (Advisory issued)	3	-	
<sup>81</sup> Mid-life display of replacement indicators	1	-	
<sup>109</sup> Low-voltage capacitor	5	-	
<b>Mechanical</b>	<b>-</b>	<b>1</b>	<b>1</b>
<sup>10</sup> Subpectoral implant (Advisory issued)	-	1	
<b>Software</b>	<b>3</b>	<b>-</b>	<b>3</b>
<sup>94</sup> Charge time limit	3	-	
<b>Other</b>	<b>2</b>	<b>-</b>	<b>2</b>
Non-patterned	-	-	
<sup>57</sup> Battery depletion	2	-	
<b>WW Confirmed Malfunctions</b>	<b>31</b>	<b>1</b>	<b>32</b>

[More details](#) about malfunctions

[References](#) cited in table above

**CONTAK RENEWAL 4 HE**

Models H197/H199

<b>U.S. Survival Probability</b>	<b>Worldwide Malfunction Details</b>	<b>Product Advisories</b>
----------------------------------	--------------------------------------	---------------------------

<b>CONTAK RENEWAL 4 HE</b> <b>Models H197/H199</b>			
<b>Worldwide Distribution: 7,000</b>			
<b>Worldwide Confirmed Malfunctions: 145</b>			
	Without Compromised Therapy	With Compromised Therapy	Total
<b>Electrical</b>	<b>129</b>	<b>2</b>	<b>131</b>
<sup>7</sup> Shortened replacement window (Advisory issued)	67	1	
<sup>9</sup> Premature battery depletion (Advisory issued)	2	-	
<sup>21</sup> Extended charge time post-mid-life	10	-	
<sup>32</sup> Capacitor	1	-	
<sup>42</sup> Integrated circuit	1	1	
<sup>64</sup> Capacitor	1	-	
<sup>81</sup> Mid-life display of replacement indicators	25	-	
<sup>82</sup> High-voltage capacitor	1	-	
<sup>109</sup> Low-voltage capacitor	21	-	
<b>Mechanical</b>	<b>6</b>	<b>4</b>	<b>10</b>
<sup>4</sup> Magnetic reed switch 2010 (Advisory issued)	-	1	
<sup>10</sup> Subpectoral implant (Advisory issued)	-	1	
<sup>33</sup> Header	1	1	
<sup>47</sup> Seal plug	2	-	
<sup>93</sup> Setscrew	1	1	
<sup>95</sup> Seal plug	1	-	
<sup>102</sup> Cracked solder joint	1	-	
<b>Software</b>	<b>-</b>	<b>-</b>	<b>0</b>
<b>Other</b>	<b>3</b>	<b>1</b>	<b>4</b>
Non-patterned	1	1	
<sup>57</sup> Battery depletion	2	-	
<b>WW Confirmed Malfunctions</b>	<b>138</b>	<b>7</b>	<b>145</b>


[More details](#) about malfunctions

[References](#) cited in table above

**CONTAK RENEWAL 4 RF**

Models H230/H235

<b>U.S. Survival Probability</b>	<b>Worldwide Malfunction Details</b>	<b>Product Advisories</b>
----------------------------------	--------------------------------------	---------------------------

<b>CONTAK RENEWAL 4 RF</b> <b>Models H230/H235</b> 			
<b>Worldwide Distribution: 8,000</b>			
<b>Worldwide Confirmed Malfunctions: 25</b>			
	Without Compromised Therapy	With Compromised Therapy	Total
<b>Electrical</b>	<b>14</b>	<b>3</b>	<b>17</b>
<sup>7</sup> Shortened replacement window (Advisory issued)	8	1	
<sup>21</sup> Extended charge time post-mid-life	1	-	
<sup>42</sup> Integrated circuit	1	2	
<sup>69</sup> Capacitor	1	-	
<sup>81</sup> Mid-life display of replacement indicators	1	-	
<sup>109</sup> Low-voltage capacitor	2	-	
<b>Mechanical</b>	<b>-</b>	<b>3</b>	<b>3</b>
<sup>4</sup> Magnetic reed switch 2010 (Advisory issued)	-	2	
<sup>33</sup> Header	-	1	
<b>Software</b>	<b>-</b>	<b>-</b>	<b>0</b>
<b>Other</b>	<b>2</b>	<b>3</b>	<b>5</b>
Non-patterned	1	-	
<sup>57</sup> Battery depletion	1	3	
<b>WW Confirmed Malfunctions</b>	<b>16</b>	<b>9</b>	<b>25</b>


[More details](#) about malfunctions

[References](#) cited in table above

**CONTAK RENEWAL 4 RF HE**

Model H239

<b>U.S. Survival Probability</b>	<b>Worldwide Malfunction Details</b>	<b>Product Advisories</b>
----------------------------------	--------------------------------------	---------------------------

<b>CONTAK RENEWAL 4 RF HE</b> <b>Model H239</b> 			
<b>Worldwide Distribution:</b> 1,000 <b>Worldwide Confirmed Malfunctions:</b> 5			
	Without Compromised Therapy	With Compromised Therapy	Total
<b>Electrical</b>	<b>5</b>	<b>-</b>	<b>5</b>
<sup>7</sup> Shortened replacement window (Advisory issued)	2	-	
<sup>42</sup> Integrated circuit	2	-	
<sup>109</sup> Low-voltage capacitor	1	-	
<b>Mechanical</b>	<b>-</b>	<b>-</b>	<b>0</b>
<b>Software</b>	<b>-</b>	<b>-</b>	<b>0</b>
<b>Other</b>	<b>-</b>	<b>-</b>	<b>0</b>
Non-patterned	-	-	
<b>WW Confirmed Malfunctions</b>	<b>5</b>	<b>0</b>	<b>5</b>

[More details](#) about malfunctions

[References](#) cited in table above



CRM PRODUCT PERFORMANCE REPORT Q1 2014

INVIVE


Models V172/V173/V182/V183/W172/  
W173

U.S. Survival Probability	Worldwide Malfunction Details	Product Advisories
---------------------------	-------------------------------	--------------------


U.S. Summary

U.S. Registered Implants: 4,000	U.S. Normal Battery Depletions: 0
U.S. Approval Date: May 2012	U.S. Unconfirmed Reports of Premature Battery Depletion : 0
U.S. Estimated Active Implants: 4,000	U.S. Malfunctions:0
	Without Compromised Therapy:0
	With Compromised Therapy:0

Battery Depletions and Malfunctions



Malfunctions Only




U.S. Survival Probability

	Year	1	2	3	4	5	6	7	8	9	10
Non Advisory Population	Depletions and Malfunctions(%) (Confidence Interval)	100.00 (-0.0/+0.0)	100.00 @ 18 mo. (-0.0/+0.0)	-	-	-	-	-	-	-	-
Registered Implants: 4000	Malfunctions Only(%) (Confidence Interval)	100.00 (-0.0/+0.0)	100.00 @ 18 mo. (-0.0/+0.0)	-	-	-	-	-	-	-	-
	Effective Sample Size	1209	215	-	-	-	-	-	-	-	-

**INVIVE**

Models V172/V173/V182/V183/W172/  
W173

U.S. Survival Probability	Worldwide Malfunction Details	Product Advisories
---------------------------	-------------------------------	--------------------

**INVIVE**  
Models V172/V173/V182/V183/W172/  
W173 

**Worldwide Distribution:** 10,000  
**Worldwide Confirmed Malfunctions:** 0

	Without Compromised Therapy	With Compromised Therapy	Total
<b>Electrical</b>	-	-	<b>0</b>
<b>Mechanical</b>	-	-	<b>0</b>
<b>Software</b>	-	-	<b>0</b>
<b>Other</b>	-	-	<b>0</b>
Non-patterned	-	-	
<b>WW Confirmed Malfunctions</b>	<b>0</b>	<b>0</b>	<b>0</b>


[More details](#) about malfunctions

[References](#) cited in table above

**CONTAK RENEWAL TR 2**

Models H140/H145

<b>U.S. Survival Probability</b>	<b>Worldwide Malfunction Details</b>	<b>Product Advisories</b>
----------------------------------	--------------------------------------	---------------------------

<b>CONTAK RENEWAL TR 2</b> <b>Models H140/H145</b> 			
<b>Worldwide Distribution:</b> 31,000			
<b>Worldwide Confirmed Malfunctions:</b> 27			
	Without Compromised Therapy	With Compromised Therapy	Total
<b>Electrical</b>	<b>1</b>	<b>-</b>	<b>1</b>
<sup>32</sup> Capacitor	1	-	
<b>Mechanical</b>	<b>4</b>	<b>-</b>	<b>4</b>
<sup>47</sup> Seal plug	1	-	
<sup>75</sup> Setscrew block	2	-	
<sup>95</sup> Seal plug	1	-	
<b>Software</b>	<b>12</b>	<b>-</b>	<b>12</b>
<sup>60</sup> Memory error	1	-	
<sup>79</sup> Stored EGMs	11	-	
<b>Other</b>	<b>9</b>	<b>1</b>	<b>10</b>
Non-patterned	8	1	
<sup>91</sup> Alert messages	1	-	
<b>WW Confirmed Malfunctions</b>	<b>26</b>	<b>1</b>	<b>27</b>

[More details](#) about malfunctions

[References](#) cited in table above

CRM PRODUCT PERFORMANCE REPORT Q1 2014

CONTAK RENEWAL TR


Models H120/H125

U.S. Survival Probability	Worldwide Malfunction Details	Product Advisories
---------------------------	-------------------------------	--------------------


**U.S. Summary**

U.S. Registered Implants: 19,000	U.S. Normal Battery Depletions: 1,398
U.S. Approval Date: January 2004	U.S. Unconfirmed Reports of Premature Battery Depletion : 10
U.S. Estimated Active Implants: 9,000	U.S. Malfunctions:42
	Without Compromised Therapy:40
	With Compromised Therapy:2

**Battery Depletions and Malfunctions**



**Malfunctions Only**



**U.S. Survival Probability**


	Year	1	2	3	4	5	6	7	8	9	10
Non Advisory Population	Depletions and Malfunctions(%)	99.91	99.47	98.46	96.27	91.31	82.56	68.13	48.65	37.77	—
	(Confidence Interval)	(-0.1/+0.0)	(-0.1/+0.1)	(-0.2/+0.2)	(-0.4/+0.4)	(-0.8/+0.7)	(-1.2/+1.2)	(-1.9/+1.9)	(-2.7/+2.7)	@ 102 mo. (-3.1/+3.2)	
Registered Implants: 19000	Malfunctions Only(%)	99.97	99.94	99.87	99.83	99.68	99.51	99.33	98.83	98.08	—
	(Confidence Interval)	(-0.0/+0.0)	(-0.1/+0.0)	(-0.1/+0.1)	(-0.1/+0.1)	(-0.2/+0.1)	(-0.3/+0.2)	(-0.4/+0.2)	(-0.8/+0.5)	@ 102 mo. (-1.7/+0.9)	
	Effective Sample Size	15593	12713	9100	5964	3635	2043	984	387	211	—
23-Jun-06 and 24-Aug-06 Low Voltage Capacitor*	Survival probability data not provided because this population does not meet report inclusion criteria (see Statistical Methodology for more details). Refer to Product Advisories for more information.										

\*Devices subject to an advisory. Refer to the Advisories for more details. Devices may be part of more than one advisory.

## CONTAK RENEWAL TR

Models H120/H125

U.S. Survival Probability	Worldwide Malfunction Details	Product Advisories
---------------------------	-------------------------------	--------------------

<b>CONTAK RENEWAL TR</b> <b>Models H120/H125</b> 			
<b>Worldwide Distribution:</b> 19,000 <b>Worldwide Confirmed Malfunctions:</b> 42			
	Without Compromised Therapy	With Compromised Therapy	Total
<b>Electrical</b>	<b>1</b>	<b>1</b>	<b>2</b>
<sup>8</sup> Low-voltage capacitor (Advisory issued)	1	-	
<sup>32</sup> Capacitor	-	1	
<b>Mechanical</b>	<b>5</b>	<b>-</b>	<b>5</b>
<sup>47</sup> Seal plug	5	-	
<b>Software</b>	<b>26</b>	<b>-</b>	<b>26</b>
<sup>79</sup> Stored EGMs	26	-	
<b>Other</b>	<b>8</b>	<b>1</b>	<b>9</b>
Non-patterned	7	1	
<sup>91</sup> Alert messages	1	-	
<b>WW Confirmed Malfunctions</b>	<b>40</b>	<b>2</b>	<b>42</b>

[More details](#) about malfunctions


[References](#) cited in table above

**INCEPTA ICD DR 4-Site**

Models E162/F162

U.S. Survival Probability	Worldwide Malfunction Details	Product Advisories
---------------------------	-------------------------------	--------------------

<b>U.S. Summary</b>	
U.S. Registered Implants: 6,000	U.S. Normal Battery Depletions: 1
U.S. Approval Date: November 2011	U.S. Unconfirmed Reports of Premature Battery Depletion : 0
U.S. Estimated Active Implants: 6,000	U.S. Malfunctions: 2
	Without Compromised Therapy: 2
	With Compromised Therapy: 0




U.S. Survival Probability		Year	1	2	3	4	5	6	7	8	9	10
Non Advisory Population	Depletions and Malfunctions (%) (Confidence Interval)	99.95 (-0.1/+0.0)	99.91 @ 23 mo. (-0.2/+0.1)	-	-	-	-	-	-	-	-	-
	Malfunctions Only (%) (Confidence Interval)	99.95 (-0.1/+0.0)	99.95 @ 23 mo. (-0.1/+0.0)	-	-	-	-	-	-	-	-	-
Registered Implants: 6000	Effective Sample Size	2653	215	-	-	-	-	-	-	-	-	-

**INCEPTA ICD DR 4-Site**

Models E162/F162

<b>U.S. Survival Probability</b>	<b>Worldwide Malfunction Details</b>	<b>Product Advisories</b>
----------------------------------	--------------------------------------	---------------------------

**INCEPTA ICD DR 4-Site**  
**Models E162/F162**



**Worldwide Distribution: 11,000**  
**Worldwide Confirmed Malfunctions: 3**

	Without Compromised Therapy	With Compromised Therapy	Total
<b>Electrical</b>	<b>1</b>	<b>-</b>	<b>1</b>
<sup>121</sup> Integrated circuit	1	-	
<b>Mechanical</b>	<b>-</b>	<b>1</b>	<b>1</b>
<sup>103</sup> Transformer	-	1	
<b>Software</b>	<b>1</b>	<b>-</b>	<b>1</b>
<sup>122</sup> Memory errors	1	-	
<b>Other</b>	<b>-</b>	<b>-</b>	<b>0</b>
Non-patterned	-	-	
<b>WW Confirmed Malfunctions</b>	<b>2</b>	<b>1</b>	<b>3</b>

[More details](#) about malfunctions


[References](#) cited in table above

**INCEPTA ICD DR**

Models E163/F163

<b>U.S. Survival Probability</b>	<b>Worldwide Malfunction Details</b>	<b>Product Advisories</b>
----------------------------------	--------------------------------------	---------------------------

<b>U.S. Summary</b>	
U.S. Registered Implants: 4,000	U.S. Normal Battery Depletions: 0
U.S. Approval Date: November 2011	U.S. Unconfirmed Reports of Premature Battery Depletion : 0
U.S. Estimated Active Implants: 4,000	U.S. Malfunctions:1
	Without Compromised Therapy:1
	With Compromised Therapy:0



<b>U.S. Survival Probability</b>		Year	1	2	3	4	5	6	7	8	9	10
Non Advisory Population	Depletions and Malfunctions(%) (Confidence Interval)	99.96 (-0.2/+0.0)	99.96 @ 21 mo. (-0.2/+0.0)	-	-	-	-	-	-	-	-	-
	Malfunctions Only(%) (Confidence Interval)	99.96 (-0.2/+0.0)	99.96 @ 21 mo. (-0.2/+0.0)	-	-	-	-	-	-	-	-	-
Registered Implants: 4000	Effective Sample Size	1482	252	-	-	-	-	-	-	-	-	-




**INCEPTA ICD DR**

Models E163/F163

<b>U.S. Survival Probability</b>	<b>Worldwide Malfunction Details</b>	<b>Product Advisories</b>
----------------------------------	--------------------------------------	---------------------------

**INCEPTA ICD DR**  
**Models E163/F163**



**Worldwide Distribution: 6,000**  
**Worldwide Confirmed Malfunctions: 2**

	Without Compromised Therapy	With Compromised Therapy	Total
<b>Electrical</b>	<b>1</b>	<b>-</b>	<b>1</b>
<sup>117</sup> Low-voltage capacitors	1	-	
<b>Mechanical</b>	<b>-</b>	<b>-</b>	<b>0</b>
<b>Software</b>	<b>-</b>	<b>-</b>	<b>0</b>
<b>Other</b>	<b>1</b>	<b>-</b>	<b>1</b>
Non-patterned	1	-	
<b>WW Confirmed Malfunctions</b>	<b>2</b>	<b>0</b>	<b>2</b>

[More details](#) about malfunctions

[References](#) cited in table above

## INCEPTA ICD VR 4-Site


Models E160/F160

U.S. Survival Probability	Worldwide Malfunction Details	Product Advisories
---------------------------	-------------------------------	--------------------


### U.S. Summary

U.S. Registered Implants: 5,000	U.S. Normal Battery Depletions: 3
U.S. Approval Date: November 2011	U.S. Unconfirmed Reports of Premature Battery Depletion : 0
U.S. Estimated Active Implants: 5,000	U.S. Malfunctions: 2
	Without Compromised Therapy: 1
	With Compromised Therapy: 1

### Battery Depletions and Malfunctions



### Malfunctions Only



### U.S. Survival Probability


	Year	1	2	3	4	5	6	7	8	9	10
Non Advisory Population	Depletions and Malfunctions (%) (Confidence Interval)	99.82 (-0.3/+0.1)	99.82 @ 22 mo. (-0.3/+0.1)	-	-	-	-	-	-	-	-
Registered Implants: 5000	Malfunctions Only (%) (Confidence Interval)	99.93 (-0.2/+0.1)	99.93 @ 22 mo. (-0.2/+0.1)	-	-	-	-	-	-	-	-
	Effective Sample Size	2182	304	-	-	-	-	-	-	-	-

**INCEPTA ICD VR 4-Site**

Models E160/F160

<b>U.S. Survival Probability</b>	<b>Worldwide Malfunction Details</b>	<b>Product Advisories</b>
----------------------------------	--------------------------------------	---------------------------

**INCEPTA ICD VR 4-Site**  
**Models E160/F160**



**Worldwide Distribution: 9,000**  
**Worldwide Confirmed Malfunctions: 2**

	Without Compromised Therapy	With Compromised Therapy	Total
<b>Electrical</b>	-	-	<b>0</b>
<b>Mechanical</b>	-	<b>1</b>	<b>1</b>
<sup>103</sup> Transformer	-	1	
<b>Software</b>	<b>1</b>	-	<b>1</b>
<sup>122</sup> Memory errors	1	-	
<b>Other</b>	-	-	<b>0</b>
Non-patterned	-	-	
<b>WW Confirmed Malfunctions</b>	<b>1</b>	<b>1</b>	<b>2</b>

[More details](#) about malfunctions


[References](#) cited in table above

**INCEPTA ICD VR**

Models E161/F161

U.S. Survival Probability	Worldwide Malfunction Details	Product Advisories
---------------------------	-------------------------------	--------------------

<b>U.S. Summary</b>	
U.S. Registered Implants: 2,000	U.S. Normal Battery Depletions: 0
U.S. Approval Date: November 2011	U.S. Unconfirmed Reports of Premature Battery Depletion : 0
U.S. Estimated Active Implants: 2,000	U.S. Malfunctions:1
	Without Compromised Therapy:0
	With Compromised Therapy:1




U.S. Survival Probability		Year	1	2	3	4	5	6	7	8	9	10
Non Advisory Population	Depletions and Malfunctions(%) (Confidence Interval)	99.96 (-0.3/+0.0)	99.96 @ 20 mo. (-0.3/+0.0)	-	-	-	-	-	-	-	-	-
	Malfunctions Only(%) (Confidence Interval)	99.96 (-0.3/+0.0)	99.96 @ 20 mo. (-0.3/+0.0)	-	-	-	-	-	-	-	-	-
Registered Implants: 2000	Effective Sample Size	979	269	-	-	-	-	-	-	-	-	-

**INCEPTA ICD VR**

Models E161/F161

<b>U.S. Survival Probability</b>	<b>Worldwide Malfunction Details</b>	<b>Product Advisories</b>
----------------------------------	--------------------------------------	---------------------------

**INCEPTA ICD VR**  
**Models E161/F161**



**Worldwide Distribution: 4,000**  
**Worldwide Confirmed Malfunctions: 1**

	Without Compromised Therapy	With Compromised Therapy	Total
<b>Electrical</b>	-	<b>1</b>	<b>1</b>
<sup>112</sup> High-voltage capacitor	-	1	
<b>Mechanical</b>	-	-	<b>0</b>
<b>Software</b>	-	-	<b>0</b>
<b>Other</b>	-	-	<b>0</b>
Non-patterned	-	-	
<b>WW Confirmed Malfunctions</b>	<b>0</b>	<b>1</b>	<b>1</b>

[More details](#) about malfunctions


[References](#) cited in table above

## ENERGEN ICD DR 4-Site

Models E142/F142

U.S. Survival Probability	Worldwide Malfunction Details	Product Advisories
---------------------------	-------------------------------	--------------------

<b>U.S. Summary</b>	
U.S. Registered Implants: 10,000	U.S. Normal Battery Depletions: 1
U.S. Approval Date: November 2011	U.S. Unconfirmed Reports of Premature Battery Depletion : 0
U.S. Estimated Active Implants: 9,000	U.S. Malfunctions:4
	Without Compromised Therapy:2
	With Compromised Therapy:2




U.S. Survival Probability		Year	1	2	3	4	5	6	7	8	9	10
Registered Implants: 10000	Non Advisory Population	Depletions and Malfunctions(%) (Confidence Interval)	99.92 (-0.1/+0.0)	99.92 @ 23 mo. (-0.1/+0.0)	-	-	-	-	-	-	-	-
		Malfunctions Only(%) (Confidence Interval)	99.93 (-0.1/+0.0)	99.93 @ 23 mo. (-0.1/+0.0)	-	-	-	-	-	-	-	-
		Effective Sample Size	4211	254	-	-	-	-	-	-	-	-

**ENERGEN ICD DR 4-Site**

Models E142/F142

<b>U.S. Survival Probability</b>	<b>Worldwide Malfunction Details</b>	<b>Product Advisories</b>
----------------------------------	--------------------------------------	---------------------------

**ENERGEN ICD DR 4-Site**  
Models E142/F142 

**Worldwide Distribution:** 13,000  
**Worldwide Confirmed Malfunctions:** 5

	Without Compromised Therapy	With Compromised Therapy	Total
<b>Electrical</b>	<b>2</b>	<b>2</b>	<b>4</b>
<sup>117</sup> Low-voltage capacitors	1	-	
<sup>121</sup> Integrated circuit	1	2	
<b>Mechanical</b>	-	-	<b>0</b>
<b>Software</b>	-	-	<b>0</b>
<b>Other</b>	<b>1</b>	-	<b>1</b>
Non-patterned	1	-	
<b>WW Confirmed Malfunctions</b>	<b>3</b>	<b>2</b>	<b>5</b>

[More details](#) about malfunctions


[References](#) cited in table above

## ENERGEN ICD DR

Models E143/F143

U.S. Survival Probability	Worldwide Malfunction Details	Product Advisories
---------------------------	-------------------------------	--------------------

<b>U.S. Summary</b>	
U.S. Registered Implants: 7,000	U.S. Normal Battery Depletions: 0
U.S. Approval Date: November 2011	U.S. Unconfirmed Reports of Premature Battery Depletion : 0
U.S. Estimated Active Implants: 6,000	U.S. Malfunctions: 1
	Without Compromised Therapy: 1
	With Compromised Therapy: 0



U.S. Survival Probability		Year	1	2	3	4	5	6	7	8	9	10
Non Advisory Population	Depletions and Malfunctions (%) (Confidence Interval)	99.98 (-0.1/+0.0)	99.98 @ 22 mo. (-0.1/+0.0)	-	-	-	-	-	-	-	-	-
	Malfunctions Only (%) (Confidence Interval)	99.98 (-0.1/+0.0)	99.98 @ 22 mo. (-0.1/+0.0)	-	-	-	-	-	-	-	-	-
Registered Implants: 7000	Effective Sample Size	2818	239	-	-	-	-	-	-	-	-	-




**ENERGEN ICD DR**

Models E143/F143

<b>U.S. Survival Probability</b>	<b>Worldwide Malfunction Details</b>	<b>Product Advisories</b>
----------------------------------	--------------------------------------	---------------------------

**ENERGEN ICD DR**  
Models E143/F143



**Worldwide Distribution: 9,000**  
**Worldwide Confirmed Malfunctions: 1**

	Without Compromised Therapy	With Compromised Therapy	Total
<b>Electrical</b>	<b>1</b>	<b>-</b>	<b>1</b>
<sup>117</sup> Low-voltage capacitors	1	-	
<b>Mechanical</b>	<b>-</b>	<b>-</b>	<b>0</b>
<b>Software</b>	<b>-</b>	<b>-</b>	<b>0</b>
<b>Other</b>	<b>-</b>	<b>-</b>	<b>0</b>
Non-patterned	-	-	
<b>WW Confirmed Malfunctions</b>	<b>1</b>	<b>0</b>	<b>1</b>

[More details](#) about malfunctions


[References](#) cited in table above

## ENERGEN ICD VR 4-Site

Models E140/F140

U.S. Survival Probability	Worldwide Malfunction Details	Product Advisories
---------------------------	-------------------------------	--------------------

<b>U.S. Summary</b>	
U.S. Registered Implants: 9,000	U.S. Normal Battery Depletions: 3
U.S. Approval Date: November 2011	U.S. Unconfirmed Reports of Premature Battery Depletion : 0
U.S. Estimated Active Implants: 9,000	U.S. Malfunctions:2
	Without Compromised Therapy:0
	With Compromised Therapy:2




U.S. Survival Probability		Year	1	2	3	4	5	6	7	8	9	10
Non Advisory Population	Depletions and Malfunctions (%) (Confidence Interval)	99.92 (-0.1/+0.0)	99.92 @ 23 mo. (-0.1/+0.0)	-	-	-	-	-	-	-	-	-
	Malfunctions Only (%) (Confidence Interval)	99.96 (-0.1/+0.0)	99.96 @ 23 mo. (-0.1/+0.0)	-	-	-	-	-	-	-	-	-
Registered Implants: 9000	Effective Sample Size	3814	229	-	-	-	-	-	-	-	-	-

**ENERGEN ICD VR 4-Site**

Models E140/F140

<b>U.S. Survival Probability</b>	<b>Worldwide Malfunction Details</b>	<b>Product Advisories</b>
----------------------------------	--------------------------------------	---------------------------

**ENERGEN ICD VR 4-Site**  
Models E140/F140 

**Worldwide Distribution:** 13,000  
**Worldwide Confirmed Malfunctions:** 3

	Without Compromised Therapy	With Compromised Therapy	Total
<b>Electrical</b>	-	-	<b>0</b>
<b>Mechanical</b>	-	<b>1</b>	<b>1</b>
<sup>103</sup> Transformer	-	1	
<b>Software</b>	<b>1</b>	-	<b>1</b>
<sup>122</sup> Memory errors	1	-	
<b>Other</b>	-	<b>1</b>	<b>1</b>
Non-patterned	-	1	
<b>WW Confirmed Malfunctions</b>	<b>1</b>	<b>2</b>	<b>3</b>

[More details](#) about malfunctions


[References](#) cited in table above

**ENERGEN ICD VR**

Models E141/F141

U.S. Survival Probability	Worldwide Malfunction Details	Product Advisories
---------------------------	-------------------------------	--------------------

<b>U.S. Summary</b>	
U.S. Registered Implants: 5,000	U.S. Normal Battery Depletions: 2
U.S. Approval Date: November 2011	U.S. Unconfirmed Reports of Premature Battery Depletion : 0
U.S. Estimated Active Implants: 5,000	U.S. Malfunctions: 1
	Without Compromised Therapy: 1
	With Compromised Therapy: 0




U.S. Survival Probability		Year	1	2	3	4	5	6	7	8	9	10
Non Advisory Population	Depletions and Malfunctions (%) (Confidence Interval)	99.89 (-0.2/+0.1)	99.89 @ 22 mo. (-0.2/+0.1)	-	-	-	-	-	-	-	-	-
	Malfunctions Only (%) (Confidence Interval)	99.97 (-0.2/+0.0)	99.97 @ 22 mo. (-0.2/+0.0)	-	-	-	-	-	-	-	-	-
Registered Implants: 5000	Effective Sample Size	2133	227	-	-	-	-	-	-	-	-	-

## ENERGEN ICD VR

Models E141/F141

<b>U.S. Survival Probability</b>	<b>Worldwide Malfunction Details</b>	<b>Product Advisories</b>
----------------------------------	--------------------------------------	---------------------------

<b>ENERGEN ICD VR</b> <b>Models E141/F141</b> 			
<b>Worldwide Distribution: 7,000</b>			
<b>Worldwide Confirmed Malfunctions: 4</b>			
	Without Compromised Therapy	With Compromised Therapy	Total
<b>Electrical</b>	-	<b>2</b>	<b>2</b>
<sup>121</sup> Integrated circuit	-	2	
<b>Mechanical</b>	-	-	<b>0</b>
<b>Software</b>	<b>1</b>	-	<b>1</b>
<sup>122</sup> Memory errors	1	-	
<b>Other</b>	<b>1</b>	-	<b>1</b>
Non-patterned	1	-	
<b>WW Confirmed Malfunctions</b>	<b>2</b>	<b>2</b>	<b>4</b>

[More details](#) about malfunctions


[References](#) cited in table above

**PUNCTUA ICD DR 4-Site**

Models E052/F052

<b>U.S. Survival Probability</b>	<b>Worldwide Malfunction Details</b>	<b>Product Advisories</b>
----------------------------------	--------------------------------------	---------------------------

**PUNCTUA ICD DR 4-Site**  
**Models E052/F052**



**Worldwide Distribution: 1,000**  
**Worldwide Confirmed Malfunctions: 0**

	Without Compromised Therapy	With Compromised Therapy	Total
<b>Electrical</b>	-	-	<b>0</b>
<b>Mechanical</b>	-	-	<b>0</b>
<b>Software</b>	-	-	<b>0</b>
<b>Other</b>	-	-	<b>0</b>
Non-patterned	-	-	
<b>WW Confirmed Malfunctions</b>	<b>0</b>	<b>0</b>	<b>0</b>

[More details](#) about malfunctions


[References](#) cited in table above

**PUNCTUA ICD DR**

Models E053/F053

<b>U.S. Survival Probability</b>	<b>Worldwide Malfunction Details</b>	<b>Product Advisories</b>
----------------------------------	--------------------------------------	---------------------------

**PUNCTUA ICD DR**  
**Models E053/F053**



**Worldwide Distribution: 3,000**  
**Worldwide Confirmed Malfunctions: 0**

	Without Compromised Therapy	With Compromised Therapy	Total
<b>Electrical</b>	-	-	<b>0</b>
<b>Mechanical</b>	-	-	<b>0</b>
<b>Software</b>	-	-	<b>0</b>
<b>Other</b>	-	-	<b>0</b>
Non-patterned	-	-	
<b>WW Confirmed Malfunctions</b>	<b>0</b>	<b>0</b>	<b>0</b>

[More details](#) about malfunctions


[References](#) cited in table above

**PUNCTUA ICD VR 4-Site**

Models E050/F050

<b>U.S. Survival Probability</b>	<b>Worldwide Malfunction Details</b>	<b>Product Advisories</b>
----------------------------------	--------------------------------------	---------------------------

**PUNCTUA ICD VR 4-Site**  
**Models E050/F050**



**Worldwide Distribution: 2,000**  
**Worldwide Confirmed Malfunctions: 1**

	Without Compromised Therapy	With Compromised Therapy	Total
<b>Electrical</b>	-	-	<b>0</b>
<b>Mechanical</b>	-	-	<b>0</b>
<b>Software</b>	-	-	<b>0</b>
<b>Other</b>	-	<b>1</b>	<b>1</b>
Non-patterned	-	1	
<b>WW Confirmed Malfunctions</b>	<b>0</b>	<b>1</b>	<b>1</b>

[More details](#) about malfunctions

[References](#) cited in table above




**PUNCTUA ICD VR**

Models E051/F051

<b>U.S. Survival Probability</b>	<b>Worldwide Malfunction Details</b>	<b>Product Advisories</b>
----------------------------------	--------------------------------------	---------------------------

**PUNCTUA ICD VR**  
Models E051/F051



**Worldwide Distribution: 4,000**  
**Worldwide Confirmed Malfunctions: 1**

	Without Compromised Therapy	With Compromised Therapy	Total
<b>Electrical</b>	-	-	<b>0</b>
<b>Mechanical</b>	-	-	<b>0</b>
<b>Software</b>	-	-	<b>0</b>
<b>Other</b>	<b>1</b>	-	<b>1</b>
Non-patterned	1	-	
<b>WW Confirmed Malfunctions</b>	<b>1</b>	<b>0</b>	<b>1</b>

[More details](#) about malfunctions

[References](#) cited in table above

## SQ-RX Pulse Generator


Model 1010

U.S. Survival Probability	Worldwide Malfunction Details	Product Advisories
---------------------------	-------------------------------	--------------------


### U.S. Summary

U.S. Approval Date: September 2012	U.S. Normal Battery Depletions: 0 U.S. Unconfirmed Reports of Premature Battery Depletion : 1 U.S. Malfunctions:4 Without Compromised Therapy:1 With Compromised Therapy:3
------------------------------------	--

### Battery Depletions and Malfunctions



### Malfunctions Only



### U.S. Survival Probability

	Year	1	2	3	4	5	6	7	8	9	10
Total Population	Depletions and Malfunctions (%)	99.61	99.13	-	-	-	-	-	-	-	-
	(Confidence Interval)	(-0.9/+0.3)	(-2.0/+0.6)								
	Malfunctions Only (%)	99.61	99.13	-	-	-	-	-	-	-	-
	(Confidence Interval)	(-0.9/+0.3)	(-2.0/+0.6)								
1-Mar-13 Unintended Fuse Activation*	Survival probability data not provided because this population does not meet report inclusion criteria (see Statistical Methodology for more details). Refer to Product Advisories for more information.										
1-Jun-11 High Cathode Condition *	Survival probability data not provided because this population does not meet report inclusion criteria (see Statistical Methodology for more details). Refer to Product Advisories for more information.										


\*Devices subject to an advisory. Refer to the Advisories for more details. Devices may be part of more than one advisory.

This product family may not conform to all AdvaMed/ISO recommended performance reporting data elements.

## SQ-RX Pulse Generator

Model 1010

<b>U.S. Survival Probability</b>	<b>Worldwide Malfunction Details</b>	<b>Product Advisories</b>
----------------------------------	--------------------------------------	---------------------------

<b>SQ-RX Pulse Generator</b> <b>Model 1010</b> 			
<b>Worldwide Confirmed Malfunctions: 32</b>			
	Without Compromised Therapy	With Compromised Therapy	Total
<b>Electrical</b>	-	<b>3</b>	<b>3</b>
<sup>2</sup> Unintended fuse activation 2013 (Advisory issued)	-	3	
<b>Mechanical</b>	<b>8</b>	<b>7</b>	<b>15</b>
<sup>3</sup> High cathode condition 2011 (Advisory issued)	1	2	
<sup>126</sup> Battery depletion	7	5	
<b>Software</b>	<b>2</b>	-	<b>2</b>
<sup>128</sup> Unintended Battery Depletion Alert	2	-	
<b>Other</b>	<b>7</b>	<b>5</b>	<b>12</b>
Non-patterned	6	4	
<sup>127</sup> Telemetry	1	1	
<b>WW Confirmed Malfunctions</b>	<b>17</b>	<b>15</b>	<b>32</b>

[More details](#) about malfunctions


[References](#) cited in table above

TELIGEN DR

Models E110/E111/F110/F111

U.S. Survival Probability	Worldwide Malfunction Details	Product Advisories
---------------------------	-------------------------------	--------------------

<b>U.S. Summary</b>	
U.S. Registered Implants: 66,000	U.S. Normal Battery Depletions: 78
U.S. Approval Date: May 2008	U.S. Unconfirmed Reports of Premature Battery Depletion : 16
U.S. Estimated Active Implants: 50,000	U.S. Malfunctions: 316
	Without Compromised Therapy: 250
	With Compromised Therapy: 66



U.S. Survival Probability		Year	1	2	3	4	5	6	7	8	9	10
Non Advisory Population Registered Implants: 33000	Depletions and Malfunctions (%) (Confidence Interval)	99.92 (-0.0/+0.0)	99.86 (-0.0/+0.0)	99.78 (-0.1/+0.1)	99.19 @ 47 mo. (-1.5/+0.5)	-	-	-	-	-	-	-
	Malfunctions Only (%) (Confidence Interval)	99.95 (-0.0/+0.0)	99.93 (-0.0/+0.0)	99.88 (-0.1/+0.0)	99.75 @ 47 mo. (-0.4/+0.2)	-	-	-	-	-	-	-
	Effective Sample Size	29279	24355	10599	233	-	-	-	-	-	-	-
Advisory: 01-Dec-09 (Subpectoral Implant 2009)* Registered Implants: 30,000	Depletions and Malfunctions (%) (Confidence Interval)	99.88 (-0.1/+0.0)	99.77 (-0.1/+0.1)	99.61 (-0.1/+0.1)	98.89 (-0.2/+0.1)	97.85 (-0.3/+0.2)	97.71 @ 65 (-0.3/+0.3)	-	-	-	-	-
	Malfunctions Only (%) (Confidence Interval)	99.89 (-0.1/+0.0)	99.82 (-0.1/+0.0)	99.72 (-0.1/+0.1)	99.11 (-0.1/+0.1)	98.22 (-0.3/+0.2)	98.11 @ 65 (-0.3/+0.3)	-	-	-	-	-
	Effective Sample Size	26752	23507	20679	17524	4200	508	-	-	-	-	-
Advisory: 29-Aug-13 (Low Voltage Capacitor 2013)* Registered Implants: 11,000	Depletions and Malfunctions (%) (Confidence Interval)	99.90 (-0.1/+0.0)	99.77 (-0.1/+0.1)	99.56 (-0.2/+0.1)	98.12 (-0.4/+0.3)	95.69 (-1.0/+0.8)	95.22 @ 65 (-1.0/+0.8)	-	-	-	-	-
	Malfunctions Only (%) (Confidence Interval)	99.90 (-0.1/+0.0)	99.81 (-0.1/+0.1)	99.67 (-0.1/+0.1)	98.35 (-0.3/+0.3)	96.09 (-1.0/+0.8)	95.61 @ 65	-	-	-	-	-
	Effective Sample Size	11000	11000	11000	11000	11000	11000	-	-	-	-	-


					(-1.2/+0.9)				
Effective Sample Size	9986	8790	7723	5114	742	242	-	-	-

\*Devices subject to an advisory. Refer to the Advisories for more details. Devices may be part of more than one advisory.

**TELIGEN DR**

Models E110/E111/F110/F111

<b>U.S. Survival Probability</b>	<b>Worldwide Malfunction Details</b>	<b>Product Advisories</b>
----------------------------------	--------------------------------------	---------------------------

<b>TELIGEN DR</b> <b>Models E110/E111/F110/F111</b> 			
<b>Worldwide Distribution:</b> 90,000			
<b>Worldwide Confirmed Malfunctions:</b> 434			
	Without Compromised Therapy	With Compromised Therapy	Total
<b>Electrical</b>	<b>302</b>	<b>36</b>	<b>338</b>
<sup>1</sup> Low Voltage Capacitor (Advisory issued)	206	8	
<sup>110</sup> Safety Core-electrocautery	3	-	
<sup>112</sup> High-voltage capacitor	1	5	
<sup>117</sup> Low-voltage capacitors	5	-	
<sup>121</sup> Integrated circuit	13	15	
<sup>124</sup> Battery	53	8	
<sup>125</sup> Low-voltage capacitor	21	-	
<b>Mechanical</b>	<b>14</b>	<b>47</b>	<b>61</b>
<sup>5</sup> Subpectoral implant 2009 (Advisory issued)	3	5	
<sup>103</sup> Transformer	-	20	
<sup>107</sup> Seal plug	2	-	
<sup>108</sup> Difficulty securing lead	8	8	
<sup>115</sup> Header contacts	1	11	
<sup>129</sup> Header	-	3	
<b>Software</b>	<b>14</b>	<b>-</b>	<b>14</b>
<sup>119</sup> Alert messages not displayed post-EOL	3	-	
<sup>122</sup> Memory errors	11	-	
<b>Other</b>	<b>16</b>	<b>5</b>	<b>21</b>
Non-patterned	16	5	
<b>WW Confirmed Malfunctions</b>	<b>346</b>	<b>88</b>	<b>434</b>

[More details](#) about malfunctions

[References](#) cited in table above

CRM PRODUCT PERFORMANCE REPORT Q1 2014


TELIGEN VR

Models E102/E103/F102/F103


U.S. Survival Probability	Worldwide Malfunction Details	Product Advisories
---------------------------	-------------------------------	--------------------

<b>U.S. Summary</b> U.S. Registered Implants: 38,000 U.S. Approval Date: May 2008 U.S. Estimated Active Implants: 29,000	U.S. Normal Battery Depletions: 43 U.S. Unconfirmed Reports of Premature Battery Depletion : 10 U.S. Malfunctions:213 Without Compromised Therapy:160 With Compromised Therapy:53
---	---

Battery Depletions and Malfunctions



Malfunctions Only



U.S. Survival Probability

	Year	1	2	3	4	5	6	7	8	9	10
Non Advisory Population Registered Implants: 21000	Depletions and Malfunctions(%) (Confidence Interval)	99.92 (-0.1/+0.0)	99.85 (-0.1/+0.0)	99.74 (-0.1/+0.1)	99.66 @ 46 mo. (-0.2/+0.1)	-	-	-	-	-	-
	Malfunctions Only(%) (Confidence Interval)	99.93 (-0.0/+0.0)	99.90 (-0.1/+0.0)	99.82 (-0.1/+0.1)	99.75 @ 46 mo. (-0.2/+0.1)	-	-	-	-	-	-
	Effective Sample Size	18579	15272	5421	273	-	-	-	-	-	-
Subpectoral Implant 2009* Registered Implants: 16000	Depletions and Malfunctions(%) (Confidence Interval)	99.78 (-0.1/+0.1)	99.66 (-0.1/+0.1)	99.50 (-0.1/+0.1)	98.69 (-0.2/+0.2)	97.55 (-0.5/+0.4)	97.41 @ 65 mo. (-0.5/+0.4)	-	-	-	-
	Malfunctions Only(%) (Confidence Interval)	99.80 (-0.1/+0.1)	99.73 (-0.1/+0.1)	99.63 (-0.1/+0.1)	98.95 (-0.2/+0.2)	98.03 (-0.4/+0.4)	97.89 @ 65 mo. (-0.4/+0.4)	-	-	-	-
	Effective Sample Size	13682	12001	10520	8883	2193	271	-	-	-	-
Low Voltage Capacitor 2013* Registered Implants: 6,000	Depletions and Malfunctions(%) (Confidence Interval)	99.82 (-0.2/+0.1)	99.74 (-0.2/+0.1)	99.53 (-0.2/+0.1)	97.39 (-0.6/+0.5)	94.97 (-1.3/+1.0)	94.43 @ 64 mo. (-1.3/+1.0)	-	-	-	-
	Malfunctions Only(%) (Confidence Interval)	99.84 (-0.1/+0.1)	99.82 (-0.2/+0.1)	99.68 (-0.2/+0.1)	97.74 (-0.6/+0.5)	95.40 (-1.3/+1.0)	94.85 @ 64 mo.	-	-	-	-
	Effective Sample Size	6,000	-	-	-	-	-	-	-	-	-

					(-1.3/+1.0)				
Effective Sample Size	5226	4585	4025	2718	501	257	-	-	-


\*Devices subject to an advisory. Refer to the Advisories for more details. Devices may be part of more than one advisory.



**TELIGEN VR**

Models E102/E103/F102/F103

<b>U.S. Survival Probability</b>	<b>Worldwide Malfunction Details</b>	<b>Product Advisories</b>
----------------------------------	--------------------------------------	---------------------------

<b>TELIGEN VR</b> <b>Models E102/E103/F102/F103</b> 			
<b>Worldwide Distribution:</b> 64,000			
<b>Worldwide Confirmed Malfunctions:</b> 340			
	Without Compromised Therapy	With Compromised Therapy	Total
<b>Electrical</b>	<b>222</b>	<b>22</b>	<b>244</b>
<sup>1</sup> Low Voltage Capacitor (Advisory issued)	141	3	
<sup>110</sup> Safety Core-electrocautery	1	1	
<sup>112</sup> High-voltage capacitor	-	2	
<sup>117</sup> Low-voltage capacitors	4	-	
<sup>121</sup> Integrated circuit	5	12	
<sup>124</sup> Battery	53	3	
<sup>125</sup> Low-voltage capacitor	18	1	
<b>Mechanical</b>	<b>14</b>	<b>55</b>	<b>69</b>
<sup>5</sup> Subpectoral implant 2009 (Advisory issued)	4	10	
<sup>65</sup> Transformer	-	1	
<sup>103</sup> Transformer	-	13	
<sup>107</sup> Seal plug	1	-	
<sup>108</sup> Difficulty securing lead	-	10	
<sup>115</sup> Header contacts	8	15	
<sup>129</sup> Header	1	6	
<b>Software</b>	<b>12</b>	<b>-</b>	<b>12</b>
<sup>6</sup> Respiratory Sensor Oversensing	1	-	
<sup>119</sup> Alert messages not displayed post-EOL	4	-	
<sup>122</sup> Memory errors	7	-	
<b>Other</b>	<b>8</b>	<b>7</b>	<b>15</b>
Non-patterned	8	7	
<b>WW Confirmed Malfunctions</b>	<b>256</b>	<b>84</b>	<b>340</b>

[More details](#) about malfunctions

[References](#) cited in table above

CRM PRODUCT PERFORMANCE REPORT Q1 2014


CONFIENT DR

Models E030/F030


U.S. Survival Probability	Worldwide Malfunction Details	Product Advisories
---------------------------	-------------------------------	--------------------

<b>U.S. Summary</b>	
U.S. Registered Implants: 7,000	U.S. Normal Battery Depletions: 28
U.S. Approval Date: February 2008	U.S. Unconfirmed Reports of Premature Battery Depletion : 2
U.S. Estimated Active Implants: 4,000	U.S. Malfunctions:7
	Without Compromised Therapy:6
	With Compromised Therapy:1

**Battery Depletions and Malfunctions**



**Malfunctions Only**



**U.S. Survival Probability**


	Year	1	2	3	4	5	6	7	8	9	10
Non Advisory Population	Depletions and Malfunctions(%) (Confidence Interval)	99.94 (-0.1/+0.0)	99.89 (-0.1/+0.1)	99.73 (-0.2/+0.1)	99.53 (-0.2/+0.2)	99.35 (-0.3/+0.2)	98.89 @ 69 mo. (-0.5/+0.3)	-	-	-	-
Registered Implants: 7000	Malfunctions Only(%) (Confidence Interval)	99.98 (-0.1/+0.0)	99.98 (-0.1/+0.0)	99.98 (-0.1/+0.0)	99.91 (-0.1/+0.1)	99.82 (-0.2/+0.1)	99.82 @ 69 mo. (-0.2/+0.1)	-	-	-	-
	Effective Sample Size	6164	5392	4568	3723	2533	497	-	-	-	-

**CONFIENT DR**

Models E030/F030

<b>U.S. Survival Probability</b>	<b>Worldwide Malfunction Details</b>	<b>Product Advisories</b>
----------------------------------	--------------------------------------	---------------------------

**CONFIENT DR**  
**Models E030/F030**



**Worldwide Distribution: 8,000**  
**Worldwide Confirmed Malfunctions: 7**

	Without Compromised Therapy	With Compromised Therapy	Total
<b>Electrical</b>	<b>5</b>	<b>-</b>	<b>5</b>
<sup>32</sup> Capacitor	1	-	
<sup>42</sup> Integrated circuit	2	-	
<sup>125</sup> Low-voltage capacitor	2	-	
<b>Mechanical</b>	<b>-</b>	<b>-</b>	<b>0</b>
<b>Software</b>	<b>-</b>	<b>-</b>	<b>0</b>
<b>Other</b>	<b>1</b>	<b>1</b>	<b>2</b>
Non-patterned	1	-	
<sup>57</sup> Battery depletion	-	1	
<b>WW Confirmed Malfunctions</b>	<b>6</b>	<b>1</b>	<b>7</b>

[More details](#) about malfunctions

[References](#) cited in table above


VITALITY 2 DR

Model T165


U.S. Survival Probability	Worldwide Malfunction Details	Product Advisories
---------------------------	-------------------------------	--------------------

<b>U.S. Summary</b>	
U.S. Registered Implants: 31,000	U.S. Normal Battery Depletions: 9,349
U.S. Approval Date: March 2004	U.S. Unconfirmed Reports of Premature Battery Depletion : 78
U.S. Estimated Active Implants: 7,000	U.S. Malfunctions: 1137
	Without Compromised Therapy: 1073
	With Compromised Therapy: 64

**Battery Depletions and Malfunctions**



**Malfunctions Only**



**U.S. Survival Probability**

		Year	1	2	3	4	5	6	7	8	9	10
Non Advisory Population	Depletions and Malfunctions(%)	99.92	99.79	99.60	98.57	93.57	58.77	14.23	-	-	-	-
	(Confidence Interval)	(-0.1/+0.0)	(-0.1/+0.1)	(-0.1/+0.1)	(-0.2/+0.2)	(-0.5/+0.5)	(-1.3/+1.2)	@ 80 mo. (-1.4/+1.5)				
	Registered Implants: 17000											
05-Apr-07 and 04-Mar-09 Shortened Replacement Window*	Depletions and Malfunctions(%)	99.86	99.39	96.63	85.51	60.66	17.99	6.98	-	-	-	-
	(Confidence Interval)	(-0.1/+0.1)	(-0.2/+0.2)	(-0.5/+0.4)	(-0.9/+0.9)	(-1.4/+1.4)	(-1.2/+1.2)	@ 77 mo. (-0.8/+0.9)				
	Registered Implants: 9000											
10-Mar-07 Product Update - Mid-life Display of Replacement Indicators*	Depletions and Malfunctions(%)	99.92	99.89	99.81	99.52	99.39	99.15	98.77	-	-	-	-
	(Confidence Interval)	(-0.0/+0.0)	(-0.1/+0.0)	(-0.1/+0.1)	(-0.1/+0.1)	(-0.2/+0.1)	(-0.2/+0.2)	@ 80 mo. (-1.1/+0.6)				
	Effective Sample Size	15245	13378	11653	9877	7837	2863	255	-	-	-	-
05-Apr-07 and 04-Mar-09 Shortened Replacement Window*	Depletions and Malfunctions(%)	99.92	99.56	97.36	91.21	86.78	84.76	84.23	-	-	-	-
	(Confidence Interval)	(-0.1/+0.0)	(-0.2/+0.1)	(-0.4/+0.4)	(-0.8/+0.7)	(-1.0/+0.9)	(-1.2/+1.1)	@ 77 mo. (-1.4/+1.3)				
	Effective Sample Size	7844	6862	5806	4451	2724	691	236	-	-	-	-
10-Mar-07 Product Update - Mid-life Display of Replacement Indicators*	Depletions and Malfunctions(%)	99.30	98.16	96.34	89.38	73.16	22.69	9.23	-	-	-	-
	(Confidence Interval)	(-0.3/+0.2)	(-0.4/+0.3)	(-0.6/+0.5)	(-1.0/+0.9)	(-1.6/+1.5)	(-1.6/+1.7)	@ 76 mo. (-1.2/+1.3)				
	Effective Sample Size	7844	6862	5806	4451	2724	691	236	-	-	-	-

Registered Implants: 6000											
	Malfunctions Only (%) (Confidence Interval)	99.34 (-0.3/+0.2)	98.28 (-0.4/+0.3)	96.80 (-0.6/+0.5)	92.35 (-0.9/+0.8)	91.26 (-1.0/+0.9)	89.30 (-1.2/+1.1)	87.90 @ 76 mo. (-1.8/+1.6)	-	-	-
	Effective Sample Size	4992	4339	3719	2980	2097	549	205	-	-	-
23-Jun-06 and 24- Aug-06 Low Voltage Capacitor*	Depletions and Malfunctions (%) (Confidence Interval)	99.84 (-1.0/+0.1)	99.42 (-1.2/+0.4)	98.25 (-1.7/+0.9)	89.37 (-3.5/+2.7)	65.57 @ 59 mo. (-5.3/+5.0)	-	-	-	-	-
Registered Implants: 1000											
	Malfunctions Only (%) (Confidence Interval)	99.84 (-1.0/+0.1)	99.63 (-1.1/+0.3)	98.69 (-1.6/+0.7)	94.80 (-2.8/+1.8)	86.94 @ 59 mo. (-4.3/+3.3)	-	-	-	-	-
	Effective Sample Size	555	472	403	321	203	-	-	-	-	-
12-May-06 Premature Battery Depletion*	Survival probability data not provided because this population does not meet report inclusion criteria (see Statistical Methodology for more details). Refer to Product Advisories for more information.										

\*Devices subject to an advisory. Refer to the Advisories for more details. Devices may be part of more than one advisory.

## VITALITY 2 DR

Model T165

U.S. Survival Probability	Worldwide Malfunction Details	Product Advisories
---------------------------	-------------------------------	--------------------

**VITALITY 2 DR**  
**Model T165**


Worldwide Distribution: 43,000

Worldwide Confirmed Malfunctions: 1366

	Without Compromised Therapy	With Compromised Therapy	Total
<b>Electrical</b>	<b>1264</b>	<b>46</b>	<b>1310</b>
<sup>7</sup> Shortened replacement window (Advisory issued)	477	24	
<sup>8</sup> Low-voltage capacitor (Advisory issued)	1	-	
<sup>9</sup> Premature battery depletion (Advisory issued)	163	1	
<sup>21</sup> Extended charge time post-mid-life	99	1	
<sup>27</sup> Integrated circuit	1	1	
<sup>29</sup> Reconfirmation after charge	1	-	
<sup>32</sup> Capacitor	1	1	
<sup>42</sup> Integrated circuit	7	11	
<sup>64</sup> Capacitor	3	1	
<sup>69</sup> Capacitor	4	-	
<sup>71</sup> Device tones	1	-	
<sup>81</sup> Mid-life display of replacement indicators	267	-	
<sup>82</sup> High-voltage capacitor	4	1	
<sup>89</sup> Integrated circuit	1	-	
<sup>100</sup> Logic errors	-	3	
<sup>109</sup> Low-voltage capacitor	234	2	
<b>Mechanical</b>	<b>7</b>	<b>6</b>	<b>13</b>
<sup>47</sup> Seal plug	4	3	
<sup>65</sup> Transformer	-	1	
<sup>95</sup> Seal plug	2	-	
<sup>130</sup> Solder joint	1	2	
<b>Software</b>	<b>2</b>	<b>2</b>	<b>4</b>
<sup>78</sup> Memory location	-	2	
<sup>80</sup> Memory location	1	-	
<sup>106</sup> Misaligned markers	1	-	
<b>Other</b>	<b>17</b>	<b>22</b>	<b>39</b>
Non-patterned	10	8	
<sup>26</sup> Firmware error	5	8	
<sup>36</sup> Battery depletion	2	5	
<sup>113</sup> Magnet rate	-	1	
<b>WW Confirmed Malfunctions</b>	<b>1290</b>	<b>76</b>	<b>1366</b>

[More details](#) about malfunctions

[References](#) cited in table above

# CRM PRODUCT PERFORMANCE REPORT Q1 2014

## VITALITY 2 EL DR


Model T167

U.S. Survival Probability	Worldwide Malfunction Details	Product Advisories
---------------------------	-------------------------------	--------------------


### U.S. Summary

<p><b>U.S. Registered Implants:</b> 8,000</p> <p><b>U.S. Approval Date:</b> March 2004</p> <p><b>U.S. Estimated Active Implants:</b> 3,000</p>	<p><b>U.S. Normal Battery Depletions:</b> 1,192</p> <p><b>U.S. Unconfirmed Reports of Premature Battery Depletion :</b> 13</p> <p><b>U.S. Malfunctions:</b>746</p> <p style="padding-left: 20px;"><b>Without Compromised Therapy:</b>734</p> <p style="padding-left: 20px;"><b>With Compromised Therapy:</b>12</p>
--	--

### Battery Depletions and Malfunctions



### Malfunctions Only



### U.S. Survival Probability

	Year	1	2	3	4	5	6	7	8	9	10
Non Advisory Population Registered Implants: 5000	Depletions and Malfunctions(%) (Confidence Interval)	99.98 (-0.1/+0.0)	99.85 (-0.2/+0.1)	99.66 (-0.2/+0.1)	99.01 (-0.4/+0.3)	91.24 (-1.1/+1.0)	81.93 (-1.6/+1.5)	69.94 (-3.0/+2.8)	65.14 @ 86 mo. (-3.6/+3.4)	-	-
	Malfunctions Only(%) (Confidence Interval)	99.98 (-0.1/+0.0)	99.90 (-0.2/+0.1)	99.82 (-0.2/+0.1)	99.50 (-0.3/+0.2)	93.34 (-1.0/+0.9)	87.16 (-1.4/+1.3)	86.33 (-1.6/+1.5)	86.33 @ 86 mo. (-1.6/+1.5)	-	-
	Effective Sample Size	4363	3831	3361	2910	2325	1271	296	209	-	-

05-Apr-07 and 04-Mar-09	Depletions and Malfunctions(%) (Confidence Interval)	99.94 (-0.4/+0.1)	99.22 (-0.6/+0.3)	98.37 (-0.8/+0.5)	93.39 (-1.5/+1.2)	77.47 (-2.6/+2.4)	57.97 (-3.2/+3.1)	31.97 (-3.2/+3.4)	28.86 @ 85 mo. (-3.1/+3.3)	-	-
Shortened Replacement Window*											
Registered Implants: 2000											
	Malfunctions Only(%) (Confidence Interval)	99.94 (-0.4/+0.1)	99.41 (-0.5/+0.3)	98.63 (-0.7/+0.5)	94.61 (-1.4/+1.1)	83.60 (-2.4/+2.1)	75.90 (-2.8/+2.6)	73.79 (-3.1/+2.9)	73.79 @ 85 mo. (-3.1/+2.9)	-	-
	Effective Sample Size	1699	1489	1289	1076	781	477	222	222	-	-
10-Mar-07	Depletions and Malfunctions(%) (Confidence Interval)	99.68 (-0.5/+0.2)	99.40 (-0.7/+0.3)	99.18 (-0.8/+0.4)	96.23 (-1.5/+1.1)	74.33 (-3.3/+3.1)	58.01 (-3.8/+3.7)	42.66 @ 82 mo. (-4.0/+4.1)	-	-	-
Product Update - Mid-life Display of Replacement Indicators*											
Registered Implants: 1000											
	Malfunctions Only(%) (Confidence Interval)	99.92 (-0.5/+0.1)	99.74 (-0.6/+0.2)	99.52 (-0.7/+0.3)	97.51 (-1.3/+0.9)	81.18 (-3.1/+2.7)	71.14 (-3.7/+3.4)	70.90 @ 82 mo. (-3.7/+3.5)	-	-	-
	Effective Sample Size	1171	1024	899	763	501	320	207	-	-	-
23-Jun-06 and 24-Aug-06	Survival probability data not provided because this population does not meet report inclusion criteria (see Statistical Methodology for more details). Refer to Product Advisories for more information.										
Low Voltage Capacitor*											


\*Devices subject to an advisory. Refer to the Advisories for more details. Devices may be part of more than one advisory.



VITALITY 2 EL DR

Model T167

U.S. Survival Probability	Worldwide Malfunction Details	Product Advisories
---------------------------	-------------------------------	--------------------

VITALITY 2 EL DR Model T167 			
<b>Worldwide Distribution: 14,000</b>			
<b>Worldwide Confirmed Malfunctions: 1018</b>			
	Without Compromised Therapy	With Compromised Therapy	Total
<b>Electrical</b>	<b>981</b>	<b>9</b>	<b>990</b>
<sup>7</sup> Shortened replacement window (Advisory issued)	143	2	
<sup>21</sup> Extended charge time post-mid-life	13	-	
<sup>32</sup> Capacitor	1	-	
<sup>42</sup> Integrated circuit	-	4	
<sup>64</sup> Capacitor	1	-	
<sup>81</sup> Mid-life display of replacement indicators	782	-	
<sup>82</sup> High-voltage capacitor	-	2	
<sup>89</sup> Integrated circuit	-	1	
<sup>109</sup> Low-voltage capacitor	41	-	
<b>Mechanical</b>	<b>7</b>	<b>3</b>	<b>10</b>
<sup>10</sup> Subpectoral implant (Advisory issued)	1	1	
<sup>33</sup> Header	1	-	
<sup>47</sup> Seal plug	5	1	
<sup>103</sup> Transformer	-	1	
<b>Software</b>	<b>7</b>	<b>1</b>	<b>8</b>
<sup>80</sup> Memory location	1	1	
<sup>106</sup> Misaligned markers	6	-	
<b>Other</b>	<b>4</b>	<b>6</b>	<b>10</b>
Non-patterned	3	3	
<sup>26</sup> Firmware error	1	3	
<b>WW Confirmed Malfunctions</b>	<b>999</b>	<b>19</b>	<b>1018</b>

[More details](#) about malfunctions

[References](#) cited in table above

CRM PRODUCT PERFORMANCE REPORT Q1 2014


VITALITY 2 EL VR

Model T177


U.S. Survival Probability	Worldwide Malfunction Details	Product Advisories
---------------------------	-------------------------------	--------------------

<b>U.S. Summary</b> U.S. Registered Implants: 7,000 U.S. Approval Date: March 2004 U.S. Estimated Active Implants: 3,000	U.S. Normal Battery Depletions: 718 U.S. Unconfirmed Reports of Premature Battery Depletion : 7 U.S. Malfunctions:1056 Without Compromised Therapy:1045 With Compromised Therapy:11
---	---

Battery Depletions and Malfunctions



Malfunctions Only



U.S. Survival Probability

	Year	1	2	3	4	5	6	7	8	9	10
Non Advisory Population Registered Implants: 4000	Depletions and Malfunctions(%) (Confidence Interval)	99.95 (-0.2/+0.0)	99.92 (-0.2/+0.1)	99.78 (-0.2/+0.1)	99.39 (-0.4/+0.2)	97.30 (-0.8/+0.6)	64.21 (-2.5/+2.5)	46.34 @ 81 mo. (-3.1/+3.1)	-	-	-
	Malfunctions Only(%) (Confidence Interval)	99.95 (-0.2/+0.0)	99.92 (-0.2/+0.1)	99.85 (-0.2/+0.1)	99.73 (-0.3/+0.1)	98.44 (-0.6/+0.4)	72.95 (-2.5/+2.3)	59.21 @ 81 mo. (-3.3/+3.2)	-	-	-
	Effective Sample Size	3631	3176	2762	2386	1975	775	229	-	-	-
05-Apr-07 and 04-Mar-09 Shortened Replacement Window* Registered Implants: 2000	Depletions and Malfunctions(%) (Confidence Interval)	99.95 (-0.3/+0.0)	99.56 (-0.5/+0.2)	98.48 (-0.8/+0.5)	95.88 (-1.2/+1.0)	85.72 (-2.2/+2.0)	61.11 (-3.2/+3.1)	41.59 (-3.4/+3.5)	35.70 @ 87 mo. (-3.4/+3.6)	-	-
	Malfunctions Only(%) (Confidence Interval)	99.95 (-0.3/+0.0)	99.56 (-0.5/+0.2)	98.84 (-0.7/+0.4)	97.00 (-1.1/+0.8)	88.94 (-2.1/+1.8)	68.84 (-3.2/+3.0)	61.45 (-3.5/+3.4)	60.90 @ 87 mo. (-3.6/+3.4)	-	-
	Effective Sample Size	1687	1474	1279	1087	822	496	273	212	-	-
10-Mar-07 Product Update - Mid-life Display of Replacement Indicators* Registered Implants: 2000	Depletions and Malfunctions(%) (Confidence Interval)	99.61 (-0.6/+0.2)	99.61 (-0.6/+0.2)	99.38 (-0.8/+0.3)	98.80 (-1.0/+0.6)	91.14 (-2.5/+2.0)	50.48 (-4.4/+4.4)	45.30 @ 74 mo. (-4.4/+4.5)	-	-	-
	Malfunctions Only(%) (Confidence Interval)	99.61 (-0.6/+0.2)	99.61 (-0.6/+0.2)	99.38 (-0.8/+0.3)	98.80 (-1.0/+0.6)	91.14 (-2.5/+2.0)	50.48 (-4.4/+4.4)	45.30 @ 74 mo. (-4.4/+4.5)	-	-	-
	Effective Sample Size	1687	1474	1279	1087	822	496	273	212	-	-

Registered Implants: 1000											
	Malfunctions Only (%)	99.72	99.72	99.48	98.90	93.22	59.83	54.67	-	-	-
	(Confidence Interval)	(-0.6/+0.2)	(-0.6/+0.2)	(-0.7/+0.3)	(-1.0/+0.5)	(-2.3/+1.8)	(-4.6/+4.4)	@ 74 mo. (-4.7/+4.6)			
	Effective Sample Size	975	854	747	647	527	240	209	-	-	-
23-Jun-06 and 24-Aug-06 Low Voltage Capacitor*	Survival probability data not provided because this population does not meet report inclusion criteria (see Statistical Methodology for more details). Refer to Product Advisories for more information.										

\*Devices subject to an advisory. Refer to the Advisories for more details. Devices may be part of more than one advisory.

VITALITY 2 EL VR

Model T177

U.S. Survival Probability	Worldwide Malfunction Details	Product Advisories
---------------------------	-------------------------------	--------------------

VITALITY 2 EL VR Model T177			
Worldwide Distribution: 16,000			
Worldwide Confirmed Malfunctions: 1549			
	Without Compromised Therapy	With Compromised Therapy	Total
<b>Electrical</b>	<b>1513</b>	<b>6</b>	<b>1519</b>
<sup>7</sup> Shortened replacement window (Advisory issued)	137	1	
<sup>8</sup> Low-voltage capacitor (Advisory issued)	2	1	
<sup>21</sup> Extended charge time post-mid-life	10	1	
<sup>42</sup> Integrated circuit	-	3	
<sup>64</sup> Capacitor	1	-	
<sup>69</sup> Capacitor	2	-	
<sup>81</sup> Mid-life display of replacement indicators	1294	-	
<sup>82</sup> High-voltage capacitor	2	-	
<sup>109</sup> Low-voltage capacitor	65	-	
<b>Mechanical</b>	<b>1</b>	<b>8</b>	<b>9</b>
<sup>10</sup> Subpectoral implant (Advisory issued)	-	5	
<sup>33</sup> Header	-	1	
<sup>47</sup> Seal plug	1	-	
<sup>103</sup> Transformer	-	2	
<b>Software</b>	<b>-</b>	<b>2</b>	<b>2</b>
<sup>78</sup> Memory location	-	1	
<sup>80</sup> Memory location	-	1	
<b>Other</b>	<b>10</b>	<b>9</b>	<b>19</b>
Non-patterned	10	7	
<sup>36</sup> Battery depletion	-	2	
<b>WW Confirmed Malfunctions</b>	<b>1524</b>	<b>25</b>	<b>1549</b>

[More details](#) about malfunctions

[References](#) cited in table above


VITALITY 2 VR

Model T175


U.S. Survival Probability	Worldwide Malfunction Details	Product Advisories
---------------------------	-------------------------------	--------------------

<b>U.S. Summary</b> U.S. Registered Implants: 21,000 U.S. Approval Date: March 2004 U.S. Estimated Active Implants: 6,000	U.S. Normal Battery Depletions: 4,147 U.S. Unconfirmed Reports of Premature Battery Depletion : 33 U.S. Malfunctions:1237 Without Compromised Therapy:1212 With Compromised Therapy:25
--	--

Battery Depletions and Malfunctions



Malfunctions Only



U.S. Survival Probability

	Year	1	2	3	4	5	6	7	8	9	10
Non Advisory Population Registered Implants: 11000	Depletions and Malfunctions(%) (Confidence Interval)	99.95 (-0.1/+0.0)	99.82 (-0.1/+0.1)	99.48 (-0.2/+0.1)	97.56 (-0.4/+0.3)	93.59 (-0.6/+0.6)	88.18 (-1.0/+0.9)	46.94 (-3.0/+3.0)	32.10 @ 86 mo. (-3.3/+3.4)	-	-
	Malfunctions Only(%) (Confidence Interval)	99.97 (-0.1/+0.0)	99.92 (-0.1/+0.0)	99.80 (-0.1/+0.1)	98.42 (-0.3/+0.3)	96.20 (-0.5/+0.5)	95.99 (-0.5/+0.5)	95.62 (-0.7/+0.6)	95.62 @ 86 mo. (-0.7/+0.6)	-	-
	Effective Sample Size	9497	8331	7129	5984	4762	2569	354	265	-	-
05-Apr-07 and 04-Mar-09 Shortened Replacement Window* Registered Implants: 6000	Depletions and Malfunctions(%) (Confidence Interval)	99.93 (-0.1/+0.0)	99.43 (-0.3/+0.2)	97.47 (-0.5/+0.4)	90.30 (-1.0/+0.9)	78.08 (-1.4/+1.3)	52.65 (-1.8/+1.8)	17.19 (-1.5/+1.6)	11.63 @ 86 mo. (-1.3/+1.5)	-	-
	Malfunctions Only(%) (Confidence Interval)	99.97 (-0.1/+0.0)	99.50 (-0.2/+0.2)	97.79 (-0.5/+0.4)	92.40 (-0.9/+0.8)	86.40 (-1.2/+1.1)	84.88 (-1.3/+1.2)	83.22 (-1.5/+1.4)	83.22 @ 86 mo. (-1.5/+1.4)	-	-
	Effective Sample Size	5391	4691	4023	3238	2378	1380	363	233	-	-
10-Mar-07 Product Update - Mid-life Display of Replacement Indicators* Registered Implants: 6000	Depletions and Malfunctions(%) (Confidence Interval)	99.34 (-0.3/+0.2)	96.85 (-0.6/+0.5)	94.72 (-0.8/+0.7)	86.62 (-1.3/+1.2)	76.43 (-1.7/+1.6)	56.33 (-2.1/+2.1)	16.19 (-1.7/+1.9)	13.87 @ 85 mo. (-1.6/+1.8)	-	-
	Malfunctions Only(%) (Confidence Interval)	99.97 (-0.1/+0.0)	99.50 (-0.2/+0.2)	97.79 (-0.5/+0.4)	92.40 (-0.9/+0.8)	86.40 (-1.2/+1.1)	84.88 (-1.3/+1.2)	83.22 (-1.5/+1.4)	83.22 @ 86 mo. (-1.5/+1.4)	-	-
	Effective Sample Size	5391	4691	4023	3238	2378	1380	363	233	-	-


Registered Implants: 4000											
	Malfunctions Only (%) (Confidence Interval)	99.39 (-0.3/+0.2)	96.95 (-0.6/+0.5)	95.13 (-0.8/+0.7)	89.20 (-1.2/+1.1)	84.05 (-1.4/+1.3)	83.38 (-1.5/+1.4)	81.62 (-1.8/+1.7)	81.62 @ 85 mo. (-1.8/+1.7)	-	-
	Effective Sample Size	3906	3330	2851	2263	1681	1062	249	207	-	-
23-Jun-06 and 24-Aug-06 Low Voltage Capacitor*	Depletions and Malfunctions (%) (Confidence Interval)	99.47 (-1.1/+0.4)	98.64 (-1.5/+0.7)	96.88 (-2.1/+1.3)	92.79 (-3.1/+2.2)	77.93 (-4.9/+4.3)	75.32 @ 62 mo. (-5.1/+4.5)	-	-	-	-
Registered Implants: 1000											
	Malfunctions Only (%) (Confidence Interval)	99.82 (-1.1/+0.2)	98.99 (-1.4/+0.6)	97.73 (-1.9/+1.0)	95.08 (-2.7/+1.8)	84.95 (-4.5/+3.6)	84.95 @ 62 mo. (-4.5/+3.6)	-	-	-	-
	Effective Sample Size	504	432	366	307	216	202	-	-	-	-
12-May-06 Premature Battery Depletion*	Survival probability data not provided because this population does not meet report inclusion criteria (see Statistical Methodology for more details). Refer to Product Advisories for more information.										

\*Devices subject to an advisory. Refer to the Advisories for more details. Devices may be part of more than one advisory.

VITALITY 2 VR

Model T175

U.S. Survival Probability	Worldwide Malfunction Details	Product Advisories
---------------------------	-------------------------------	--------------------

<b>VITALITY 2 VR</b> <b>Model T175</b> 			
<b>Worldwide Distribution:</b> 37,000 <b>Worldwide Confirmed Malfunctions:</b> 1576			
	Without Compromised Therapy	With Compromised Therapy	Total
<b>Electrical</b>	<b>1524</b>	<b>26</b>	<b>1550</b>
<sup>7</sup> Shortened replacement window (Advisory issued)	347	9	
<sup>8</sup> Low-voltage capacitor (Advisory issued)	-	1	
<sup>9</sup> Premature battery depletion (Advisory issued)	219	6	
<sup>21</sup> Extended charge time post-mid-life	59	-	
<sup>27</sup> Integrated circuit	-	1	
<sup>32</sup> Capacitor	1	-	
<sup>42</sup> Integrated circuit	4	7	
<sup>64</sup> Capacitor	1	-	
<sup>69</sup> Capacitor	4	-	
<sup>81</sup> Mid-life display of replacement indicators	769	-	
<sup>82</sup> High-voltage capacitor	-	1	
<sup>109</sup> Low-voltage capacitor	120	1	
<b>Mechanical</b>	<b>2</b>	<b>1</b>	<b>3</b>
<sup>47</sup> Seal plug	2	1	
<b>Software</b>	<b>-</b>	<b>1</b>	<b>1</b>
<sup>80</sup> Memory location	-	1	
<b>Other</b>	<b>16</b>	<b>6</b>	<b>22</b>
Non-patterned	14	6	
<sup>36</sup> Battery depletion	2	-	
<b>WW Confirmed Malfunctions</b>	<b>1542</b>	<b>34</b>	<b>1576</b>

[More details](#) about malfunctions


[References](#) cited in table above

ADVANTIO DR

Models J063/J066/K063/K066/K083

U.S. Survival Probability	Worldwide Malfunction Details	Product Advisories
---------------------------	-------------------------------	--------------------

<b>U.S. Summary</b>	
U.S. Registered Implants: 29,000	U.S. Normal Battery Depletions: 3
U.S. Approval Date: May 2012	U.S. Unconfirmed Reports of Premature Battery Depletion : 1
U.S. Estimated Active Implants: 28,000	U.S. Malfunctions:3
	Without Compromised Therapy:3
	With Compromised Therapy:0



<b>U.S. Survival Probability</b>		Year	1	2	3	4	5	6	7	8	9	10
Non Advisory Population	Depletions and Malfunctions(%) (Confidence Interval)	99.97 (-0.0/+0.0)	99.95 @ 19 mo. (-0.1/+0.0)	-	-	-	-	-	-	-	-	-
	Malfunctions Only(%) (Confidence Interval)	99.98 (-0.1/+0.0)	99.97 @ 19 mo. (-0.1/+0.0)	-	-	-	-	-	-	-	-	-
Registered Implants: 29000	Effective Sample Size	8890	366	-	-	-	-	-	-	-	-	-




**ADVANTIO DR**

Models J063/J066/K063/K066/K083

<b>U.S. Survival Probability</b>	<b>Worldwide Malfunction Details</b>	<b>Product Advisories</b>
----------------------------------	--------------------------------------	---------------------------

**ADVANTIO DR**  
**Models J063/J066/K063/K066/K083**



**Worldwide Distribution:** 36,000  
**Worldwide Confirmed Malfunctions:** 4

	Without Compromised Therapy	With Compromised Therapy	Total
<b>Electrical</b>	<b>2</b>	<b>-</b>	<b>2</b>
<sup>117</sup> Low-voltage capacitors	1	-	
<sup>121</sup> Integrated circuit	1	-	
<b>Mechanical</b>	<b>-</b>	<b>-</b>	<b>0</b>
<b>Software</b>	<b>1</b>	<b>-</b>	<b>1</b>
<sup>122</sup> Memory errors	1	-	
<b>Other</b>	<b>1</b>	<b>-</b>	<b>1</b>
Non-patterned	1	-	
<b>WW Confirmed Malfunctions</b>	<b>4</b>	<b>0</b>	<b>4</b>

[More details](#) about malfunctions


[References](#) cited in table above

**ADVANTIO EL DR**

Models J064/K064/K067/K084

<b>U.S. Survival Probability</b>	<b>Worldwide Malfunction Details</b>	<b>Product Advisories</b>
----------------------------------	--------------------------------------	---------------------------

**ADVANTIO EL DR**  
**Models J064/K064/K067/K084**



**Worldwide Distribution: 3,000**  
**Worldwide Confirmed Malfunctions: 1**

	Without Compromised Therapy	With Compromised Therapy	Total
<b>Electrical</b>	-	1	1
<sup>117</sup> Low-voltage capacitors	-	1	
<b>Mechanical</b>	-	-	0
<b>Software</b>	-	-	0
<b>Other</b>	-	-	0
Non-patterned	-	-	
<b>WW Confirmed Malfunctions</b>	<b>0</b>	<b>1</b>	<b>1</b>

[More details](#) about malfunctions


[References](#) cited in table above

ADVANTIO SR

Models J062/J065/K062/K065/K082

U.S. Survival Probability	Worldwide Malfunction Details	Product Advisories
---------------------------	-------------------------------	--------------------

<b>U.S. Summary</b>	
U.S. Registered Implants: 6,000	U.S. Normal Battery Depletions: 1
U.S. Approval Date: May 2012	U.S. Unconfirmed Reports of Premature Battery Depletion : 0
U.S. Estimated Active Implants: 6,000	U.S. Malfunctions:0
	Without Compromised Therapy:0
	With Compromised Therapy:0




<b>U.S. Survival Probability</b>		Year	1	2	3	4	5	6	7	8	9	10
Non Advisory Population	Depletions and Malfunctions (%) (Confidence Interval)	99.98 (-0.1/+0.0)	99.98 @ 18 mo. (-0.1/+0.0)	-	-	-	-	-	-	-	-	-
	Malfunctions Only (%) (Confidence Interval)	100.00 (-0.0/+0.0)	100.00 @ 18 mo. (-0.0/+0.0)	-	-	-	-	-	-	-	-	-
Registered Implants: 6000	Effective Sample Size	1842	214	-	-	-	-	-	-	-	-	-

**ADVANTIO SR**

Models J062/J065/K062/K065/K082

<b>U.S. Survival Probability</b>	<b>Worldwide Malfunction Details</b>	<b>Product Advisories</b>
----------------------------------	--------------------------------------	---------------------------

**ADVANTIO SR**  
**Models J062/J065/K062/K065/K082**



**Worldwide Distribution:** 12,000  
**Worldwide Confirmed Malfunctions:** 1

	Without Compromised Therapy	With Compromised Therapy	Total
<b>Electrical</b>	-	1	1
<sup>121</sup> Integrated circuit	-	1	
<b>Mechanical</b>	-	-	0
<b>Software</b>	-	-	0
<b>Other</b>	-	-	0
Non-patterned	-	-	
<b>WW Confirmed Malfunctions</b>	<b>0</b>	<b>1</b>	<b>1</b>

[More details](#) about malfunctions

[References](#) cited in table above


**INGENIO DR**

Models J173/J176/K173/K176/K183


<b>U.S. Survival Probability</b>	<b>Worldwide Malfunction Details</b>	<b>Product Advisories</b>
----------------------------------	--------------------------------------	---------------------------

<b>U.S. Summary</b>	
U.S. Registered Implants: 32,000	U.S. Normal Battery Depletions: 4
U.S. Approval Date: May 2012	U.S. Unconfirmed Reports of Premature Battery Depletion : 1
U.S. Estimated Active Implants: 31,000	U.S. Malfunctions:4
	Without Compromised Therapy:4
	With Compromised Therapy:0

**Battery Depletions and Malfunctions**



**Malfunctions Only**



**U.S. Survival Probability**


	Year	1	2	3	4	5	6	7	8	9	10
Non Advisory Population	Depletions and Malfunctions(%) (Confidence Interval)	99.97 (-0.0/+0.0)	99.94 @ 19 mo. (-0.1/+0.0)	-	-	-	-	-	-	-	-
	Malfunctions Only(%) (Confidence Interval)	99.99 (-0.0/+0.0)	99.97 @ 19 mo. (-0.1/+0.0)	-	-	-	-	-	-	-	-
Registered Implants: 32000	Effective Sample Size	9218	680	-	-	-	-	-	-	-	-

**INGENIO DR**

Models J173/J176/K173/K176/K183

<b>U.S. Survival Probability</b>	<b>Worldwide Malfunction Details</b>	<b>Product Advisories</b>
----------------------------------	--------------------------------------	---------------------------

**INGENIO DR**  
**Models J173/J176/K173/K176/K183**



**Worldwide Distribution:** 51,000  
**Worldwide Confirmed Malfunctions:** 5

	Without Compromised Therapy	With Compromised Therapy	Total
<b>Electrical</b>	<b>1</b>	<b>-</b>	<b>1</b>
<sup>117</sup> Low-voltage capacitors	1	-	
<b>Mechanical</b>	<b>-</b>	<b>-</b>	<b>0</b>
<b>Software</b>	<b>1</b>	<b>-</b>	<b>1</b>
<sup>122</sup> Memory errors	1	-	
<b>Other</b>	<b>3</b>	<b>-</b>	<b>3</b>
Non-patterned	3	-	
<b>WW Confirmed Malfunctions</b>	<b>5</b>	<b>0</b>	<b>5</b>

[More details](#) about malfunctions

[References](#) cited in table above

**INGENIO EL DR**

Models J174/J177/K174/K177/K184

<b>U.S. Survival Probability</b>	<b>Worldwide Malfunction Details</b>	<b>Product Advisories</b>
----------------------------------	--------------------------------------	---------------------------

**INGENIO EL DR**  
Models J174/J177/K174/K177/K184 

**Worldwide Distribution:** 13,000

**Worldwide Confirmed Malfunctions:** 2

	Without Compromised Therapy	With Compromised Therapy	Total
<b>Electrical</b>	<b>1</b>	<b>-</b>	<b>1</b>
<sup>117</sup> Low-voltage capacitors	1	-	
<b>Mechanical</b>	<b>-</b>	<b>-</b>	<b>0</b>
<b>Software</b>	<b>-</b>	<b>-</b>	<b>0</b>
<b>Other</b>	<b>1</b>	<b>-</b>	<b>1</b>
Non-patterned	1	-	
<b>WW Confirmed Malfunctions</b>	<b>2</b>	<b>0</b>	<b>2</b>

[More details](#) about malfunctions

[References](#) cited in table above


**INGENIO SR**

Models J172/J175/K172/K175/K182


<b>U.S. Survival Probability</b>	<b>Worldwide Malfunction Details</b>	<b>Product Advisories</b>
----------------------------------	--------------------------------------	---------------------------

<b>U.S. Summary</b>	
U.S. Registered Implants: 6,000	U.S. Normal Battery Depletions: 0
U.S. Approval Date: May 2012	U.S. Unconfirmed Reports of Premature Battery Depletion : 0
U.S. Estimated Active Implants: 5,000	U.S. Malfunctions:0
	Without Compromised Therapy:0
	With Compromised Therapy:0

**Battery Depletions and Malfunctions**



**Malfunctions Only**



<b>U.S. Survival Probability</b>		Year	1	2	3	4	5	6	7	8	9	10
Non Advisory Population	Depletions and Malfunctions (%) (Confidence Interval)	100.00 (-0.0/+0.0)	100.00 @ 18 mo. (-0.0/+0.0)	-	-	-	-	-	-	-	-	-
	Malfunctions Only (%) (Confidence Interval)	100.00 (-0.0/+0.0)	100.00 @ 18 mo. (-0.0/+0.0)	-	-	-	-	-	-	-	-	-
Registered Implants: 6000	Effective Sample Size	1533	230	-	-	-	-	-	-	-	-	-




## INGENIO SR

Models J172/J175/K172/K175/K182

U.S. Survival Probability	Worldwide Malfunction Details	Product Advisories
---------------------------	-------------------------------	--------------------

**INGENIO SR**  
Models J172/J175/K172/K175/K182



**Worldwide Distribution:** 15,000  
**Worldwide Confirmed Malfunctions:** 0

	Without Compromised Therapy	With Compromised Therapy	Total
<b>Electrical</b>	-	-	<b>0</b>
<b>Mechanical</b>	-	-	<b>0</b>
<b>Software</b>	-	-	<b>0</b>
<b>Other</b>	-	-	<b>0</b>
Non-patterned	-	-	
<b>WW Confirmed Malfunctions</b>	<b>0</b>	<b>0</b>	<b>0</b>

[More details](#) about malfunctions

[References](#) cited in table above

CRM PRODUCT PERFORMANCE REPORT Q1 2014

ALTRUA 60 DR


Model S602

U.S. Survival Probability	Worldwide Malfunction Details	Product Advisories
---------------------------	-------------------------------	--------------------


U.S. Summary

U.S. Registered Implants: 22,000	U.S. Normal Battery Depletions: 90
U.S. Approval Date: April 2008	U.S. Unconfirmed Reports of Premature Battery Depletion : 2
U.S. Estimated Active Implants: 16,000	U.S. Malfunctions:3
	Without Compromised Therapy:2
	With Compromised Therapy:1

Battery Depletions and Malfunctions



Malfunctions Only




U.S. Survival Probability

		Year	1	2	3	4	5	6	7	8	9	10
Non Advisory Population	Depletions and Malfunctions(%) (Confidence Interval)		99.99 (-0.0/+0.0)	99.94 (-0.1/+0.0)	99.84 (-0.1/+0.1)	99.54 (-0.1/+0.1)	99.08 (-0.2/+0.2)	98.83 @ 67 mo. (-0.4/+0.3)	-	-	-	-
	Malfunctions Only(%) (Confidence Interval)		99.99 (-0.0/+0.0)	99.98 (-0.0/+0.0)	99.98 (-0.0/+0.0)	99.98 (-0.0/+0.0)	99.98 (-0.0/+0.0)	99.98 @ 67 mo. (-0.0/+0.0)	-	-	-	-
Registered Implants: 22000		Effective Sample Size	18944	16179	13025	9872	4049	206	-	-	-	-

## ALTRUA 60 DR

Model S602

<b>U.S. Survival Probability</b>	<b>Worldwide Malfunction Details</b>	<b>Product Advisories</b>
----------------------------------	--------------------------------------	---------------------------

**ALTRUA 60 DR**  
**Model S602**


**Worldwide Distribution:** 55,000  
**Worldwide Confirmed Malfunctions:** 5

	Without Compromised Therapy	With Compromised Therapy	Total
<b>Electrical</b>	<b>1</b>	<b>-</b>	<b>1</b>
<sup>32</sup> Capacitor	1	-	
<b>Mechanical</b>	<b>1</b>	<b>1</b>	<b>2</b>
<sup>39</sup> Capacitor array	1	-	
<sup>108</sup> Difficulty securing lead	-	1	
<b>Software</b>	<b>-</b>	<b>-</b>	<b>0</b>
<b>Other</b>	<b>1</b>	<b>1</b>	<b>2</b>
Non-patterned	1	1	
<b>WW Confirmed Malfunctions</b>	<b>3</b>	<b>2</b>	<b>5</b>

[More details](#) about malfunctions

[References](#) cited in table above

CRM PRODUCT PERFORMANCE REPORT Q1 2014


ALTRUA 60 DR (Downsize)

Model S603


U.S. Survival Probability	Worldwide Malfunction Details	Product Advisories
---------------------------	-------------------------------	--------------------

<b>U.S. Summary</b>	
U.S. Registered Implants: 90,000	U.S. Normal Battery Depletions: 661
U.S. Approval Date: April 2008	U.S. Unconfirmed Reports of Premature Battery Depletion : 23
U.S. Estimated Active Implants: 69,000	U.S. Malfunctions:16
	Without Compromised Therapy:9
	With Compromised Therapy:7

Battery Depletions and Malfunctions



Malfunctions Only



U.S. Survival Probability


	Year	1	2	3	4	5	6	7	8	9	10
Non Advisory Population	Depletions and Malfunctions(%) (Confidence Interval)	99.98 (-0.0/+0.0)	99.91 (-0.0/+0.0)	99.66 (-0.1/+0.0)	98.98 (-0.1/+0.1)	97.14 (-0.3/+0.3)	95.24 @ 67 mo. (-0.7/+0.6)	-	-	-	-
Registered Implants: 90000	Malfunctions Only(%) (Confidence Interval)	99.99 (-0.0/+0.0)	99.99 (-0.0/+0.0)	99.98 (-0.0/+0.0)	99.97 (-0.0/+0.0)	99.97 (-0.0/+0.0)	99.97 @ 67 mo. (-0.0/+0.0)	-	-	-	-
	Effective Sample Size	78249	61459	41538	23348	7029	299	-	-	-	-

**ALTRUA 60 DR (Downsize)**

Model S603

<b>U.S. Survival Probability</b>	<b>Worldwide Malfunction Details</b>	<b>Product Advisories</b>
----------------------------------	--------------------------------------	---------------------------

**ALTRUA 60 DR (Downsize)**  
**Model S603**



**Worldwide Distribution: 131,000**  
**Worldwide Confirmed Malfunctions: 18**

	Without Compromised Therapy	With Compromised Therapy	Total
<b>Electrical</b>	<b>5</b>	<b>5</b>	<b>10</b>
<sup>32</sup> Capacitor	4	4	
<sup>89</sup> Integrated circuit	1	1	
<b>Mechanical</b>	<b>2</b>	<b>-</b>	<b>2</b>
<sup>105</sup> Connector block	1	-	
<sup>108</sup> Difficulty securing lead	1	-	
<b>Software</b>	<b>-</b>	<b>-</b>	<b>0</b>
<b>Other</b>	<b>3</b>	<b>3</b>	<b>6</b>
Non-patterned	-	2	
<sup>72</sup> Battery depletion	2	1	
<sup>120</sup> Battery status	1	-	
<b>WW Confirmed Malfunctions</b>	<b>10</b>	<b>8</b>	<b>18</b>

[More details](#) about malfunctions

[References](#) cited in table above



CRM PRODUCT PERFORMANCE REPORT Q1 2014

ALTRUA 60 DR EL

Model S606

U.S. Survival Probability	Worldwide Malfunction Details	Product Advisories
---------------------------	-------------------------------	--------------------

<b>U.S. Summary</b>	
U.S. Registered Implants: 59,000	U.S. Normal Battery Depletions: 59
U.S. Approval Date: April 2008	U.S. Unconfirmed Reports of Premature Battery Depletion : 5
U.S. Estimated Active Implants: 51,000	U.S. Malfunctions:6
	Without Compromised Therapy:4
	With Compromised Therapy:2



**U.S. Survival Probability**


	Year	1	2	3	4	5	6	7	8	9	10
Non Advisory Population	Depletions and Malfunctions(%) (Confidence Interval)	99.98 (-0.0/+0.0)	99.94 (-0.0/+0.0)	99.85 (-0.1/+0.0)	99.71 (-0.1/+0.1)	99.48 @ 56 mo. (-0.3/+0.2)	-	-	-	-	-
Registered Implants: 59000	Malfunctions Only(%) (Confidence Interval)	99.99 (-0.0/+0.0)	99.99 (-0.0/+0.0)	99.99 (-0.0/+0.0)	99.99 (-0.0/+0.0)	99.99 @ 56 mo. (-0.0/+0.0)	-	-	-	-	-
	Effective Sample Size	51664	37415	19428	5606	283	-	-	-	-	-

**ALTRUA 60 DR EL**

Model S606

<b>U.S. Survival Probability</b>	<b>Worldwide Malfunction Details</b>	<b>Product Advisories</b>
----------------------------------	--------------------------------------	---------------------------

**ALTRUA 60 DR EL**  
**Model S606**



**Worldwide Distribution:** 89,000  
**Worldwide Confirmed Malfunctions:** 7

	Without Compromised Therapy	With Compromised Therapy	Total
<b>Electrical</b>	<b>4</b>	<b>-</b>	<b>4</b>
<sup>32</sup> Capacitor	3	-	
<sup>42</sup> Integrated circuit	1	-	
<b>Mechanical</b>	<b>-</b>	<b>1</b>	<b>1</b>
<sup>108</sup> Difficulty securing lead	-	1	
<b>Software</b>	<b>-</b>	<b>-</b>	<b>0</b>
<b>Other</b>	<b>1</b>	<b>1</b>	<b>2</b>
Non-patterned	1	-	
<sup>72</sup> Battery depletion	-	1	
<b>WW Confirmed Malfunctions</b>	<b>5</b>	<b>2</b>	<b>7</b>

[More details](#) about malfunctions

[References](#) cited in table above



CRM PRODUCT PERFORMANCE REPORT Q1 2014

ALTRUA 60 SR

Model S601

U.S. Survival Probability	Worldwide Malfunction Details	Product Advisories
---------------------------	-------------------------------	--------------------

<b>U.S. Summary</b>	
U.S. Registered Implants: 32,000	U.S. Normal Battery Depletions: 74
U.S. Approval Date: April 2008	U.S. Unconfirmed Reports of Premature Battery Depletion : 0
U.S. Estimated Active Implants: 22,000	U.S. Malfunctions:2
	Without Compromised Therapy:0
	With Compromised Therapy:2




U.S. Survival Probability		Year	1	2	3	4	5	6	7	8	9	10
Non Advisory Population	Depletions and Malfunctions(%) (Confidence Interval)	99.95 (-0.0/+0.0)	99.89 (-0.0/+0.0)	99.80 (-0.1/+0.1)	99.65 (-0.1/+0.1)	99.16 (-0.3/+0.2)	98.68 @ 66 mo. (-0.6/+0.4)	-	-	-	-	-
	Malfunctions Only(%) (Confidence Interval)	100.00 (-0.0/+0.0)	100.00 (-0.0/+0.0)	99.99 (-0.0/+0.0)	99.99 (-0.0/+0.0)	99.99 (-0.0/+0.0)	99.99 @ 66 mo. (-0.0/+0.0)	-	-	-	-	-
Registered Implants: 32000	Effective Sample Size	26283	19907	12652	6501	1761	202	-	-	-	-	-



## ALTRUA 60 SR

Model S601

<b>U.S. Survival Probability</b>	<b>Worldwide Malfunction Details</b>	<b>Product Advisories</b>
----------------------------------	--------------------------------------	---------------------------

**ALTRUA 60 SR**  
**Model S601**


**Worldwide Distribution:** 68,000  
**Worldwide Confirmed Malfunctions:** 8

	Without Compromised Therapy	With Compromised Therapy	Total
<b>Electrical</b>	<b>2</b>	<b>3</b>	<b>5</b>
<sup>32</sup> Capacitor	2	1	
<sup>89</sup> Integrated circuit	-	2	
<b>Mechanical</b>	-	-	<b>0</b>
<b>Software</b>	-	-	<b>0</b>
<b>Other</b>	-	<b>3</b>	<b>3</b>
Non-patterned	-	2	
<sup>72</sup> Battery depletion	-	1	
<b>WW Confirmed Malfunctions</b>	<b>2</b>	<b>6</b>	<b>8</b>


[More details](#) about malfunctions

[References](#) cited in table above

## ALTRUA 50 SR

Model S501

<b>U.S. Survival Probability</b>	<b>Worldwide Malfunction Details</b>	<b>Product Advisories</b>
----------------------------------	--------------------------------------	---------------------------

**ALTRUA 50 SR**  
**Model S501**


**Worldwide Distribution:** 23,000  
**Worldwide Confirmed Malfunctions:** 5

	Without Compromised Therapy	With Compromised Therapy	Total
<b>Electrical</b>	<b>1</b>	<b>2</b>	<b>3</b>
<sup>32</sup> Capacitor	1	2	
<b>Mechanical</b>	-	-	<b>0</b>
<b>Software</b>	-	-	<b>0</b>
<b>Other</b>	-	<b>2</b>	<b>2</b>
Non-patterned	-	1	
<sup>72</sup> Battery depletion	-	1	
<b>WW Confirmed Malfunctions</b>	<b>1</b>	<b>4</b>	<b>5</b>

[More details](#) about malfunctions


[References](#) cited in table above

**ALTRUA 50 DR (Downsize)**

Model S502

<b>U.S. Survival Probability</b>	<b>Worldwide Malfunction Details</b>	<b>Product Advisories</b>
----------------------------------	--------------------------------------	---------------------------

**ALTRUA 50 DR (Downsize)**  
**Model S502**



**Worldwide Distribution:** 42,000  
**Worldwide Confirmed Malfunctions:** 5

	Without Compromised Therapy	With Compromised Therapy	Total
<b>Electrical</b>	<b>3</b>	<b>-</b>	<b>3</b>
<sup>32</sup> Capacitor	2	-	
<sup>89</sup> Integrated circuit	1	-	
<b>Mechanical</b>	<b>-</b>	<b>1</b>	<b>1</b>
<sup>108</sup> Difficulty securing lead	-	1	
<b>Software</b>	<b>-</b>	<b>-</b>	<b>0</b>
<b>Other</b>	<b>1</b>	<b>-</b>	<b>1</b>
Non-patterned	-	-	
<sup>72</sup> Battery depletion	1	-	
<b>WW Confirmed Malfunctions</b>	<b>4</b>	<b>1</b>	<b>5</b>

[More details](#) about malfunctions


[References](#) cited in table above

**ALTRUA 50 DDD (Downsize)**

Model S503

<b>U.S. Survival Probability</b>	<b>Worldwide Malfunction Details</b>	<b>Product Advisories</b>
----------------------------------	--------------------------------------	---------------------------

**ALTRUA 50 DDD (Downsize)**  
**Model S503**



**Worldwide Distribution:** 10,000  
**Worldwide Confirmed Malfunctions:** 3

	Without Compromised Therapy	With Compromised Therapy	Total
<b>Electrical</b>	-	-	<b>0</b>
<b>Mechanical</b>	-	-	<b>0</b>
<b>Software</b>	-	-	<b>0</b>
<b>Other</b>	<b>1</b>	<b>2</b>	<b>3</b>
Non-patterned	-	-	
<sup>72</sup> Battery depletion	-	2	
<sup>120</sup> Battery status	1	-	
<b>WW Confirmed Malfunctions</b>	<b>1</b>	<b>2</b>	<b>3</b>

[More details](#) about malfunctions


[References](#) cited in table above

**ALTRUA 50 VDD (Downsize)**

Model S504

<b>U.S. Survival Probability</b>	<b>Worldwide Malfunction Details</b>	<b>Product Advisories</b>
----------------------------------	--------------------------------------	---------------------------

**ALTRUA 50 VDD (Downsize)**  
**Model S504**



**Worldwide Distribution: 6,000**  
**Worldwide Confirmed Malfunctions: 0**

	Without Compromised Therapy	With Compromised Therapy	Total
<b>Electrical</b>	-	-	<b>0</b>
<b>Mechanical</b>	-	-	<b>0</b>
<b>Software</b>	-	-	<b>0</b>
<b>Other</b>	-	-	<b>0</b>
Non-patterned	-	-	
<b>WW Confirmed Malfunctions</b>	<b>0</b>	<b>0</b>	<b>0</b>


[More details](#) about malfunctions

[References](#) cited in table above

## ALTRUA 50 SSI

Model S508

U.S. Survival Probability	Worldwide Malfunction Details	Product Advisories
---------------------------	-------------------------------	--------------------

**ALTRUA 50 SSI**  
**Model S508**


**Worldwide Distribution:** 5,000  
**Worldwide Confirmed Malfunctions:** 1

	Without Compromised Therapy	With Compromised Therapy	Total
<b>Electrical</b>	-	-	<b>0</b>
<b>Mechanical</b>	-	-	<b>0</b>
<b>Software</b>	-	-	<b>0</b>
<b>Other</b>	-	<b>1</b>	<b>1</b>
Non-patterned	-	-	
<sup>72</sup> Battery depletion	-	<b>1</b>	
<b>WW Confirmed Malfunctions</b>	<b>0</b>	<b>1</b>	<b>1</b>

[More details](#) about malfunctions

[References](#) cited in table above

CRM PRODUCT PERFORMANCE REPORT Q1 2014


ALTRUA 40 DR

Model S402


U.S. Survival Probability	Worldwide Malfunction Details	Product Advisories
---------------------------	-------------------------------	--------------------

<b>U.S. Summary</b>	
U.S. Registered Implants: 2,000	U.S. Normal Battery Depletions: 4
U.S. Approval Date: April 2008	U.S. Unconfirmed Reports of Premature Battery Depletion : 1
U.S. Estimated Active Implants: 1,000	U.S. Malfunctions:0
	Without Compromised Therapy:0
	With Compromised Therapy:0

**Battery Depletions and Malfunctions**



**Malfunctions Only**



**U.S. Survival Probability**


	Year	1	2	3	4	5	6	7	8	9	10
Non Advisory Population	Depletions and Malfunctions(%) (Confidence Interval)	99.93 (-0.4/+0.1)	99.93 (-0.4/+0.1)	99.93 (-0.4/+0.1)	99.66 (-0.6/+0.2)	99.66 (-0.6/+0.2)	99.66 @ 64 mo. (-0.6/+0.2)	-	-	-	-
Registered Implants: 2000	Malfunctions Only(%) (Confidence Interval)	100.00 (-0.0/+0.0)	100.00 (-0.0/+0.0)	100.00 (-0.0/+0.0)	100.00 (-0.0/+0.0)	100.00 (-0.0/+0.0)	100.00 @ 64 mo. (-0.0/+0.0)	-	-	-	-
	Effective Sample Size	1517	1346	1195	1053	560	228	-	-	-	-

**ALTRUA 40 DR**

Model S402

<b>U.S. Survival Probability</b>	<b>Worldwide Malfunction Details</b>	<b>Product Advisories</b>
----------------------------------	--------------------------------------	---------------------------

**ALTRUA 40 DR**  
**Model S402**



**Worldwide Distribution: 3,000**  
**Worldwide Confirmed Malfunctions: 1**

	Without Compromised Therapy	With Compromised Therapy	Total
<b>Electrical</b>	-	-	<b>0</b>
<b>Mechanical</b>	-	-	<b>0</b>
<b>Software</b>	-	-	<b>0</b>
<b>Other</b>	-	<b>1</b>	<b>1</b>
Non-patterned	-	-	
<sup>72</sup> Battery depletion	-	<b>1</b>	
<b>WW Confirmed Malfunctions</b>	<b>0</b>	<b>1</b>	<b>1</b>

[More details](#) about malfunctions

[References](#) cited in table above



CRM PRODUCT PERFORMANCE REPORT Q1 2014


ALTRUA 40 DR (downsize)

Model S403


U.S. Survival Probability	Worldwide Malfunction Details	Product Advisories
---------------------------	-------------------------------	--------------------

<b>U.S. Summary</b>	
U.S. Registered Implants: 14,000	U.S. Normal Battery Depletions: 102
U.S. Approval Date: April 2008	U.S. Unconfirmed Reports of Premature Battery Depletion : 1
U.S. Estimated Active Implants: 11,000	U.S. Malfunctions:2
	Without Compromised Therapy:2
	With Compromised Therapy:0

Battery Depletions and Malfunctions



Malfunctions Only



U.S. Survival Probability


	Year	1	2	3	4	5	6	7	8	9	10
Non Advisory Population	Depletions and Malfunctions(%) (Confidence Interval)	99.98 (-0.0/+0.0)	99.93 (-0.1/+0.0)	99.69 (-0.1/+0.1)	98.88 (-0.3/+0.3)	96.70 (-0.9/+0.7)	96.02 @ 64 mo. (-1.2/+0.9)	-	-	-	-
Registered Implants: 14000	Malfunctions Only(%) (Confidence Interval)	99.99 (-0.0/+0.0)	99.99 (-0.0/+0.0)	99.99 (-0.0/+0.0)	99.98 (-0.1/+0.0)	99.98 (-0.1/+0.0)	99.98 @ 64 mo. (-0.1/+0.0)	-	-	-	-
	Effective Sample Size	12502	10096	6526	3343	841	330	-	-	-	-

**ALTRUA 40 DR (downsize)**

Model S403

<b>U.S. Survival Probability</b>	<b>Worldwide Malfunction Details</b>	<b>Product Advisories</b>
----------------------------------	--------------------------------------	---------------------------

**ALTRUA 40 DR (downsize)**  
**Model S403**



**Worldwide Distribution: 22,000**  
**Worldwide Confirmed Malfunctions: 2**

	Without Compromised Therapy	With Compromised Therapy	Total
<b>Electrical</b>	-	-	<b>0</b>
<b>Mechanical</b>	<b>2</b>	-	<b>2</b>
<sup>107</sup> Seal plug	1	-	
<sup>108</sup> Difficulty securing lead	1	-	
<b>Software</b>	-	-	<b>0</b>
<b>Other</b>	-	-	<b>0</b>
Non-patterned	-	-	
<b>WW Confirmed Malfunctions</b>	<b>2</b>	<b>0</b>	<b>2</b>

[More details](#) about malfunctions


[References](#) cited in table above

ALTRUA 40 DR EL

Model S404

U.S. Survival Probability	Worldwide Malfunction Details	Product Advisories
---------------------------	-------------------------------	--------------------

<b>U.S. Summary</b>	
U.S. Registered Implants: 5,000	U.S. Normal Battery Depletions: 6
U.S. Approval Date: April 2008	U.S. Unconfirmed Reports of Premature Battery Depletion : 0
U.S. Estimated Active Implants: 4,000	U.S. Malfunctions:0
	Without Compromised Therapy:0
	With Compromised Therapy:0




<b>U.S. Survival Probability</b>		Year	1	2	3	4	5	6	7	8	9	10
Non Advisory Population	Depletions and Malfunctions (%) (Confidence Interval)	100.00 (-0.0/+0.0)	99.97 (-0.2/+0.0)	99.93 (-0.2/+0.1)	99.57 (-0.6/+0.3)	99.57 @ 53 mo. (-0.6/+0.3)	-	-	-	-	-	-
	Malfunctions Only (%) (Confidence Interval)	100.00 (-0.0/+0.0)	100.00 (-0.0/+0.0)	100.00 (-0.0/+0.0)	100.00 (-0.0/+0.0)	100.00 @ 53 mo. (-0.0/+0.0)	-	-	-	-	-	-
Registered Implants: 5000		Effective Sample Size	4449	3484	2045	740	255	-	-	-	-	-

**ALTRUA 40 DR EL**

Model S404

<b>U.S. Survival Probability</b>	<b>Worldwide Malfunction Details</b>	<b>Product Advisories</b>
----------------------------------	--------------------------------------	---------------------------

**ALTRUA 40 DR EL**  
**Model S404**



**Worldwide Distribution: 9,000**  
**Worldwide Confirmed Malfunctions: 1**

	Without Compromised Therapy	With Compromised Therapy	Total
<b>Electrical</b>	<b>1</b>	<b>-</b>	<b>1</b>
<sup>32</sup> Capacitor	1	-	
<b>Mechanical</b>	<b>-</b>	<b>-</b>	<b>0</b>
<b>Software</b>	<b>-</b>	<b>-</b>	<b>0</b>
<b>Other</b>	<b>-</b>	<b>-</b>	<b>0</b>
Non-patterned	-	-	
<b>WW Confirmed Malfunctions</b>	<b>1</b>	<b>0</b>	<b>1</b>

[More details](#) about malfunctions

[References](#) cited in table above


# ALTRUA 40 SR

Model S401


<b>U.S. Survival Probability</b>	<b>Worldwide Malfunction Details</b>	<b>Product Advisories</b>
----------------------------------	--------------------------------------	---------------------------

<b>U.S. Summary</b>	
U.S. Registered Implants: 5,000	U.S. Normal Battery Depletions: 7
U.S. Approval Date: April 2008	U.S. Unconfirmed Reports of Premature Battery Depletion : 0
U.S. Estimated Active Implants: 3,000	U.S. Malfunctions: 1
	Without Compromised Therapy: 1
	With Compromised Therapy: 0

## Battery Depletions and Malfunctions



## Malfunctions Only




<b>U.S. Survival Probability</b>		Year	1	2	3	4	5	6	7	8	9	10
Non Advisory Population	Depletions and Malfunctions (%)	99.93	99.93	99.93	99.87	99.64	99.64	-	-	-	-	-
	(Confidence Interval)	(-0.1/+0.0)	(-0.1/+0.0)	(-0.1/+0.0)	(-0.2/+0.1)	(-0.6/+0.2)	@ 61 mo. (-0.6/+0.2)					
Registered Implants: 5000	Malfunctions Only (%)	99.98	99.98	99.98	99.98	99.98	99.98	-	-	-	-	-
	(Confidence Interval)	(-0.2/+0.0)	(-0.2/+0.0)	(-0.2/+0.0)	(-0.2/+0.0)	(-0.2/+0.0)	@ 61 mo. (-0.2/+0.0)					
	Effective Sample Size	3954	3104	1962	1037	267	222	-	-	-	-	-

## ALTRUA 40 SR

Model S401

U.S. Survival Probability	Worldwide Malfunction Details	Product Advisories
---------------------------	-------------------------------	--------------------

**ALTRUA 40 SR**  
**Model S401**


**Worldwide Distribution:** 9,000  
**Worldwide Confirmed Malfunctions:** 2

	Without Compromised Therapy	With Compromised Therapy	Total
<b>Electrical</b>	<b>1</b>	<b>1</b>	<b>2</b>
<sup>32</sup> Capacitor	1	-	
<sup>89</sup> Integrated circuit	-	1	
<b>Mechanical</b>	-	-	<b>0</b>
<b>Software</b>	-	-	<b>0</b>
<b>Other</b>	-	-	<b>0</b>
Non-patterned	-	-	
<b>WW Confirmed Malfunctions</b>	<b>1</b>	<b>1</b>	<b>2</b>

[More details](#) about malfunctions


[References](#) cited in table above

**ALTRUA 20 DDD**

Model S207

<b>U.S. Survival Probability</b>	<b>Worldwide Malfunction Details</b>	<b>Product Advisories</b>
----------------------------------	--------------------------------------	---------------------------

**ALTRUA 20 DDD**  
**Model S207**



**Worldwide Distribution:** 1,000  
**Worldwide Confirmed Malfunctions:** 0

	Without Compromised Therapy	With Compromised Therapy	Total
<b>Electrical</b>	-	-	<b>0</b>
<b>Mechanical</b>	-	-	<b>0</b>
<b>Software</b>	-	-	<b>0</b>
<b>Other</b>	-	-	<b>0</b>
Non-patterned	-	-	
<b>WW Confirmed Malfunctions</b>	<b>0</b>	<b>0</b>	<b>0</b>


[More details](#) about malfunctions

[References](#) cited in table above

## ALTRUA 20 SSI

Model S206

<b>U.S. Survival Probability</b>	<b>Worldwide Malfunction Details</b>	<b>Product Advisories</b>
----------------------------------	--------------------------------------	---------------------------

**ALTRUA 20 SSI**  
**Model S206**


**Worldwide Distribution: 7,000**  
**Worldwide Confirmed Malfunctions: 0**

	Without Compromised Therapy	With Compromised Therapy	Total
<b>Electrical</b>	-	-	<b>0</b>
<b>Mechanical</b>	-	-	<b>0</b>
<b>Software</b>	-	-	<b>0</b>
<b>Other</b>	-	-	<b>0</b>
Non-patterned	-	-	
<b>WW Confirmed Malfunctions</b>	<b>0</b>	<b>0</b>	<b>0</b>

[More details](#) about malfunctions

[References](#) cited in table above



CRM PRODUCT PERFORMANCE REPORT Q1 2014


ALTRUA 20 DR

Models S202/S205


U.S. Survival Probability	Worldwide Malfunction Details	Product Advisories
---------------------------	-------------------------------	--------------------

<b>U.S. Summary</b>	
U.S. Registered Implants: 2,000	U.S. Normal Battery Depletions: 11
U.S. Approval Date: April 2008	U.S. Unconfirmed Reports of Premature Battery Depletion : 0
U.S. Estimated Active Implants: 1,000	U.S. Malfunctions:0
	Without Compromised Therapy:0
	With Compromised Therapy:0

**Battery Depletions and Malfunctions**



**Malfunctions Only**




<b>U.S. Survival Probability</b>		Year	1	2	3	4	5	6	7	8	9	10
Non Advisory Population	Depletions and Malfunctions(%) (Confidence Interval)	100.00 (-0.0/+0.0)	100.00 (-0.0/+0.0)	99.83 (-0.5/+0.1)	99.41 (-0.7/+0.3)	98.70 (-1.1/+0.6)	98.70 @ 63 mo. (-1.1/+0.6)	-	-	-	-	
	Malfunctions Only(%) (Confidence Interval)	100.00 (-0.0/+0.0)	100.00 (-0.0/+0.0)	100.00 (-0.0/+0.0)	100.00 (-0.0/+0.0)	100.00 (-0.0/+0.0)	100.00 @ 63 mo. (-0.0/+0.0)	-	-	-	-	
Registered Implants: 2000	Effective Sample Size	1446	1254	1049	853	423	233	-	-	-	-	

**ALTRUA 20 DR**

Models S202/S205

<b>U.S. Survival Probability</b>	<b>Worldwide Malfunction Details</b>	<b>Product Advisories</b>
----------------------------------	--------------------------------------	---------------------------

**ALTRUA 20 DR**  
**Models S202/S205**



**Worldwide Distribution: 3,000**  
**Worldwide Confirmed Malfunctions: 0**

	Without Compromised Therapy	With Compromised Therapy	Total
<b>Electrical</b>	-	-	<b>0</b>
<b>Mechanical</b>	-	-	<b>0</b>
<b>Software</b>	-	-	<b>0</b>
<b>Other</b>	-	-	<b>0</b>
Non-patterned	-	-	
<b>WW Confirmed Malfunctions</b>	<b>0</b>	<b>0</b>	<b>0</b>

[More details](#) about malfunctions

[References](#) cited in table above



CRM PRODUCT PERFORMANCE REPORT Q1 2014

ALTRUA 20 DR (downsize)

Model S203

U.S. Survival Probability	Worldwide Malfunction Details	Product Advisories
---------------------------	-------------------------------	--------------------

<b>U.S. Summary</b>	
U.S. Registered Implants: 5,000	U.S. Normal Battery Depletions: 33
U.S. Approval Date: April 2008	U.S. Unconfirmed Reports of Premature Battery Depletion : 2
U.S. Estimated Active Implants: 4,000	U.S. Malfunctions:0
	Without Compromised Therapy:0
	With Compromised Therapy:0



**U.S. Survival Probability**


	Year	1	2	3	4	5	6	7	8	9	10
Non Advisory Population	Depletions and Malfunctions(%) (Confidence Interval)	99.98 (-0.1/+0.0)	99.82 (-0.2/+0.1)	99.46 (-0.3/+0.2)	98.88 (-0.5/+0.4)	97.97 (-1.1/+0.7)	97.63 @ 62 mo. (-1.4/+0.9)	-	-	-	-
Registered Implants: 5000	Malfunctions Only(%) (Confidence Interval)	100.00 (-0.0/+0.0)	100.00 (-0.0/+0.0)	100.00 (-0.0/+0.0)	100.00 (-0.0/+0.0)	100.00 (-0.0/+0.0)	100.00 @ 62 mo. (-0.0/+0.0)	-	-	-	-
	Effective Sample Size	4417	3574	2369	1274	341	233	-	-	-	-

**ALTRUA 20 DR (downsize)**

Model S203

<b>U.S. Survival Probability</b>	<b>Worldwide Malfunction Details</b>	<b>Product Advisories</b>
----------------------------------	--------------------------------------	---------------------------

**ALTRUA 20 DR (downsize)**  
**Model S203**



**Worldwide Distribution:** 16,000  
**Worldwide Confirmed Malfunctions:** 3

	Without Compromised Therapy	With Compromised Therapy	Total
<b>Electrical</b>	<b>2</b>	<b>-</b>	<b>2</b>
<sup>32</sup> Capacitor	2	-	
<b>Mechanical</b>	<b>-</b>	<b>-</b>	<b>0</b>
<b>Software</b>	<b>-</b>	<b>-</b>	<b>0</b>
<b>Other</b>	<b>-</b>	<b>1</b>	<b>1</b>
Non-patterned	-	-	
<sup>72</sup> Battery depletion	-	1	
<b>WW Confirmed Malfunctions</b>	<b>2</b>	<b>1</b>	<b>3</b>

[More details](#) about malfunctions

[References](#) cited in table above


CRM PRODUCT PERFORMANCE REPORT Q1 2014

ALTRUA 20 DR EL

Model S208

U.S. Survival Probability	Worldwide Malfunction Details	Product Advisories
---------------------------	-------------------------------	--------------------

<b>U.S. Summary</b>	
U.S. Registered Implants: 3,000	U.S. Normal Battery Depletions: 5
U.S. Approval Date: April 2008	U.S. Unconfirmed Reports of Premature Battery Depletion : 0
U.S. Estimated Active Implants: 2,000	U.S. Malfunctions:1
	Without Compromised Therapy:0
	With Compromised Therapy:1




<b>U.S. Survival Probability</b>		Year	1	2	3	4	5	6	7	8	9	10
Non Advisory Population	Depletions and Malfunctions(%) (Confidence Interval)	99.93 (-0.2/+0.0)	99.85 (-0.3/+0.1)	99.71 (-0.4/+0.2)	99.71 (-0.4/+0.2)	99.71 @ 52 mo. (-0.4/+0.2)	-	-	-	-	-	-
	Malfunctions Only(%) (Confidence Interval)	99.97 (-0.2/+0.0)	99.97 (-0.2/+0.0)	99.97 (-0.2/+0.0)	99.97 (-0.2/+0.0)	99.97 @ 52 mo. (-0.2/+0.0)	-	-	-	-	-	-
Registered Implants: 3000	Effective Sample Size	2770	2114	1206	450	243	-	-	-	-	-	-

**ALTRUA 20 DR EL**

Model S208

<b>U.S. Survival Probability</b>	<b>Worldwide Malfunction Details</b>	<b>Product Advisories</b>
----------------------------------	--------------------------------------	---------------------------

**ALTRUA 20 DR EL**  
**Model S208**



**Worldwide Distribution:** 10,000  
**Worldwide Confirmed Malfunctions:** 1

	Without Compromised Therapy	With Compromised Therapy	Total
<b>Electrical</b>	-	1	1
<sup>32</sup> Capacitor	-	1	
<b>Mechanical</b>	-	-	0
<b>Software</b>	-	-	0
<b>Other</b>	-	-	0
Non-patterned	-	-	
<b>WW Confirmed Malfunctions</b>	<b>0</b>	<b>1</b>	<b>1</b>

[More details](#) about malfunctions

[References](#) cited in table above


**ALTRUA 20 SR**

Models S201/S204


U.S. Survival Probability	Worldwide Malfunction Details	Product Advisories
---------------------------	-------------------------------	--------------------

<b>U.S. Summary</b>	
U.S. Registered Implants: 4,000	U.S. Normal Battery Depletions: 16
U.S. Approval Date: April 2008	U.S. Unconfirmed Reports of Premature Battery Depletion : 1
U.S. Estimated Active Implants: 3,000	U.S. Malfunctions:0
	Without Compromised Therapy:0
	With Compromised Therapy:0

**Battery Depletions and Malfunctions**



**Malfunctions Only**




U.S. Survival Probability		Year	1	2	3	4	5	6	7	8	9	10
Non Advisory Population	Depletions and Malfunctions(%) (Confidence Interval)	Registered Implants: 4000	99.95 (-0.2/+0.0)	99.88 (-0.2/+0.1)	99.63 (-0.4/+0.2)	99.20 (-0.6/+0.3)	98.81 (-0.9/+0.5)	-	-	-	-	-
	Malfunctions Only(%) (Confidence Interval)	Effective Sample Size	100.00 (-0.0/+0.0)	100.00 (-0.0/+0.0)	100.00 (-0.0/+0.0)	100.00 (-0.0/+0.0)	100.00 (-0.0/+0.0)	-	-	-	-	-
			3507	2715	1695	855	233	-	-	-	-	-

**ALTRUA 20 SR**

Models S201/S204

<b>U.S. Survival Probability</b>	<b>Worldwide Malfunction Details</b>	<b>Product Advisories</b>
----------------------------------	--------------------------------------	---------------------------

**ALTRUA 20 SR**  
**Models S201/S204**



**Worldwide Distribution:** 23,000  
**Worldwide Confirmed Malfunctions:** 2

	Without Compromised Therapy	With Compromised Therapy	Total
<b>Electrical</b>	<b>1</b>	<b>-</b>	<b>1</b>
<sup>32</sup> Capacitor	1	-	
<b>Mechanical</b>	<b>-</b>	<b>-</b>	<b>0</b>
<b>Software</b>	<b>-</b>	<b>-</b>	<b>0</b>
<b>Other</b>	<b>-</b>	<b>1</b>	<b>1</b>
Non-patterned	-	1	
<b>WW Confirmed Malfunctions</b>	<b>1</b>	<b>1</b>	<b>2</b>

[More details](#) about malfunctions

[References](#) cited in table above




**INSIGNIA Ultra DR**

Model 1291

<b>U.S. Survival Probability</b>	<b>Worldwide Malfunction Details</b>	<b>Product Advisories</b>
----------------------------------	--------------------------------------	---------------------------

<b>U.S. Summary</b>	
U.S. Registered Implants: 32,000	U.S. Normal Battery Depletions: 1,182
U.S. Approval Date: November 2003	U.S. Unconfirmed Reports of Premature Battery Depletion : 17
U.S. Estimated Active Implants: 17,000	U.S. Malfunctions:108
	Without Compromised Therapy:99
	With Compromised Therapy:9



<b>U.S. Survival Probability</b>		Year	1	2	3	4	5	6	7	8	9	10	
Non Advisory Population Registered Implants: 24000	Depletions and Malfunctions(%) (Confidence Interval)	99.98 (-0.0/+0.0)	99.94 (-0.0/+0.0)	99.84 (-0.1/+0.0)	99.51 (-0.1/+0.1)	98.73 (-0.2/+0.2)	97.39 (-0.3/+0.3)	94.92 (-0.5/+0.5)	87.66 (-1.6/+1.4)	76.53 @ 104 mo. (-3.6/+3.3)	—	—	
	Malfunctions Only(%) (Confidence Interval)	99.99 (-0.0/+0.0)	99.97 (-0.0/+0.0)	99.95 (-0.0/+0.0)	99.92 (-0.1/+0.0)	99.82 (-0.1/+0.1)	99.71 (-0.1/+0.1)	99.50 (-0.2/+0.1)	99.25 (-0.4/+0.3)	99.02 @ 104 mo. (-0.7/+0.4)	—	—	
	Effective Sample Size	21004	18658	16561	14650	12886	8440	3158	784	203	—	—	
23-Jun-06 and 24-Aug-06 Low Voltage Capacitor* Registered Implants: 2000	Depletions and Malfunctions(%) (Confidence Interval)	99.90 (-0.3/+0.1)	99.79 (-0.3/+0.1)	99.59 (-0.5/+0.2)	99.30 (-0.6/+0.3)	98.80 (-0.7/+0.5)	97.57 (-1.0/+0.7)	94.73 (-1.5/+1.2)	91.09 @ 94 mo. (-2.1/+1.7)	—	—	—	
	Malfunctions Only(%) (Confidence Interval)	99.95 (-0.3/+0.0)	99.95 (-0.3/+0.0)	99.95 (-0.3/+0.0)	99.95 (-0.3/+0.0)	99.95 (-0.3/+0.0)	99.95 (-0.3/+0.0)	99.63 (-0.6/+0.2)	99.36 @ 94 mo. (-0.8/+0.4)	—	—	—	
	Effective Sample Size	1878	1659	1461	1287	1134	988	850	302	—	—	—	
22-Sep-05 Crystal Timing Component (Failure Mode 2)* Registered Implants:	Depletions and Malfunctions(%) (Confidence Interval)	100.00 (-0.0/+0.0)	99.92 (-0.1/+0.0)	99.77 (-0.2/+0.1)	99.37 (-0.3/+0.2)	98.52 (-0.4/+0.3)	96.63 (-0.7/+0.5)	93.50 (-0.9/+0.8)	87.26 (-1.3/+1.2)	73.04 (-2.3/+2.2)	68.57 @ 112 mo. (-2.8/+2.7)	—	—
	Malfunctions Only(%) (Confidence Interval)	100.00 (-0.0/+0.0)	99.92 (-0.1/+0.0)	99.77 (-0.2/+0.1)	99.37 (-0.3/+0.2)	98.52 (-0.4/+0.3)	96.63 (-0.7/+0.5)	93.50 (-0.9/+0.8)	87.26 (-1.3/+1.2)	73.04 (-2.3/+2.2)	68.57 @ 112 mo. (-2.8/+2.7)	—	—
	Effective Sample Size	—	—	—	—	—	—	—	—	—	—	—	

6000											
	Malfunctions Only (%) (Confidence Interval)	100.00 (-0.0/+0.0)	99.98 (-0.1/+0.0)	99.96 (-0.1/+0.0)	99.89 (-0.2/+0.1)	99.78 (-0.2/+0.1)	99.62 (-0.3/+0.2)	99.15 (-0.4/+0.3)	98.58 (-0.5/+0.4)	97.83 (-0.9/+0.6)	97.83 @ 112 mo. (-0.9/+0.6)
	Effective Sample Size	5702	5045	4467	3940	3453	2980	2556	2098	545	275


\*Devices subject to an advisory. Refer to the Advisories for more details. Devices may be part of more than one advisory.

Data are representative of Boston Scientific INSIGNIA Ultra and Intermedics NEXUS Ultra device performance.

**INSIGNIA Ultra DR**

Model 1291

<b>U.S. Survival Probability</b>	<b>Worldwide Malfunction Details</b>	<b>Product Advisories</b>
----------------------------------	--------------------------------------	---------------------------

<b>INSIGNIA Ultra DR</b> <b>Model 1291</b> 			
<b>Worldwide Distribution:</b> 51,000			
<b>Worldwide Confirmed Malfunctions:</b> 133			
	Without Compromised Therapy	With Compromised Therapy	Total
<b>Electrical</b>	<b>7</b>	<b>5</b>	<b>12</b>
<sup>8</sup> Low-voltage capacitor (Advisory issued)	-	2	
<sup>28</sup> Capacitor	1	-	
<sup>32</sup> Capacitor	4	2	
<sup>89</sup> Integrated circuit	2	1	
<b>Mechanical</b>	<b>7</b>	<b>5</b>	<b>12</b>
<sup>47</sup> Seal plug	5	4	
<sup>48</sup> Header	1	1	
<sup>93</sup> Setscrew	1	-	
<b>Software</b>	<b>3</b>	<b>-</b>	<b>3</b>
<sup>97</sup> Underestimation of battery status	2	-	
<sup>99</sup> Pacing rate limit	1	-	
<b>Other</b>	<b>102</b>	<b>4</b>	<b>106</b>
Non-patterned	6	3	
<sup>22</sup> Longevity labeling	67	-	
<sup>50</sup> Magnet response	1	-	
<sup>72</sup> Battery depletion	2	1	
<sup>120</sup> Battery status	26	-	
<b>WW Confirmed Malfunctions</b>	<b>119</b>	<b>14</b>	<b>133</b>

[More details](#) about malfunctions


[References](#) cited in table above

**INSIGNIA Ultra DR (downsize)**

Model 1290

<b>U.S. Survival Probability</b>	<b>Worldwide Malfunction Details</b>	<b>Product Advisories</b>
----------------------------------	--------------------------------------	---------------------------

<b>U.S. Summary</b>	
U.S. Registered Implants: 76,000	U.S. Normal Battery Depletions: 11,548
U.S. Approval Date: November 2003	U.S. Unconfirmed Reports of Premature Battery Depletion : 102
U.S. Estimated Active Implants: 27,000	U.S. Malfunctions:383
	Without Compromised Therapy:372
	With Compromised Therapy:11



<b>U.S. Survival Probability</b>		Year	1	2	3	4	5	6	7	8	9	10
Non Advisory Population Registered Implants: 54000	Depletions and Malfunctions(%) (Confidence Interval)	99.96 (-0.0/+0.0)	99.86 (-0.0/+0.0)	99.49 (-0.1/+0.1)	98.55 (-0.1/+0.1)	95.94 (-0.2/+0.2)	88.79 (-0.4/+0.4)	72.25 (-0.7/+0.7)	50.24 (-1.4/+1.4)	35.81 @ 104 mo. (-2.3/+2.4)	—	—
	Malfunctions Only(%) (Confidence Interval)	99.99 (-0.0/+0.0)	99.97 (-0.0/+0.0)	99.94 (-0.0/+0.0)	99.83 (-0.0/+0.0)	99.61 (-0.1/+0.1)	99.37 (-0.1/+0.1)	99.16 (-0.1/+0.1)	99.06 (-0.2/+0.1)	99.06 @ 104 mo. (-0.2/+0.1)	—	—
	Effective Sample Size	47640	42293	37435	32962	28467	18096	6095	1135	211	—	—
23-Jun-06 and 24-Aug-06 Low Voltage Capacitor* Registered Implants: 5000	Depletions and Malfunctions(%) (Confidence Interval)	99.88 (-0.2/+0.1)	99.75 (-0.2/+0.1)	99.51 (-0.3/+0.2)	98.41 (-0.5/+0.4)	95.67 (-0.8/+0.7)	88.21 (-1.4/+1.2)	73.29 (-1.9/+1.8)	57.61 @ 94 mo. (-2.4/+2.3)	—	—	—
	Malfunctions Only(%) (Confidence Interval)	99.95 (-0.1/+0.0)	99.95 (-0.1/+0.0)	99.95 (-0.1/+0.0)	99.78 (-0.2/+0.1)	99.37 (-0.4/+0.2)	99.10 (-0.5/+0.3)	98.76 (-0.6/+0.4)	98.68 @ 94 mo. (-0.6/+0.4)	—	—	—
	Effective Sample Size	4025	3553	3142	2733	2340	1910	1385	281	—	—	—
22-Sep-05 Crystal Timing Component (Failure Mode 1)*	Survival probability data not provided because this population does not meet report inclusion criteria (see Statistical Methodology for more details). Refer to Product Advisories for more information.											
22-Sep-05 Crystal Timing	Depletions and Malfunctions(%)	99.96 (-0.0/+0.0)	99.87 (-0.1/+0.0)	99.38 (-0.2/+0.1)	97.77 (-0.3/+0.3)	93.64 (-0.5/+0.5)	82.84 (-0.8/+0.8)	64.60 (-1.1/+1.1)	44.77 (-1.2/+1.2)	27.00 (-1.5/+1.5)	24.46 @ 110 mo.	—

Component (Failure Mode 2)*	(Confidence Interval)											(-1.6/+1.7)
Registered Implants: 17000												
Malfunctions Only (%)	99.98	99.98	99.95	99.73	99.09	98.67	98.45	98.25	97.94	97.94		
(Confidence Interval)	(-0.0/+0.0)	(-0.0/+0.0)	(-0.1/+0.0)	(-0.1/+0.1)	(-0.2/+0.2)	(-0.3/+0.2)	(-0.3/+0.2)	(-0.3/+0.3)	(-0.7/+0.5)	@ 110 mo. (-0.7/+0.5)		
Effective Sample Size	14976	13297	11732	10224	8613	6646	4429	2608	401	227		


\*Devices subject to an advisory. Refer to the Advisories for more details. Devices may be part of more than one advisory.

Data are representative of Boston Scientific INSIGNIA Ultra and Intermedics NEXUS Ultra device performance.

**INSIGNIA Ultra DR (downsize)**

Model 1290

<b>U.S. Survival Probability</b>	<b>Worldwide Malfunction Details</b>	<b>Product Advisories</b>
----------------------------------	--------------------------------------	---------------------------

INSIGNIA Ultra DR (downsize) Model 1290 			
<b>Worldwide Distribution: 124,000</b>			
<b>Worldwide Confirmed Malfunctions: 518</b>			
	Without Compromised Therapy	With Compromised Therapy	Total
<b>Electrical</b>	<b>9</b>	<b>9</b>	<b>18</b>
<sup>8</sup> Low-voltage capacitor (Advisory issued)	1	5	
<sup>32</sup> Capacitor	7	3	
<sup>89</sup> Integrated circuit	1	1	
<b>Mechanical</b>	<b>4</b>	<b>2</b>	<b>6</b>
<sup>13</sup> Crystal timing component Failure Mode 2 (Advisory issued)	-	1	
<sup>31</sup> Setscrew thread depth	1	-	
<sup>47</sup> Seal plug	2	1	
<sup>66</sup> Circuit connection	1	-	
<b>Software</b>	<b>12</b>	<b>-</b>	<b>12</b>
<sup>59</sup> Rate fault declaration	1	-	
<sup>60</sup> Memory error	2	-	
<sup>97</sup> Underestimation of battery status	8	-	
<sup>99</sup> Pacing rate limit	1	-	
<b>Other</b>	<b>475</b>	<b>7</b>	<b>482</b>
Non-patterned	22	5	
<sup>22</sup> Longevity labeling	388	-	
<sup>72</sup> Battery depletion	6	2	
<sup>120</sup> Battery status	59	-	
<b>WW Confirmed Malfunctions</b>	<b>500</b>	<b>18</b>	<b>518</b>

[More details](#) about malfunctions

[References](#) cited in table above


**INSIGNIA Ultra SR**

Model 1190

<b>U.S. Survival Probability</b>	<b>Worldwide Malfunction Details</b>	<b>Product Advisories</b>
----------------------------------	--------------------------------------	---------------------------

**U.S. Summary**

U.S. Registered Implants: 24,000	U.S. Normal Battery Depletions: 1,017
U.S. Approval Date: November 2003	U.S. Unconfirmed Reports of Premature Battery Depletion : 8
U.S. Estimated Active Implants: 8,000	U.S. Malfunctions:29
	Without Compromised Therapy:25
	With Compromised Therapy:4



**U.S. Survival Probability**

	Year	1	2	3	4	5	6	7	8	9	10
Non Advisory Population Registered Implants: 17000	Depletions and Malfunctions(%) (Confidence Interval)	99.97 (-0.0/+0.0)	99.93 (-0.1/+0.0)	99.72 (-0.1/+0.1)	99.42 (-0.2/+0.1)	98.74 (-0.3/+0.2)	97.57 (-0.4/+0.3)	92.50 (-0.9/+0.8)	81.58 (-2.4/+2.1)	74.94 @ 100 mo. (-3.6/+3.3)	—
	Malfunctions Only(%) (Confidence Interval)	99.99 (-0.0/+0.0)	99.99 (-0.0/+0.0)	99.98 (-0.1/+0.0)	99.97 (-0.1/+0.0)	99.94 (-0.1/+0.0)	99.91 (-0.1/+0.0)	99.78 (-0.2/+0.1)	99.71 (-0.3/+0.1)	99.71 @ 100 mo. (-0.3/+0.1)	—
	Effective Sample Size	14155	12095	10319	8881	7697	5166	1975	458	206	—
23-Jun-06 and 24-Aug-06 Low Voltage Capacitor* Registered Implants: 1000	Depletions and Malfunctions(%) (Confidence Interval)	99.83 (-0.5/+0.1)	99.73 (-0.6/+0.2)	99.51 (-0.7/+0.3)	99.11 (-0.9/+0.5)	98.47 (-1.2/+0.7)	97.22 (-1.6/+1.0)	93.07 (-2.5/+1.9)	86.83 @ 93 mo. (-3.5/+2.9)	—	—
	Malfunctions Only(%) (Confidence Interval)	99.91 (-0.5/+0.1)	99.91 (-0.5/+0.1)	99.91 (-0.5/+0.1)	99.77 (-0.7/+0.2)	99.77 (-0.7/+0.2)	99.59 (-0.9/+0.3)	99.59 (-0.9/+0.3)	99.59 @ 93 mo. (-0.9/+0.3)	—	—
	Effective Sample Size	1148	964	813	701	590	504	422	201	—	—
22-Sep-05 Crystal Timing Component (Failure Mode 1)*	Survival probability data not provided because this population does not meet report inclusion criteria (see Statistical Methodology for more details). Refer to Product Advisories for more information.										
22-Sep-05 Crystal Timing	Depletions and Malfunctions(%) (Confidence Interval)	100.00 (-0.0/+0.0)	99.98 (-0.1/+0.0)	99.85 (-0.2/+0.1)	99.27 (-0.4/+0.3)	98.32 (-0.6/+0.4)	96.29 (-0.9/+0.7)	89.42 (-1.5/+1.3)	76.50 (-2.2/+2.0)	63.73 (-3.1/+3.0)	—

Component (Failure Mode 2)*	(Confidence Interval)										
Registered Implants: 5000											
Malfunctions Only (%)	100.00	99.98	99.98	99.87	99.83	99.78	99.48	99.40	98.86	-	
(Confidence Interval)	(-0.0/+0.0)	(-0.1/+0.0)	(-0.1/+0.0)	(-0.2/+0.1)	(-0.2/+0.1)	(-0.3/+0.1)	(-0.4/+0.2)	(-0.5/+0.3)	(-1.4/+0.6)		
Effective Sample Size	4144	3558	3004	2532	2115	1772	1426	1051	230	-	

\*Devices subject to an advisory. Refer to the Advisories for more details. Devices may be part of more than one advisory.


Data are representative of Boston Scientific INSIGNIA Ultra and Intermedics NEXUS Ultra device performance.



**INSIGNIA Ultra SR**

Model 1190

<b>U.S. Survival Probability</b>	<b>Worldwide Malfunction Details</b>	<b>Product Advisories</b>
----------------------------------	--------------------------------------	---------------------------

<b>INSIGNIA Ultra SR</b> <b>Model 1190</b> 			
<b>Worldwide Distribution: 48,000</b>			
<b>Worldwide Confirmed Malfunctions: 47</b>			
	Without Compromised Therapy	With Compromised Therapy	Total
<b>Electrical</b>	<b>2</b>	<b>5</b>	<b>7</b>
<sup>8</sup> Low-voltage capacitor (Advisory issued)	1	3	
<sup>32</sup> Capacitor	1	-	
<sup>89</sup> Integrated circuit	-	2	
<b>Mechanical</b>	<b>3</b>	<b>1</b>	<b>4</b>
<sup>47</sup> Seal plug	3	-	
<sup>48</sup> Header	-	1	
<b>Software</b>	<b>1</b>	<b>-</b>	<b>1</b>
<sup>60</sup> Memory error	1	-	
<b>Other</b>	<b>35</b>	<b>-</b>	<b>35</b>
Non-patterned	1	-	
<sup>22</sup> Longevity labeling	23	-	
<sup>72</sup> Battery depletion	1	-	
<sup>120</sup> Battery status	10	-	
<b>WW Confirmed Malfunctions</b>	<b>41</b>	<b>6</b>	<b>47</b>

[More details](#) about malfunctions


[References](#) cited in table above

**INSIGNIA Entra DR**

Models 1294/1295

<b>U.S. Survival Probability</b>	<b>Worldwide Malfunction Details</b>	<b>Product Advisories</b>
----------------------------------	--------------------------------------	---------------------------

<b>U.S. Summary</b>	
U.S. Registered Implants: 17,000	U.S. Normal Battery Depletions: 760
U.S. Approval Date: March 2002	U.S. Unconfirmed Reports of Premature Battery Depletion : 10
U.S. Estimated Active Implants: 6,000	U.S. Malfunctions:55
	Without Compromised Therapy:49
	With Compromised Therapy:6



<b>U.S. Survival Probability</b>		Year	1	2	3	4	5	6	7	8	9	10	
Non Advisory Population	Depletions and Malfunctions(%) (Confidence Interval)	99.97 (-0.1/+0.0)	99.87 (-0.1/+0.1)	99.75 (-0.2/+0.1)	99.51 (-0.2/+0.2)	98.72 (-0.4/+0.3)	97.04 (-0.6/+0.5)	95.58 (-0.8/+0.7)	92.95 (-1.5/+1.3)	91.82 @ 101 mo. (-2.1/+1.7)	—	—	
	Malfunctions Only(%) (Confidence Interval)	100.00 (-0.0/+0.0)	99.97 (-0.1/+0.0)	99.91 (-0.1/+0.1)	99.91 (-0.1/+0.1)	99.83 (-0.2/+0.1)	99.77 (-0.2/+0.1)	99.72 (-0.2/+0.1)	99.23 (-0.9/+0.4)	99.23 @ 101 mo. (-0.9/+0.4)	—	—	
	Effective Sample Size	6258	5546	4914	4355	3771	2668	1291	453	226	—	—	
23-Jun-06 and 24-Aug-06 Low Voltage Capacitor*	Depletions and Malfunctions(%) (Confidence Interval)	100.00 (-0.0/+0.0)	100.00 (-0.0/+0.0)	99.45 (-1.1/+0.4)	99.24 (-1.3/+0.5)	98.76 (-1.5/+0.7)	97.66 (-2.0/+1.1)	97.33 (-2.1/+1.2)	94.95 @ 92 mo. (-2.9/+1.9)	—	—	—	
	Malfunctions Only(%) (Confidence Interval)	100.00 (-0.0/+0.0)	100.00 (-0.0/+0.0)	99.82 (-1.1/+0.2)	99.61 (-1.2/+0.3)	99.61 (-1.2/+0.3)	99.61 (-1.2/+0.3)	99.61 (-1.2/+0.3)	99.61 @ 92 mo. (-1.2/+0.3)	—	—	—	
	Effective Sample Size	693	607	529	452	394	338	295	237	—	—	—	
22-Sep-05 Crystal Timing Component (Failure Mode 1)*	Depletions and Malfunctions(%) (Confidence Interval)	99.83 (-0.4/+0.1)	99.69 (-0.4/+0.2)	99.46 (-0.5/+0.3)	99.19 (-0.7/+0.4)	98.09 (-1.0/+0.7)	96.06 (-1.4/+1.1)	93.81 (-1.8/+1.4)	91.04 (-2.2/+1.8)	85.50 (-2.9/+2.5)	75.81 (-3.8/+3.4)	—	—
	Malfunctions Only(%) (Confidence Interval)	100.00 (-0.0/+0.0)	100.00 (-0.0/+0.0)	100.00 (-0.0/+0.0)	100.00 (-0.0/+0.0)	100.00 (-0.0/+0.0)	100.00 (-0.0/+0.0)	100.00 (-0.0/+0.0)	100.00 (-0.0/+0.0)	100.00 (-0.0/+0.0)	100.00 (-0.0/+0.0)	100.00 (-0.0/+0.0)	100.00 (-0.0/+0.0)
	Effective Sample Size	—	—	—	—	—	—	—	—	—	—	—	

2000											
	Malfunctions Only (%) (Confidence Interval)	99.83 (-0.4/+0.1)	99.83 (-0.4/+0.1)	99.83 (-0.4/+0.1)	99.83 (-0.4/+0.1)	99.83 (-0.4/+0.1)	99.83 (-0.4/+0.1)	99.67 (-0.6/+0.2)	99.19 (-1.0/+0.5)	99.19 (-1.0/+0.5)	98.94 (-1.2/+0.6)
	Effective Sample Size	1675	1453	1213	1063	923	785	662	555	446	293
22-Sep-05 Crystal Timing Component (Failure Mode 2)*	Depletions and Malfunctions (%) (Confidence Interval)	99.97 (-0.1/+0.0)	99.90 (-0.1/+0.1)	99.74 (-0.2/+0.1)	99.34 (-0.3/+0.2)	98.43 (-0.4/+0.3)	96.95 (-0.6/+0.5)	94.63 (-0.8/+0.7)	91.60 (-1.0/+0.9)	85.38 (-1.5/+1.4)	75.94 @ 119 mo. (-2.8/+2.6)
Registered Implants: 7000	Malfunctions Only (%) (Confidence Interval)	99.99 (-0.1/+0.0)	99.95 (-0.1/+0.0)	99.95 (-0.1/+0.0)	99.91 (-0.1/+0.1)	99.83 (-0.2/+0.1)	99.71 (-0.2/+0.1)	99.47 (-0.3/+0.2)	99.31 (-0.4/+0.2)	98.98 (-0.5/+0.3)	98.84 @ 119 mo. (-0.6/+0.4)
	Effective Sample Size	6210	5482	4824	4230	3695	3189	2683	2279	1156	254


\*Devices subject to an advisory. Refer to the Advisories for more details. Devices may be part of more than one advisory.

Data are representative of Boston Scientific INSIGNIA Entra and Intermedics NEXUS Entra device performance.

**INSIGNIA Entra DR**

Models 1294/1295

<b>U.S. Survival Probability</b>	<b>Worldwide Malfunction Details</b>	<b>Product Advisories</b>
----------------------------------	--------------------------------------	---------------------------

<b>INSIGNIA Entra DR</b> <b>Models 1294/1295</b> 			
<b>Worldwide Distribution:</b> 36,000			
<b>Worldwide Confirmed Malfunctions:</b> 65			
	Without Compromised Therapy	With Compromised Therapy	Total
<b>Electrical</b>	-	<b>3</b>	<b>3</b>
<sup>27</sup> Integrated circuit	-	1	
<sup>32</sup> Capacitor	-	1	
<sup>89</sup> Integrated circuit	-	1	
<b>Mechanical</b>	<b>3</b>	<b>7</b>	<b>10</b>
<sup>12</sup> Crystal timing component Failure Mode 1 (Advisory issued)	-	5	
<sup>47</sup> Seal plug	3	-	
<sup>48</sup> Header	-	2	
<b>Software</b>	-	-	<b>0</b>
<b>Other</b>	<b>51</b>	<b>1</b>	<b>52</b>
Non-patterned	4	1	
<sup>22</sup> Longevity labeling	45	-	
<sup>120</sup> Battery status	2	-	
<b>WW Confirmed Malfunctions</b>	<b>54</b>	<b>11</b>	<b>65</b>

[More details](#) about malfunctions


[References](#) cited in table above

**INSIGNIA Entra DR (downsize)**

Model 1296

<b>U.S. Survival Probability</b>	<b>Worldwide Malfunction Details</b>	<b>Product Advisories</b>
----------------------------------	--------------------------------------	---------------------------

<b>U.S. Summary</b>	
U.S. Registered Implants: 24,000	U.S. Normal Battery Depletions: 3,735
U.S. Approval Date: March 2002	U.S. Unconfirmed Reports of Premature Battery Depletion : 25
U.S. Estimated Active Implants: 5,000	U.S. Malfunctions:96
	Without Compromised Therapy:90
	With Compromised Therapy:6



<b>U.S. Survival Probability</b>		Year	1	2	3	4	5	6	7	8	9	10	
Non Advisory Population	Depletions and Malfunctions(%)	99.96	99.85	99.43	98.43	96.60	93.07	83.52	64.68	55.38	—	—	
	(Confidence Interval)	(-0.1/+0.0)	(-0.1/+0.1)	(-0.2/+0.2)	(-0.4/+0.3)	(-0.5/+0.5)	(-0.8/+0.7)	(-1.5/+1.4)	(-2.9/+2.8)	(-3.7/+3.6)	@ 101 mo.	(-0.4/+0.3)	
	Registered Implants: 8000												
23-Jun-06 and 24-Aug-06 Low Voltage Capacitor*	Depletions and Malfunctions(%)	100.00	99.70	99.20	97.84	95.61	91.90	83.05	73.35	—	—	—	
	(Confidence Interval)	(-0.0/+0.0)	(-0.9/+0.2)	(-1.1/+0.5)	(-1.6/+0.9)	(-2.2/+1.5)	(-3.0/+2.2)	(-4.2/+3.5)	(-5.1/+4.5)	@ 91 mo.	(-1.7/+0.4)	(-3.2/+3.2)	(-3.2/+3.3)
	Registered Implants: 1000												
22-Sep-05 Crystal Timing Component (Failure Mode 1)*	Depletions and Malfunctions(%)	100.00	100.00	100.00	100.00	99.75	99.75	99.43	99.43	—	—	—	
	(Confidence Interval)	(-0.0/+0.0)	(-0.0/+0.0)	(-0.0/+0.0)	(-0.0/+0.0)	(-1.5/+0.2)	(-1.5/+0.2)	(-1.7/+0.4)	(-1.7/+0.4)	@ 91 mo.	(-1.7/+0.4)	(-3.2/+3.3)	
	Registered Implants: 763												
		Effective Sample Size	7139	6282	5501	4786	4099	2900	1336	384	201	—	
		Effective Sample Size	763	657	563	476	402	330	254	203	—	—	

3000											
	Malfunctions Only (%) (Confidence Interval)	99.90 (-0.2/+0.1)	99.90 (-0.2/+0.1)	99.90 (-0.2/+0.1)	99.90 (-0.2/+0.1)	99.84 (-0.3/+0.1)	99.68 (-0.4/+0.2)	99.59 (-0.5/+0.2)	99.48 (-0.6/+0.3)	99.48 (-0.6/+0.3)	99.48 @ 119 mo. (-0.6/+0.3)
	Effective Sample Size	2736	2405	2071	1813	1515	1228	935	599	362	205
22-Sep-05 Crystal Timing Component (Failure Mode 2)*	Depletions and Malfunctions (%) (Confidence Interval)	99.96 (-0.1/+0.0)	99.86 (-0.1/+0.1)	99.56 (-0.2/+0.1)	98.52 (-0.3/+0.3)	96.46 (-0.5/+0.4)	90.85 (-0.8/+0.7)	79.04 (-1.2/+1.1)	61.35 (-1.5/+1.5)	44.97 (-1.7/+1.7)	35.80 @ 119 mo. (-2.1/+2.1)
Registered Implants: 11000											
	Malfunctions Only (%) (Confidence Interval)	99.98 (-0.1/+0.0)	99.98 (-0.1/+0.0)	99.97 (-0.1/+0.0)	99.83 (-0.1/+0.1)	99.52 (-0.2/+0.1)	99.26 (-0.3/+0.2)	99.14 (-0.3/+0.2)	98.92 (-0.4/+0.3)	98.92 (-0.4/+0.3)	98.92 @ 119 mo. (-0.4/+0.3)
	Effective Sample Size	9584	8452	7365	6367	5506	4515	3345	2192	904	240


\*Devices subject to an advisory. Refer to the Advisories for more details. Devices may be part of more than one advisory.

Data are representative of Boston Scientific INSIGNIA Entra and Intermedics NEXUS Entra device performance.

**INSIGNIA Entra DR (downsize)**

Model 1296

<b>U.S. Survival Probability</b>	<b>Worldwide Malfunction Details</b>	<b>Product Advisories</b>
----------------------------------	--------------------------------------	---------------------------

INSIGNIA Entra DR (downsize) Model 1296 			
<b>Worldwide Distribution:</b> 47,000			
<b>Worldwide Confirmed Malfunctions:</b> 117			
	Without Compromised Therapy	With Compromised Therapy	Total
<b>Electrical</b>	<b>1</b>	<b>4</b>	<b>5</b>
<sup>8</sup> Low-voltage capacitor (Advisory issued)	-	1	
<sup>32</sup> Capacitor	1	-	
<sup>89</sup> Integrated circuit	-	3	
<b>Mechanical</b>	<b>-</b>	<b>3</b>	<b>3</b>
<sup>12</sup> Crystal timing component Failure Mode 1 (Advisory issued)	-	2	
<sup>23</sup> Solder bond	-	1	
<b>Software</b>	<b>4</b>	<b>-</b>	<b>4</b>
<sup>44</sup> Memory error	1	-	
<sup>97</sup> Underestimation of battery status	1	-	
<sup>98</sup> Interrupted telemetry	2	-	
<b>Other</b>	<b>103</b>	<b>2</b>	<b>105</b>
Non-patterned	4	2	
<sup>22</sup> Longevity labeling	95	-	
<sup>72</sup> Battery depletion	1	-	
<sup>120</sup> Battery status	3	-	
<b>WW Confirmed Malfunctions</b>	<b>108</b>	<b>9</b>	<b>117</b>

[More details](#) about malfunctions


[References](#) cited in table above

**INSIGNIA Entra SR**

Models 1195/1198

<b>U.S. Survival Probability</b>	<b>Worldwide Malfunction Details</b>	<b>Product Advisories</b>
----------------------------------	--------------------------------------	---------------------------

<b>U.S. Summary</b>	
U.S. Registered Implants: 14,000	U.S. Normal Battery Depletions: 401
U.S. Approval Date: March 2002	U.S. Unconfirmed Reports of Premature Battery Depletion : 10
U.S. Estimated Active Implants: 3,000	U.S. Malfunctions:9
	Without Compromised Therapy:7
	With Compromised Therapy:2



<b>U.S. Survival Probability</b>		Year	1	2	3	4	5	6	7	8	9	10	
Non Advisory Population	Depletions and Malfunctions(%) (Confidence Interval)	99.90 (-0.1/+0.1)	99.85 (-0.2/+0.1)	99.77 (-0.2/+0.1)	99.53 (-0.3/+0.2)	99.36 (-0.4/+0.2)	98.72 (-0.5/+0.4)	98.15 (-0.7/+0.5)	95.87 (-1.8/+1.3)	94.99 @ 97 mo. (-2.3/+1.6)	-	-	
	Malfunctions Only(%) (Confidence Interval)	99.94 (-0.1/+0.0)	99.94 (-0.1/+0.0)	99.91 (-0.1/+0.1)	99.91 (-0.1/+0.1)	99.91 (-0.1/+0.1)	99.91 (-0.1/+0.1)	99.77 (-0.4/+0.1)	99.77 (-0.4/+0.1)	99.77 @ 97 mo. (-0.4/+0.1)	-	-	
	Effective Sample Size	4710	3876	3258	2755	2312	1564	747	254	217	-	-	
23-Jun-06 and 24-Aug-06 Low Voltage Capacitor*	Depletions and Malfunctions(%) (Confidence Interval)	99.78 (-1.3/+0.2)	99.10 (-1.9/+0.6)	98.39 (-2.2/+0.9)	97.94 (-2.5/+1.1)	-	-	-	-	-	-	-	
	Malfunctions Only(%) (Confidence Interval)	99.78 (-1.3/+0.2)	99.78 (-1.3/+0.2)	99.78 (-1.3/+0.2)	99.78 (-1.3/+0.2)	-	-	-	-	-	-	-	
	Effective Sample Size	348	284	237	204	-	-	-	-	-	-	-	
22-Sep-05 Crystal Timing Component (Failure Mode 1)*	Depletions and Malfunctions(%) (Confidence Interval)	99.93 (-0.4/+0.1)	99.84 (-0.5/+0.1)	99.50 (-0.7/+0.3)	99.21 (-0.9/+0.4)	98.19 (-1.3/+0.8)	96.96 (-1.8/+1.1)	95.73 (-2.1/+1.4)	93.95 (-2.6/+1.9)	89.42 (-3.7/+2.8)	85.84 @ 113 mo. (-4.3/+3.4)	-	-
	Malfunctions Only(%) (Confidence Interval)	99.93 (-0.4/+0.1)	99.84 (-0.5/+0.1)	99.50 (-0.7/+0.3)	99.21 (-0.9/+0.4)	98.19 (-1.3/+0.8)	96.96 (-1.8/+1.1)	95.73 (-2.1/+1.4)	93.95 (-2.6/+1.9)	89.42 (-3.7/+2.8)	85.84 @ 113 mo. (-4.3/+3.4)	-	-
	Effective Sample Size	2000	-	-	-	-	-	-	-	-	-	-	



	Malfunctions Only(%) (Confidence Interval)	100.00 (-0.0/+0.0)	100.00 (-0.0/+0.0)	100.00 (-0.0/+0.0)	100.00 (-0.0/+0.0)	100.00 (-0.0/+0.0)	100.00 (-0.0/+0.0)	100.00 (-0.0/+0.0)	100.00 (-0.0/+0.0)	100.00 (-0.0/+0.0)	100.00 @ 113 mo. (-0.0/+0.0)
	Effective Sample Size	1216	999	807	662	550	447	356	298	240	204
22-Sep-05 Crystal Timing Component (Failure Mode 2)*	Depletions and Malfunctions(%) (Confidence Interval)	99.98 (-0.1/+0.0)	99.91 (-0.2/+0.1)	99.62 (-0.3/+0.2)	98.84 (-0.4/+0.3)	97.99 (-0.6/+0.5)	97.00 (-0.8/+0.6)	95.32 (-1.0/+0.8)	92.26 (-1.3/+1.2)	84.63 (-2.1/+1.9)	77.91 @ 119 mo. (-3.1/+2.8)
Registered Implants: 6000	Malfunctions Only(%) (Confidence Interval)	100.00 (-0.0/+0.0)	100.00 (-0.0/+0.0)	100.00 (-0.0/+0.0)	100.00 (-0.0/+0.0)	99.96 (-0.2/+0.0)	99.96 (-0.2/+0.0)	99.90 (-0.3/+0.1)	99.90 (-0.3/+0.1)	99.90 (-0.3/+0.1)	99.90 @ 119 mo. (-0.3/+0.1)
	Effective Sample Size	4579	3830	3179	2644	2186	1833	1544	1293	697	219


\*Devices subject to an advisory. Refer to the Advisories for more details. Devices may be part of more than one advisory.

Data are representative of Boston Scientific INSIGNIA Entra and Intermedics NEXUS Entra device performance.

**INSIGNIA Entra SR**

Models 1195/1198

<b>U.S. Survival Probability</b>	<b>Worldwide Malfunction Details</b>	<b>Product Advisories</b>
----------------------------------	--------------------------------------	---------------------------

<b>INSIGNIA Entra SR</b> <b>Models 1195/1198</b> 			
<b>Worldwide Distribution: 52,000</b>			
<b>Worldwide Confirmed Malfunctions: 25</b>			
	Without Compromised Therapy	With Compromised Therapy	Total
<b>Electrical</b>	<b>3</b>	<b>4</b>	<b>7</b>
<sup>8</sup> Low-voltage capacitor (Advisory issued)	-	2	
<sup>32</sup> Capacitor	2	2	
<sup>89</sup> Integrated circuit	1	-	
<b>Mechanical</b>	<b>1</b>	<b>6</b>	<b>7</b>
<sup>12</sup> Crystal timing component Failure Mode 1 (Advisory issued)	1	-	
<sup>13</sup> Crystal timing component Failure Mode 2 (Advisory issued)	-	1	
<sup>39</sup> Capacitor array	-	2	
<sup>47</sup> Seal plug	-	2	
<sup>95</sup> Seal plug	-	1	
<b>Software</b>	<b>-</b>	<b>-</b>	<b>0</b>
<b>Other</b>	<b>10</b>	<b>1</b>	<b>11</b>
Non-patterned	1	1	
<sup>22</sup> Longevity labeling	6	-	
<sup>120</sup> Battery status	3	-	
<b>WW Confirmed Malfunctions</b>	<b>14</b>	<b>11</b>	<b>25</b>

[More details](#) about malfunctions


[References](#) cited in table above

INSIGNIA Plus DR

Model 1297

U.S. Survival Probability	Worldwide Malfunction Details	Product Advisories
---------------------------	-------------------------------	--------------------

<b>U.S. Summary</b>	
U.S. Registered Implants: 27,000	U.S. Normal Battery Depletions: 2,721
U.S. Approval Date: March 2002	U.S. Unconfirmed Reports of Premature Battery Depletion : 18
U.S. Estimated Active Implants: 10,000	U.S. Malfunctions:116
	Without Compromised Therapy:107
	With Compromised Therapy:9



<b>U.S. Survival Probability</b>		Year	1	2	3	4	5	6	7	8	9	10	
Non Advisory Population	Depletions and Malfunctions(%)	99.98	99.97	99.77	99.27	98.53	97.21	94.65	88.87	81.60	—	—	
	(Confidence Interval)	(-0.1/+0.0)	(-0.1/+0.0)	(-0.2/+0.1)	(-0.3/+0.2)	(-0.4/+0.3)	(-0.6/+0.5)	(-0.9/+0.8)	(-2.0/+1.7)	@ 103 mo. (-3.5/+3.0)	—	—	
	Registered Implants: 7000												
23-Jun-06 and 24-Aug-06 Low Voltage Capacitor*	Depletions and Malfunctions(%)	100.00	100.00	100.00	99.19	99.19	97.25	95.65	90.11	—	—	—	
	(Confidence Interval)	(-0.0/+0.0)	(-0.0/+0.0)	(-0.0/+0.0)	(-1.3/+0.5)	(-1.3/+0.5)	(-2.2/+1.2)	(-2.7/+1.7)	@ 92 mo. (-3.9/+2.9)	—	—	—	
	Registered Implants: 1000												
22-Sep-05 Crystal Timing Component (Failure Mode 1)*	Depletions and Malfunctions(%)	100.00	100.00	100.00	100.00	100.00	99.73	99.73	99.73	—	—	—	
	(Confidence Interval)	(-0.0/+0.0)	(-0.0/+0.0)	(-0.0/+0.0)	(-0.0/+0.0)	(-0.0/+0.0)	(-1.6/+0.2)	(-1.6/+0.2)	@ 92 mo. (-1.6/+0.2)	—	—	—	
	Registered Implants: 1000												
Non Advisory Population	Effective Sample Size	6559	5831	5161	4545	3987	2822	1326	509	212	—	—	
	23-Jun-06 and 24-Aug-06 Low Voltage Capacitor*	Effective Sample Size	664	580	510	442	387	334	287	214	—	—	—
		22-Sep-05 Crystal Timing Component (Failure Mode 1)*	Effective Sample Size	664	580	510	442	387	334	287	214	—	—
Depletions and Malfunctions(%)			99.95	99.86	99.46	98.92	97.90	96.22	93.55	88.13	77.84	60.59	—
(Confidence Interval)	(-0.2/+0.0)		(-0.2/+0.1)	(-0.3/+0.2)	(-0.5/+0.3)	(-0.7/+0.5)	(-0.9/+0.7)	(-1.2/+1.0)	(-1.7/+1.5)	(-2.2/+2.1)	(-2.8/+2.8)	—	
Registered Implants:													

4000	Malfunctions Only (%) (Confidence Interval)	99.97 (-0.2/+0.0)	99.97 (-0.2/+0.0)	99.94 (-0.2/+0.0)	99.90 (-0.2/+0.1)	99.81 (-0.3/+0.1)	99.48 (-0.4/+0.2)	99.42 (-0.4/+0.3)	99.04 (-0.6/+0.4)	98.57 (-0.8/+0.5)	98.45 (-0.8/+0.5)
	Effective Sample Size	3515	3073	2598	2281	1973	1705	1459	1212	925	548
22-Sep-05 Crystal Timing Component (Failure Mode 2)*	Depletions and Malfunctions (%) (Confidence Interval)	99.95 (-0.1/+0.0)	99.89 (-0.1/+0.0)	99.55 (-0.1/+0.1)	99.10 (-0.2/+0.2)	98.17 (-0.3/+0.3)	96.51 (-0.4/+0.4)	94.07 (-0.6/+0.5)	87.80 (-0.8/+0.8)	76.28 (-1.2/+1.2)	58.47 (-1.7/+1.7)
Registered Implants: 14000	Malfunctions Only (%) (Confidence Interval)	99.99 (-0.0/+0.0)	99.99 (-0.0/+0.0)	99.98 (-0.1/+0.0)	99.94 (-0.1/+0.0)	99.88 (-0.1/+0.1)	99.58 (-0.2/+0.1)	99.38 (-0.2/+0.2)	99.05 (-0.3/+0.2)	98.79 (-0.3/+0.3)	98.70 (-0.4/+0.3)
	Effective Sample Size	12751	11249	9910	8721	7618	6598	5634	4622	2842	1004


\*Devices subject to an advisory. Refer to the Advisories for more details. Devices may be part of more than one advisory.

Data are representative of Boston Scientific INSIGNIA Plus and Intermedics NEXUS Plus device performance.

**INSIGNIA Plus DR**

Model 1297

<b>U.S. Survival Probability</b>	<b>Worldwide Malfunction Details</b>	<b>Product Advisories</b>
----------------------------------	--------------------------------------	---------------------------

<b>INSIGNIA Plus DR</b> <b>Model 1297</b> 			
<b>Worldwide Distribution: 47,000</b>			
<b>Worldwide Confirmed Malfunctions: 136</b>			
	Without Compromised Therapy	With Compromised Therapy	Total
<b>Electrical</b>	<b>3</b>	<b>3</b>	<b>6</b>
<sup>8</sup> Low-voltage capacitor (Advisory issued)	1	1	
<sup>32</sup> Capacitor	2	1	
<sup>89</sup> Integrated circuit	-	1	
<b>Mechanical</b>	<b>12</b>	<b>7</b>	<b>19</b>
<sup>12</sup> Crystal timing component Failure Mode 1 (Advisory issued)	1	2	
<sup>13</sup> Crystal timing component Failure Mode 2 (Advisory issued)	-	1	
<sup>23</sup> Solder bond	1	-	
<sup>39</sup> Capacitor array	1	-	
<sup>47</sup> Seal plug	5	-	
<sup>48</sup> Header	4	4	
<b>Software</b>	<b>6</b>	<b>-</b>	<b>6</b>
<sup>97</sup> Underestimation of battery status	3	-	
<sup>98</sup> Interrupted telemetry	2	-	
<sup>99</sup> Pacing rate limit	1	-	
<b>Other</b>	<b>103</b>	<b>2</b>	<b>105</b>
Non-patterned	5	2	
<sup>22</sup> Longevity labeling	84	-	
<sup>72</sup> Battery depletion	2	-	
<sup>120</sup> Battery status	12	-	
<b>WW Confirmed Malfunctions</b>	<b>124</b>	<b>12</b>	<b>136</b>

[More details](#) about malfunctions

[References](#) cited in table above


**INSIGNIA Plus DR (downsize)**

Model 1298


<b>U.S. Survival Probability</b>	<b>Worldwide Malfunction Details</b>	<b>Product Advisories</b>
----------------------------------	--------------------------------------	---------------------------

<b>U.S. Summary</b>	
U.S. Registered Implants: 90,000	U.S. Normal Battery Depletions: 23,404
U.S. Approval Date: March 2002	U.S. Unconfirmed Reports of Premature Battery Depletion : 113
U.S. Estimated Active Implants: 15,000	U.S. Malfunctions:368

**Battery Depletions and Malfunctions**



**Malfunctions Only**



**U.S. Survival Probability**

	Year	1	2	3	4	5	6	7	8	9	10
Non Advisory Population	Depletions and Malfunctions(%) (Confidence Interval)	99.94 (-0.1/+0.0)	99.88 (-0.1/+0.0)	99.41 (-0.1/+0.1)	98.27 (-0.2/+0.2)	95.60 (-0.4/+0.4)	88.10 (-0.7/+0.6)	70.17 (-1.2/+1.1)	49.24 (-1.8/+1.8)	33.96 @ 104 mo. (-2.4/+2.5)	-
	Registered Implants: 19000										
	Malfunctions Only(%) (Confidence Interval)	99.98 (-0.0/+0.0)	99.98 (-0.0/+0.0)	99.90 (-0.1/+0.0)	99.76 (-0.1/+0.1)	99.54 (-0.1/+0.1)	99.34 (-0.2/+0.1)	99.21 (-0.2/+0.2)	99.11 (-0.3/+0.2)	99.11 @ 104 mo. (-0.3/+0.2)	-
	Effective Sample Size	16863	14980	13240	11653	10051	6900	2874	770	219	-
23-Jun-06 and 24-Aug-06 Low Voltage Capacitor*	Depletions and Malfunctions(%) (Confidence Interval)	99.93 (-0.4/+0.1)	99.86 (-0.4/+0.1)	99.77 (-0.5/+0.2)	98.12 (-1.0/+0.7)	95.81 (-1.5/+1.1)	84.89 (-2.6/+2.3)	66.02 (-3.5/+3.3)	52.52 @ 92 mo. (-3.8/+3.7)	-	-
	Registered Implants: 2000										
	Malfunctions Only(%) (Confidence Interval)	100.00 (-0.0/+0.0)	100.00 (-0.0/+0.0)	100.00 (-0.0/+0.0)	99.71 (-0.6/+0.2)	99.37 (-0.8/+0.3)	99.25 (-0.8/+0.4)	98.91 (-1.0/+0.5)	98.91 @ 92 mo. (-1.0/+0.5)	-	-
	Effective Sample Size	1422	1251	1114	965	827	644	437	260	-	-
22-Sep-05 Crystal Timing Component (Failure Mode 1)*	Depletions and Malfunctions(%) (Confidence Interval)	99.90 (-0.1/+0.0)	99.65 (-0.1/+0.1)	99.13 (-0.2/+0.2)	98.03 (-0.3/+0.2)	95.05 (-0.5/+0.4)	86.19 (-0.8/+0.7)	67.67 (-1.1/+1.1)	47.02 (-1.3/+1.3)	31.98 (-1.3/+1.3)	22.24 (-1.2/+1.3)
	Registered Implants: 16000										
	Malfunctions Only(%) (Confidence Interval)	99.93 (-0.1/+0.0)	99.88 (-0.1/+0.0)	99.81 (-0.1/+0.1)	99.79 (-0.1/+0.1)	99.57 (-0.2/+0.1)	99.38 (-0.2/+0.1)	99.32 (-0.2/+0.2)	99.19 (-0.2/+0.2)	99.13 (-0.3/+0.2)	99.03 (-0.4/+0.3)

	Effective Sample Size	13683	12073	10374	9054	7729	6117	4097	2353	1312	726
22-Sep-05 Crystal Timing Component (Failure Mode 2)*	Depletions and Malfunctions(%) (Confidence Interval)	99.95 (-0.0/+0.0)	99.80 (-0.0/+0.0)	99.26 (-0.1/+0.1)	97.91 (-0.2/+0.1)	94.48 (-0.3/+0.2)	84.14 (-0.4/+0.4)	64.22 (-0.6/+0.6)	44.25 (-0.7/+0.7)	29.99 (-0.7/+0.7)	20.37 (-0.7/+0.7)
Registered Implants: 54000	Malfunctions Only(%) (Confidence Interval)	99.99 (-0.0/+0.0)	99.96 (-0.0/+0.0)	99.95 (-0.0/+0.0)	99.83 (-0.0/+0.0)	99.61 (-0.1/+0.1)	99.39 (-0.1/+0.1)	99.30 (-0.1/+0.1)	99.22 (-0.1/+0.1)	99.14 (-0.1/+0.1)	99.14 (-0.1/+0.1)
	Effective Sample Size	47027	41686	36744	32067	27291	21121	13703	7831	3569	1259

\*Devices subject to an advisory. Refer to the Advisories for more details. Devices may be part of more than one advisory.

Data are representative of Boston Scientific INSIGNIA Plus and Intermedics NEXUS Plus device performance.

**INSIGNIA Plus DR (downsize)**

Model 1298

<b>U.S. Survival Probability</b>	<b>Worldwide Malfunction Details</b>	<b>Product Advisories</b>
----------------------------------	--------------------------------------	---------------------------

**INSIGNIA Plus DR (downsize)  
Model 1298**



**Worldwide Distribution: 140,000**  
**Worldwide Confirmed Malfunctions: 442**

	Without Compromised Therapy	With Compromised Therapy	Total
<b>Electrical</b>	<b>11</b>	<b>10</b>	<b>21</b>
<sup>8</sup> Low-voltage capacitor (Advisory issued)	-	3	
<sup>28</sup> Capacitor	-	1	
<sup>32</sup> Capacitor	6	2	
<sup>42</sup> Integrated circuit	-	1	
<sup>89</sup> Integrated circuit	5	3	
<b>Mechanical</b>	<b>21</b>	<b>22</b>	<b>43</b>
<sup>12</sup> Crystal timing component Failure Mode 1 (Advisory issued)	5	20	
<sup>13</sup> Crystal timing component Failure Mode 2 (Advisory issued)	3	-	
<sup>23</sup> Solder bond	1	-	
<sup>39</sup> Capacitor array	3	1	
<sup>47</sup> Seal plug	3	1	
<sup>48</sup> Header	5	-	
<sup>95</sup> Seal plug	1	-	
<b>Software</b>	<b>11</b>	<b>-</b>	<b>11</b>
<sup>60</sup> Memory error	1	-	
<sup>96</sup> Interrogation at EOL	2	-	
<sup>97</sup> Underestimation of battery status	6	-	
<sup>98</sup> Interrupted telemetry	1	-	
<sup>99</sup> Pacing rate limit	1	-	
<b>Other</b>	<b>356</b>	<b>11</b>	<b>367</b>
Non-patterned	27	9	
<sup>22</sup> Longevity labeling	310	-	
<sup>46</sup> Battery depletion	2	1	
<sup>50</sup> Magnet response	1	-	
<sup>72</sup> Battery depletion	11	1	
<sup>120</sup> Battery status	5	-	
<b>WW Confirmed Malfunctions</b>	<b>399</b>	<b>43</b>	<b>442</b>

[More details](#) about malfunctions

[References](#) cited in table above




INSIGNIA Plus SR

Model 1194

U.S. Survival Probability	Worldwide Malfunction Details	Product Advisories
---------------------------	-------------------------------	--------------------

<b>U.S. Summary</b>	
U.S. Registered Implants: 27,000	U.S. Normal Battery Depletions: 2,407
U.S. Approval Date: March 2002	U.S. Unconfirmed Reports of Premature Battery Depletion : 7
U.S. Estimated Active Implants: 5,000	U.S. Malfunctions:27
	Without Compromised Therapy:19
	With Compromised Therapy:8



<b>U.S. Survival Probability</b>		Year	1	2	3	4	5	6	7	8	9	10
Non Advisory Population	Depletions and Malfunctions(%)	99.96	99.92	99.61	99.33	98.47	97.18	93.84	84.46	82.53	-	-
	(Confidence Interval)	(-0.1/+0.0)	(-0.1/+0.1)	(-0.3/+0.2)	(-0.3/+0.2)	(-0.5/+0.4)	(-0.7/+0.6)	(-1.3/+1.1)	(-3.0/+2.6)	@ 98 mo. (-3.4/+3.0)	-	-
	Registered Implants: 6000											
23-Jun-06 and 24-Aug-06 Low Voltage Capacitor*	Depletions and Malfunctions(%)	100.00	99.66	99.27	98.34	-	-	-	-	-	-	-
	(Confidence Interval)	(-0.0/+0.0)	(-2.0/+0.3)	(-2.2/+0.5)	@ 45 mo. (-2.7/+1.0)	-	-	-	-	-	-	-
	Registered Implants: 400											
22-Sep-05 Crystal Timing Component (Failure Mode 1)*	Depletions and Malfunctions(%)	100.00	99.66	99.66	99.19	-	-	-	-	-	-	-
	(Confidence Interval)	(-0.0/+0.0)	(-2.0/+0.3)	(-2.0/+0.3)	@ 45 mo. (-2.4/+0.6)	-	-	-	-	-	-	-
	Registered Implants: 6000											
22-Sep-05 Crystal Timing Component (Failure Mode 1)*	Depletions and Malfunctions(%)	99.92	99.89	99.74	99.37	98.74	98.00	94.86	84.91	71.10	60.91	-
	(Confidence Interval)	(-0.2/+0.1)	(-0.2/+0.1)	(-0.3/+0.1)	(-0.4/+0.2)	(-0.6/+0.4)	(-0.7/+0.5)	(-1.2/+1.0)	(-2.1/+1.9)	(-2.8/+2.7)	(-3.2/+3.1)	-
	Registered Implants: 6000											

4000	Malfunctions Only (%) (Confidence Interval)	99.92 (-0.2/+0.1)	99.92 (-0.2/+0.1)	99.88 (-0.2/+0.1)	99.88 (-0.2/+0.1)	99.88 (-0.2/+0.1)	99.76 (-0.3/+0.1)	99.68 (-0.4/+0.2)	99.68 (-0.4/+0.2)	99.68 (-0.4/+0.2)	99.68 (-0.4/+0.2)
	Effective Sample Size	3454	2919	2422	2071	1744	1437	1173	880	620	422
22-Sep-05 Crystal Timing Component (Failure Mode 2)*	Depletions and Malfunctions (%) (Confidence Interval)	99.94 (-0.1/+0.0)	99.87 (-0.1/+0.0)	99.65 (-0.1/+0.1)	99.24 (-0.2/+0.2)	98.49 (-0.3/+0.2)	97.09 (-0.4/+0.4)	92.87 (-0.7/+0.6)	81.89 (-1.1/+1.0)	69.79 (-1.4/+1.3)	59.09 (-1.7/+1.7)
Registered Implants: 17000	Malfunctions Only (%) (Confidence Interval)	99.99 (-0.0/+0.0)	99.99 (-0.0/+0.0)	99.96 (-0.1/+0.0)	99.95 (-0.1/+0.0)	99.94 (-0.1/+0.0)	99.91 (-0.1/+0.0)	99.89 (-0.1/+0.1)	99.89 (-0.1/+0.1)	99.82 (-0.2/+0.1)	99.82 (-0.2/+0.1)
	Effective Sample Size	13686	11696	10067	8522	7166	6028	4926	3672	2178	965


\*Devices subject to an advisory. Refer to the Advisories for more details. Devices may be part of more than one advisory.

Data are representative of Boston Scientific INSIGNIA Plus and Intermedics NEXUS Plus device performance.

**INSIGNIA Plus SR**

Model 1194

<b>U.S. Survival Probability</b>	<b>Worldwide Malfunction Details</b>	<b>Product Advisories</b>
----------------------------------	--------------------------------------	---------------------------

<b>INSIGNIA Plus SR</b> <b>Model 1194</b> 			
<b>Worldwide Distribution:</b> 51,000			
<b>Worldwide Confirmed Malfunctions:</b> 35			
	Without Compromised Therapy	With Compromised Therapy	Total
<b>Electrical</b>	<b>4</b>	<b>5</b>	<b>9</b>
<sup>8</sup> Low-voltage capacitor (Advisory issued)	1	2	
<sup>32</sup> Capacitor	2	2	
<sup>42</sup> Integrated circuit	-	1	
<sup>89</sup> Integrated circuit	1	-	
<b>Mechanical</b>	<b>1</b>	<b>6</b>	<b>7</b>
<sup>12</sup> Crystal timing component Failure Mode 1 (Advisory issued)	-	5	
<sup>39</sup> Capacitor array	1	-	
<sup>47</sup> Seal plug	-	1	
<b>Software</b>	<b>1</b>	<b>-</b>	<b>1</b>
<sup>99</sup> Pacing rate limit	1	-	
<b>Other</b>	<b>17</b>	<b>1</b>	<b>18</b>
Non-patterned	4	-	
<sup>22</sup> Longevity labeling	10	-	
<sup>46</sup> Battery depletion	-	1	
<sup>72</sup> Battery depletion	1	-	
<sup>120</sup> Battery status	2	-	
<b>WW Confirmed Malfunctions</b>	<b>23</b>	<b>12</b>	<b>35</b>

[More details](#) about malfunctions


[References](#) cited in table above

**INSIGNIA AVT**

Models 0482/0882/0982/1192/1292

<b>U.S. Survival Probability</b>	<b>Worldwide Malfunction Details</b>	<b>Product Advisories</b>
----------------------------------	--------------------------------------	---------------------------

**INSIGNIA AVT**  
**Models 0482/0882/0982/1192/1292**



**Worldwide Distribution:** 51,000  
**Worldwide Confirmed Malfunctions:** 60

	Without Compromised Therapy	With Compromised Therapy	Total
<b>Electrical</b>	-	5	5
<sup>8</sup> Low-voltage capacitor (Advisory issued)	-	3	
<sup>32</sup> Capacitor	-	1	
<sup>89</sup> Integrated circuit	-	1	
<b>Mechanical</b>	1	-	1
<sup>47</sup> Seal plug	1	-	
<b>Software</b>	-	-	0
<b>Other</b>	52	2	54
Non-patterned	1	1	
<sup>22</sup> Longevity labeling	33	-	
<sup>72</sup> Battery depletion	-	1	
<sup>120</sup> Battery status	18	-	
<b>WW Confirmed Malfunctions</b>	<b>53</b>	<b>7</b>	<b>60</b>

[More details](#) about malfunctions

[References](#) cited in table above

# CRM PRODUCT PERFORMANCE REPORT Q1 2014

## Confirmed Malfunction Details: Pulse Generators

## References

Descriptions listed below provide an overview of the clinical observations and/or analysis findings associated with each pulse generator confirmed malfunction pattern listed in this report.

All of the patterns listed are thoroughly investigated and analyzed. As part of Boston Scientific's process of continuous improvement, when possible, changes have been or will be implemented in response to identified malfunction patterns. "Improvements implemented" may include product design changes in existing or subsequent generations, manufacturing process modifications, software updates, educational communications, labeling changes, etc. Improvement implementation may vary by geography due to various factors, including regulatory review timing, and may not completely mitigate or eliminate the potential for additional malfunctions.

1. **Low Voltage Capacitor**— *August 29, 2013 Voluntary Physician Advisory*. Alert message during followup, beeping tones, premature battery depletion. Diminished low voltage capacitor performance.
2. **Unintended fuse activation 2013**— *March 1, 2013 Voluntary Physician Advisory*. Inability to interrogate, no magnet response, permanent loss of therapy without warning. Improvement implemented.
3. **High cathode condition**— *June 1, 2011 Voluntary Physician Advisory*. Premature battery depletion. Misaligned battery component. Improvement implemented.
4. **Magnetic reed switch 2010**— *July 21, 2010 Voluntary Physician Advisory*. Upon magnet removal, magnetic reed switch remains closed and device tones do not cease. Tachy therapy unavailable if the Enable Magnet Use feature is programmed ON.
5. **Subpectoral implant 2009**— *December 01, 2009 Voluntary Physician Advisory*. Noise, oversensing, inappropriate shocks, pacing inhibition, high impedance when implanted subpectorally. Weakened bond between header and titanium case. Improvement implemented.
6. **Respiratory Sensor Oversensing**— *March 23, 2009 Voluntary Physician Advisory*. Oversensing, noise, inappropriate shock, pacing inhibition. When Respiratory Sensor is ON, RV lead or system complications may cause oversensing or noise. Improvement implemented.
7. **Shortened replacement window**— *April 05, 2007 and March 04, 2009 Voluntary Physician Advisory*. Accelerated battery depletion may reduce time between elective replacement (ERI) and end of life (EOL) indicators to less than three months. Device replacement indicators continue to function normally. Degradation of low-voltage capacitor. Improvement implemented.
8. **Low-voltage capacitor**— *June 23, 2006 and August 24, 2006 Voluntary Physician Advisory*. Premature battery depletion, no output, no interrogation. Failed low-voltage capacitor. Improvement implemented.
9. **Premature battery depletion**— *May 12, 2006 Voluntary Physician Advisory*. Premature battery depletion, significantly shortened longevity and duration between elective replacement (ERI) and end of life (EOL) indicators. Gradual, premature battery depletion most common; in rare instances, rapid depletion occurred with no therapy available. Failed low-voltage capacitor. Improvement implemented.
10. **Subpectoral implant**— *May 12, 2006 and January 04, 2008 Voluntary Physician Advisory*. Beeping tones and programmer warning screen upon interrogation, loss of telemetry communications, loss of pacing (intermittent or permanent) and/or loss of shock therapy. Repetitive mechanical stress-induced component damage, only when implanted subpectorally with the serial number facing the ribs. Improvement implemented.
11. **Hermetic sealing component Second Population**— *January 21, 2006 Voluntary Physician Advisory*. Premature battery depletion resulting in loss of telemetry and/or loss of pacing output without warning, appearance of a reset warning message upon interrogation, inappropriate early display of replacement indicators, inappropriate accelerometer function resulting in sustained pacing at the programmed maximum sensor rate, or lack of appropriate accelerometer rate response during activity. Gradual degradation of hermetic sealing component with higher than normal moisture content in pacemaker case late in device service life. Improvement implemented.
12. **Crystal timing component Failure Mode 1**— *September 22, 2005 Voluntary Physician Advisory*. Intermittent or permanent loss of pacing output without warning, intermittent or permanent loss of telemetry, reversion to VVI mode or appearance of a reset warning message upon interrogation. Foreign material within a crystal timing component. Improvement implemented.
13. **Crystal timing component Failure Mode 2**— *September 22, 2005 Voluntary Physician Advisory*. At implant procedure or during pre-implant testing: Intermittent or permanent loss of pacing output without warning, intermittent or permanent loss of telemetry, reversion to VVI mode, or appearance of a reset warning message upon interrogation. Microscopic particle within a crystal timing component. Three failures have been reported following confirmation of successful implantation. No currently distributed devices are subject to this peri-implant failure mode. Improvement implemented.
14. **Hermetic sealing component Original Population**— *July 18, 2005 and January 21, 2006 Voluntary Physician Advisory*. Premature battery depletion resulting in loss of telemetry and/or loss of pacing output without warning, appearance of a reset warning message upon interrogation, inappropriate early display of replacement indicators, inappropriate accelerometer function resulting in sustained pacing at the programmed maximum sensor rate or lack of appropriate accelerometer rate response during activity. Gradual degradation of hermetic sealing component with higher than normal moisture content in pacemaker case late in device service life. Improvement implemented.
15. **Magnetic switch**— *June 23, 2005 Voluntary Physician Advisory*. Inhibition of tachyarrhythmia therapy (with no impact on bradycardia pacing) and affecting battery longevity. Magnetic switch stuck in closed position.

Improvement implemented.

16. **Shorting under header**— June 17, 2005 Voluntary Physician Advisory. Permanent loss of shock and pacing therapy. Deterioration of wire insulator (in conjunction with other factors), electrical short. Improvement implemented.
17. **Shorting in header**— June 17, 2005 Voluntary Physician Advisory. Permanent loss of shock and pacing therapy. Deterioration of wire insulator (in conjunction with other factors), electrical short. Improvement implemented.
18. **Functional latching**— June 17, 2005 and July 22, 2005 Voluntary Physician Advisory. Limited therapy availability. Functional "latching." Original June 17 advisory recommendations revised because new information indicated programming Atrial Tachy Episode Data Storage to 0% caused latching in subset of devices with previously stored atrial episode data. Reference July 22, 2005 advisory for more details. New software is now available worldwide to prevent functional latching. Improvement implemented.
19. **Safety Mode**— April 23, 2001 Voluntary Physician Advisory. Switch to Safety Mode due to rare interaction between device and specific memory component; beeping tones emitted to alert patient. Affected devices still provide full output shock delivery in Safety Mode. Improvement implemented.
20. **Integrated circuit chips**— March 29, 1999 Voluntary Physician Advisory. Loss of pacing, premature replacement indicators, backup mode operation, disruption of diagnostic measurements. High current drain resulting from a short circuit between two integrated circuit chips. Improvement implemented.
21. **Extended charge time post-mid-life**— Extended charge time post-mid-life resulting in decreased longevity, shortened ERI to EOL or no time stamp, due to increased battery impedance. Improvement implemented.
22. **Longevity labeling**— Battery longevity inconsistent with longevity labeling. Device battery status indicators are accurate and no loss of therapy has been reported.
23. **Solder bond**— Loss of device output, loss of sensing. Separation of component solder from substrate. Improvement implemented.
24. **Longevity Remaining error**— When near ERT, Longevity Remaining Estimate is incorrect when amplitude is reprogrammed to a higher value. Gas gauge display is not affected.
25. **Parameter errors**— During RF interrogation, parameter errors occur, requiring manual parameter correction. Corruption in telemetry logic. Improvement implemented.
26. **Firmware error**— Device beeping unaffected by magnet application, missing EGMs, loss of telemetry, loss of tachy therapy. Multiple resets due to firmware error. Improvement implemented.
27. **Integrated circuit**— Power on Reset state, loss of telemetry, safety mode operation or loss of output. Failed digital integrated circuit.
28. **Capacitor**— Premature battery depletion, inability to interrogate. Damage to low-voltage capacitor.
29. **Reconfirmation after charge**— Tachy therapy delayed. Redetection and therapy occur after lack of reconfirmation after initial charge. Timing conflict involving specific features. Improvement implemented.
30. **Pacing wire weld**— Loss of telemetry, loss of pacing. Weld failure between header and internal circuitry.
31. **Setscrew thread depth**— No pacing or pauses in pacing, intermittent or lack of setscrew contact with lead. Incorrect setscrew thread depth. Improvement implemented.
32. **Capacitor**— No telemetry, no pacing, premature battery depletion. Gradual, premature battery depletion most common; in rare instances, rapid depletion occurred with no therapy available. Failed low-voltage capacitor.
33. **Header**— Loosened header at pulse generator replacement or lead revision due to process variability. Improvement implemented.
34. **Short circuit**— Permanent loss of shock and pacing therapy. Deterioration of wire insulator (in conjunction with other factors), electrical short. Improvement implemented.
35. **Feedthrough wires**— High impedance and/or loss of pacing therapy. Broken wire connecting header to internal circuitry. Improvement implemented.
36. **Battery depletion**— Premature battery depletion.
37. **Power on reset**— Power on Reset state for which tachy and brady therapy are available at preset parameters.
38. **Battery depletion**— Premature battery depletion. Failed battery.
39. **Capacitor array**— Loss of device output, loss of capture, inability to accurately measure charge times causing elective replacement indicator declaration. Damage to capacitor array. Improvement implemented.
40. **High current drain**— Loss of pacing, premature replacement indicators, backup mode operation, disruption of diagnostic measurements. High current drain resulting from a short circuit between two integrated circuit chips. Improvement implemented.
41. **Short circuit**— Permanent loss of shock and pacing therapy, electrical short. Insulation degradation due to incorrect wire routing. Improvement implemented.
42. **Integrated circuit**— No telemetry, premature battery depletion. Integrated circuit issue within high-voltage transistor.
43. **Impedance measurements**— High impedance value (generally >3000 ohms) recorded and displayed as weekly average impedance. Low daily impedance recordings cause error in calculating weekly average impedance due to memory storage limitations. Device fully capable of therapy delivery as programmed. Improvement implemented.
44. **Memory error**— Pacing not as expected. Memory map error. Improvement implemented.
45. **Shortened ERI to EOL**— Time from ERI to EOL less than expected. Increased battery impedance near EOL.
46. **Battery depletion**— Premature battery depletion and loss of capture.
47. **Seal plug**— Non-cardiac signals on electrograms leading to inhibition of pacing and/or inappropriate shock delivery. Damaged seal plug. Improvement implemented.
48. **Header**— High impedance, compromised header bonding identified during lead revision procedures. Insufficient medical adhesive bonding between header and case. Improvement implemented.
49. **Telemetry or atrial noise**— Noise during telemetry and/or atrial sensing. Inappropriate contact between telemetry coil and device case. Improvement implemented.

50. **Magnet response**— No magnet response. Particulate material in component. Improvement implemented.
51. **Diagnostic data error**— No ventricular sense (VS) markers displayed on real-time EGMs when hysteresis is on. Improvement implemented.
52. **Overestimation of battery status**— *May 06,2003 Voluntary Physician Advisory*. Improvement in battery status between follow-up visits and/or overestimation of remaining longevity. Very specific conditions involving the use of an atrial tachycardia response feature. Software upgrade distributed late November 2003 eliminated the possibility of battery status overestimation. Improvement implemented.
53. **Impedance**— Atrial and/or ventricular pacing impedances >2500 ohms in unipolar and bipolar modes.
54. **Oscillator circuit**— Beeping and/or alert messages during interrogation shortly after implant. Oscillator circuit operates at higher frequency due to increase in temperature. Improvement implemented.
55. **Integrated circuit**— Lack of ventricular markers and reversion to Safety Mode, due to IC fault. Improvement implemented.
56. **Battery weld**— No pacing output and/or inability to interrogate. Battery weld. Improvement implemented.
57. **Battery depletion**— Premature battery depletion.
58. **Resistor**— Alert messages upon interrogation. Damaged resistor. Improvement implemented
59. **Rate fault declaration**— Inappropriate pacing due to timing interaction when autocapture is programmed on. Improvement implemented.
60. **Memory error**— Device resets (including pacing at reset parameters) and inability to interrogate. Errors in device memory.
61. **Adhesive consistency**— Non-cardiac signals on electrograms leading to inhibition of pacing and/or inappropriate shock delivery. Bubbles or voids in adhesive. Improvement implemented.
62. **Reset during charge**— Power on reset state during therapeutic shock charging attempt due to firmware issue. Improvement implemented.
63. **Hybrid circuit**— Alert messages or loss of output. Failed solder joints on device hybrid circuit. Improvement implemented.
64. **Capacitor**— Premature battery depletion, no telemetry. Damage to capacitor. Improvement implemented.
65. **Transformer**— Charge time alert message and/or end of life (EOL) indicator displayed, loss of shock therapy. Damaged transformer. Improvement implemented.
66. **Circuit connection**— Loss of telemetry, no magnet tones, loss of pacing, loss of shock therapy. Damaged internal circuit connection. Improvement implemented.
67. **Memory address**— Inability to interrogate. Memory address error. Improvement implemented.
68. **Telemetry coil**— No pacing output and/or an inability to interrogate. Short circuit between pulse generator feedthrough wires and telemetry coil. Improvement implemented.
69. **Capacitor**— Premature battery depletion, significantly shortened ERI to EOL time. Gradual, premature battery depletion most common; in rare instances, rapid depletion occurred with no therapy available. Failed low-voltage capacitor. Improvement implemented.
70. **Battery depletion**— Premature, gradual depletion of battery; in rare instances, rapid depletion with no therapy available.
71. **Device tones**— Inappropriately sustained device tone. Magnet removal or initialization of programming event during specific timing window. Device fully capable of therapy delivery as programmed. Improvement implemented.
72. **Battery depletion**— Loss of therapy, inability to interrogate, no magnet response, premature battery depletion.
73. **Solder bond**— Inability to interrogate, no magnet response, no pacing output. Broken solder bond between wire mounting surface and internal circuitry. Improvement implemented.
74. **Internal device connection**— Intermittent or no telemetry. Telemetry coil connection. Improvement implemented.
75. **Setscrew block**— No pacing or pauses in pacing, intermittent or lack of setscrew contact with lead. Incorrect setscrew block. Improvement implemented.
76. **Memory location**— Inappropriate early display of elective replacement indicator (ERI). Incorrect data within a specific memory location.
77. **Feedthrough filter capacitor**— Inability to interrogate device following shock delivery. High voltage build up between feedthrough leads on capacitor surface. Improvement implemented.
78. **Memory location**— Inability to perform diagnostic testing, loss of shock therapy. Incorrect data within a specific memory location. Improvement implemented.
79. **Stored EGMs**— Inability to view stored EGMs. Incorrect EGM index location.
80. **Memory location**— Loss of pacing and/or shock therapy, premature battery depletion, inability to perform diagnostic testing, loss of telemetry, and/or incorrect diagnostic information. Incorrect data within a specific memory location.
81. **Mid-life display of replacement indicators**— Extended charge time resulting in early occurrence of ERI or EOL, or shortened ERI to EOL time. Please reference the ERI Charge Time Limit Extended During Mid-Life Product Update for more details. Improvement implemented.
82. **High-voltage capacitor**— In most cases, temporary long capform charge time; in some cases delayed therapy or loss of tachy therapy. Extended charge time could prompt EOL declaration. High-voltage capacitor issue.
83. **Battery post**— Inability to interrogate, no pacing output. Bent battery post. Improvement implemented.
84. **Sensing**— Oversensing and/or delivery of inappropriate shocks. Physiologic and/or mechanical signals. Improvement implemented.
85. **Early ERI declaration**— Early appearance of ERI. Increased battery impedance prompts ERI declaration. End of life indicators operate as designed.
86. **Software download**— Safety Mode operation at predetermined brady and tachy parameters. Incomplete

- software download. Restoration tool available. Improvement implemented.
87. **A/D module**— Inability to obtain telemetry, reversion to Safety Mode, device beeping. Failure within Analog to Digital (A/D) module.
  88. **Early ERI declaration**— Early appearance of elective replacement indicator (ERI). Increased battery impedance extends charge time and prompts ERI declaration. Therapy availability unaffected, end of life indicators operate as designed. Longevity estimation relabeled. Improvement implemented.
  89. **Integrated circuit**— Premature battery depletion, loss of pacing output, inability to interrogate, loss of sensing, high-rate pacing, loss of shock therapy. Damage to integrated circuit. Improvement implemented.
  90. **Battery depletion**— Premature battery depletion due to current drain.
  91. **Alert messages**— During programmer interactions, alert messages appear which are able to be cleared. In one case, an alert message occurred with two memory errors after multiple device resets.
  92. **Diagnostic data error**— Potential inability to view daily measurements and/or inappropriate indication of BOL. Rate fault reset. Improvement implemented.
  93. **Setscrew**— Inability to tighten or loosen setscrews during implant or replacement procedure due to process variability. Improvement implemented.
  94. **Charge time limit**— Early appearance of elective replacement indicator (ERI). Incorrect extended charge time limit. Improvement implemented.
  95. **Seal plug**— Lifted or missing seal plugs. Inadequate medical adhesive bond. Improvement implemented
  96. **Interrogation at EOL**— No interrogation at end of life (EOL). Improvement implemented.
  97. **Underestimation of battery status**— Underestimation of remaining longevity due to invalid charge time measurement. Improvement implemented.
  98. **Interrupted telemetry**— Early appearance of Elective Replacement Time (ERT) indicator, unexpected impedance measurements (>2500 ohms). Interruption in telemetry sequence during software upgrade. Improvement implemented.
  99. **Pacing rate limit**— Inability to interrogate. Inappropriate pacing due to feature interaction. Improvement implemented.
  100. **Logic errors**— Unable to complete diagnostic, cap reform or daily shock lead measurement processes; loss of pacing output, loss of shock therapy. Logic errors when device is performing capacitor reforms or shock impedance measurements.
  101. **Reed switch**— While implanted, continuous device tone or beeping occurs. During interrogation, magnet presence dialog box appears. Tachy therapy unavailable if the Enable Magnet Use feature is programmed ON. Reed switch stuck in closed position. Improvement implemented.
  102. **Cracked solder joint**— Safety mode operation, beeping tones. Cracked solder joint.
  103. **Transformer**— Inability to interrogate, loss of pacing and shock therapy. Failed transformer.
  104. **Atrial pacing alert message**— Atrial pacing alert message, beeping tones due to software design. No effect on therapy availability. Improvement implemented.
  105. **Connector block**— Connector block can be moved out of alignment or displaced from header. Prolonged implant procedure, high impedance, no pacing, no sensing. Improvement implemented.
  106. **Misaligned markers**— Stored episode markers do not match recorded EGM. Software error when ventricular episode begins during ATR episode. New software was released in 2006 which prevents misaligned markers. Improvement implemented.
  107. **Seal plug**— Non-cardiac signals on electrograms may result in loss of pacing or inappropriate shocks. Seal plug allows air in lead port to escape.
  108. **Difficulty securing lead**— Noise, high impedance, inappropriate shocks or loss of therapy due to crossthreaded setscrews, intermittent or lack of contact between lead and header. Improvement implemented.
  109. **Low-voltage capacitor**— Premature battery depletion, early appearance of elective replacement indicator (ERI). Failed low-voltage capacitor. Improvement implemented.
  110. **Safety Core-electrocautery**— During electrocautery, device may enter Safety Core. Circuitry response to noise caused by electrocautery. Improvement implemented.
  111. **Resistor**— Alert messages upon interrogation, beeping tones or premature battery depletion. Resistor material oxidation. Improvement implemented.
  112. **High-voltage capacitor**— Alert message upon interrogation, extended charge time. Damaged high voltage capacitor.
  113. **Magnet rate**— During interrogation, magnet rate remains after removal of magnet. Reed switch stuck in closed position. Improvement implemented.
  114. **Battery status**— Battery status readings below BOL at or soon after implant caused by exposure to below room temperatures before implant. Actual battery status and longevity are not affected. Improvement implemented.
  115. **Header contacts**— Noise, oversensing, inappropriate shock, high pacing impedance, possible loss of pacing and sensing. Poor header connection with lead terminals due to contacts.
  116. **Safety Core-programming**— Device enters Safety Core after three consecutive invalid programming attempts, due to firmware issue. Improvement implemented.
  117. **Low-voltage capacitors**— Premature battery depletion, voltage alert during followup, device beeping. Capacitor failure.
  118. **Bent flex circuit**— Charge time-out and/or EOL declaration, due to bent flex circuit. Improvement implemented.
  119. **Alert messages not displayed post-EOL**— No alert message display after EOL declaration. Improvement implemented.
  120. **Battery status**— Longevity remaining, battery status, gas gauge and/or magnet rate do not align or are inconsistent.
  121. **Integrated circuit**— Loss of telemetry, premature battery depletion, alert message during followup. Integrated circuit issue. Improvement implemented.



122. **Memory errors**— Safety mode operation, inaccurately labeled pacing data. Errors in device memory
123. **High voltage circuit**— Alert message after implant, loss of shock therapy. Failed output module.
124. **Battery**— Beeping tones and alert message upon interrogation. Reduced battery voltage. Improvement implemented.
125. **Low-voltage capacitor**— Alert message during followup, beeping tones, premature battery depletion. Capacitor failure.
126. **Battery depletion**— Premature, gradual depletion of battery; in rare instances, rapid depletion with no therapy available. Improvement implemented.
127. **Telemetry**— Inability to interrogate, premature battery depletion.
128. **Unintended Battery Depletion Alert**— Beeping tones, Battery Depletion alert during followup despite normal battery depletion. Alert may be cleared without impact to battery status or therapy availability. Improvement implemented.
129. **Header**— Noise, oversensing, inappropriate shocks, pacing inhibition, high impedance when implanted subcutaneously. Weakened bond between header and titanium case. Improvement Implemented.
130. **Solder joint**— Beeping tones and programmer warning screen upon interrogation, loss of telemetry communications. Cracked solder joint due to repetitive mechanical stress-induced component damage when implanted subpectorally with the serial number facing the ribs.

## Before/During Implant Procedure - Worldwide Malfunctions: Pulse Generators

This section of the report depicts the number of product malfunctions that occurred worldwide either before implant (prior to opening the sterile product packaging) or during implant (once the sterile product packaging has been opened). In all cases, the product in question must be returned to Boston Scientific CRM and confirmed through laboratory analysis to have operated or exhibited a problem outside the specified performance limits established by Boston Scientific. Damage incurred during shipping/transit or due to external factors warned against in labeling (e.g. radiation) is not reported as device malfunction here.

The Electrical category is comprised of confirmed malfunctions involving electrical components such as batteries and capacitors, and also includes fault codes encountered at implant. The majority of before/during implant pulse generator confirmed malfunctions in the Mechanical category are issues occurring within the connector block (e.g. stuck setscrews, seal plug/ring issues). The Software category consists primarily of confirmed malfunctions that result in telemetry issues. Confirmed malfunctions in the Labeling and Packaging categories include product labeling/identification issues and damage to sterile packaging, respectively. The Other category is comprised of non-patterned confirmed malfunctions.

CRT-D/Model	Worldwide Distribution	Electrical	Mechanical	Software	Other	Labeling	Packaging
INCEPTA CRT-D 4-Site N160/N162/P162	11000	0	0	0	0	0	0
INCEPTA CRT-D N161/N163/N164/N165/P163/P165	11000	1	0	0	0	0	0
ENERGEN CRT-D 4-Site N140/N142/P142	14000	2	0	0	3	0	0
ENERGEN CRT-D N141/N143/P143	12000	2	0	0	3	0	0
PUNCTUA CRT-D 4-Site N050/N052/P052	2000	0	0	0	0	0	0
PUNCTUA CRT-D N051/N053/P053	3000	1	0	0	1	0	0
COGNIS N118/N119/N120/P106/P107/P108	108000	23	50	3	27	0	0
LIVIAN HE H227/H229/H247/H249	7000	3	1	0	2	0	0
LIVIAN H220/H225/H240/H245	6000	0	1	0	2	0	0

<b>CRT-D/Model (cont.)</b>	Worldwide Distribution	Electrical	Mechanical	Software	Other	Labeling	Packaging
CONTAK RENEWAL 4 RF HE H239	1000	14	0	0	0	0	0
CONTAK RENEWAL 4 RF H230/H235	8000	45	2	0	1	0	0
CONTAK RENEWAL 4 HE H197/H199	7000	7	7	0	0	0	0
CONTAK RENEWAL 4 H190/H195	18000	1	13	1	2	0	0
CONTAK RENEWAL 4 AVT HE M177/M179	1000	0	2	0	1	0	0
CONTAK RENEWAL 4 AVT M170/M175	2000	1	0	0	1	0	0
CONTAK RENEWAL 3 RF HE H217/H219	18000	368	4	5	5	0	0
CONTAK RENEWAL 3 RF H210/H215	21000	493	9	1	7	0	0
CONTAK RENEWAL 3 HE H177/H179	23000	60	45	0	8	0	0
CONTAK RENEWAL 3 H170/H175	34000	46	65	0	15	0	0
<b>CRT-P/Model</b>	Worldwide Distribution	Electrical	Mechanical	Software	Other	Labeling	Packaging
INVIVE V172/V173/V182/V183/W172/W173	10000	0	0	0	2	0	0
CONTAK RENEWAL TR 2 H140/H145	31000	1	8	0	2	0	0
CONTAK RENEWAL TR H120/H125	19000	0	11	0	5	0	0

ICD/Model	Worldwide Distribution	Electrical	Mechanical	Software	Other	Labeling	Packaging
INCEPTA ICD VR 4-Site E160/F160	9000	0	1	0	2	0	0
INCEPTA ICD DR 4-Site E162/F162	11000	0	0	0	0	0	0
INCEPTA ICD VR E161/F161	4000	0	0	0	0	0	0
INCEPTA ICD DR E163/F163	6000	0	0	0	0	0	0
ENERGEN ICD VR 4-Site E140/F140	13000	1	0	0	2	0	0
ENERGEN ICD DR 4-Site E142/F142	13000	1	1	0	0	0	0
ENERGEN ICD VR E141/F141	7000	2	0	0	1	0	0
ENERGEN ICD DR E143/F143	9000	2	1	0	1	0	0
PUNCTUA ICD VR 4-Site E050/F050	2000	0	0	0	0	0	0
PUNCTUA ICD DR 4-Site E052/F052	1000	1	0	0	0	0	0
PUNCTUA ICD VR E051/F051	4000	0	0	0	1	0	0
PUNCTUA ICD DR E053/F053	3000	1	0	0	0	0	0
TELIGEN VR E102/E103/F102/F103	64000	7	32	2	18	0	0
TELIGEN DR E110/E111/F110/F111	90000	6	42	1	24	0	0
CONFIENT DR E030/F030	8000	1	3	0	0	0	0

Pacemaker/Model	Worldwide Distribution	Electrical	Mechanical	Software	Other	Labeling	Packaging
ADVANTIO EL DR J064/K064/K067/K084	3000	0	0	0	0	0	0
ADVANTIO SR J062/J065/K062/K065/K082	12000	0	0	0	0	0	0
ADVANTIO DR J063/J066/K063/K066/K083	36000	3	0	0	2	0	0
INGENIO EL DR J174/J177/K174/K177/K184	13000	1	0	0	0	0	0
INGENIO SR J172/J175/K172/K175/K182	15000	0	0	0	1	0	0
INGENIO DR J173/J176/K173/K176/K183	51000	0	0	0	2	0	0
ALTRUA 60 SR S601	68000	1	6	0	2	0	0
ALTRUA 60 DR EL S606	89000	0	8	0	2	0	0
ALTRUA 60 DR (Downsize) S603	131000	1	22	0	4	0	0
ALTRUA 60 DR S602	55000	1	11	0	2	0	0
ALTRUA 50 SR S501	23000	0	1	0	0	0	0
ALTRUA 50 DR (Downsize) S502	42000	0	2	0	1	0	0
ALTRUA 50 DDD (Downsize) S503	10000	1	1	0	0	0	0
ALTRUA 50 VDD (Downsize) S504	6000	0	0	0	0	0	0
ALTRUA 50 SSI S508	5000	0	0	0	0	0	0
ALTRUA 40 SR S401	9000	0	2	0	0	0	0

<b>Pacemaker/Model (cont.)</b>	Worldwide Distribution	Electrical	Mechanical	Software	Other	Labeling	Packaging
ALTRUA 40 DR (Downsize) S403	22000	0	4	0	2	0	0
ALTRUA 40 DR S402	3000	1	3	0	0	0	0
ALTRUA 40 DR EL S404	9000	0	0	0	0	0	0
ALTRUA 20 SR S201/S204	23000	0	2	0	0	0	0
ALTRUA 20 DR (Downsize) S203	16000	0	1	0	0	0	0
ALTRUA 20 DR S202/S205	3000	1	0	0	1	0	0
ALTRUA 20 DR EL S208	10000	0	0	0	0	0	0
ALTRUA 20 DDD S207	1000	0	0	0	0	0	0
ALTRUA 20 SSI S206	7000	0	0	0	0	0	0
INSIGNIA Ultra SR 1190*	48000	3	3	1	4	0	0
INSIGNIA Ultra DR (Downsize) 1290*	124000	3	3	1	6	0	0
INSIGNIA Ultra DR 1291*	51000	1	8	1	4	0	0
INSIGNIA Entra SR 1195/1198*	52000	1	0	3	5	0	0
INSIGNIA Entra DR (Downsize) 1296*	47000	1	7	6	3	0	0
INSIGNIA Entra DR 1294/1295*	36000	0	6	3	9	0	0

<b>Pacemaker/Model (cont.)</b>	Worldwide Distribution	Electrical	Mechanical	Software	Other	Labeling	Packaging
INSIGNIA Plus SR 1194*	51000	1	5	13	5	0	0
INSIGNIA Plus DR (Downsize) 1298*	140000	3	16	30	7	0	1
INSIGNIA Plus DR 1297*	47000	0	7	8	6	0	1
INSIGNIA AVT 0482/0882/0982/1192/1292*	51000	1	1	0	3	0	0

\*Counts consist of Boston Scientific and Intermedics co-branded pacemaker data.

## U.S. Reason for Out of Service

As requested by the Heart Rhythm Society Task Force on Device Performance Policies and Guidelines, Boston Scientific provides reasons for device explant or out of service, if known. The reasons consist of normal battery depletion, unconfirmed premature battery depletion, device upgrade, device malfunction (which includes devices under advisory that have experienced a malfunction), complication related to another system component or clinical condition, (such as infection), or "other," a category consisting of patient death, prophylactic device explant, elective replacement, general product dissatisfaction, other observation/complication, unspecified, or unknown.

The counts for normal battery depletion, unconfirmed premature battery depletion, and device malfunction are reflected in the U.S. survival probability data. Reason for device explant or out of service may either be confirmed through laboratory analysis (as in the case of device malfunction) or it may be reported to Boston Scientific with no associated device return or laboratory analysis. Although a device may be indicated by the health care provider to have been taken out of service for more than one reason, the table below indicates only one reason per device in category counts.

CRT-D/Model	U.S. Registered Implants	Normal Battery Depletion	Unconfirmed Premature Battery Depletion	Device Upgrade	Device Malfunction <sup>1</sup>	Complication related to another system component or clinical condition <sup>2</sup>	Other <sup>3</sup>
INCEPTA CRT-D 4-Site N160/N162/P162	5000	1	0	0	0	52	263
INCEPTA CRT-D N161/N163/N164/N165/P163/P165	8000	0	1	0	2	82	442
ENERGEN CRT-D 4-Site N140/N142/P142	9000	0	1	2	0	94	502
ENERGEN CRT-D N141/N143/P143	9000	0	0	1	7	82	597
COGNIS N118/N119/N120/P106/P107/P108	75000	290	20	6	317	1392	20286
LIVIAN HE H227/H229/H247/H249	6000	528	4	1	4	172	2252
LIVIAN H220/H225/H240/H245	5000	398	0	3	8	118	1880
CONTAK RENEWAL 3 RF HE H217/H219	18000	6110	24	15	141	563	8922
CONTAK RENEWAL 3 RF H210/H215	21000	6309	27	13	175	524	10701
CONTAK RENEWAL 3 HE H177/H179	24000	8088	92	18	875	614	12350
CONTAK RENEWAL 3 H170/H175	34000	11863	72	29	973	776	18187



<b>CRT-P/Model</b>	U.S. Registered Implants	Normal Battery Depletion	Unconfirmed Premature Battery Depletion	Device Upgrade	Device Malfunction <sup>1</sup>	Complication related to another system component or clinical condition <sup>2</sup>	Other <sup>3</sup>
INVIVE V172/V173/V182/V183/W172/W173	4000	0	0	0	0	14	236
CONTAK RENEWAL TR H120/H125	19000	1398	10	131	42	239	8210
<b>S-ICD/Model</b>		Normal Battery Depletion	Unconfirmed Premature Battery Depletion	Device Upgrade	Device Malfunction <sup>1</sup>	Complication related to another system component or clinical condition <sup>2</sup>	Other <sup>3</sup>
SQ-RX Pulse Generator 1010		0	1	0	7	12	67
<b>ICD/Model</b>	U.S. Registered Implants	Normal Battery Depletion	Unconfirmed Premature Battery Depletion	Device Upgrade	Device Malfunction <sup>1</sup>	Complication related to another system component or clinical condition <sup>2</sup>	Other <sup>3</sup>
INCEPTA ICD VR 4-Site E160/F160	5000	3	0	3	2	34	196
INCEPTA ICD DR 4-Site E162/F162	6000	1	0	4	2	55	247
INCEPTA ICD VR E161/F161	2000	0	0	1	1	25	93
INCEPTA ICD DR E163/F163	4000	0	0	1	1	25	125
ENERGEN ICD VR 4-Site E140/F140	9000	3	0	3	2	71	362
ENERGEN ICD DR 4-Site E142/F142	10000	1	0	5	4	85	416
ENERGEN ICD VR E141/F141	5000	2	0	1	1	27	209
ENERGEN ICD DR E143/F143	7000	0	0	3	1	38	301

<b>ICD/Model, continued...</b>	U.S. Registered Implants	Normal Battery Depletion	Unconfirmed Premature Battery Depletion	Device Upgrade	Device Malfunction <sup>1</sup>	Complication related to another system component or clinical condition <sup>2</sup>	Other <sup>3</sup>
TELIGEN VR E102/E103/F102/F103	38000	43	10	306	213	532	7769
TELIGEN DR E110/E111/F110/F111	66000	78	16	435	316	956	14340
CONFIENT DR E030/F030	7000	28	2	91	7	136	2261
VITALITY 2 EL VR T177	7000	718	7	146	1056	106	2375
VITALITY 2 EL DR T167	8000	1192	13	141	747	129	2943
VITALITY 2 VR T175	21000	4147	33	377	1237	292	8663
VITALITY 2 DR T165	31000	9349	78	526	1137	449	12732
VITALITY DR HE T180	13000	1456	13	229	400	298	6020
VITALITY DS DR T125	22000	7553	67	362	1181	303	9938
VITALITY DS VR T135	19000	4945	40	317	1553	252	8545
VITALITY EL T127	4000	806	9	60	617	69	1489

<b>Pacemaker/Model</b>	U.S. Registered Implants	Normal Battery Depletion	Unconfirmed Premature Battery Depletion	Device Upgrade	Device Malfunction <sup>1</sup>	Complication related to another system component or clinical condition <sup>2</sup>	Other <sup>3</sup>
ADVANTIO SR J062/J065/K062/K065/K082	6000	1	0	4	0	26	365
ADVANTIO DR J063/J066/K063/K066/K083	29000	3	1	6	3	119	869
INGENIO SR J172/J175/K172/K175/K182	6000	0	0	4	0	19	301
INGENIO DR J173/J176/K173/K176/K183	32000	4	1	6	4	98	805

Pacemaker/Model, continued...	U.S. Registered Implants	Normal Battery Depletion	Unconfirmed Premature Battery Depletion	Device Upgrade	Device Malfunction <sup>1</sup>	Complication related to another system component or clinical condition <sup>2</sup>	Other <sup>3</sup>
ALTRUA 60 SR S601	32000	74	0	157	2	148	9335
ALTRUA 60 DR (Downsize) S603	90000	661	23	340	16	509	18489
ALTRUA 60 DR S602	22000	90	2	120	3	163	4866
ALTRUA 60 DR EL S606	59000	59	5	175	6	344	8057
ALTRUA 40 SR S401	5000	7	0	14	1	20	1588
ALTRUA 40 DR (downsize) S403	14000	102	1	33	2	70	3059
ALTRUA 40 DR S402	2000	4	1	14	0	6	527
ALTRUA 40 DR EL S404	5000	6	0	20	0	38	961
ALTRUA 20 SR S201/S204	4000	16	1	15	0	33	1672
ALTRUA 20 DR (downsize) S203	5000	33	2	17	0	33	1407
ALTRUA 20 DR S202/S205	2000	11	0	5	0	11	542
ALTRUA 20 DR EL S208	3000	5	0	12	1	7	687
INSIGNIA Ultra SR 1190 <sup>4</sup>	24000	1017	8	193	30	137	14697
INSIGNIA Ultra DR (Downsize) 1290 <sup>4</sup>	76000	11548	102	525	396	577	35109
INSIGNIA Ultra DR 1291 <sup>4</sup>	32000	1182	17	281	108	288	12919

<b>Pacemaker/Model (cont.)</b>	U.S. Registered Implants	Normal Battery Depletion	Unconfirmed Premature Battery Depletion	Device Upgrade	Device Malfunction <sup>1</sup>	Complication related to another system component or clinical condition <sup>2</sup>	Other <sup>3</sup>
INSIGNIA Entra SR 1195/1198 <sup>4</sup>	14000	401	10	80	9	72	9752
INSIGNIA Entra DR (Downsize) 1296 <sup>4</sup>	24000	3735	25	128	96	146	14404
INSIGNIA Entra DR 1294/1295 <sup>4</sup>	17000	760	10	114	55	176	9567
INSIGNIA Plus SR 1194 <sup>4</sup>	27000	2407	7	221	27	156	19629
INSIGNIA Plus DR (Downsize) 1298 <sup>4</sup>	90000	23404	113	528	371	691	49839
INSIGNIA Plus DR 1297 <sup>4</sup>	27000	2721	18	245	118	251	13570
PULSAR MAX II SR (Downsize) 1180 <sup>4</sup>	7000	1525	8	35	4	30	5226
PULSAR MAX II DR 1280 <sup>4</sup>	29000	9008	18	175	186	217	17148
DISCOVERY II SR (Downsize) 1184 <sup>4</sup>	13000	1965	6	36	5	72	10192
DISCOVERY II SR 1186/1187 <sup>4</sup>	3000	306	1	20	2	23	2560
DISCOVERY II DR (Downsize) 1283 <sup>4</sup>	33000	9676	54	93	29	229	20725
DISCOVERY II DR 1284/1286 <sup>4</sup>	23000	5761	9	123	21	168	14908
PULSAR MAX DR 1270 <sup>4</sup>	41000	11357	55	185	221	309	25995
MERIDIAN DR 1276	16000	2880	17	39	41	122	11952

<sup>1</sup> Device malfunction consists of all U.S. confirmed malfunctions for a product/product grouping. These include confirmed malfunctions for advisory populations, as well as any other type of malfunction in which a device was returned and confirmed by laboratory analysis to have malfunctioned.. U.S. confirmed malfunction counts are reflected in U.S. survival probability.

<sup>2</sup> System component and/or clinical condition complications may include, for example: infection, erosion, lead-to-PG interface.

<sup>3</sup> Other consists of: patient death, elective replacement, general product dissatisfaction, other observation/complication, unspecified, or unknown.

<sup>4</sup> Counts consist of Boston Scientific and Intermedics co-branded pacemaker data.

## ACUITY Spiral


Models 4591/4592/4593

<b>U.S. Survival Probability</b>	<b>Worldwide Malfunction Details</b>	<b>Product Advisories</b>	<b>Longitude Survival Probability</b>
----------------------------------	--------------------------------------	---------------------------	---------------------------------------

### U.S. Summary

U.S. Registered Implants: 19,000	U.S. Chronic Lead Complications: 229
U.S. Approval Date: May 2008	U.S. Malfunctions:87
U.S. Estimated Active Implants: 14,000	Without Compromised Therapy:3
	With Compromised Therapy:84

### Complications and Malfunctions




### U.S. Survival Probability

Year	1	2	3	4	5	6	7	8	9	10
Non Advisory Population	98.54 (-0.2/+0.2)	98.20 (-0.2/+0.2)	97.96 (-0.2/+0.2)	97.74 (-0.3/+0.3)	97.55 (-0.4/+0.3)	97.55 @ 66 mo. (-0.4/+0.3)	—	—	—	—
Registered Implants: 18000										
Effective Sample Size	13617	9859	6399	3531	1123	254	—	—	—	—

## ACUITY Spiral

Models 4591/4592/4593

U.S. Survival Probability	Worldwide Malfunction Details	Product Advisories	Longitude Survival Probability
---------------------------	-------------------------------	--------------------	--------------------------------

ACUITY Spiral Models 4591/4592/4593 			
<b>Worldwide Distribution:</b> 35,000			
<b>Worldwide Confirmed Malfunctions:</b> 97			
	Without Compromised Therapy	With Compromised Therapy	Total
<b>Conductor</b>	<b>1</b>	<b>3</b>	<b>4</b>
<sup>28</sup> Non-patterned, Conductor	1	3	
<b>Crimp/Weld/Bond</b>	-	-	<b>0</b>
<b>Extrinsic</b>	-	<b>90</b>	<b>90</b>
<sup>30</sup> Unconfirmed Extrinsic	-	90	
<b>Insulation</b>	-	<b>1</b>	<b>1</b>
<sup>25</sup> Non-patterned, Insulation	-	1	
<b>Other</b>	<b>2</b>	-	<b>2</b>
<sup>27</sup> Non-patterned, Other	2	-	
<b>WW Confirmed Malfunctions</b>	<b>3</b>	<b>94</b>	<b>97</b>

[More details](#) about malfunctions

[References](#) cited in table above

## ACUITY Spiral Longitude\*


Models 4591/4592/4593

U.S. Survival Probability	Worldwide Malfunction Details	Product Advisories	Longitude Survival Probability
---------------------------	-------------------------------	--------------------	--------------------------------

### Longitude Registry Summary Data

Leads Enrolled: 1185	Chronic Lead Complications: 14
Leads Active: 921	Malfunctions:10
Cumulative Followup Months : 38,805	Without Compromised Therapy:0
	With Compromised Therapy:10

### Complications and Malfunctions



### Longitude Registry Survival Probability

Year	1	2	3	4	5	6	7	8	9	10
Longitude	98.14 (-1.0/+0.7)	97.63 (-1.2/+0.8)	97.27 (-1.3/+0.9)	97.27 (-1.3/+0.9)	97.27 @ 50 mo. (-1.3/+0.9)	--	--	--	--	--
Registered Implants: 1185										
Effective Sample Size	901	653	391	108	62	--	--	--	--	--

## ACUITY Steerable


Models 4554/4555/4556

U.S. Survival Probability	Worldwide Malfunction Details	Product Advisories
---------------------------	-------------------------------	--------------------

### U.S. Summary

U.S. Registered Implants: 26,000	U.S. Chronic Lead Complications: 324
U.S. Approval Date: May 2008	U.S. Malfunctions: 179
U.S. Estimated Active Implants: 17,000	Without Compromised Therapy: 7
	With Compromised Therapy: 172

### Complications and Malfunctions



### U.S. Survival Probability


Year	1	2	3	4	5	6	7	8	9	10
Non Advisory Population	98.49 (-0.2/+0.1)	98.13 (-0.2/+0.2)	97.83 (-0.2/+0.2)	97.61 (-0.2/+0.2)	97.22 (-0.3/+0.3)	96.93 (-0.4/+0.3)	96.93 @ 78 mo. (-0.4/+0.3)	—	—	—
Registered Implants: 26000										
Effective Sample Size	19654	15481	11612	8122	4568	1539	324	—	—	—



**ACUITY Steerable**

Models 4554/4555/4556

<b>U.S. Survival Probability</b>	<b>Worldwide Malfunction Details</b>	<b>Product Advisories</b>
----------------------------------	--------------------------------------	---------------------------

<b>ACUITY Steerable</b> <b>Models 4554/4555/4556</b> 			
<b>Worldwide Distribution:</b> 55,000			
<b>Worldwide Confirmed Malfunctions:</b> 233			
	Without Compromised Therapy	With Compromised Therapy	Total
<b>Conductor</b>	<b>3</b>	<b>36</b>	<b>39</b>
<sup>28</sup> Non-patterned, Conductor	1	9	
<sup>35</sup> Extracardiac fracture	2	27	
<b>Crimp/Weld/Bond</b>	<b>-</b>	<b>-</b>	<b>0</b>
<b>Extrinsic</b>	<b>2</b>	<b>185</b>	<b>187</b>
<sup>30</sup> Unconfirmed Extrinsic	-	185	
<sup>31</sup> Inconclusive Extrinsic	2	-	
<b>Insulation</b>	<b>-</b>	<b>1</b>	<b>1</b>
<sup>29</sup> Non-patterned, Insulation	-	1	
<b>Other</b>	<b>5</b>	<b>1</b>	<b>6</b>
<sup>27</sup> Non-patterned, Other	5	1	
<b>WW Confirmed Malfunctions</b>	<b>10</b>	<b>223</b>	<b>233</b>

[More details](#) about malfunctions

[References](#) cited in table above

## EASYTRAK 3

Models 4522/4524/4525/4527/4548/  
4549/4550


<b>U.S. Survival Probability</b>	<b>Worldwide Malfunction Details</b>	<b>Product Advisories</b>
----------------------------------	--------------------------------------	---------------------------

### U.S. Summary

U.S. Registered Implants: 20,000  
U.S. Approval Date: November 2004  
U.S. Estimated Active Implants: 11,000

U.S. Chronic Lead Complications: 310  
U.S. Malfunctions: 100  
Without Compromised Therapy: 4  
With Compromised Therapy: 96

### Complications and Malfunctions



### U.S. Survival Probability

Year	1	2	3	4	5	6	7	8	9	10
Non Advisory Population	98.68 (-0.2/+0.2)	98.49 (-0.2/+0.2)	98.25 (-0.2/+0.2)	97.90 (-0.2/+0.2)	97.46 (-0.3/+0.3)	97.12 (-0.3/+0.3)	96.66 (-0.4/+0.3)	96.60 (-0.4/+0.4)	96.32 (-0.5/+0.5)	96.32 @ 110 mo. (-0.5/+0.5)
Registered Implants: 20000										
Effective Sample Size	15913	13113	10820	8752	6737	5062	3406	1756	436	271

### EASYTRAK 3

Models 4522/4524/4525/4527/4548/  
4549/4550

U.S. Survival Probability	Worldwide Malfunction Details	Product Advisories
---------------------------	-------------------------------	--------------------

EASYTRAK 3 Models 4522/4524/4525/4527/4548/ 4549/4550			
Worldwide Distribution: 39,000			
Worldwide Confirmed Malfunctions: 127			
	Without Compromised Therapy	With Compromised Therapy	Total
<b>Conductor</b>	<b>5</b>	<b>34</b>	<b>39</b>
<sup>28</sup> Non-patterned, Conductor	3	5	
<sup>35</sup> Extracardiac fracture	2	29	
<b>Crimp/Weld/Bond</b>	-	-	<b>0</b>
<b>Extrinsic</b>	-	<b>83</b>	<b>83</b>
<sup>30</sup> Unconfirmed Extrinsic	-	83	
<b>Insulation</b>	<b>3</b>	<b>1</b>	<b>4</b>
<sup>29</sup> Non-patterned, Insulation	3	1	
<b>Other</b>	<b>1</b>	-	<b>1</b>
<sup>27</sup> Non-patterned, Other	1	-	
<b>WW Confirmed Malfunctions</b>	<b>9</b>	<b>118</b>	<b>127</b>

[More details](#) about malfunctions

[References](#) cited in table above

## EASYTRAK 2

Models 4515/4517/4518/4520/4542/  
4543/4544


<b>U.S. Survival Probability</b>	<b>Worldwide Malfunction Details</b>	<b>Product Advisories</b>
----------------------------------	--------------------------------------	---------------------------

### U.S. Summary

U.S. Registered Implants: 89,000  
U.S. Approval Date: August 2004  
U.S. Estimated Active Implants: 47,000

U.S. Chronic Lead Complications: 1,507  
U.S. Malfunctions: 614  
Without Compromised Therapy: 24  
With Compromised Therapy: 590

### Complications and Malfunctions




### U.S. Survival Probability

Year	1	2	3	4	5	6	7	8	9	10
Non Advisory Population	98.77 (-0.1/+0.1)	98.30 (-0.1/+0.1)	97.88 (-0.1/+0.1)	97.52 (-0.1/+0.1)	97.00 (-0.1/+0.1)	96.40 (-0.2/+0.2)	95.93 (-0.2/+0.2)	95.62 (-0.2/+0.2)	95.37 (-0.3/+0.3)	95.31 @ 113 mo. (-0.3/+0.3)
Registered Implants: 89000										
Effective Sample Size	71148	58858	47946	37889	28271	20341	13343	7183	1993	254

**EASYTRAK 2**

Models 4515/4517/4518/4520/4542/  
4543/4544

U.S. Survival Probability	Worldwide Malfunction Details	Product Advisories
---------------------------	-------------------------------	--------------------

<b>EASYTRAK 2</b> Models 4515/4517/4518/4520/4542/ 4543/4544 			
Worldwide Distribution: 161,000			
Worldwide Confirmed Malfunctions: 791			
	Without Compromised Therapy	With Compromised Therapy	Total
<b>Conductor</b>	<b>12</b>	<b>350</b>	<b>362</b>
<sup>26</sup> Conductor fracture	10	305	
<sup>28</sup> Non-patterned, Conductor	2	45	
<b>Crimp/Weld/Bond</b>	<b>-</b>	<b>-</b>	<b>0</b>
<b>Extrinsic</b>	<b>4</b>	<b>403</b>	<b>407</b>
<sup>30</sup> Unconfirmed Extrinsic	-	388	
<sup>31</sup> Inconclusive Extrinsic	4	15	
<b>Insulation</b>	<b>9</b>	<b>2</b>	<b>11</b>
<sup>29</sup> Non-patterned, Insulation	9	2	
<b>Other</b>	<b>6</b>	<b>5</b>	<b>11</b>
<sup>27</sup> Non-patterned, Other	6	5	
<b>WW Confirmed Malfunctions</b>	<b>31</b>	<b>760</b>	<b>791</b>

[More details](#) about malfunctions

[References](#) cited in table above

## EASYTRAK

Models 4510/4511/4512/4513/4535/  
4536/4537/4538


U.S. Survival Probability	Worldwide Malfunction Details	Product Advisories
---------------------------	-------------------------------	--------------------

### U.S. Summary

U.S. Registered Implants: 38,000  
U.S. Approval Date: May 2002  
U.S. Estimated Active Implants: 9,000

U.S. Chronic Lead Complications: 711  
U.S. Malfunctions: 185  
Without Compromised Therapy: 10  
With Compromised Therapy: 175

### Complications and Malfunctions



### U.S. Survival Probability

Year	1	2	3	4	5	6	7	8	9	10
Non Advisory Population	98.94 (-0.1/+0.1)	98.57 (-0.1/+0.1)	98.15 (-0.2/+0.1)	97.84 (-0.2/+0.2)	97.36 (-0.2/+0.2)	97.00 (-0.2/+0.2)	96.74 (-0.2/+0.2)	96.34 (-0.3/+0.3)	96.09 (-0.3/+0.3)	95.94 (-0.3/+0.3)
Registered Implants: 38000										
Effective Sample Size	30541	26257	22529	19365	16524	13960	11566	9650	8062	4402

## EASYTRAK

Models 4510/4511/4512/4513/4535/  
4536/4537/4538

U.S. Survival Probability	Worldwide Malfunction Details	Product Advisories
---------------------------	-------------------------------	--------------------

EASYTRAK Models 4510/4511/4512/4513/4535/ 4536/4537/4538			
Worldwide Distribution: 53,000			
Worldwide Confirmed Malfunctions: 202			
	Without Compromised Therapy	With Compromised Therapy	Total
<b>Conductor</b>	-	12	12
<sup>28</sup> Non-patterned, Conductor	-	12	
<b>Crimp/Weld/Bond</b>	-	-	0
<b>Extrinsic</b>	-	176	176
<sup>30</sup> Unconfirmed Extrinsic	-	175	
<sup>31</sup> Inconclusive Extrinsic	-	1	
<b>Insulation</b>	3	3	6
<sup>29</sup> Non-patterned, Insulation	3	3	
<b>Other</b>	7	1	8
<sup>27</sup> Non-patterned, Other	7	1	
<b>WW Confirmed Malfunctions</b>	<b>10</b>	<b>192</b>	<b>202</b>

[More details](#) about malfunctions

[References](#) cited in table above

## RELIANCE G 4-FRONT Dual Coil Active Fixation

Models 0658/0695/0696

U.S. Survival Probability	Worldwide Malfunction Details	Product Advisories
---------------------------	-------------------------------	--------------------

**RELIANCE G 4-FRONT Dual Coil Active Fixation Models 0658/0695/0696**



**Worldwide Distribution: 2,000**

**Worldwide Confirmed Malfunctions: 4**

	Without Compromised Therapy	With Compromised Therapy	Total
<b>Conductor</b>	-	-	<b>0</b>
<b>Crimp/Weld/Bond</b>	-	-	<b>0</b>
<b>Extrinsic</b>	-	<b>4</b>	<b>4</b>
<sup>30</sup> Unconfirmed Extrinsic	-	4	
<b>Insulation</b>	-	-	<b>0</b>
<b>Other</b>	-	-	<b>0</b>
<b>WW Confirmed Malfunctions</b>	<b>0</b>	<b>4</b>	<b>4</b>

[More details](#) about malfunctions

[References](#) cited in table above




## RELIANCE SG 4-FRONT Single Coil Active Fixation

Models 0657/0692/0693

U.S. Survival Probability	Worldwide Malfunction Details	Product Advisories
---------------------------	-------------------------------	--------------------

**RELIANCE SG 4-FRONT**  
**Single Coil Active Fixation**  
**Models 0657/0692/0693**



**Worldwide Distribution: 4,000**  
**Worldwide Confirmed Malfunctions: 0**

	Without Compromised Therapy	With Compromised Therapy	Total
<b>Conductor</b>	-	-	<b>0</b>
<b>Crimp/Weld/Bond</b>	-	-	<b>0</b>
<b>Extrinsic</b>	-	-	<b>0</b>
<b>Insulation</b>	-	-	<b>0</b>
<b>Other</b>	-	-	<b>0</b>
<b>WW Confirmed Malfunctions</b>	<b>0</b>	<b>0</b>	<b>0</b>

[More details](#) about malfunctions


[References](#) cited in table above

**Q-TRAK SQ Electrode**

Model 3010

<b>U.S. Survival Probability</b>	<b>Worldwide Malfunction Details</b>	<b>Product Advisories</b>
----------------------------------	--------------------------------------	---------------------------

U.S. Summary	
U.S. Approval Date: September 2012	U.S. Chronic Lead Complications: 0 U.S. Malfunctions:0 Without Compromised Therapy:0 With Compromised Therapy:0




**U.S. Survival Probability**

Year	1	2	3	4	5	6	7	8	9	10
Non Advisory Population	100.00 (-0.0/+0.0)	100.00 (-0.0/+0.0)	100.00 @ 25 mo. (-0.0/+0.0)	-	-	-	-	-	-	-
Registered Implants:	-	-	-	-	-	-	-	-	-	-

**Q-TRAK SQ Electrode**

Model 3010

<b>U.S. Survival Probability</b>	<b>Worldwide Malfunction Details</b>	<b>Product Advisories</b>
----------------------------------	--------------------------------------	---------------------------

Q-TRAK SQ Electrode Model 3010 			
<b>Worldwide Confirmed Malfunctions: 1</b>			
	Without Compromised Therapy	With Compromised Therapy	Total
<b>Conductor</b>	-	-	<b>0</b>
<b>Crimp/Weld/Bond</b>	-	<b>1</b>	<b>1</b>
<sup>37</sup> Weld fracture	-	1	
<b>Extrinsic</b>	-	-	<b>0</b>
<b>Insulation</b>	-	-	<b>0</b>
<b>Other</b>	-	-	<b>0</b>
<b>WW Confirmed Malfunctions</b>	<b>0</b>	<b>1</b>	<b>1</b>

[More details](#) about malfunctions

[References](#) cited in table above

**ENDOTAK RELIANCE G 4-Site  
Dual Coil, Active Fixation**


Models 0295/0296

U.S. Survival Probability	Worldwide Malfunction Details	Product Advisories
---------------------------	-------------------------------	--------------------

**U.S. Summary**

U.S. Registered Implants: 27,000	U.S. Chronic Lead Complications: 30
U.S. Approval Date: November 2010	U.S. Malfunctions: 18
U.S. Estimated Active Implants: 26,000	Without Compromised Therapy: 0
	With Compromised Therapy: 18

**Complications and Malfunctions**



**U.S. Survival Probability**


Year	1	2	3	4	5	6	7	8	9	10
Non Advisory Population	99.77 (-0.1/+0.1)	99.76 (-0.1/+0.1)	-	-	-	-	-	-	-	-
Registered Implants: 27000										
Effective Sample Size	12565	316	-	-	-	-	-	-	-	-

**ENDOTAK RELIANCE G 4-Site  
Dual Coil, Active Fixation**

Models 0295/0296

U.S. Survival Probability	Worldwide Malfunction Details	Product Advisories
---------------------------	-------------------------------	--------------------

**ENDOTAK RELIANCE G 4-Site  
Dual Coil, Active Fixation  
Models 0295/0296**



**Worldwide Distribution:** 51,000  
**Worldwide Confirmed Malfunctions:** 75

	Without Compromised Therapy	With Compromised Therapy	Total
<b>Conductor</b>	-	1	1
<sup>28</sup> Non-patterned, Conductor	-	1	
<b>Crimp/Weld/Bond</b>	-	-	0
<b>Extrinsic</b>	-	56	56
<sup>30</sup> Unconfirmed Extrinsic	-	56	
<b>Insulation</b>	7	8	15
<sup>29</sup> Non-patterned, Insulation	7	8	
<b>Other</b>	2	1	3
<sup>27</sup> Non-patterned, Other	2	1	
<b>WW Confirmed Malfunctions</b>	<b>9</b>	<b>66</b>	<b>75</b>

[More details](#) about malfunctions

[References](#) cited in table above

**ENDOTAK RELIANCE G 4-Site  
Dual Coil, Passive Fixation**


Models 0285/0286

U.S. Survival Probability	Worldwide Malfunction Details	Product Advisories
---------------------------	-------------------------------	--------------------

**U.S. Summary**

U.S. Registered Implants: 1,000	U.S. Chronic Lead Complications: 2
U.S. Approval Date: November 2010	U.S. Malfunctions: 1
U.S. Estimated Active Implants: 1,000	Without Compromised Therapy: 0
	With Compromised Therapy: 1

**Complications and Malfunctions**



**U.S. Survival Probability**


Year	1	2	3	4	5	6	7	8	9	10
Non Advisory Population	99.72 (-0.8/+0.2)	99.72 @ 17 mo. (-0.8/+0.2)	--	--	--	--	--	--	--	--
Registered Implants: 1000										
Effective Sample Size	386	209	--	--	--	--	--	--	--	--

**ENDOTAK RELIANCE G 4-Site  
Dual Coil, Passive Fixation**

Models 0285/0286

U.S. Survival Probability	Worldwide Malfunction Details	Product Advisories
---------------------------	-------------------------------	--------------------

**ENDOTAK RELIANCE G 4-Site  
Dual Coil, Passive Fixation  
Models 0285/0286**



**Worldwide Distribution:** 6,000  
**Worldwide Confirmed Malfunctions:** 3

	Without Compromised Therapy	With Compromised Therapy	Total
<b>Conductor</b>	-	-	<b>0</b>
<b>Crimp/Weld/Bond</b>	-	-	<b>0</b>
<b>Extrinsic</b>	-	<b>3</b>	<b>3</b>
<sup>30</sup> Unconfirmed Extrinsic	-	3	
<b>Insulation</b>	-	-	<b>0</b>
<b>Other</b>	-	-	<b>0</b>
<b>WW Confirmed Malfunctions</b>	<b>0</b>	<b>3</b>	<b>3</b>

[More details](#) about malfunctions

[References](#) cited in table above

**ENDOTAK RELIANCE 4-Site  
Dual Coil, Active Fixation**


Models 0275/0276

U.S. Survival Probability	Worldwide Malfunction Details	Product Advisories
---------------------------	-------------------------------	--------------------

**U.S. Summary**

U.S. Registered Implants: 2,000	U.S. Chronic Lead Complications: 3
U.S. Approval Date: November 2010	U.S. Malfunctions: 3
U.S. Estimated Active Implants: 2,000	Without Compromised Therapy: 0
	With Compromised Therapy: 3

**Complications and Malfunctions**



**U.S. Survival Probability**

Year	1	2	3	4	5	6	7	8	9	10
Non Advisory Population	99.78 (-0.3/+0.1)	99.63 @ 21 mo. (-0.6/+0.2)	--	--	--	--	--	--	--	--
Registered Implants: 2000										
Effective Sample Size	1284	280	--	--	--	--	--	--	--	--




**ENDOTAK RELIANCE 4-Site  
Dual Coil, Active Fixation**

Models 0275/0276

U.S. Survival Probability	Worldwide Malfunction Details	Product Advisories
---------------------------	-------------------------------	--------------------

**ENDOTAK RELIANCE 4-Site  
Dual Coil, Active Fixation  
Models 0275/0276**



**Worldwide Distribution:** 3,000  
**Worldwide Confirmed Malfunctions:** 3

	Without Compromised Therapy	With Compromised Therapy	Total
<b>Conductor</b>	-	-	<b>0</b>
<b>Crimp/Weld/Bond</b>	-	-	<b>0</b>
<b>Extrinsic</b>	-	<b>3</b>	<b>3</b>
<sup>30</sup> Unconfirmed Extrinsic	-	3	
<b>Insulation</b>	-	-	<b>0</b>
<b>Other</b>	-	-	<b>0</b>
<b>WW Confirmed Malfunctions</b>	<b>0</b>	<b>3</b>	<b>3</b>

[More details](#) about malfunctions


[References](#) cited in table above

**ENDOTAK RELIANCE 4-Site  
Dual Coil, Passive Fixation**

Models 0265/0266

U.S. Survival Probability	Worldwide Malfunction Details	Product Advisories
---------------------------	-------------------------------	--------------------

**ENDOTAK RELIANCE 4-Site  
Dual Coil, Passive Fixation  
Models 0265/0266**



**Worldwide Distribution: 500**  
**Worldwide Confirmed Malfunctions: 0**

	Without Compromised Therapy	With Compromised Therapy	Total
<b>Conductor</b>	-	-	<b>0</b>
<b>Crimp/Weld/Bond</b>	-	-	<b>0</b>
<b>Extrinsic</b>	-	-	<b>0</b>
<b>Insulation</b>	-	-	<b>0</b>
<b>Other</b>	-	-	<b>0</b>
<b>WW Confirmed Malfunctions</b>	<b>0</b>	<b>0</b>	<b>0</b>

[More details](#) about malfunctions

[References](#) cited in table above

## ENDOTAK RELIANCE SG 4-Site Single Coil, Active Fixation


Models 0292/0293

U.S. Survival Probability	Worldwide Malfunction Details	Product Advisories
---------------------------	-------------------------------	--------------------

### U.S. Summary

U.S. Registered Implants: 18,000	U.S. Chronic Lead Complications: 20
U.S. Approval Date: November 2010	U.S. Malfunctions: 7
U.S. Estimated Active Implants: 17,000	Without Compromised Therapy: 1
	With Compromised Therapy: 6

### Complications and Malfunctions



### U.S. Survival Probability


Year	1	2	3	4	5	6	7	8	9	10
Non Advisory Population	99.81 (-0.1/+0.1)	99.72 @ 23 mo. (-0.2/+0.1)	--	--	--	--	--	--	--	--
Registered Implants: 18000										
Effective Sample Size	6385	405	--	--	--	--	--	--	--	--

**ENDOTAK RELIANCE SG 4-Site  
Single Coil, Active Fixation**

Models 0292/0293

U.S. Survival Probability	Worldwide Malfunction Details	Product Advisories
---------------------------	-------------------------------	--------------------

**ENDOTAK RELIANCE SG 4-Site  
Single Coil, Active Fixation  
Models 0292/0293**



**Worldwide Distribution:** 38,000  
**Worldwide Confirmed Malfunctions:** 24

	Without Compromised Therapy	With Compromised Therapy	Total
<b>Conductor</b>	-	1	1
<sup>28</sup> Non-patterned, Conductor	-	1	
<b>Crimp/Weld/Bond</b>	-	-	0
<b>Extrinsic</b>	-	17	17
<sup>30</sup> Unconfirmed Extrinsic	-	17	
<b>Insulation</b>	2	4	6
<sup>29</sup> Non-patterned, Insulation	2	4	
<b>Other</b>	-	-	0
<b>WW Confirmed Malfunctions</b>	2	22	24

[More details](#) about malfunctions

[References](#) cited in table above

**ENDOTAK RELIANCE SG 4-Site  
Single Coil, Passive Fixation**

Models 0282/0283

U.S. Survival Probability	Worldwide Malfunction Details	Product Advisories
---------------------------	-------------------------------	--------------------

**ENDOTAK RELIANCE SG 4-Site  
Single Coil, Passive Fixation  
Models 0282/0283**



**Worldwide Distribution: 2,000**  
**Worldwide Confirmed Malfunctions: 3**

	Without Compromised Therapy	With Compromised Therapy	Total
<b>Conductor</b>	-	-	<b>0</b>
<b>Crimp/Weld/Bond</b>	-	-	<b>0</b>
<b>Extrinsic</b>	<b>2</b>	<b>1</b>	<b>3</b>
<sup>30</sup> Unconfirmed Extrinsic	-	1	
<sup>31</sup> Inconclusive Extrinsic	2	-	
<b>Insulation</b>	-	-	<b>0</b>
<b>Other</b>	-	-	<b>0</b>
<b>WW Confirmed Malfunctions</b>	<b>2</b>	<b>1</b>	<b>3</b>

[More details](#) about malfunctions

[References](#) cited in table above

## ENDOTAK RELIANCE G Dual Coil, Active Fixation


Models 0164/0165/0166/0167/0184/  
0185/0186/0187

U.S. Survival Probability	Worldwide Malfunction Details	Product Advisories	Longitude Survival Probability
---------------------------	-------------------------------	--------------------	--------------------------------

### U.S. Summary

U.S. Registered Implants: 186,000	U.S. Chronic Lead Complications: 552
U.S. Approval Date: May 2004	U.S. Malfunctions: 584
U.S. Estimated Active Implants: 118,000	Without Compromised Therapy: 86
	With Compromised Therapy: 498

### Complications and Malfunctions



### U.S. Survival Probability

Year	1	2	3	4	5	6	7	8	9	10
Non Advisory Population	99.77 (-0.0/+0.0)	99.66 (-0.0/+0.0)	99.57 (-0.0/+0.0)	99.45 (-0.0/+0.0)	99.29 (-0.1/+0.0)	99.08 (-0.1/+0.1)	98.90 (-0.1/+0.1)	98.67 (-0.1/+0.1)	98.51 (-0.1/+0.1)	98.39 @ 117 mo. (-0.2/+0.2)
Registered Implants: 186000										
Effective Sample Size	162289	141037	112594	87526	62704	41778	26348	14564	4643	340

**ENDOTAK RELIANCE G  
Dual Coil, Active Fixation**

Models 0164/0165/0166/0167/0184/  
0185/0186/0187

U.S. Survival Probability	Worldwide Malfunction Details	Product Advisories	Longitude Survival Probability
---------------------------	-------------------------------	--------------------	--------------------------------

**ENDOTAK RELIANCE G  
Dual Coil, Active Fixation  
Models 0164/0165/0166/0167/0184/  
0185/0186/0187**



**Worldwide Distribution: 250,000  
Worldwide Confirmed Malfunctions: 847**

	Without Compromised Therapy	With Compromised Therapy	Total
<b>Conductor</b>	<b>2</b>	<b>77</b>	<b>79</b>
<sup>25</sup> Conductor fracture	-	49	
<sup>28</sup> Non-patterned, Conductor	2	28	
<b>Crimp/Weld/Bond</b>	<b>2</b>	<b>-</b>	<b>2</b>
<sup>32</sup> Non-patterned, Crimp, Weld, Bond	2	-	
<b>Extrinsic</b>	<b>10</b>	<b>562</b>	<b>572</b>
<sup>30</sup> Unconfirmed Extrinsic	-	560	
<sup>31</sup> Inconclusive Extrinsic	10	2	
<b>Insulation</b>	<b>113</b>	<b>47</b>	<b>160</b>
<sup>29</sup> Non-patterned, Insulation	113	47	
<b>Other</b>	<b>21</b>	<b>13</b>	<b>34</b>
<sup>27</sup> Non-patterned, Other	21	13	
<b>WW Confirmed Malfunctions</b>	<b>148</b>	<b>699</b>	<b>847</b>

[More details](#) about malfunctions

[References](#) cited in table above

**ENDOTAK RELIANCE G  
Dual Coil, Active Fixation Longitude\***


Models 0164/0165/0166/0167/0184/  
0185/0186/0187

U.S. Survival Probability	Worldwide Malfunction Details	Product Advisories	Longitude Survival Probability
---------------------------	-------------------------------	--------------------	--------------------------------

**Longitude Registry Summary Data**

Leads Enrolled: 619	Chronic Lead Complications: 1
Leads Active: 476	Malfunctions: 1
Cumulative Followup Months : 23,265	Without Compromised Therapy: 0
	With Compromised Therapy: 1

**Complications and Malfunctions**



**Longitude Registry Survival Probability**

Year	1	2	3	4	5	6	7	8	9	10
Longitude	100.00 (-0.0/+0.0)	99.58 (-1.3/+0.3)	99.58 (-1.3/+0.3)	99.58 (-1.3/+0.3)	99.58 @ 60 mo. (-1.3/+0.3)	—	—	—	—	—
Registered Implants: 619										
Effective Sample Size	522	440	308	128	50	—	—	—	—	—



**ENDOTAK RELIANCE G  
Dual Coil, Passive Fixation**


Models 0174/0175/0176/0177

U.S. Survival Probability	Worldwide Malfunction Details	Product Advisories
---------------------------	-------------------------------	--------------------

**U.S. Summary**

U.S. Registered Implants: 14,000	U.S. Chronic Lead Complications: 81
U.S. Approval Date: May 2004	U.S. Malfunctions: 54
U.S. Estimated Active Implants: 8,000	Without Compromised Therapy: 8
	With Compromised Therapy: 46

**Complications and Malfunctions**



**U.S. Survival Probability**

Year	1	2	3	4	5	6	7	8	9	10
Non Advisory Population	99.72 (-0.1/+0.1)	99.57 (-0.1/+0.1)	99.37 (-0.2/+0.1)	99.17 (-0.2/+0.2)	98.90 (-0.2/+0.2)	98.62 (-0.3/+0.2)	98.17 (-0.4/+0.3)	97.79 (-0.5/+0.4)	97.39 (-0.7/+0.5)	97.10 @ 115 mo. (-1.0/+0.7)
Registered Implants: 14000										
Effective Sample Size	11661	10028	8214	6569	4959	3543	2394	1441	541	205

**ENDOTAK RELIANCE G  
Dual Coil, Passive Fixation**

Models 0174/0175/0176/0177

U.S. Survival Probability	Worldwide Malfunction Details	Product Advisories
---------------------------	-------------------------------	--------------------

**ENDOTAK RELIANCE G  
Dual Coil, Passive Fixation  
Models 0174/0175/0176/0177**



**Worldwide Distribution: 38,000  
Worldwide Confirmed Malfunctions: 146**

	Without Compromised Therapy	With Compromised Therapy	Total
<b>Conductor</b>	-	<b>16</b>	<b>16</b>
<sup>25</sup> Conductor fracture	-	13	
<sup>28</sup> Non-patterned, Conductor	-	3	
<b>Crimp/Weld/Bond</b>	-	<b>1</b>	<b>1</b>
<sup>36</sup> Conductor connection	-	1	
<b>Extrinsic</b>	<b>7</b>	<b>91</b>	<b>98</b>
<sup>30</sup> Unconfirmed Extrinsic	-	87	
<sup>31</sup> Inconclusive Extrinsic	7	4	
<b>Insulation</b>	<b>15</b>	<b>9</b>	<b>24</b>
<sup>29</sup> Non-patterned, Insulation	15	9	
<b>Other</b>	<b>7</b>	-	<b>7</b>
<sup>27</sup> Non-patterned, Other	7	-	
<b>WW Confirmed Malfunctions</b>	<b>29</b>	<b>117</b>	<b>146</b>

[More details](#) about malfunctions

[References](#) cited in table above

**ENDOTAK RELIANCE  
Dual Coil, Active Fixation**


Models 0157/0158/0159

<b>U.S. Survival Probability</b>	<b>Worldwide Malfunction Details</b>	<b>Product Advisories</b>
----------------------------------	--------------------------------------	---------------------------

**U.S. Summary**

U.S. Registered Implants: 97,000	U.S. Chronic Lead Complications: 374
U.S. Approval Date: July 2002	U.S. Malfunctions: 242
U.S. Estimated Active Implants: 43,000	Without Compromised Therapy: 31
	With Compromised Therapy: 211

**Complications and Malfunctions**



**U.S. Survival Probability**

Year	1	2	3	4	5	6	7	8	9	10
Non Advisory Population	99.84 (-0.0/+0.0)	99.78 (-0.0/+0.0)	99.68 (-0.0/+0.0)	99.57 (-0.1/+0.0)	99.46 (-0.1/+0.1)	99.32 (-0.1/+0.1)	99.11 (-0.1/+0.1)	98.96 (-0.1/+0.1)	98.80 (-0.1/+0.1)	98.66 (-0.1/+0.1)
Registered Implants: 97000										
Effective Sample Size	84651	74908	64747	55362	46282	38286	31219	24879	18447	10161

**ENDOTAK RELIANCE  
Dual Coil, Active Fixation**

Models 0157/0158/0159

U.S. Survival Probability	Worldwide Malfunction Details	Product Advisories
---------------------------	-------------------------------	--------------------

**ENDOTAK RELIANCE  
Dual Coil, Active Fixation  
Models 0157/0158/0159**



**Worldwide Distribution:** 113,000

**Worldwide Confirmed Malfunctions:** 274

	Without Compromised Therapy	With Compromised Therapy	Total
<b>Conductor</b>	-	<b>16</b>	<b>16</b>
<sup>25</sup> Conductor fracture	-	12	
<sup>28</sup> Non-patterned, Conductor	-	4	
<b>Crimp/Weld/Bond</b>	<b>3</b>	<b>1</b>	<b>4</b>
<sup>5</sup> Seal rings	2	1	
<sup>32</sup> Non-patterned, Crimp, Weld, Bond	1	-	
<b>Extrinsic</b>	<b>1</b>	<b>193</b>	<b>194</b>
<sup>30</sup> Unconfirmed Extrinsic	-	192	
<sup>31</sup> Inconclusive Extrinsic	1	1	
<b>Insulation</b>	<b>30</b>	<b>19</b>	<b>49</b>
<sup>29</sup> Non-patterned, Insulation	30	19	
<b>Other</b>	<b>8</b>	<b>3</b>	<b>11</b>
<sup>27</sup> Non-patterned, Other	8	3	
<b>WW Confirmed Malfunctions</b>	<b>42</b>	<b>232</b>	<b>274</b>

[More details](#) about malfunctions

[References](#) cited in table above

**ENDOTAK RELIANCE  
Dual Coil, Passive Fixation**


Models 0147/0148/0149

<b>U.S. Survival Probability</b>	<b>Worldwide Malfunction Details</b>	<b>Product Advisories</b>
----------------------------------	--------------------------------------	---------------------------

**U.S. Summary**

<b>U.S. Registered Implants:</b> 33,000 <b>U.S. Approval Date:</b> October 2000 <b>U.S. Estimated Active Implants:</b> 12,000	<b>U.S. Chronic Lead Complications:</b> 215 <b>U.S. Malfunctions:</b> 90 <b>Without Compromised Therapy:</b> 7 <b>With Compromised Therapy:</b> 83
---	---

**Complications and Malfunctions**




**U.S. Survival Probability**

Year	1	2	3	4	5	6	7	8	9	10
Non Advisory Population	99.79 (-0.1/+0.0)	99.73 (-0.1/+0.1)	99.60 (-0.1/+0.1)	99.45 (-0.1/+0.1)	99.26 (-0.1/+0.1)	99.08 (-0.1/+0.1)	98.88 (-0.2/+0.1)	98.66 (-0.2/+0.2)	98.55 (-0.2/+0.2)	98.41 (-0.2/+0.2)
Registered Implants: 33000										
Effective Sample Size	28507	25370	22470	19795	17351	15133	13085	11210	9445	7420

**ENDOTAK RELIANCE**  
**Dual Coil, Passive Fixation**

Models 0147/0148/0149

U.S. Survival Probability	Worldwide Malfunction Details	Product Advisories
---------------------------	-------------------------------	--------------------

<b>ENDOTAK RELIANCE</b> <b>Dual Coil, Passive Fixation</b> <b>Models 0147/0148/0149</b> 			
<b>Worldwide Distribution:</b> 67,000			
<b>Worldwide Confirmed Malfunctions:</b> 191			
	Without Compromised Therapy	With Compromised Therapy	Total
<b>Conductor</b>	-	<b>10</b>	<b>10</b>
<sup>25</sup> Conductor fracture	-	3	
<sup>28</sup> Non-patterned, Conductor	-	7	
<b>Crimp/Weld/Bond</b>	-	<b>2</b>	<b>2</b>
<sup>36</sup> Conductor connection	-	2	
<b>Extrinsic</b>	<b>8</b>	<b>120</b>	<b>128</b>
<sup>30</sup> Unconfirmed Extrinsic	-	118	
<sup>31</sup> Inconclusive Extrinsic	8	2	
<b>Insulation</b>	<b>21</b>	<b>24</b>	<b>45</b>
<sup>29</sup> Non-patterned, Insulation	21	24	
<b>Other</b>	<b>2</b>	<b>4</b>	<b>6</b>
<sup>6</sup> Manufacturing material	-	1	
<sup>27</sup> Non-patterned, Other	2	3	
<b>WW Confirmed Malfunctions</b>	<b>31</b>	<b>160</b>	<b>191</b>

[More details](#) about malfunctions

[References](#) cited in table above

**ENDOTAK RELIANCE SG**  
**Single Coil, Active Fixation**


Models 0160/0161/0162/0180/0181/  
 0182

U.S. Survival Probability	Worldwide Malfunction Details	Product Advisories
---------------------------	-------------------------------	--------------------

**U.S. Summary**

U.S. Registered Implants: 25,000	U.S. Chronic Lead Complications: 69
U.S. Approval Date: May 2004	U.S. Malfunctions: 112
U.S. Estimated Active Implants: 20,000	Without Compromised Therapy: 17
	With Compromised Therapy: 95

**Complications and Malfunctions**



**U.S. Survival Probability**

Year	1	2	3	4	5	6	7	8	9	10
Non Advisory Population	99.67 (-0.1/+0.1)	99.46 (-0.1/+0.1)	99.32 (-0.1/+0.1)	99.00 (-0.2/+0.2)	98.72 (-0.3/+0.2)	98.33 (-0.4/+0.3)	97.94 (-0.5/+0.4)	97.81 (-0.6/+0.5)	97.54 @ 104 mo. (-0.9/+0.7)	-
Registered Implants: 25000										
Effective Sample Size	19941	15801	9742	6110	3420	1687	855	438	228	-

**ENDOTAK RELIANCE SG**  
**Single Coil, Active Fixation**

Models 0160/0161/0162/0180/0181/  
 0182

U.S. Survival Probability	Worldwide Malfunction Details	Product Advisories
---------------------------	-------------------------------	--------------------

**ENDOTAK RELIANCE SG**  
**Single Coil, Active Fixation**  
 Models 0160/0161/0162/0180/0181/  
 0182



**Worldwide Distribution: 52,000**  
**Worldwide Confirmed Malfunctions: 244**

	Without Compromised Therapy	With Compromised Therapy	Total
<b>Conductor</b>	<b>1</b>	<b>39</b>	<b>40</b>
<sup>25</sup> Conductor fracture	1	33	
<sup>28</sup> Non-patterned, Conductor	-	6	
<b>Crimp/Weld/Bond</b>	<b>-</b>	<b>-</b>	<b>0</b>
<b>Extrinsic</b>	<b>1</b>	<b>137</b>	<b>138</b>
<sup>30</sup> Unconfirmed Extrinsic	-	137	
<sup>31</sup> Inconclusive Extrinsic	1	-	
<b>Insulation</b>	<b>42</b>	<b>13</b>	<b>55</b>
<sup>29</sup> Non-patterned, Insulation	42	13	
<b>Other</b>	<b>5</b>	<b>6</b>	<b>11</b>
<sup>27</sup> Non-patterned, Other	5	6	
<b>WW Confirmed Malfunctions</b>	<b>49</b>	<b>195</b>	<b>244</b>

[More details](#) about malfunctions

[References](#) cited in table above



**ENDOTAK RELIANCE SG**  
**Single Coil, Passive Fixation**


Models 0170/0171/0172/0173

U.S. Survival Probability	Worldwide Malfunction Details	Product Advisories
---------------------------	-------------------------------	--------------------

**U.S. Summary**

U.S. Registered Implants: 1,000	U.S. Chronic Lead Complications: 1
U.S. Approval Date: May 2004	U.S. Malfunctions:3
U.S. Estimated Active Implants: 1,000	Without Compromised Therapy:1
	With Compromised Therapy:2

**Complications and Malfunctions**



**U.S. Survival Probability**

Year	1	2	3	4	5	6	7	8	9	10
Non Advisory Population	99.85 (-0.9/+0.1)	99.26 (-1.6/+0.5)	99.26 @ 30 mo. (-1.6/+0.5)	--	--	--	--	--	--	--
Registered Implants: 1000										
Effective Sample Size	451	276	204	--	--	--	--	--	--	--

**ENDOTAK RELIANCE SG**  
**Single Coil, Passive Fixation**

Models 0170/0171/0172/0173

U.S. Survival Probability	Worldwide Malfunction Details	Product Advisories
---------------------------	-------------------------------	--------------------

**ENDOTAK RELIANCE SG**  
**Single Coil, Passive Fixation**  
**Models 0170/0171/0172/0173**



**Worldwide Distribution: 3,000**  
**Worldwide Confirmed Malfunctions: 19**

	Without Compromised Therapy	With Compromised Therapy	Total
<b>Conductor</b>	<b>1</b>	<b>2</b>	<b>3</b>
<sup>25</sup> Conductor fracture	1	2	
<b>Crimp/Weld/Bond</b>	<b>-</b>	<b>-</b>	<b>0</b>
<b>Extrinsic</b>	<b>3</b>	<b>10</b>	<b>13</b>
<sup>30</sup> Unconfirmed Extrinsic	-	9	
<sup>31</sup> Inconclusive Extrinsic	3	1	
<b>Insulation</b>	<b>3</b>	<b>-</b>	<b>3</b>
<sup>29</sup> Non-patterned, Insulation	3	-	
<b>Other</b>	<b>-</b>	<b>-</b>	<b>0</b>
<b>WW Confirmed Malfunctions</b>	<b>7</b>	<b>12</b>	<b>19</b>

[More details](#) about malfunctions

[References](#) cited in table above

# ENDOTAK RELIANCE S Single Coil, Active Fixation


Models 0137/0138

U.S. Survival Probability	Worldwide Malfunction Details	Product Advisories
---------------------------	-------------------------------	--------------------

## U.S. Summary

<p>U.S. Registered Implants: 2,000                  U.S. Approval Date: July 2002                  U.S. Estimated Active Implants: 1,000</p>	<p>U.S. Chronic Lead Complications: 7                  U.S. Malfunctions: 8                  Without Compromised Therapy: 2                  With Compromised Therapy: 6</p>
--	--

## Complications and Malfunctions



## U.S. Survival Probability

Year	1	2	3	4	5	6	7	8	9	10
Non Advisory Population	99.82 <small>(-0.3/+0.1)</small>	99.70 <small>(-0.4/+0.2)</small>	99.62 <small>(-0.4/+0.2)</small>	99.53 <small>(-0.5/+0.2)</small>	99.40 <small>(-0.6/+0.3)</small>	99.11 <small>(-0.8/+0.4)</small>	99.11 <small>(-0.8/+0.4)</small>	98.90 <small>(-1.0/+0.5)</small>	98.59 <small>(-1.3/+0.7)</small>	97.79 @ <small>118 mo. (-2.0/+1.1)</small>
Registered Implants: 2000										
Effective Sample Size	1990	1620	1296	1015	771	597	462	370	284	206

**ENDOTAK RELIANCE S**  
**Single Coil, Active Fixation**

Models 0137/0138

U.S. Survival Probability	Worldwide Malfunction Details	Product Advisories
---------------------------	-------------------------------	--------------------

**ENDOTAK RELIANCE S**  
**Single Coil, Active Fixation**  
**Models 0137/0138**



**Worldwide Distribution:** 5,000  
**Worldwide Confirmed Malfunctions:** 17

	Without Compromised Therapy	With Compromised Therapy	Total
<b>Conductor</b>	-	<b>2</b>	<b>2</b>
<sup>25</sup> Conductor fracture	-	2	
<b>Crimp/Weld/Bond</b>	-	-	<b>0</b>
<b>Extrinsic</b>	-	<b>8</b>	<b>8</b>
<sup>30</sup> Unconfirmed Extrinsic	-	8	
<b>Insulation</b>	<b>5</b>	<b>1</b>	<b>6</b>
<sup>26</sup> Non-patterned, Insulation	5	1	
<b>Other</b>	<b>1</b>	-	<b>1</b>
<sup>27</sup> Non-patterned, Other	1	-	
<b>WW Confirmed Malfunctions</b>	<b>6</b>	<b>11</b>	<b>17</b>

[More details](#) about malfunctions

[References](#) cited in table above

## ENDOTAK RELIANCE S Single Coil, Passive Fixation


Models 0127/0128

U.S. Survival Probability	Worldwide Malfunction Details	Product Advisories
---------------------------	-------------------------------	--------------------

### U.S. Summary

U.S. Registered Implants: 1,000	U.S. Chronic Lead Complications: 4
U.S. Approval Date: October 2000	U.S. Malfunctions: 4
U.S. Estimated Active Implants: 200	Without Compromised Therapy: 0
	With Compromised Therapy: 4

### Complications and Malfunctions



### U.S. Survival Probability

Year	1	2	3	4	5	6	7	8	9	10
Non Advisory Population	99.67 (-1.0/+0.3)	99.67 (-1.0/+0.3)	99.67 (-1.0/+0.3)	99.43 (-1.2/+0.4)	99.43 (-1.2/+0.4)	99.43 (-1.2/+0.4)	99.03 (-1.8/+0.6)	99.03 @ 95 mo. (-1.8/+0.6)	-	-
Registered Implants: 1000										
Effective Sample Size	560	488	428	369	329	278	236	202	-	-

## ENDOTAK RELIANCE S Single Coil, Passive Fixation

Models 0127/0128

U.S. Survival Probability	Worldwide Malfunction Details	Product Advisories
---------------------------	-------------------------------	--------------------

### ENDOTAK RELIANCE S Single Coil, Passive Fixation Models 0127/0128



Worldwide Distribution: 4,000  
Worldwide Confirmed Malfunctions: 21

	Without Compromised Therapy	With Compromised Therapy	Total
<b>Conductor</b>	-	2	2
<sup>28</sup> Non-patterned, Conductor	-	2	
<b>Crimp/Weld/Bond</b>	-	-	0
<b>Extrinsic</b>	1	12	13
<sup>30</sup> Unconfirmed Extrinsic	-	12	
<sup>31</sup> Inconclusive Extrinsic	1	-	
<b>Insulation</b>	3	2	5
<sup>29</sup> Non-patterned, Insulation	3	2	
<b>Other</b>	1	-	1
<sup>27</sup> Non-patterned, Other	1	-	
<b>WW Confirmed Malfunctions</b>	<b>5</b>	<b>16</b>	<b>21</b>

[More details](#) about malfunctions

[References](#) cited in table above

## INGEVITY Positive Fixation

Models 7640/7641/7642/7740/7741/  
7742

U.S. Survival Probability	Worldwide Malfunction Details	Product Advisories
---------------------------	-------------------------------	--------------------

INGEVITY Positive Fixation Models 7640/7641/7642/7740/7741/ 7742			
Worldwide Distribution: 2,000			
Worldwide Confirmed Malfunctions: 0			
	Without Compromised Therapy	With Compromised Therapy	Total
<b>Conductor</b>	-	-	<b>0</b>
<b>Crimp/Weld/Bond</b>	-	-	<b>0</b>
<b>Extrinsic</b>	-	-	<b>0</b>
<b>Insulation</b>	-	-	<b>0</b>
<b>Other</b>	-	-	<b>0</b>
<b>WW Confirmed Malfunctions</b>	<b>0</b>	<b>0</b>	<b>0</b>


[More details](#) about malfunctions

[References](#) cited in table above

## INGEVITY Passive Fixation

Models 7631/7632/7731/7732

<b>U.S. Survival Probability</b>	<b>Worldwide Malfunction Details</b>	<b>Product Advisories</b>
----------------------------------	--------------------------------------	---------------------------

**INGEVITY Passive Fixation**  
**Models 7631/7632/7731/7732**


**Worldwide Distribution: 1,000**  
**Worldwide Confirmed Malfunctions: 0**

	Without Compromised Therapy	With Compromised Therapy	Total
<b>Conductor</b>	-	-	<b>0</b>
<b>Crimp/Weld/Bond</b>	-	-	<b>0</b>
<b>Extrinsic</b>	-	-	<b>0</b>
<b>Insulation</b>	-	-	<b>0</b>
<b>Other</b>	-	-	<b>0</b>
<b>WW Confirmed Malfunctions</b>	<b>0</b>	<b>0</b>	<b>0</b>

[More details](#) about malfunctions


[References](#) cited in table above



## INGEVITY Atrial J Passive Fixation

Models 7635/7636/7735/7736

<b>U.S. Survival Probability</b>	<b>Worldwide Malfunction Details</b>	<b>Product Advisories</b>
----------------------------------	--------------------------------------	---------------------------

**INGEVITY Atrial J Passive Fixation**  
**Models 7635/7636/7735/7736**


**Worldwide Distribution: 1,000**  
**Worldwide Confirmed Malfunctions: 0**

	Without Compromised Therapy	With Compromised Therapy	Total
<b>Conductor</b>	-	-	<b>0</b>
<b>Crimp/Weld/Bond</b>	-	-	<b>0</b>
<b>Extrinsic</b>	-	-	<b>0</b>
<b>Insulation</b>	-	-	<b>0</b>
<b>Other</b>	-	-	<b>0</b>
<b>WW Confirmed Malfunctions</b>	<b>0</b>	<b>0</b>	<b>0</b>


[More details](#) about malfunctions

[References](#) cited in table above

**FLEXTEND 2 Active Fixation**

Models 4095/4096/4097

<b>U.S. Survival Probability</b>	<b>Worldwide Malfunction Details</b>	<b>Product Advisories</b>
----------------------------------	--------------------------------------	---------------------------

<b>FLEXTEND 2 Active Fixation</b> <b>Models 4095/4096/4097</b> 			
<b>Worldwide Distribution: 148,000</b>			
<b>Worldwide Confirmed Malfunctions: 193</b>			
	Without Compromised Therapy	With Compromised Therapy	Total
<b>Conductor</b>	<b>3</b>	<b>32</b>	<b>35</b>
<sup>7</sup> Lead conductor	2	18	
<sup>33</sup> Conductor damage	1	14	
<b>Crimp/Weld/Bond</b>	<b>-</b>	<b>-</b>	<b>0</b>
<b>Extrinsic</b>	<b>1</b>	<b>100</b>	<b>101</b>
<sup>30</sup> Unconfirmed Extrinsic	-	100	
<sup>31</sup> Inconclusive Extrinsic	1	-	
<b>Insulation</b>	<b>38</b>	<b>7</b>	<b>45</b>
<sup>2</sup> Inner insulation abrasion	3	-	
<sup>29</sup> Non-patterned, Insulation	4	-	
<sup>34</sup> Insulation damage	31	7	
<b>Other</b>	<b>11</b>	<b>1</b>	<b>12</b>
<sup>27</sup> Non-patterned, Other	10	1	
<b>WW Confirmed Malfunctions</b>	<b>53</b>	<b>140</b>	<b>193</b>

[More details](#) about malfunctions

[References](#) cited in table above

## FLEXTEND Active Fixation

Models 4086/4087/4088


<b>U.S. Survival Probability</b>	<b>Worldwide Malfunction Details</b>	<b>Product Advisories</b>
----------------------------------	--------------------------------------	---------------------------

### U.S. Summary

U.S. Registered Implants: 224,000  
 U.S. Approval Date: February 2002  
 U.S. Estimated Active Implants: 107,000

U.S. Chronic Lead Complications: 2,313  
 U.S. Malfunctions: 789  
 Without Compromised Therapy: 112  
 With Compromised Therapy: 677

### Complications and Malfunctions



### U.S. Survival Probability

Year	1	2	3	4	5	6	7	8	9	10
Non Advisory Population	99.60 (-0.0/+0.0)	99.39 (-0.0/+0.0)	99.19 (-0.0/+0.0)	98.92 (-0.1/+0.0)	98.63 (-0.1/+0.1)	98.28 (-0.1/+0.1)	97.91 (-0.1/+0.1)	97.55 (-0.1/+0.1)	97.22 (-0.1/+0.1)	96.95 (-0.1/+0.1)
Registered Implants: 224000										
Effective Sample Size	187392	161591	138192	116424	96877	79820	61378	42935	27055	14112

## FLEXTEND Active Fixation

Models 4086/4087/4088

<b>U.S. Survival Probability</b>	<b>Worldwide Malfunction Details</b>	<b>Product Advisories</b>
----------------------------------	--------------------------------------	---------------------------

### FLEXTEND Active Fixation Models 4086/4087/4088



**Worldwide Distribution:** 271,000

**Worldwide Confirmed Malfunctions:** 874

	Without Compromised Therapy	With Compromised Therapy	Total
<b>Conductor</b>	<b>6</b>	<b>169</b>	<b>175</b>
<sup>7</sup> Lead conductor	2	79	
<sup>28</sup> Non-patterned, Conductor	-	7	
<sup>33</sup> Conductor damage	4	83	
<b>Crimp/Weld/Bond</b>	<b>-</b>	<b>-</b>	<b>0</b>
<b>Extrinsic</b>	<b>2</b>	<b>562</b>	<b>564</b>
<sup>30</sup> Unconfirmed Extrinsic	-	560	
<sup>31</sup> Inconclusive Extrinsic	2	2	
<b>Insulation</b>	<b>97</b>	<b>21</b>	<b>118</b>
<sup>2</sup> Inner insulation abrasion	19	4	
<sup>29</sup> Non-patterned, Insulation	8	-	
<sup>34</sup> Insulation damage	70	17	
<b>Other</b>	<b>14</b>	<b>3</b>	<b>17</b>
<sup>27</sup> Non-patterned, Other	14	3	
<b>WW Confirmed Malfunctions</b>	<b>119</b>	<b>755</b>	<b>874</b>

[More details](#) about malfunctions

[References](#) cited in table above

CRM PRODUCT PERFORMANCE REPORT Q1 2014

**FINELINE II EZ/FINELINE II Sterox EZ  
Positive Fixation (Polyurethane)**


Models 4463/4464/4465/4469/4470/  
4471

U.S. Survival Probability	Worldwide Malfunction Details	Product Advisories	Longitude Survival Probability
---------------------------	-------------------------------	--------------------	--------------------------------

**U.S. Summary**

U.S. Registered Implants: 394,000	U.S. Chronic Lead Complications: 1,644
U.S. Approval Date: January 2000	U.S. Malfunctions: 414
U.S. Estimated Active Implants: 236,000	Without Compromised Therapy: 15
	With Compromised Therapy: 399

**Complications and Malfunctions**



**U.S. Survival Probability**

Year	1	2	3	4	5	6	7	8	9	10
Non Advisory Population	99.82 (-0.0/+0.0)	99.73 (-0.0/+0.0)	99.65 (-0.0/+0.0)	99.55 (-0.0/+0.0)	99.43 (-0.0/+0.0)	99.28 (-0.0/+0.0)	99.10 (-0.0/+0.0)	98.95 (-0.1/+0.1)	98.80 (-0.1/+0.1)	98.66 (-0.1/+0.1)
Registered Implants: 393000										
Effective Sample Size	326512	270728	221169	176910	136831	105084	78653	55955	37273	22383


Data are representative of Boston Scientific FINELINE II (poly) and Intermedics Thinline II (poly) lead performance.

**FINELINE II EZ/FINELINE II Sterox EZ  
Positive Fixation (Polyurethane)**

Models 4463/4464/4465/4469/4470/  
4471

U.S. Survival Probability	Worldwide Malfunction Details	Product Advisories	Longitude Survival Probability
---------------------------	-------------------------------	--------------------	--------------------------------

**FINELINE II EZ/FINELINE II Sterox EZ  
Positive Fixation (Polyurethane)  
Models 4463/4464/4465/4469/4470/  
4471**



**Worldwide Distribution: 582,000**  
**Worldwide Confirmed Malfunctions: 463**

	Without Compromised Therapy	With Compromised Therapy	Total
<b>Conductor</b>	<b>3</b>	<b>109</b>	<b>112</b>
<sup>7</sup> Lead conductor	2	51	
<sup>28</sup> Non-patterned, Conductor	-	6	
<sup>33</sup> Conductor damage	1	52	
<b>Crimp/Weld/Bond</b>	<b>-</b>	<b>2</b>	<b>2</b>
<sup>23</sup> Terminal weld	-	1	
<sup>32</sup> Non-patterned, Crimp, Weld, Bond	-	1	
<b>Extrinsic</b>	<b>-</b>	<b>325</b>	<b>325</b>
<sup>30</sup> Unconfirmed Extrinsic	-	319	
<sup>31</sup> Inconclusive Extrinsic	-	6	
<b>Insulation</b>	<b>9</b>	<b>6</b>	<b>15</b>
<sup>34</sup> Insulation damage	9	6	
<b>Other</b>	<b>7</b>	<b>2</b>	<b>9</b>
<sup>27</sup> Non-patterned, Other	7	2	
<b>WW Confirmed Malfunctions</b>	<b>19</b>	<b>444</b>	<b>463</b>

[More details](#) about malfunctions

[References](#) cited in table above

**FINELINE II EZ  
Positive Fixation (poly) Longitude\***


Models 4463/4464/4465/4469/4470/  
4471

U.S. Survival Probability	Worldwide Malfunction Details	Product Advisories	Longitude Survival Probability
---------------------------	-------------------------------	--------------------	--------------------------------

**Longitude Registry Summary Data**

Leads Enrolled: 524	Chronic Lead Complications: 0
Leads Active: 430	Malfunctions:1
Cumulative Followup Months : 15,894	Without Compromised Therapy:0
	With Compromised Therapy:1

**Complications and Malfunctions**



**Longitude Registry Survival Probability**

Year	1	2	3	4	5	6	7	8	9	10
Longitude	99.77 (-1.4/+0.2)	99.77 (-1.4/+0.2)	99.77 (-1.4/+0.2)	99.77 (-1.4/+0.2)	99.77 @ 54 (-1.4/+0.2)	-	-	-	-	-
Registered Implants: 524										
Effective Sample Size	374	295	195	96	51	-	-	-	-	-

**FINELINE II EZ/FINELINE II Sterox EZ  
Positive Fixation (Silicone)**


Models 4466/4467/4468/4472/4473/  
4474

<b>U.S. Survival Probability</b>	<b>Worldwide Malfunction Details</b>	<b>Product Advisories</b>
----------------------------------	--------------------------------------	---------------------------

**U.S. Summary**

<b>U.S. Registered Implants:</b> 48,000	<b>U.S. Chronic Lead Complications:</b> 447
<b>U.S. Approval Date:</b> January 2000	<b>U.S. Malfunctions:</b> 168
<b>U.S. Estimated Active Implants:</b> 25,000	<b>Without Compromised Therapy:</b> 13
	<b>With Compromised Therapy:</b> 155

**Complications and Malfunctions**



**U.S. Survival Probability**

Year	1	2	3	4	5	6	7	8	9	10
Non Advisory Population	99.74 (-0.1/+0.0)	99.57 (-0.1/+0.1)	99.38 (-0.1/+0.1)	99.15 (-0.1/+0.1)	98.81 (-0.1/+0.1)	98.36 (-0.2/+0.2)	97.91 (-0.2/+0.2)	97.47 (-0.2/+0.2)	97.25 (-0.3/+0.2)	96.94 (-0.3/+0.3)
Registered Implants: 48000										
Effective Sample Size	40722	34770	29309	24344	19905	16064	12725	9624	6856	4397

Data are representative of Boston Scientific FINELINE II (silicone) and Intermedics Thinline II (silicone) lead performance.




**FINELINE II EZ/FINELINE II Sterox EZ  
Positive Fixation (Silicone)**

Models 4466/4467/4468/4472/4473/  
4474

U.S. Survival Probability	Worldwide Malfunction Details	Product Advisories
---------------------------	-------------------------------	--------------------

**FINELINE II EZ/FINELINE II Sterox EZ  
Positive Fixation (Silicone)**  
Models 4466/4467/4468/4472/4473/  
4474



**Worldwide Distribution:** 128,000  
**Worldwide Confirmed Malfunctions:** 215

	Without Compromised Therapy	With Compromised Therapy	Total
<b>Conductor</b>	<b>1</b>	<b>119</b>	<b>120</b>
<sup>7</sup> Lead conductor	1	73	
<sup>28</sup> Non-patterned, Conductor	-	2	
<sup>33</sup> Conductor damage	-	44	
<b>Crimp/Weld/Bond</b>	<b>-</b>	<b>-</b>	<b>0</b>
<b>Extrinsic</b>	<b>-</b>	<b>70</b>	<b>70</b>
<sup>30</sup> Unconfirmed Extrinsic	-	68	
<sup>31</sup> Inconclusive Extrinsic	-	2	
<b>Insulation</b>	<b>7</b>	<b>8</b>	<b>15</b>
<sup>29</sup> Non-patterned, Insulation	2	-	
<sup>34</sup> Insulation damage	5	8	
<b>Other</b>	<b>5</b>	<b>2</b>	<b>7</b>
<sup>27</sup> Non-patterned, Other	5	2	
<b>WW Confirmed Malfunctions</b>	<b>13</b>	<b>202</b>	<b>215</b>

[More details](#) about malfunctions

[References](#) cited in table above

**FINELINE II/FINELINE II  
Sterox Atrial J (Polyurethane)**


Models 4477/4478/4479/4480

<b>U.S. Survival Probability</b>	<b>Worldwide Malfunction Details</b>	<b>Product Advisories</b>
----------------------------------	--------------------------------------	---------------------------

**U.S. Summary**

U.S. Registered Implants: 55,000	U.S. Chronic Lead Complications: 484
U.S. Approval Date: January 2000	U.S. Malfunctions: 87
U.S. Estimated Active Implants: 30,000	Without Compromised Therapy: 18
	With Compromised Therapy: 69

**Complications and Malfunctions**



**U.S. Survival Probability**


Year	1	2	3	4	5	6	7	8	9	10
Non Advisory Population	99.41 (-0.1/+0.1)	99.24 (-0.1/+0.1)	99.11 (-0.1/+0.1)	98.99 (-0.1/+0.1)	98.88 (-0.1/+0.1)	98.70 (-0.1/+0.1)	98.53 (-0.1/+0.1)	98.38 (-0.2/+0.1)	98.16 (-0.2/+0.2)	98.06 (-0.2/+0.2)
Registered Implants: 55000										
Effective Sample Size	44933	37613	31314	25528	20378	16132	12508	9431	6778	4261

Data are representative of Boston Scientific FINELINE II (poly) and Intermedics Thinline II (poly) lead performance.

**FINELINE II/FINELINE II  
Sterox Atrial J (Polyurethane)**

Models 4477/4478/4479/4480

<b>U.S. Survival Probability</b>	<b>Worldwide Malfunction Details</b>	<b>Product Advisories</b>
----------------------------------	--------------------------------------	---------------------------

<b>FINELINE II/FINELINE II Sterox Atrial J (Polyurethane) Models 4477/4478/4479/4480</b> 			
<b>Worldwide Distribution: 244,000</b>			
<b>Worldwide Confirmed Malfunctions: 145</b>			
	Without Compromised Therapy	With Compromised Therapy	Total
<b>Conductor</b>	<b>2</b>	<b>10</b>	<b>12</b>
<sup>7</sup> Lead conductor	-	3	
<sup>33</sup> Conductor damage	2	7	
<b>Crimp/Weld/Bond</b>	-	-	<b>0</b>
<b>Extrinsic</b>	-	<b>96</b>	<b>96</b>
<sup>30</sup> Unconfirmed Extrinsic	-	95	
<sup>31</sup> Inconclusive Extrinsic	-	1	
<b>Insulation</b>	-	<b>1</b>	<b>1</b>
<sup>34</sup> Insulation damage	-	1	
<b>Other</b>	<b>32</b>	<b>4</b>	<b>36</b>
<sup>22</sup> J-shape	30	4	
<sup>27</sup> Non-patterned, Other	2	-	
<b>WW Confirmed Malfunctions</b>	<b>34</b>	<b>111</b>	<b>145</b>

[More details](#) about malfunctions

[References](#) cited in table above

**FINELINE II/FINELINE II Sterox  
Passive Fixation (Polyurethane)**


Models 4452/4453/4456/4457

<b>U.S. Survival Probability</b>	<b>Worldwide Malfunction Details</b>	<b>Product Advisories</b>
----------------------------------	--------------------------------------	---------------------------

**U.S. Summary**

U.S. Registered Implants: 169,000	U.S. Chronic Lead Complications: 811
U.S. Approval Date: January 2000	U.S. Malfunctions: 112
U.S. Estimated Active Implants: 86,000	Without Compromised Therapy: 5
	With Compromised Therapy: 107

**Complications and Malfunctions**



**U.S. Survival Probability**

Year	1	2	3	4	5	6	7	8	9	10
Non Advisory Population	99.81 (-0.0/+0.0)	99.72 (-0.0/+0.0)	99.63 (-0.0/+0.0)	99.54 (-0.0/+0.0)	99.43 (-0.0/+0.0)	99.31 (-0.1/+0.1)	99.14 (-0.1/+0.1)	99.01 (-0.1/+0.1)	98.83 (-0.1/+0.1)	98.74 (-0.1/+0.1)
Registered Implants: 169000										
Effective Sample Size	138896	116627	96854	79042	63183	50019	38991	29510	21476	14274


Data are representative of Boston Scientific FINELINE II (poly) and Intermedics Thinline II (poly) lead performance.

**FINELINE II/FINELINE II Sterox  
Passive Fixation (Polyurethane)**

Models 4452/4453/4456/4457

U.S. Survival Probability	Worldwide Malfunction Details	Product Advisories
---------------------------	-------------------------------	--------------------

**FINELINE II/FINELINE II Sterox  
Passive Fixation (Polyurethane)  
Models 4452/4453/4456/4457**



**Worldwide Distribution:** 450,000  
**Worldwide Confirmed Malfunctions:** 149

	Without Compromised Therapy	With Compromised Therapy	Total
<b>Conductor</b>	<b>1</b>	<b>40</b>	<b>41</b>
<sup>7</sup> Lead conductor	-	13	
<sup>28</sup> Non-patterned, Conductor	-	3	
<sup>33</sup> Conductor damage	1	24	
<b>Crimp/Weld/Bond</b>	<b>-</b>	<b>-</b>	<b>0</b>
<b>Extrinsic</b>	<b>1</b>	<b>93</b>	<b>94</b>
<sup>30</sup> Unconfirmed Extrinsic	-	91	
<sup>31</sup> Inconclusive Extrinsic	1	2	
<b>Insulation</b>	<b>2</b>	<b>7</b>	<b>9</b>
<sup>34</sup> Insulation damage	2	7	
<b>Other</b>	<b>4</b>	<b>-</b>	<b>4</b>
<sup>27</sup> Non-patterned, Other	4	-	
<b>WW Confirmed Malfunctions</b>	<b>8</b>	<b>141</b>	<b>149</b>

[More details](#) about malfunctions

[References](#) cited in table above

**FINELINE II/FINELINE II Sterox  
Passive Fixation (Silicone)**

Models 4454/4455/4458/4459


<b>U.S. Survival Probability</b>	<b>Worldwide Malfunction Details</b>	<b>Product Advisories</b>
----------------------------------	--------------------------------------	---------------------------

**U.S. Summary**

U.S. Registered Implants: 14,000  
 U.S. Approval Date: January 2000  
 U.S. Estimated Active Implants: 6,000

U.S. Chronic Lead Complications: 159  
 U.S. Malfunctions:31  
 Without Compromised Therapy:0  
 With Compromised Therapy:31

**Complications and Malfunctions**



**U.S. Survival Probability**

Year	1	2	3	4	5	6	7	8	9	10
Non Advisory Population	99.67 (-0.1/+0.1)	99.49 (-0.1/+0.1)	99.17 (-0.2/+0.2)	98.88 (-0.2/+0.2)	98.70 (-0.3/+0.2)	98.38 (-0.3/+0.3)	97.99 (-0.4/+0.3)	97.65 (-0.4/+0.3)	97.44 (-0.4/+0.4)	97.28 (-0.5/+0.4)
Registered Implants: 14000										
Effective Sample Size	11697	10103	8601	7188	6055	5044	4164	3363	2570	1836


Data are representative of Boston Scientific FINELINE II (silicone) and Intermedics Thinline II (silicone) lead performance.

**FINELINE II/FINELINE II Sterox  
Passive Fixation (Silicone)**

Models 4454/4455/4458/4459

U.S. Survival Probability	Worldwide Malfunction Details	Product Advisories
---------------------------	-------------------------------	--------------------

**FINELINE II/FINELINE II Sterox  
Passive Fixation (Silicone)  
Models 4454/4455/4458/4459**



Worldwide Distribution: 97,000

Worldwide Confirmed Malfunctions: 67

	Without Compromised Therapy	With Compromised Therapy	Total
<b>Conductor</b>	-	<b>38</b>	<b>38</b>
<sup>7</sup> Lead conductor	-	15	
<sup>33</sup> Conductor damage	-	23	
<b>Crimp/Weld/Bond</b>	-	-	<b>0</b>
<b>Extrinsic</b>	-	<b>20</b>	<b>20</b>
<sup>30</sup> Unconfirmed Extrinsic	-	20	
<b>Insulation</b>	<b>2</b>	<b>4</b>	<b>6</b>
<sup>34</sup> Insulation damage	2	4	
<b>Other</b>	-	<b>3</b>	<b>3</b>
<sup>27</sup> Non-patterned, Other	-	3	
<b>WW Confirmed Malfunctions</b>	<b>2</b>	<b>65</b>	<b>67</b>

[More details](#) about malfunctions

[References](#) cited in table above

## CRM PRODUCT PERFORMANCE REPORT Q1 2014

### Confirmed Malfunction Details: Leads

#### References

Descriptions listed below provide an overview of the clinical observations and/or analysis findings associated with each confirmed lead malfunction pattern listed in this report.

All of the patterns listed are thoroughly investigated and analyzed. As part of Boston Scientific's process of continuous improvement, when possible, changes have been or will be implemented in response to identified malfunction patterns. "Improvements implemented" may include product design changes in existing or subsequent generations, manufacturing process modifications, educational communications, labeling changes, etc. Improvement implementation may vary by geography due to various factors, including regulatory review timing, and may not completely mitigate or eliminate the potential for additional malfunctions.

1. **IS-1 terminal pin**— Compromised insulation and/or conductor integrity if lead is bent sharply away from the header block when placed in implant pocket or if pulse generator migrates from implant site. Improvement implemented.
2. **Inner insulation abrasion**— Loss of capture, decreasing impedance, increased pacing thresholds, noisy signals, oversensing. Abrasion of inner insulation.
3. **Lead body**— Insulation abrasion due to lead-on-lead or lead-on-can contact combined with damage attributed to application of compressive or torsional loads which may be due to clavicle-first rib entrapment. Damage to lead body may expose conductor.
4. **Terminal leg insulation**— Loss of sensing, loss of pacing, loss of defibrillation therapy. Abraded insulation on terminal leg portion of lead due to lead-on-lead or lead-on-can contact. Improvement implemented.
5. **Seal rings**— Insertion difficulty at implant, difficulty removing lead from header post-implant. Proximal silicone seal rings not fully adhered to lead terminal. Improvement implemented.
6. **Manufacturing material**— Loss of sensing, loss of pacing, noisy signals. Manufacturing material embedded in lead body.
7. **Lead conductor**— Loss of capture, inability to deliver therapy. Fatigue of lead conductor due to repeated flexing.
8. **Lead body**— Lead fracture, inappropriate shocks, oversensing. Insulation damage resulting from implant stresses or manufacturing variability.
9. **Lead conductor**— Loss of sensing, loss of pacing. Physical damage to lead body due to repeated flexing.
10. **Lead conductor**— Physical damage to lead conductor, inappropriate shocks, oversensing. Displacement of yoke component may lead to fatigue of high-voltage lead conductor. Improvement implemented.
11. **Lead connector**— Insulation damage resulting from bending or tension at the terminal connector. May lead to inappropriate shocks, oversensing.
12. **Lead conductor**— High impedance, loss of sensing, loss of pacing. Variability in wire conductor material. Improvement implemented.
13. **Serial number label**— Loss of sensing, loss of pacing. Sharp edge in serial number label resulting in breach in outer lead insulation. Improvement implemented.
14. **Conductor connection**— Loss of sensing, loss of pacing, loss of defibrillation output. Improper conductor wire connection. Improvement implemented.
15. **Electrode tip**— Separation between electrode tip and lead body.
16. **Lead body**— Physical damage to lead body, inappropriate shocks. Abraded insulation due to contact with patient anatomy.
17. **Terminal component**— Loss of sensing, loss of pacing, terminal pin separation from terminal ring during implant or ICD replacement. Improvement implemented.
18. **IS-1 terminal pin**— Compromised insulation and/or conductor integrity if lead is bent sharply away from the header block when placed in implant pocket or if pulse generator migrates from implant site. Improvement implemented.
19. **Serial number label**— Loss of sensing, loss of pacing. Broken serial number label due to either sharp bend away from header at implant or repetitive movement during implant.
20. **DF-1 terminal pin**— Loss of sensing, loss of pacing, loss of defibrillation output. Compromised insulation and/or conductor integrity from sharp or excessive bending. Improvement implemented.
21. **Yoke component**— Noise, impedance anomalies, threshold variation. Use of multiple or pre-formed stylets may cause component within lead yoke to dislodge. Improvement implemented.
22. **J-shape**— Placement difficulty, dislodgement. Elevated temperatures resulting in a relaxation of pre-formed J-shape. Improvement implemented.
23. **Terminal weld**— Impedance rise, loss of pacing. Loss of connection on terminal weld. Improvement implemented.
24. **Lead conductor**— Muscle stimulation, inappropriate shocks, oversensing, high pacing impedance, inability to deliver therapy. Repeated flexing leading to fatigue of lead conductor.
25. **Conductor fracture**— High impedance, loss of capture, loss of pacing, inappropriate shocks. Flex fatigue leading to discontinuity of pace/sense conductor.



26. **Conductor fracture**— High impedance, loss of LV capture, loss of LV pacing. Flex fatigue leading to discontinuity of conductor.
27. **Non-patterned, Other**— Confirmed malfunction for which the root cause does not fit within other categories and is not associated with other malfunctions, or has not yet been identified.
28. **Non-patterned, Conductor**— Conductor malfunction (including clavicle fatigue or crush damage) where the root cause is not associated with other malfunctions.
29. **Non-patterned, Insulation**— Lead insulation breach (including damage due to lead-on-lead or lead-on-anatomy contact, clavicle fatigue or crush) where the root cause is not associated with other malfunctions.
30. **Unconfirmed Extrinsic**— Lead complication after 30 days of implant time with lead return, where analysis could not identify an out of specification condition. Includes complications such as dislodgement, perforation or failure to capture.
31. **Inconclusive Extrinsic**— Lead complication after 30 days of implant time with lead return, where analysis was inconclusive. Includes partial lead returns and leads damaged by the explantation process.
32. **Non-patterned, Crimp, Weld, Bond**— Interruption in conductor or lead body associated with a point of connection where the root cause is not associated with other malfunctions.
33. **Conductor damage**— Noise, oversensing, inappropriate shocks, possible loss of therapy. Conductor damage attributed to application of compressive or torsional loads which may be due to clavicle-first rib entrapment.
34. **Insulation damage**— Low pacing impedance, noise, possible loss of therapy. Insulation abrasion due to lead-on-lead or lead-on-can contact, or due to application of compressive or torsional loads which may be due to clavicle-first rib entrapment. Damage to lead body may expose conductor.
35. **Extracardiac fracture**— High impedance, loss of LV capture, loss of LV pacing. Flex fatigue near suture sleeve, not including clavicle-first rib damage, leading to discontinuity of conductor.
36. **Conductor connection**— Loss of sensing, loss of pacing, loss of defibrillation output. Improper conductor wire connection. Improvement implemented.
37. **Weld fracture**— Noise, loss of sensing. Fractured weld.

## U.S. Chronic Lead Complications (Occurring After the First Month of Service)

Boston Scientific strives to provide meaningful detail in describing the performance of our products. To be included in the Chronic Lead Complications table, a lead must be successfully implanted, with clinical observations occurring after the first month of implant, and have been removed from service surgically or electronically, but not returned for laboratory analysis. The categories are consistent with the AdvaMed guidance for *Uniform Reporting of Clinical Performance of Cardiac Rhythm Management Pulse Generators and Leads*. Although multiple complications are possible for any given lead, only one complication is reported per lead. The complication reported is determined by an observation hierarchy, indicated by the order of the categories from left to right in the table. The number of U.S. Registered Implants is also provided as context for the data. Chronic Lead Complications are included in the calculation of survival probability. Complications for both the U.S. Registered Implant population and for leads enrolled in the Longitude Surveillance Registry (bottom of table) are provided.

Pacing Leads/Model	U.S. Registered Implants	Cardiac Perforation	Conductor fracture/ helix damage	Lead dislodgement	Failure to capture	Oversensing	Failure to sense	Insulation breach	Abnormal pacing impedance	Abnormal defibrillation impedance	Extracardiac stimulation
FLEXTEND Active Fixation 4086/4087/4088	224000	47	551	597	479	149	57	116	265	0	52
FINELINE II/FINELINE II Sterox Passive Fixation (Polyurethane) 4452/4453/4456/4457	169000	0	239	143	133	17	14	128	119	0	18
FINELINE II EZ/FINELINE II Sterox EZ Positive Fixation (Polyurethane) 4463/4464/4465/4469/4470/4471	394000	11	351	442	217	24	53	305	218	0	23
FINELINE II/FINELINE II Sterox Atrial J (Polyurethane) 4477/4478/4479/4480	55000	0	67	242	88	4	8	42	27	0	6
FINELINE II/FINELINE II Sterox Passive Fixation (Silicone) 4454/4455/4458/4459	14000	1	75	15	33	7	2	12	13	0	1
FINELINE II/FINELINE II Sterox EZ Positive Fixation (Silicone) 4466/4467/4468/4472/5573/4474	48000	0	170	55	58	23	8	58	73	0	2
CRT Leads/Model	U.S. Registered Implants	Cardiac Perforation	Conductor fracture/ helix damage	Lead dislodgement	Failure to capture	Oversensing	Failure to sense	Insulation breach	Abnormal pacing impedance	Abnormal defibrillation impedance	Extracardiac stimulation
ACUITY Steerable 4554/4555/4556	26000	1	10	213	14	1	1	3	8	0	73
ACUITY Spiral 4591/4592/4593	19000	0	7	114	12	0	1	0	4	0	92

<b>CRT Leads/Model (cont.)</b>	U.S. Registered Implants	Cardiac Perforation	Conductor fracture/ helix damage	Lead dislodgement	Failure to capture	Oversensing	Failure to sense	Insulation breach	Abnormal pacing impedance	Abnormal defibrillation impedance	Extracardiac stimulation
EASYTRAK 3 4522/4524/4525/4527/4548/4549/4550	20000	1	20	178	28	0	1	5	5	0	72
EASYTRAK 2 4515/4517/4518/4520/4542/4543/4544	89000	0	162	763	140	1	2	32	46	0	361
EASYTRAK 4510/4511/4512/4513/4535/4536/ 4537/4538	38000	1	43	307	75	1	0	30	20	0	234

<b>Defibrillation Leads/Model</b>	U.S. Registered Implants	Cardiac Perforation	Conductor fracture/ helix damage	Lead dislodgement	Failure to capture	Oversensing	Failure to sense	Insulation breach	Abnormal pacing impedance	Abnormal defibrillation impedance	Extracardiac stimulation
ENDOTAK RELIANCE G 4-Site Dual Coil, Active Fixation 0295/0296	27000	1	3	18	3	2	3	0	0	0	0
ENDOTAK RELIANCE G 4-Site Dual Coil, Passive Fixation 0285/0286	1000	0	0	0	0	0	0	0	0	0	0
ENDOTAK RELIANCE SG 4-Site Single Coil, Active Fixation 0292/0293	18000	4	0	8	3	2	1	0	1	1	0
ENDOTAK RELIANCE 4-Site Dual Coil, Active Fixation 0275/0276	2000	1	1	0	1	0	0	0	0	0	0
ENDOTAK RELIANCE G Dual Coil, Active Fixation 0164/0165/0167/0184/0185/0186/0187	186000	10	106	149	43	87	20	33	48	41	15
ENDOTAK RELIANCE G Dual Coil, Passive Fixation 0174/0175/0176/0177	14000	0	13	20	8	4	2	4	21	7	2
ENDOTAK RELIANCE SG Single Coil, Active Fixation 0160/0161/0162/0180/0181/0182	25000	3	15	18	7	13	0	3	8	2	0
ENDOTAK RELIANCE SG Single Coil, Passive Fixation 0170/0171/0172/0173	1000	0	0	0	0	0	0	0	1	0	0
ENDOTAK RELIANCE Dual Coil, Active Fixation 0157/0158/0159	97000	4	86	52	23	86	14	30	55	19	5
ENDOTAK RELIANCE Dual Coil, Passive Fixation 0147/0148/0149	33000	3	44	27	19	29	2	18	60	11	2

**Defibrillation Leads/Model  
(cont.)**

	U.S. Registered Implants	Cardiac Perforation	Conductor fracture/ helix damage	Lead dislodgement	Failure to capture	Oversensing	Failure to sense	Insulation breach	Abnormal pacing impedance	Abnormal defibrillation impedance	Extracardiac stimulation
ENDOTAK RELIANCE S Single Coil, Active Fixation 0137/0138	2000	0	4	0	0	0	0	0	2	0	1
ENDOTAK RELIANCE S Single Coil, Passive Fixation 0127/0128	1000	0	0	2	1	1	0	0	0	0	0

**S-ICD Electrodes/Model**

		Cardiac Perforation	Conductor fracture/ helix damage	Lead dislodgement	Failure to capture	Oversensing	Failure to sense	Insulation breach	Abnormal pacing impedance	Abnormal defibrillation impedance	Extracardiac stimulation
Q-TRAK SQ Electrode 3400		0	0	0	0	0	0	0	0	0	0

**Longitude Surveillance Registry**

	Leads Enrolled	Cardiac Perforation	Conductor fracture/ helix damage	Lead dislodgement	Failure to capture	Oversensing	Failure to sense	Insulation breach	Abnormal pacing impedance	Abnormal defibrillation impedance	Extracardiac stimulation
ACUITY Spiral 4591/4592/4593	1185	0	0	9	0	0	0	0	0	0	5
ENDOTAK RELIANCE G Dual Coil, Active Fixation 0164/0165/0167/0184/0185/0186/0187	619	0	0	0	0	0	0	0	1	0	0
FINELINE II EZ/FINELINE II Sterox EZ Positive Fixation (Polyurethane) 4463/4464/4465/4469/4470/4471	524	0	0	0	0	0	0	0	0	0	0

## U.S. Acute Lead Observations

Boston Scientific strives to provide meaningful detail reflective of real-world product experience. In the first weeks following lead implantation, physiologic responses and lead performance can vary until chronic lead stability is attained. Acute lead performance may be subject to a number of factors, including patient-specific anatomy, clinical conditions and/or varying implant conditions/techniques.

Because acute implant time contributes to overall clinical experience, Boston Scientific provides specific information regarding acute lead performance. To be included in the Acute Lead Observations table, a lead must first be successfully implanted, with clinical observations occurring within the first month of implant. These reports may or may not have resulted in clinical action and/or product return to Boston Scientific. The categories are consistent with the AdvaMed guidance for *Uniform Reporting of Clinical Performance of Cardiac Rhythm Management Pulse Generators and Leads*. Although multiple observations are possible for any given lead, only one observation is reported per lead. The observation reported is determined by an observation hierarchy, indicated by the order of the categories from left to right in the table. The number of U.S. Registered Implants is also provided as context for the data. Acute Lead Observations are not included in calculation of lead survival probability. Observations for both the U.S. Registered Implant population and for leads enrolled in the Longitude Surveillance Registry (bottom of table) are included.

Pacing Leads/Model	U.S. Registered Implants	Cardiac Perforation	Conductor fracture/ helix damage	Lead dislodgement	Failure to capture	Oversensing	Failure to sense	Insulation breach	Abnormal pacing impedance	Abnormal defibrillation impedance	Extracardiac stimulation
FLEXTEND Active Fixation 4086/4087/4088	224000	220	188	1306	410	65	84	54	202	0	50
FINELINE II/FINELINE II Sterox Passive Fixation (Polyurethane) 4452/4453/4456/4457	169000	14	13	403	158	5	26	22	36	0	19
FINELINE II EZ/FINELINE II Sterox EZ Positive Fixation (Polyurethane) 4463/4464/4465/4469/4470/4471	394000	69	77	612	222	91	83	57	219	0	37
FINELINE II/FINELINE II Sterox Atrial J (Polyurethane) 4477/4478/4479/4480	55000	1	18	420	85	5	29	17	18	0	6
FINELINE II/FINELINE II Sterox Passive Fixation (Silicone) 4454/4455/4458/4459	14000	1	4	33	15	1	3	6	5	0	0
FINELINE II/FINELINE II Sterox EZ Positive Fixation (Silicone) 4466/4467/4468/4472/5573/4474	48000	2	15	95	24	9	9	21	12	0	6
CRT Leads/Model	U.S. Registered Implants	Cardiac Perforation	Conductor fracture/ helix damage	Lead dislodgement	Failure to capture	Oversensing	Failure to sense	Insulation breach	Abnormal pacing impedance	Abnormal defibrillation impedance	Extracardiac stimulation
ACUITY Steerable 4554/4555/4556	26000	1	2	303	38	23	2	7	127	0	222
ACUITY Spiral 4591/4592/4593	19000	5	4	179	53	8	2	10	35	0	202

<b>CRT Leads/Model (cont.)</b>	U.S. Registered Implants	Cardiac Perforation	Conductor fracture/ helix damage	Lead dislodgement	Failure to capture	Oversensing	Failure to sense	Insulation breach	Abnormal pacing impedance	Abnormal defibrillation impedance	Extracardiac stimulation
EASYTRAK 3 4522/4524/4525/4527/4548/4549/4550	20000	4	2	257	37	9	2	7	45	0	173
EASYTRAK 2 4515/4517/4518/4520/4542/4543/4544	89000	13	7	852	115	45	9	24	188	0	670
EASYTRAK 4510/4511/4512/4513/4535/4536/ 4537/4538	38000	7	8	196	38	15	1	17	34	0	185

<b>Defibrillation Leads/Model</b>	U.S. Registered Implants	Cardiac Perforation	Conductor fracture/ helix damage	Lead dislodgement	Failure to capture	Oversensing	Failure to sense	Insulation breach	Abnormal pacing impedance	Abnormal defibrillation impedance	Extracardiac stimulation
ENDOTAK RELIANCE G 4-Site Dual Coil, Active Fixation 0295/0296	27000	20	21	75	50	34	5	4	40	7	3
ENDOTAK RELIANCE G 4-Site Dual Coil, Passive Fixation 0285/0286	1000	1	0	2	0	2	0	1	10	1	0
ENDOTAK RELIANCE SG 4-Site Single Coil, Active Fixation 0292/0293	27000	10	28	46	17	22	4	2	31	37	7
ENDOTAK RELIANCE 4-Site Dual Coil, Active Fixation 0275/0276	2000	2	2	1	2	3	0	0	8	0	1
ENDOTAK RELIANCE G Dual Coil, Active Fixation 0164/0165/0167/0184/0185/0186/0187	186000	118	125	472	125	246	33	45	253	189	61
ENDOTAK RELIANCE G Dual Coil, Passive Fixation 0174/0175/0176/0177	14000	4	1	48	28	15	3	0	102	15	1
ENDOTAK RELIANCE SG Single Coil, Active Fixation 0160/0161/0162/0180/0181/0182	25000	25	15	69	25	29	12	3	47	109	7
ENDOTAK RELIANCE SG Single Coil, Passive Fixation 0170/0171/0172/0173	1000	0	0	1	1	0	1	0	9	1	0
ENDOTAK RELIANCE Dual Coil, Active Fixation 0157/0158/0159	97000	30	63	162	43	116	20	24	102	43	18
ENDOTAK RELIANCE Dual Coil, Passive Fixation 0147/0148/0149	33000	20	21	75	50	34	5	4	40	7	3

<b>Defibrillation Leads/Model continued...</b>	U.S. Registered Implants	Cardiac Perforation	Conductor fracture/ helix damage	Lead dislodgement	Failure to capture	Oversensing	Failure to sense	Insulation breach	Abnormal pacing impedance	Abnormal defibrillation impedance	Extracardiac stimulation
ENDOTAK RELIANCE S Single Coil, Active Fixation 0137/0138	2000	0	1	2	2	2	1	0	4	7	0
ENDOTAK RELIANCE S Single Coil, Passive Fixation 0127/0128	1000	0	0	1	0	1	1	0	1	0	0

<b>S-ICD Electrodes/Model</b>	Cardiac Perforation	Conductor fracture/ helix damage	Lead dislodgement	Failure to capture	Oversensing	Failure to sense	Insulation breach	Abnormal pacing impedance	Abnormal defibrillation impedance	Extracardiac stimulation
Q-TRAK SQ Electrode 3400	0	0	1	0	7	0	0	0	0	0

<b>Longitude Surveillance Registry</b>	Leads Enrolled	Cardiac Perforation	Conductor fracture/ helix damage	Lead dislodgement	Failure to capture	Oversensing	Failure to sense	Insulation breach	Abnormal pacing impedance	Abnormal defibrillation impedance	Extracardiac stimulation
ACUITY Spiral 4591/4592/4593	1185	0	0	11	8	1	1	0	3	0	38
ENDOTAK RELIANCE G Dual Coil, Active Fixation 0164/0165/0167/0184/0185/0186/0187	619	0	0	1	0	1	0	1	0	0	0
FINELINE II EZ/FINELINE II Sterox EZ Positive Fixation (Polyurethane) 4463/4464/4465/4469/4470/4471	524	0	0	1	0	0	0	0	0	0	0

## Before/During Implant Procedure - Worldwide Malfunctions: Leads

This section of the report depicts the number of product malfunctions that occurred worldwide either before implant (prior to opening the sterile product packaging) or during implant (once the sterile product packaging has been opened). In all cases, the product in question must be returned to Boston Scientific CRM and confirmed through laboratory analysis to have operated or exhibited a problem outside the specified performance limits established by Boston Scientific. Damage incurred during shipping/transit or due to external factors warned against in labeling is not reported as device malfunction here.

The Conductor category includes any conductor break or damage with complete or intermittent loss of continuity that could interrupt current flow, including clavicle fatigue or crush damage. The Insulation category includes any lead insulation breach, such as damage due to lead-on-lead or lead-on-anatomy contact, or clavicle fatigue or crush. The Crimp/Weld/Bond category includes any interruption in the conductor or lead body associated with a point of connection. The Other category includes malfunctions for which the root cause does not fit within other categories or has not yet been determined. The Labeling and Packaging categories include product identification issues and damage to sterile packaging, respectively. The Implant Accessory category includes lead malfunctions due to catheter, guidewire or sheath issues.

CRT Leads/Model	Worldwide Distribution	Conductor	Insulation	Crimp/Weld/Bond	Other	Labeling	Packaging	Implant Accessory
ACUITY Steerable 4554/4555/4556	55000	0	0	0	4	0	2	0
ACUITY Spiral 4591/4592/4593	35000	0	0	0	2	1	0	0
EASYTRAK 3 4522/4524/4525/4527/4548/4549/4550	39000	0	2	0	17	0	0	2
EASYTRAK 2 4515/4517/4518/4520/4542/4543/4544	161000	0	5	0	7	1	1	1
EASYTRAK 4510/4511/4512/4513/4535/4536/4537/4538	53000	0	0	0	4	0	0	3

Defibrillation Leads/Model	Worldwide Distribution	Conductor	Insulation	Crimp/Weld/Bond	Other	Labeling	Packaging	Implant Accessory
RELIANCE G 4-FRONT Dual Coil Active Fixation 0658/0695/0696	2000	0	0	0	0	0	0	0
RELIANCE SG 4-FRONT Single Coil Active Fixation 0657/0692/0693	4000	0	0	0	1	0	0	0
ENDOTAK RELIANCE G 4-Site Dual Coil Active Fixation 0295/0296	51000	0	0	0	42	0	1	0
ENDOTAK RELIANCE G 4-Site Dual Coil Passive Fixation 0285/0286	6000	0	0	0	3	0	1	0



<b>Defibrillation Leads/Model (cont.)</b>	Worldwide Distribution	Conductor	Insulation	Crimp/Weld/Bond	Other	Labeling	Packaging	Implant Accessory
ENDOTAK RELIANCE SG 4-Site Single Coil Active Fixation 0292/0293	38000	0	0	0	8	0	1	0
ENDOTAK RELIANCE SG 4-Site Single Coil Passive Fixation 0282/0283	2000	0	0	0	0	0	0	0
ENDOTAK RELIANCE 4-Site Dual Coil Active Fixation 0275/0276	3000	0	0	0	6	0	0	0
ENDOTAK RELIANCE 4-Site Dual Coil Passive Fixation 0265/0266	500	0	0	0	0	0	0	0
ENDOTAK RELIANCE G Dual Coil Active Fixation 0164/0165/0166/0167/0184/0185/0186/0187	250000	0	0	27	354	0	3	14
ENDOTAK RELIANCE G Dual Coil Passive Fixation 0174/0175/0176/0177	38000	0	0	3	55	0	2	0
ENDOTAK RELIANCE SG Single Coil Active Fixation 0160/0161/0162/0180/0181/0182	52000	0	0	6	57	0	1	3
ENDOTAK RELIANCE SG Single Coil Passive Fixation 0170/0171/0172/0173	3000	0	0	0	2	0	0	0
ENDOTAK RELIANCE Dual Coil Active Fixation 4515/4517/4518/4520/4542/4543/4544	113000	0	0	16	130	0	0	0
ENDOTAK RELIANCE Dual Coil Passive Fixation 4510/4511/4512/4513/4535/4536/4537/4538	67000	0	1	0	30	0	1	0
ENDOTAK RELIANCE S Single Coil Active Fixation 4515/4517/4518/4520/4542/4543/4544	5000	0	0	1	3	0	0	0
ENDOTAK RELIANCE S Single Coil Passive Fixation 4510/4511/4512/4513/4535/4536/4537/4538	4000	0	0	0	0	0	0	0

Pacing Leads/Model	Worldwide Distribution	Conductor	Insulation	Crimp/Weld/Bond	Other	Labeling	Packaging	Implant Accessory
INGEVITY Passive Fixation 7631/7632/7731/7732	1000	0	0	0	0	0	0	0
INGEVITY Positive Fixation 7640/7641/7642/7740/7741/7742	2000	0	0	0	0	0	0	0
INGEVITY Atrial J Passive Fixation 7635/7636/7735/7736	1000	0	0	0	0	0	0	0
FLEXTEND 2 Active Fixation 4095/4096/4097	148000	0	0	8	102	1	0	0
FLEXTEND Active Fixation 4086/4087/4088	271000	0	0	54	573	0	1	4
FINELINE II/FINELINE II Sterox Passive Fixation (polyurethane) 4452/4453/4456/4457*	450000	1	0	2	6	3	25	0
FINELINE II/FINELINE II Sterox EZ Positive Fixation (polyurethane) 4463/4464/4465/4469/4470/4471*	582000	0	0	7	55	1	44	2
FINELINE II/FINELINE II Sterox Atrial J (poly) 4477/4478/4479/4480*	244000	0	0	7	55	1	44	2
FINELINE II/FINELINE II Sterox Passive Fixation (silicone) 4454/4455/4458/4459*	97000	0	0	2	1	1	1	0
FINELINE II/FINELINE II Sterox EZ Positive Fixation (silicone) 4466/4467/4468/4472/5573/4474*	97000	0	0	2	1	1	1	0

\*Counts consist of Boston Scientific and Intermedics co-branded pacing leads data.

## Product Advisories

A Product Advisory is a voluntary letter issued to inform physicians of an anomalous device behavior identified by Boston Scientific's Quality System. A Product Advisory is issued when there is a material elevation in risk to patient safety with potential for compromised lifesaving therapy, or when Boston Scientific can provide meaningful guidance to improve patient outcomes or device performance. Boston Scientific considers many perspectives in the decision to issue a Product Advisory, including internal expertise and guidance from an independent Patient Safety Advisory Board (PSAB). The Board includes cardiac electrophysiology, engineering, statistics, risk management and bioethics experts.

This report section includes summaries of Product Advisories for which significant, active U.S. device populations exist. In general, this includes advisories for which the estimated active U.S. advisory population is at least 200. Physician and patient letters, as well as Advisory Updates, are available at [www.bostonscientific.com](http://www.bostonscientific.com). With respect to the number of reported events listed in the summaries below, Boston Scientific recognizes that the actual number of clinical malfunctions may be greater than the number actually reported. Additionally, rate projections are provided with the acknowledgment that predictive modeling is inherently uncertain due to its dependence on the device age distribution of reported events and resultant statistical approximations and assumptions. Advisory notifications may vary by geography, based upon local regulatory requirements. Please contact the local Boston Scientific office for more information. Not all products may be approved for use in all geographies, as product approval is geography specific.

PRODUCT	<b>ORIGINAL COMMUNICATION 29-Aug-13 — Low Voltage Capacitor 2013</b>
<p>A serialized search tool to determine if a specific device is affected by this product advisory is available at <a href="http://www.bostonscientific.com">www.bostonscientific.com</a>.</p>	<p>Voluntary Physician Advisory FDA Classification: Class II</p> <p>Boston Scientific has identified a low voltage (LV) capacitor component that, in some devices, may experience diminished performance after two or more years of implant time. This can increase battery use and eventually trigger one or more Safety Architecture alerts, accompanied by patient-audible beeping. Safety Architecture alerts have proven effective in identifying instances of unexpected battery use before therapy becomes unavailable. The most common alert is a yellow screen displayed on the programmer upon initial interrogation which states: "Voltage is too low for projected remaining capacity. Contact Technical Services with Code 1003". In other instances, diminished LV capacitor performance can result in an unanticipated "Explant" ("ERI") battery status alert and a replacement window that may be less than 3 months.</p>
<p><b>COGNIS</b> Models N106/N107/N118/N119/ P106/P107</p>	<p>All devices that experience diminished LV capacitor performance require replacement. If not replaced, increased current drain could deplete the battery and compromise therapy or telemetry. If device beeping or a Safety Architecture alert is observed, call Technical Services for an analysis of "save-to-disk" information, which will clarify how much time is available to replace the device.</p>
<p><b>TELIGEN VR</b> Models E102/F102</p>	<p><i>Rate of Occurrence</i> A total of approximately 264,000 COGNIS and TELIGEN defibrillators have been distributed and implanted since May of 2008. However, a subset of ~38,500 devices (15% of the total) that were manufactured prior to December 2009 has experienced a higher number of LV capacitor malfunctions (approximately 0.67% or 1 in 150). Devices from this subset have not been available for implant since November of 2010. In contrast to the subset population, the rate of diminished LV capacitor performance for other COGNIS/TELIGEN devices (225,500) is approximately 0.0093% or 1 in 10,700.</p>
<p><b>TELIGEN DR</b> Models E110/F110</p>	<p>Please refer to Appendix A of the physician letter for US Survival Probability for the Low Voltage Capacitor 2013 subset and devices not in the subset.</p>
<p>Physician and patient letters are available at <a href="http://www.bostonscientific.com">www.bostonscientific.com</a>.</p>	<p><b>CURRENT STATUS 17-Jan-14</b></p>
	<p>No devices in the advisory population remain available for implant.</p>
	<p><i>Confirmed Malfunctions (worldwide)</i></p>
	<p>For confirmed malfunctions, refer to the Worldwide Malfunction Details tab of the Product Performance Report for COGNIS and TELIGEN devices and see pattern titled "Low Voltage Capacitor 2013."</p>
	<p>There have been no reported patient deaths associated with this advisory.</p>
	<p><i>Projected Rate of Occurrence</i></p>
	<p>For current performance of a specific product family, refer to the U.S. Survival Probability tab of the Product Performance Report and see population titled "29-Aug-13 Low Voltage Capacitor 2013."</p>

**29-Aug-13 — Low Voltage Capacitor 2013, continued...**

**CURRENT RECOMMENDATION 17-Jan-14**

---

There are no additional clinical recommendations beyond the current standard of patient care and normal device monitoring, which are described in device labeling.

- As always, instruct patients to contact your clinic if beeping is heard from their device. Note that “Beep When Explant is Indicated” is nominally programmed “On” when shipped from the factory.
- Physicians should promptly investigate alerts and unanticipated replacement indicator messages.
- Following a Safety Architecture alert, contact Boston Scientific Technical Services as directed on programmer alert screens. Technical Services can facilitate an evaluation of “save-to-disk” information (while still implanted) to help clarify available replacement time. Note that “Approximate time to Explant” and “Time Remaining” estimates displayed on the programmer are not accurate following a Safety Architecture low voltage alert.
- Boston Scientific’s LATITUDE® Patient Management System (remote monitoring) can improve detection time for battery performance messages by conveying Safety Architecture alerts and/or notifying when scheduled check-ups have not occurred. Note that weekly voltage alerts are nominally configured “On” in LATITUDE.

---

Standard Warranty program available, please contact your local representative for terms and conditions.

PRODUCT	<b>ORIGINAL COMMUNICATION 01-Mar-13 — Unintended Fuse Activation 2013</b>
A serialized search tool to determine if a specific device is affected by this product advisory is available at <a href="http://www.bostonscientific.com">www.bostonscientific.com</a> .	Voluntary Physician Advisory FDA Classification: Class II
<b>SQ-RX S-ICD</b> Model1010	Boston Scientific has identified a rare condition in which an internal protective fuse can be unintentionally activated while the device is charging its capacitors for shock delivery or induction. Should this occur, the defibrillator would not be able to deliver therapy or communicate with the Q-TECH Model 2020 programmer, and would be unable to emit tones or otherwise respond to magnet application. No patient deaths have been reported as a result of this behavior; affected devices were replaced without the need for emergency medical care. A non-invasive, software-based mitigation has been developed to protect the fuse from unintended activation.
Physician and patient letters are available at <a href="http://www.bostonscientific.com">www.bostonscientific.com</a> .	<p><i>Rate of Occurrence</i> The fuse has been unintentionally activated once during an implant procedure and three times post-implant out of approximately 1,900 devices implanted worldwide. All three post-implant events occurred within one month of implant. Engineering analysis also indicates this condition is more likely to occur early in device life.</p>
	<b>CURRENT STATUS 17-Jan-14</b>
	No devices in the advisory population remain available for implant.
	<i>Confirmed Malfunctions (worldwide)</i>
	Six (6) malfunctions have been confirmed worldwide of devices experiencing Unintended Fuse Activation.
	There have been no reported patient deaths associated with this advisory.
	<p><i>Projected Rate of Occurrence</i> The projected rate of occurrence for SQ-RX advisory devices is 0.002% at 60 months.</p>
	<b>CURRENT RECOMMENDATION 17-Jan-14</b>
	– Confirm that your Q-TECH Model 2020 programmers have been upgraded with software version 1.95.0 or later.
	– To access the software version directly from the programmer, turn the programmer ON, select the "Programmer Settings" button, and then select the "About Programmer" button. Programmer software version can also be viewed on the printed report from a device follow-up.
	– Schedule a follow-up visit for each of your S-ICD System patients to update their device with new software: For patients whose device has been implanted for three months or more, ensure the next scheduled visit occurs within three months of the previous visit, as recommended in device labeling.
	– At the next follow-up visit, interrogate each patient's device using a programmer with version 1.95.0 or later software. Interrogation with an updated programmer will automatically add new software to the implanted device to protect the fuse from unintended activation.
	– Resume normal patient follow-up monitoring and programming as directed in device labeling. Devices interrogated using a programmer with version 1.95.0 or later software are no longer subject to this advisory.
	Standard Warranty program available, please contact your local representative for terms and conditions.

PRODUCT	ORIGINAL COMMUNICATION 01-Jun-11 — High Cathode Condition
<p>A serialized search tool to determine if a specific device is affected by this product advisory is available at <a href="http://www.bostonscientific.com">www.bostonscientific.com</a>.</p>	<p>Voluntary Physician Advisory FDA Classification: Pending</p>
<p><b>SQ-RX S-ICD</b> Model1010</p>	<p>Boston Scientific has determined that a specific subset of SQ-RX pulse generators may not achieve the five (5) year typical longevity due to premature battery depletion. Although the longevity of these devices may be shortened, therapy remains available through the onset of the Elective Replacement Indicator (ERI). This subset of devices also has the potential for the time between the onset of the ERI and End of Life (EOL) indicators to be less than the nominal three (3) months. The risk for premature battery depletion is due to a specific condition within an individual battery cell. No affected devices remain available for implant.</p>
<p>Physician letter is available at <a href="http://www.bostonscientific.com">www.bostonscientific.com</a>.</p>	<p><i>Rate of Occurrence</i> Cameron Health has confirmed one (1) occurrence of a device experiencing premature battery depletion due to this condition.</p> <p>Cameron Health conducted a systematic analysis of all implanted devices worldwide and identified two populations at risk for premature battery depletion due to this condition:</p> <ul style="list-style-type: none"> <li>– Population I consists of 18 devices that were confirmed through manufacturing records to contain the condition within a battery cell. Bench testing and engineering analysis estimate that 33.1% [+20.8/-19.5] of devices in this population may experience premature battery depletion due to this condition over the five (5) year typical device longevity. There has been one (1) confirmed occurrence in this population to date.</li> <li>– Population II consists of 386 devices for which manufacturing records cannot exclude the possibility of the condition existing within a battery cell. Bench testing and engineering analysis estimate that 3.3% [+2.2/-1.8] of devices in this population may experience premature battery depletion due to this condition over the five (5) year typical device longevity. There have been zero (0) confirmed occurrences in this population to date.</li> </ul>
	<p><b>CURRENT STATUS 17-Jan-14</b></p>
	<p>No devices in the advisory population remain available for implant.</p> <p><i>Confirmed Malfunctions (worldwide)</i></p> <p>Three (3) malfunctions have been confirmed worldwide of devices experiencing High Cathode Condition.</p> <p>There have been no reported patient deaths associated with this advisory.</p> <p><i>Projected Rate of Occurrence</i></p> <ul style="list-style-type: none"> <li>– Population I – Bench testing and engineering analysis estimate that 33.1% [+20.8/-19.5] of devices in this population may experience premature battery depletion due to this condition over the five (5) year typical device longevity.</li> <li>– Population II - Bench testing and engineering analysis estimate that 3.3% [+2.2/-1.8] of devices in this population may experience premature battery depletion due to this condition over the five (5) year typical device longevity.</li> </ul>
	<p><b>CURRENT RECOMMENDATION 17-Jan-14</b></p>
	<ul style="list-style-type: none"> <li>– If a device experiences an unexpected decline in battery voltage leading to ERI, schedule replacement as soon as possible.</li> <li>– Communicate to patients that, consistent with device labeling, an emitted audible tone requires immediate follow-up by their physician. Once triggered, the audible tone will sound for 16 seconds every 9 hours until the trigger condition has been resolved. As a reminder, a Cameron Health Model 4520 magnet placed over the device may be used to demonstrate the audible tone.</li> </ul> <p>For patients implanted with devices in Population I, Cameron Health recommends this communication occurs as soon as practical. For patients implanted with devices in Population II, Cameron Health recommends this communication occurs at the next follow-up.</p>
	<p>Standard Warranty program available, please contact your local representative for terms and conditions.</p>

PRODUCT	ORIGINAL COMMUNICATION July 2010— Magnetic Reed Switch 2010
<p>A serialized search tool to determine if a specific device is affected by this product advisory is available at <a href="http://www.bostonscientific.com">www.bostonscientific.com</a>.</p>	<p>Voluntary Physician Advisory FDA Classification: Class II</p>
<p><b>CONTAK RENEWAL 3</b> Models H170/H175</p>	<p>Some Boston Scientific defibrillators include a component referred to as a “magnetic reed switch,” designed to sense the presence of a magnet. If Enable Magnet Use is programmed to On (nominally On) and a magnet is applied in emergent situations or during a medical/surgical procedure, the switch is designed to close and temporarily prevent delivery of undesired tachy therapy. When the magnet is removed, the magnetic switch is designed to open and thereby restore ability to deliver programmed tachy therapy.</p>
<p><b>CONTAK RENEWAL 3 HE</b> Models H177/H179</p>	<p>Magnetic reed switch technology has historically demonstrated a very low but non-zero rate of failing to open upon magnet removal. However, certain Boston Scientific CRT-Ds and ICDs manufactured between January 2006 and November 2007 have exhibited a somewhat higher rate of magnetic reed switch failure. Approximately 34,000 of these devices remain actively implanted; no devices in this population are available for implant. Devices manufactured after November of 2007 have returned to historic performance rates and are not included in this advisory</p>
<p><b>CONTAK RENEWAL 3 RF</b> Models H210/H215</p>	<p>No patient deaths or injuries have been reported as a result of this issue, although some devices have been replaced. Most devices with a magnetic reed switch confirmed to be stuck in a closed position have remained implanted after “Enable Magnet Use” was programmed to Off (see Recommendations).</p>
<p><b>CONTAK RENEWAL 3 RF HE</b> Models H217/H219</p>	<p>Rate of Occurrence A rate of one failure per 670 devices (0.0015) has been observed to date in the advisory population (average implant time of 38 months). However, with rapid identification and reprogramming, the probability of patient harm (therapy not available when needed due to a stuck magnetic reed switch) is estimated to be less than one in one million for a 60-month device service life.</p>
<p><b>CONTAK RENEWAL 4</b> Models H190/H195/H197/H199</p>	<p>Rate of Occurrence A rate of one failure per 670 devices (0.0015) has been observed to date in the advisory population (average implant time of 38 months). However, with rapid identification and reprogramming, the probability of patient harm (therapy not available when needed due to a stuck magnetic reed switch) is estimated to be less than one in one million for a 60-month device service life.</p>
<p><b>CONTAK RENEWAL 4 AVT/AVT HE</b> Models M170/M175/M177/M179</p>	<p><b>CURRENT STATUS 17-Jan-14</b></p>
<p><b>CONTAK RENEWAL 4 RF</b> Models H230/H235/H239</p>	<p>There have been no reported patient deaths associated with this advisory. <i>Projected Rate of Occurrence</i> The projected rate of occurrence for the advisory device population is 0.0029 at 60 months.</p>
<p><b>VITALITY DR HE</b> Model T180</p>	<p><b>CURRENT RECOMMENDATION 17-Jan-14</b></p>
<p>Physician and patient letters are available at <a href="http://www.bostonscientific.com">www.bostonscientific.com</a>.</p>	<p>Consistent with physician instructions for use and patient manual labeling, physicians should continue routine follow-up sessions and patients should be reminded to contact their clinic or go to the hospital emergency room immediately if they hear tones/beeps from their device. In addition, Boston Scientific recommends:</p> <ol style="list-style-type: none"> <li>1) In a hospital/clinic/surgery setting, if tones are heard upon magnet application but do not cease upon magnet removal, the device should be interrogated with a programmer and checked per normal standard of care.</li> <li>2) In the United States, use of the LATITUDE remote patient monitoring system may help identify loss of daily measurements and thereby facilitate timely detection of a stuck reed switch. [NOTE: A pop-up message and/or LATITUDE alert do not appear for missing Daily Measurements.]</li> </ol>

**July 2010— Magnetic Reed Switch 2010, continued...**

**CURRENT RECOMMENDATION, continued...**

---

3) If a stuck magnetic switch is confirmed, program the Enable Magnet Use feature to Off, which ensures that programmed therapy will be provided to treat tachyarrhythmias. However, if Enable Magnet Use is programmed Off:

- A magnet will no longer inhibit tachy therapy.
- The Patient Triggered Monitor feature will no longer be available.

Contact Boston Scientific Technical Services (1-800-CARDIAC) for assistance to re-activate Daily Measurements for devices with a stuck magnetic switch.

4) After consultation with our Independent Patient Safety Advisory Board, **Boston Scientific does not recommend prophylactic explant**. We further advise that physicians **do not routinely program Enable Magnet Use to Off in the absence of a confirmed stuck magnetic reed switch** because the benefits of magnet use to disable tachy therapy in emergent situations outweigh the probability of patient harm associated with a stuck reed switch.

---

Standard Warranty program available, please contact your local representative for terms and conditions.



PRODUCT	ORIGINAL COMMUNICATION 01-Dec-09 — Subpectoral Implant 2009
<p>A serialized search tool to determine if a specific device is affected by this product advisory is available at <a href="http://www.bostonscientific.com">www.bostonscientific.com</a>.</p>	<p>Voluntary Physician Advisory            FDA Classification: Class II</p> <p>This advisory is limited to devices identified in the product model list that were implanted subpectorally. Devices implanted subcutaneously are not included in this advisory.</p>
<p><i>This advisory is limited to those models listed below implanted subpectorally.</i></p>	<p>Boston Scientific has determined that the bond between the header and case could be weakened by significant forces associated with a subpectoral implant procedure or when a device in a subpectoral position is pushed against a rib during contraction of the pectoralis muscle. A weakened header bond may alter lead impedance and introduce noise that may inhibit pacing therapy or initiate inappropriate tachy therapy. Additional mechanical stress applied to a weakened bond may eventually cause header connection wires to fracture, resulting in loss of therapy.</p>
<p><b>COGNIS</b>            Models            N106/N107/N108/N118/N119            P106/P107/P108</p>	<p>A weakened header bond can result in one or more of the following device behaviors:</p> <ul style="list-style-type: none"> <li>– Significant changes in measured lead impedance</li> <li>– Noise on real-time or stored electrograms</li> <li>– Intermittent inhibition of pacing</li> <li>– Inappropriate anti-tachy pacing or shock therapy</li> </ul>
<p><b>TELIGEN VR</b>            Models E102/F102</p>	<ul style="list-style-type: none"> <li>– Loss of pacing therapy</li> <li>– Loss of anti-tachy pacing and shock therapy</li> </ul>
<p><b>TELIGEN DR</b>            Models E110/E111/F110/F111</p>	<p>No patient deaths related to this behavior have been reported. Patients have required early device replacement due to inappropriate shocks and/or noise induced by pocket manipulation or arm movement.</p>
<p><i>Physician and patient letters are available at <a href="http://www.bostonscientific.com">www.bostonscientific.com</a>.</i></p>	<p><i>Rate of Occurrence</i></p> <p>The implant orientation of devices is not reported to Boston Scientific, making it difficult to provide rate of occurrence and prediction information. Two (2) reports have been received worldwide of subpectoral implants with weakened header bonds. An estimated 5% of approximately 77,000 COGNIS and TELIGEN devices worldwide have been implanted in a subpectoral location.</p> <p>The following factors may also impact the risk of failure if implanted in a subpectoral location:</p> <ul style="list-style-type: none"> <li>– Exact location of the patient’s ribs relative to the device</li> <li>– Body size and/or muscle mass of the patient (risk may increase for larger/muscular patients)</li> <li>– Activity level and/or occupation of the patient (risk may increase for more active patients)</li> </ul>
	<p><b>CURRENT STATUS 17-Jan-14</b></p>
	<p><b>COGNIS and TELIGEN devices are now available with improved header bond strength in the U.S. and the EU.</b> The stronger bond allows physicians to position the devices in a subpectoral position, if desired.</p>
	<p><i>Reported events (worldwide)</i>            Seventy-eight (78) reports have been received worldwide of subpectoral implants with weakened header bonds. An estimated 10% of approximately 105,000 COGNIS and TELIGEN devices worldwide have been implanted in a subpectoral location.</p>
	<p>There have been no reported patient deaths associated with this advisory.</p>
	<p><i>Rate of Occurrence</i></p>
	<p>An estimated 10% of COGNIS and TELIGEN devices worldwide have been implanted in a subpectoral position. The rate of occurrence for subpectoral implants of COGNIS advisory devices is 1.95% at 60 months. The rate of occurrence for subpectoral implants of TELIGEN advisory devices is 0.6% at 60 months.</p>

01-Dec-09 — Subpectoral Implant 2009, continued...

**CURRENT RECOMMENDATION 17-Jan-14**

---

**If a patient's device was implanted subcutaneously, it is excluded from this advisory** and no change to current patient management is recommended.

**For affected devices implanted in a subpectoral location:**

- Follow patient at least once every three months as recommended in device instructions for use.
- Consider advising patients to contact their physician or clinic if they receive shocks, in order to ensure timely review of associated electrograms and other device data via in-clinic or remote interrogation.
- Where available, consider using the LATITUDE® Patient Management System to facilitate remote device checks between in-clinic follow-ups.

**COGNIS and TELIGEN devices are now available with improved header bond strength in the U.S. and the EU.** The stronger bond allows physicians to position the devices in a subpectoral position, if desired.

---

Standard Warranty program available, please contact your local representative for terms and conditions.

PRODUCT	ORIGINAL COMMUNICATION 05-Apr-07 and 04-Mar-09— Shortened Replacement Window
<p>A serialized search tool to determine if a specific device is affected by this product advisory is available at <a href="http://www.bostonscientific.com">www.bostonscientific.com</a>.</p>	<p>Voluntary Physician Advisory            FDA Classification: Class II</p> <p>Low-voltage capacitors may be subject to degradation. These capacitors may cause accelerated battery depletion and may reduce the time between Elective Replacement Indicator (ERI) and End Of Life (EOL) to less than three months. Device replacement indicators continue to function normally.</p>
<p><b>CONTAK RENEWAL 4 RF HE</b>            Model H239</p>	<p>In April 2007, Boston Scientific CRM communicated with physicians regarding a population of devices subject to this failure mechanism. As of March 2009, the April 2007 advisory population has not experienced any clinically significant changes to either the rate of occurrence or patient management recommendations.</p>
<p><b>CONTAK RENEWAL 4 RF</b>            Models H230/H235</p>	<p>In March 2009, a second population was identified of 856 active ICDs and CRT-Ds manufactured with capacitors from the same supplier that may be subject to the same failure mechanism. The cumulative failure rate for accelerated depletion within this population is approximately 6% at 42 months and is projected to increase. Recommendations described in April 2007 have been 99.9% successful in identifying susceptible devices and ensuring replacement at ERI in the original population, and will minimize patient risk associated with a shortened replacement window when applied to this second population. No devices from this population have been registered as implanted after April 2007. No devices in this subset remain available for implant.</p>
<p><b>CONTAK RENEWAL 4 HE</b>            Models H197/H199</p>	
<p><b>CONTAK RENEWAL 4</b>            Models H190/H195</p>	
<p><b>CONTAK RENEWAL 4            AVT / AVT HE</b>            Models M170/M175/M177/M179</p>	<p><b>CURRENT STATUS 17-Jan-14</b></p>
<p><b>CONTAK RENEWAL 3 RF HE</b>            Models H217/H219</p>	<p><i>Confirmed Malfunctions (worldwide)</i>  <b>April 2007 Population</b>            2,563 malfunctions have been confirmed out of an advisory population of approximately 75,000 devices.</p>
<p><b>CONTAK RENEWAL 3 RF</b>            Models H210/H215</p>	<p>115 of these devices exhibited a shortened ERI to EOL replacement window (less than 90 days).</p>
<p><b>CONTAK RENEWAL 3 HE</b>            Models H177/H179</p>	<p><b>March 2009 Population</b>            116 malfunctions have been confirmed out of an advisory population of 856 active devices. Two of those devices exhibited a shortened ERI to EOL replacement window (less than 90 days).</p>
<p><b>CONTAK RENEWAL 3</b>            Models H170/H175</p>	<p>There have been no reported patient deaths associated with either advisory population.</p>
<p><b>CONTAK RENEWAL 3            AVT / AVT HE</b>            Models M155/M159</p>	<p>No devices currently being distributed are susceptible to this malfunction mode.</p>
<p><b>VITALITY 2 EL VR/DR</b>            Models T177/T167</p>	<p><i>Rate of Occurrence</i>  <b>April 2007 Population</b>            The cumulative rate of occurrence for accelerated battery depletion for the April 2007 advisory population is approximately 5.0% at 60 months.</p>
<p><b>VITALITY 2 VR/DR</b>            Models T175/T165</p>	<p><b>March 2009 Population</b>            The cumulative rate of occurrence for accelerated battery depletion for the March 2009 advisory population is approximately 15.8% at 60 months.</p>
<p><b>VITALITY DR HE</b>            Model T180</p>	<p>Following monitoring recommendations below will minimize patient risk associated with a shortened replacement window.</p>
<p><b>VITALITY DS VR/DR</b>            Model T135/T125</p>	

**05-Apr-07 and 04-Mar-09— Shortened Replacement Window, continued...**

**CURRENT RECOMMENDATION 17-Jan-14**

**VITALITY EL**

Model T127

**VITALITY AVT A155**

Model A155

*Physician and patient letters are available at [www.bostonscientific.com](http://www.bostonscientific.com).*

Patient management recommendations from the April 5, 2007 physician communication remain unchanged.

If a patient has a device with a degraded capacitor, the time from implant to 2.65 volts (Middle of Life 2 / MOL2) will be reduced.

To determine whether a patient may be at risk for a reduced ERI to EOL time, note when 2.65 volts (MOL2) was observed. For each patient:

1. Review patient records to assess battery voltage.
2. If battery voltage is **above** 2.65 volts (MOL2), continue to follow patient every three months per device labeling.
3. If battery voltage is **at or below** 2.65 volts (MOL2), determine the time between device implant and this observation.
4. If the time from implant to 2.65 volts (MOL2) is greater than 27 months (32 months for VITALITY® EL / 2 EL / HE devices), the patient is not at risk for a shortened ERI to EOL time, and **this advisory no longer applies.**
5. If the time from implant to 2.65 volts (MOL2) is 27 months or less (32 months for VITALITY® EL / 2 EL / HE devices), **the patient should be followed monthly until ERI.** For devices that require monthly follow-up, replace the device within 30 days after ERI is displayed as ERI to EOL time may be shortened.

NOTE: If it is not clear when a battery voltage of less than 2.65 volts (MOL2) was reached, conduct a memory "Save to Disk" and return (mail or e-mail) to Boston Scientific CRM for prompt analysis. Contact your local Boston Scientific representative or Technical Services for further assistance

In geographies where available, the LATITUDE® Patient Management System can facilitate remote patient monitoring and provide automatic notification when the device reaches a battery status of ERI.

Standard Warranty program available, please contact your local representative for terms and conditions.

PRODUCT	ORIGINAL COMMUNICATION 10-Mar-07 — Product Update — Mid-Life Display of Replacement Indicators
A serialized search tool to determine if a specific device is affected by this product advisory is available at <a href="http://www.bostonscientific.com">www.bostonscientific.com</a> .	FDA Classification: Devices in Table 1, Column 1 of this <i>Product Update</i> were classified as Class II (27-November-07)
<b>CONTAK RENEWAL 4 RF HE</b> Model H239	Certain devices may display ERI or EOL during mid-life (typically 24–48 months), even though battery voltage (typically greater than or equal to 2.65 volts) and capacity remain available. This behavior is caused by high battery impedance rather than low battery voltage.
<b>CONTAK RENEWAL 4 RF / HE</b> Models H230/H235/H197/H199	Devices that have triggered charge time-based ERI or EOL during mid-life have several months, and in most cases more than one year of remaining battery voltage and capacity, which allows the devices to continue to provide brady and LV pacing and maximum energy shocks. However, if ERI or EOL is triggered, device replacement should be scheduled.
<b>CONTAK RENEWAL 4 and 4 AVT / AVT HE</b> Models H190/H195/M170/M175/M177/M179	<i>Rate Projection</i> Certain devices, typically implanted prior to July 2005 (Table 1, Column 1 of the Product Update) are projected to exhibit Mid-Life Display of Replacement Indicators as indicated below:
<b>CONTAK RENEWAL 3 RF HE</b> Models H217/H219	– VITALITY AVT (Model A135), VITALITY VR/DR, VITALITY DR+ ( <b>Projected rate: 8–10%</b> ) – VITALITY AVT (Model A155), VITALITY DS VR/DR, VITALITY 2 VR/DR, ASSURE ( <b>Projected rate: 4–7%</b> )
<b>CONTAK RENEWAL 3 RF / HE</b> Models H210/H215/H177/H179	– VITALITY EL; VITALITY 2 EL DR/VR; VITALITY DR HE; CONTAK RENEWAL 3/4/3HE/4HE; CONTAK RENEWAL 3 RF/4RF/3RF HE/4RF HE; CONTAK RENEWAL 3 AVT/4AVT/3AVT HE/4AVT HE ( <b>Projected rate: 1–2%</b> )
<b>CONTAK RENEWAL 3 and 3 AVT / AVT HE</b> Models H170/H175/M155/M159	Continuous manufacturing improvements intended to reduce variability in battery performance have been implemented by our battery supplier, which mitigate the occurrence of mid-life display of replacement indicators.
<b>VITALITY 2 EL VR/DR</b> Models T177/T167	<b>CURRENT STATUS 17-Jan-14</b>
<b>VITALITY 2 VR/DR</b> Models T175/T165	<i>Confirmed Malfunctions (worldwide)</i> For confirmed malfunction counts related to a specific product family, refer to the Confirmed Malfunction Details section of the Product Performance Report and see pattern titled “Mid-life display of replacement indicators.”
<b>VITALITY DR HE and EL</b> Model T180 and Model T127	<i>Projected Rate of Occurrence</i> For projected rates of occurrence see device-specific ranges listed above. Some performance differences have been observed between product families. For example, dual chamber devices have generally performed better than single chamber devices within the same product family. For current performance of a specific product family, refer to the U.S. Survival Probability section of the Product Performance Report and see population titled “10–Mar-07 Product Update — Mid-Life Display of Replacement Indicators.”
<b>VITALITY DS VR/DR</b> Model T135/T125	
<b>VITALITY AVT A135 / A155</b> Models A135/A155	
<b>VITALITY VR/DR and DR+</b> Models 1871/1870/1872	<b>CURRENT RECOMMENDATION 17-Jan-14</b>
<b>ASSURE</b> Model B301	<u>Patient management recommendations from the March 10, 2007 Product Update remain unchanged.</u>
<i>The Product Update and patient letter are available at <a href="http://www.bostonscientific.com">www.bostonscientific.com</a>.</i>	<i>Patient Management Considerations</i> – Normal follow-up. If ERI or EOL is triggered, device replacement should be scheduled. – Physicians can consider individual patient needs relative to the potential device behaviors associated with mid-life display of ERI or EOL. – Activating the programmable feature “Beep When ERI is Reached” (nominally ON) will provide audible tones when the device reaches ERI. – Last measured charge time and date are stored in device memory and are available during device interrogation. Commanding a manual capacitor reform may be helpful in characterizing the current charge time.
	Standard Warranty program available, please contact your local representative for terms and conditions.

PRODUCT	<b>ORIGINAL COMMUNICATION 23-Jun-06 and 24-Aug-06— Low Voltage Capacitor</b>
<p>A serialized search tool to determine if a specific device is affected by this product advisory is available at <a href="http://www.bostonscientific.com">www.bostonscientific.com</a>.</p>	<p>Voluntary Physician Advisory FDA Classification: Class II</p>
<p><b>INSIGNIA Ultra SR</b> Models 1190/1390</p>	<p>Devices within a well-defined subset manufactured using low-voltage capacitors from a single component supplier may perform in a manner that leads to device malfunction, including intermittent or permanent loss of output or telemetry, or premature battery depletion. At the time of the original June 23, 2006 communication, approximately 49,800 devices had been distributed, and approximately 27,200 devices had been implanted worldwide. Boston Scientific initiated retrieval of all non-implanted devices within this subset from hospital and sales force inventory. An Advisory Update was issued on August 24, 2006, with a revised estimation of the implanted population to be approximately 31,000. All product currently being shipped and available for implant is not susceptible to this issue.</p>
<p><b>INSIGNIA Ultra DR and Ultra DR Downsize</b> Models 1291/1491/1290/1490</p>	<p><i>Reported Events (worldwide)</i></p>
<p><b>INSIGNIA Entra SR</b> Models 1195/1198/1395/1398</p>	<p>At the time of the original June 23, 2006 communication, a total of five (5) devices had been confirmed to have malfunctioned. As reported in the August 24, 2006 Advisory Update, five (5) additional malfunctions were confirmed since the original June 23, 2006 communication. A total of 10 confirmed malfunctions represented 0.032% of the implanted population of approximately 31,000 devices. Seven (7) of 10 malfunctions were identified while implanted, and three were identified prior to the implant procedure. There were no reports of patient death associated with this issue. There were a total of three (3) reports of patients experiencing syncope associated with loss of pacing.</p>
<p><b>INSIGNIA Entra DR (downsize)</b> Models 1296/1466</p>	<p><i>Projected Rate of Occurrence</i></p>
<p><b>INSIGNIA Entra DR</b> Models 1294/1295/1494/1495</p>	<p>While a statistically significant projection of expected failures for implanted devices was not possible, testing suggested that the frequency of new malfunctions would continue to decrease in the future.</p>
<p><b>INSIGNIA Entra SSI</b> Models 0484/0485/1325/1326</p>	<p><b>CURRENT STATUS 17-Jan-14</b></p>
<p><b>INSIGNIA Entra DDD</b> Models 0985/0986/1426</p>	<p><i>Confirmed Malfunctions (worldwide)</i> 46 malfunctions have been confirmed from the advisory population. 35 of these were identified while implanted. There were an estimated 32,000 advisory devices implanted. 11 malfunctions were identified prior to implantation.</p>
<p><b>INSIGNIA Plus SR</b> Models 1194/1394</p>	<p>There have been no reported patient deaths associated with this advisory. No devices currently being distributed are susceptible to this malfunction mode.</p>
<p><b>INSIGNIA Plus DR and Plus DR Downsize</b> Models 1297/1467/1298/1468</p>	<p><i>Projected Rate of Occurrence</i></p>
<p><b>INSIGNIA AVT</b> Models 0482/0882/0982 1192/1292/1392/1428/1432/1492</p>	<p>The rate of occurrence is projected to range between 0.10% and 0.22%.</p>
<p><b>CONTAK RENEWAL TR / TR2</b> Models H120/H125/H140/H145</p>	<p><b>CURRENT RECOMMENDATION 17-Jan-14</b> <u>Patient management recommendations from the August 24, 2006 Advisory Update remain unchanged.</u></p>
<p><b>VITALITY 2 EL VR/DR</b> Models T177/T167</p>	<p>– Normal follow-up. – Physicians should consider the low and declining failure rate in addition to the unique needs of individual patients when making medical decisions regarding patient management.</p>
<p><b>VITALITY 2 VR/DR</b> Models T175/T165</p>	<p>As always, advise patients to seek attention immediately if they experience syncope or lightheadedness.</p>
<p><b>VITALITY DR HE</b> Model T180</p>	<p>– Should the device exhibit symptoms described below, please contact your local sales representative or Technical Services for assistance with device evaluation.</p>
<p><b>VITALITY DS VR/DR</b> Models T135/T125</p>	<p><b>Device Behavior</b> <b>Pacemakers: INSIGNIA/NEXUS</b> – Intermittent or permanent loss of pacing output – Inability to interrogate – Erased values in Daily Measurements – ERT or EOL indicator message displayed earlier than expected – A gas gauge less than BOL within six months of implant</p>

**23-Jun-06 and 24-Aug-06— Low Voltage Capacitor, continued...**

**VITALITY VR/DR and EL**  
Models 1870/1871/T127

**VENTAK PRIZM 2 VR/DR**  
Models 1860/1861

*Physician and patient  
letters are available at  
[www.bostonscientific.com](http://www.bostonscientific.com).*

**CURRENT RECOMMENDATION, continued...**

**CRT-Ps: RENEWAL TR/TR2**

- ERI or EOL indicator message displayed earlier than expected
- Fault Code 11 message (high current indicator)
- A gas gauge less than BOL within six months of implant

**ICDs: VENTAK PRIZM 2, VITALITY and VITALITY 2**

- ERI or EOL indicator message displayed earlier than expected
- A battery voltage **less than 3.10V** within six months of implant

---

Standard Warranty program available, please contact your local representative for terms and conditions.

PRODUCT	ORIGINAL COMMUNICATION 12-May-06 and 04-Jan-08 Subpectoral Implant
<p>A serialized search tool to determine if a specific device is affected by this product advisory is available at <a href="http://www.bostonscientific.com">www.bostonscientific.com</a>.</p> <p><i>This advisory is limited to those models listed below implanted subpectorally with the serial number facing the ribs..</i></p>	<p>Voluntary Physician Advisory FDA Classification: Class II</p> <p>Accelerated life testing has confirmed that repetitive mechanical stress applied to a specific area of the titanium case can induce component damage and device malfunction only if the device is implanted subpectorally with the serial number facing the ribs (leads exiting the pulse generator in a clockwise fashion). An anterior/posterior (AP) radiograph can be used to determine device orientation. Due to component location, damage associated with this subpectoral failure mode will not occur in a subcutaneous position or in a position with the serial number facing up.</p>
<p><b>CONTAK RENEWAL 4 HE</b> Models H197/H199</p>	<p>This failure mechanism can result in one or more of the following device behaviors:</p> <ul style="list-style-type: none"> <li>- Loss of shock therapy</li> <li>- Loss of pacing therapy (intermittent or permanent)</li> <li>- Loss of telemetry communications</li> <li>- Beeping (16 tones every six hours), and a programmer warning screen upon interrogation</li> </ul>
<p><b>CONTAK RENEWAL 4</b> Models H190/H195</p>	
<p><b>CONTAK RENEWAL 4 AVT / AVT HE</b> Models M170/M175/M177/M179</p>	<p><i>Reported Events</i> Two (2) reports of device malfunction associated with subpectoral implantation in an uncommon orientation (serial number facing ribs) were received. No patient deaths related to this advisory were reported. One patient required external pacing and immediate device replacement due to lack of pacing therapy. The vast majority of affected devices are implanted subcutaneously and are not subject to this failure mechanism.</p>
<p><b>CONTAK RENEWAL 3 HE</b> Models H177/H179</p>	
<p><b>CONTAK RENEWAL 3</b> Models H170/H175</p>	<p><i>Rate of Occurrence</i> The implant orientation of devices is not reported. For this reason, no rate of occurrence or failure rate projection was provided. However, based on available information, it is estimated that the number of devices implanted in a susceptible orientation is likely less than 1% of the total population.</p>
<p><b>CONTAK RENEWAL 3 AVT / AVT HE</b> Models M155/M159</p>	<p><b>CURRENT STATUS 17-Jan-14</b></p>
<p><b>VITALITY 2 EL VR/DR</b> Models T177/T167</p>	<p><i>Confirmed Malfunctions (worldwide)</i> <u>May 12, 2006 Population</u> Nineteen (19) malfunctions have been confirmed from an estimated 700 devices implanted in the susceptible orientation.</p>
<p><b>VITALITY DR HE</b> Model T180</p>	<p><u>January 4, 2008 Population</u> Seven (7) malfunctions have been confirmed from an estimated 330 devices implanted in the susceptible orientation.</p>
<p><b>VITALITY EL</b> Model T127</p>	
<p><b>VITALITY DR+</b> Model 1872</p>	<p>There have been no reported patient deaths associated with this advisory.</p>
<p><i>Physician and patient letters are available at <a href="http://www.bostonscientific.com">www.bostonscientific.com</a>.</i></p>	<p><i>Projected Rate of Occurrence</i> The projected rate of occurrence for devices implanted in the susceptible orientation is estimated to be 3% to 4% at 60 months.</p>
	<p><b>CURRENT RECOMMENDATION 17-Jan-14</b></p> <p><u>Patient management recommendations for both populations remain unchanged from the May 12, 2006 physician communication.</u></p> <ul style="list-style-type: none"> <li>- For patients implanted with a model listed in the advisory, review records to determine if the device was implanted subpectorally. Devices implanted subcutaneously are not subject to this advisory.</li> <li>- For subpectoral implants, use an AP radiograph to determine specific device orientation. <ul style="list-style-type: none"> <li>- If the leads exit the pulse generator in a counter clockwise direction (serial number facing away from the ribs), this advisory does not apply and no change to current patient management is necessary.</li> </ul> </li> </ul>



**12-May-06 and 04-Jan-08 Subpectoral Implant, continued...**

**CURRENT RECOMMENDATION, continued...**

If the device is in a susceptible orientation (serial number facing the ribs):

- Advise patient of the potential for device failure.
- Follow patient at 3 month intervals in accordance with device labeling.
- Consider device repositioning or replacement for physically active patients or for patients who regularly need device therapy.
- For future implants, when considering subpectoral implantation, orient the device with the serial number facing away from the ribs.

---

Standard Warranty program available, please contact your local representative for terms and conditions.

PRODUCT	<b>ORIGINAL COMMUNICATION 22-Sep-05 — Crystal Timing Component</b>
Identifiable by serial number. Not all serial numbers are affected.	Voluntary Physician Advisory FDA Classification: Class II
A serialized search tool to determine if a specific device is affected by this product advisory is available at <a href="http://www.bostonscientific.com">www.bostonscientific.com</a> .	Two separate failure modes were identified that may result in intermittent or permanent loss of pacing output without warning, intermittent or permanent loss of telemetry, and/or reversion to VVI mode or appearance of a Reset warning message upon interrogation. The root cause of the first failure mode is foreign material within a crystal timing component. As of September 22, 2005, the root cause of the second failure mode had not yet been determined and analysis was ongoing. As of the December 12, 2005 Advisory Update, root cause had been identified as a microscopic particle within the crystal timing component.
<b>INSIGNIA Ultra SR</b> Models 1190/1390	
<b>INSIGNIA Ultra DR and Ultra DR Downsize</b> Models 1291/1491/1290/1490	<i>Reported Events</i> Failure Mode 1—As of September 6, 2005, 36 malfunctions have been confirmed out of 49,500 devices distributed worldwide (0.073%). The majority of malfunctions occurred early in life, with a mean implant time of seven (7) months. There were no reported patient deaths. The supplier of the crystal timing component used in this subset of devices has eliminated foreign material within the crystal chamber, and no malfunctions were observed in any devices shipped after March 12, 2004.
<b>INSIGNIA Entra SR</b> Models 1195/1198/1395/1398	
<b>INSIGNIA Entra DR (downsize)</b> Models 1296/1466	Failure Mode 2—As of September 6, 2005, 16 malfunctions were confirmed out of 341,000 devices distributed worldwide (0.0047%). All 16 devices exhibited a no-output condition at the implant procedure or during pre-implant testing. There were no reported patient deaths.
<b>INSIGNIA Entra DR</b> Models 1294/1295/1494/1495	<i>Rate Projection</i> Failure Mode 1—As of the September 22, 2005 communication, Guidant's modeling, based on field experience and statistical analysis, predicted the malfunction rate for the active device population of 41,000 to be between 0.017% to 0.037% over the remaining device lifetime.
<b>INSIGNIA Entra SSI</b> Models 0484/0485/1325/1326	
<b>INSIGNIA Entra DDD</b> Models 0985/0986/1426	<b>CURRENT STATUS 17-Jan-14</b>
<b>INSIGNIA Plus SR</b> Models 1194/1394	<i>Confirmed Malfunctions (worldwide)</i> Failure Mode 1— 62 malfunctions out of approximately 49,500 advisory population devices have been confirmed. There have been no reported patient deaths associated with this advisory.
<b>INSIGNIA Plus DR and Plus DR Downsize</b> Models 1297/1467/1298/1468	Failure Mode 2— 26 malfunctions out of approximately 257,000 (0.010%) devices distributed have been confirmed. Twenty-two (22) malfunctions were identified before or during the implant procedure and four (4) were identified after implant. There have been no reported patient deaths associated with this advisory.
<b>INSIGNIA AVT</b> Models 0482/0882/0982 1192/1292/1392/1428/1432/1492	None of the INSIGNIA or NEXUS devices currently being distributed are susceptible to this malfunction mode.
<i>Physician and patient letters are available at <a href="http://www.bostonscientific.com">www.bostonscientific.com</a>.</i>	<i>Projected Rate of Occurrence</i> Failure Mode 1— The rate of occurrence for the estimated worldwide active advisory device population of 6,000 is projected to range between 0.027% and 0.038%.
	<b>CURRENT RECOMMENDATION 17-Jan-14</b>
	Failure Mode 1— <u>Patient management recommendations from the September 22, 2005 physician communication remain unchanged.</u>
	Failure Mode 2— <u>Patient management recommendations supersede those originally communicated on September 22, 2005.</u>
	– Normal follow-up for both Failure Mode 1 and Failure Mode 2 devices.
	– Specific to Failure Mode 1, physicians should consider the projected low and declining malfunction rate in addition to the unique needs of individual patients in their medical decisions regarding patient management.
	As always, advise patients to seek attention immediately if they experience syncope or lightheadedness.
	Standard Warranty program available, please contact your local representative for terms and conditions.

PRODUCT	ORIGINAL COMMUNICATION 18-Jul-05 and 21-Jan-06 — Hermetic Sealing Component
Identifiable by serial number. Not all serial numbers are affected.	Voluntary Physician Advisory (18-Jul-05) FDA Classification: Class I
A serialized search tool to determine if a specific device is affected by this product advisory is available at <a href="http://www.bostonscientific.com">www.bostonscientific.com</a> .	Voluntary Physician Advisory (21-Jan-06) FDA Classification: Class I
<b>CONTAK TR</b> Model 1241	A hermetic sealing component utilized in a subset of pacemakers may experience a gradual degradation, resulting in higher than normal moisture content within the pacemaker case late in the device's service life; this could lead to a variety of inappropriate clinical behaviors.
<b>DISCOVERY II SR (downsize)</b> Models 1184/1384	The original July 18, 2005 physician communication bounded the population to approximately 78,000 devices manufactured between November 25, 1997 and October 26, 2000; this number was further refined to 77,500 devices manufactured between October 27, 1997 and October 26, 2000.
<b>DISCOVERY II SR</b> Models 1186/1187/1385	The original July 18, 2005 communication predicted the rate of malfunction in the remaining active implanted devices (estimated at that time to be 28,000 worldwide) to be between 0.17% and 0.51% over the remaining device lifetime, based on field experience and statistical life-table analysis.
<b>DISCOVERY II DR (downsize)</b> Models 1283/1483	The original July 18, 2005 communication predicted the rate of malfunction in the remaining active implanted devices (estimated at that time to be 28,000 worldwide) to be between 0.17% and 0.51% over the remaining device lifetime, based on field experience and statistical life-table analysis.
<b>DISCOVERY II DR</b> Models 1284/1286/1484/1485	A Second Population of 54,000 devices was subsequently identified to be at risk of hermetic seal degradation (but at a much lower rate than the original population). This was communicated in the January 21, 2006 letter.
<b>DISCOVERY II SSI (downsize)</b> Models 0481/1349	Original Population— <u>Patient management recommendations from the July 18, 2005 physician letter remain unchanged and are provided below under CURRENT RECOMMENDATION</u> ; however, physicians should reassess patients in light of the increased projected rate of occurrence (detailed below).
<b>DISCOVERY II DDD</b> Models 0981/1285/1499	Original Population— <u>Patient management recommendations from the July 18, 2005 physician letter remain unchanged and are provided below under CURRENT RECOMMENDATION</u> ; however, physicians should reassess patients in light of the increased projected rate of occurrence (detailed below).
<b>PULSAR MAX II SR (downsize)</b> Models 1180/1380	Second Population—Physicians should consider the Original Population recommendations while taking into account the lower projected rate of occurrence.
<b>PULSAR MAX II SR / DR</b> Models 1181/1290/1480	<i>Rate Projection</i>
<b>DISCOVERY SR/SR (downsize)</b> Models 1174/1175	Refined Original Population—The predicted failure rate for the estimated worldwide active device population of 16,000 had increased from the July 18, 2005 estimate as communicated in the January 21, 2006 letter and was projected to range between 0.31% and 0.88% over the remaining device lifetime.
<b>DISCOVERY DR/DR (downsize)</b> Models 1274/1275/1273	Second Population—For the remaining lifetime of the estimated worldwide 19,300 active devices, the projected rate of occurrence for reported events was estimated to be between 0.02% and 0.06%.
<b>PULSAR MAX SR (downsize)</b> Model 1170	<b>CURRENT STATUS 17-Jan-14</b>
<b>PULSAR MAX SR / DR</b> Model 1171/1270	<i>Reported Events (worldwide)</i> Refined Original Population— 342 malfunctions have been confirmed out of the 77,500 advisory population devices.
<b>PULSAR</b> Models 1272/0470/0870/0970/ 0972/1172	Second Population— 13 malfunctions have been confirmed out of the 54,000 advisory population devices.
<b>MERIDIAN SSI / DDD</b> Models 0476/0976	<i>Projected Rate of Occurrence</i> Refined Original Population—The rate of occurrence for the estimated worldwide active device population of 3,000 is projected to range between 0.31% and 0.88% over the remaining device lifetime, as communicated in the January 21, 2006 Advisory Update letter.
<b>MERIDIAN SR / DR</b> Models 1176/1276	Second Population—The rate of occurrence for the estimated worldwide active device population of 2,000 is projected to range between 0.02% and 0.06%, as communicated in the January 21, 2006 Advisory Update letter.

**18-Jul-05 and 21-Jan-06 — Hermetic Sealing Component, continued...**

Physician and patient letters are available at [www.bostonscientific.com](http://www.bostonscientific.com).

**CURRENT RECOMMENDATION 17-Jan-14**

---

Original Population— **Patient management recommendations from the July 18, 2005 physician letter remain unchanged**; however, physicians should reassess patients in light of the increased projected rate of occurrence communicated in the January 21, 2006 Advisory Update letter.

Second Population—Physicians should consider the Original Population recommendations while taking into account the lower projected rate of occurrence.

- Consider replacing devices for pacemaker-dependent patients.
- Advise patients to seek attention immediately if they notice a prolonged rapid heart rate, experience syncope or lightheadedness, or have new or increased symptoms of heart failure.
- Select a suitable Maximum Sensor Rate (MSR) setting, given the rare possibility that inappropriate sustained pacing at MSR can occur

OR

- Consider programming the accelerometer OFF to prevent inappropriate sustained pacing at MSR.
- Consider increasing the frequency of programmer follow-ups. This increases the likelihood of detecting a malfunction that has already occurred, but does not guarantee that the device will not exhibit this malfunction mode in the future. At each patient follow-up:

- Evaluate for the clinical behaviors described in the July 18, 2005 letter.
- Evaluate battery status for signs of early or rapid depletion between sequential follow-up visits.
- Evaluate the accelerometer rate response (for devices with this feature).
  - Accelerometer ON:
    - Look for inappropriate MSR pacing or pacing higher than the programmed lower rate limit (LRL) while the patient is at rest.
    - Look for lack of rate response with activity (i.e., isometrics, short hall walk).
  - Accelerometer OFF:
    - *Temporarily* program the accelerometer ON and evaluate as described above

- Consider increasing the frequency of transtelephonic monitoring to detect inappropriate sustained MSR pacing and/or loss of pacing output.

- If any of these device behaviors are observed, contact your local representative or Technical Services for troubleshooting and recommendations.

---

Standard Warranty program available, please contact your local representative for terms and conditions.

# Trademarks

The following are trademarks of Boston Scientific Corporation, CRM Division (doing business as Cardiac Pacemakers, Inc., a Boston Scientific Company) used in connection with the goods or services indicated:

ACUITY	ENDOTAK DSP	LIVIAN
ACUITY X4	ENDOTAK ENDURANCE	MERIDIAN
ADVANTIO	ENDOTAK ENDURANCE EZ	ORIGEN
ALTRUA	ENDOTAK ENDURANCE RX	PULSAR
AUTOGEN	ENDOTAK RELIANCE	PUNCTUA
ACT	ENERGEN	RELIANCE 4-FRONT
CHFD	FINELINE	SELUTE
COGNIS	FLEXTEND	SWEET PICOTIP
CONFIENT	FORMIO	SWEET TIP
CONTAK	INSIGNIA	TELIGEN
CONTAK RENEWAL	INGENIO	ULTRA 1
CONTAK RENEWAL TR	INGEVITY	VENTAK
CONTAK TR	INCEPTA	VENTAK PRIZM
DISCOVERY	INLIVEN	VITALIO
DYNAGEN	INOGEN	VITALITY
EASYTRAK	INTUA	4-SITE
ENDOTAK	INVIVE	
EQUIO		

The following marks are registered trademarks for Intermedics, Inc and Cameron Health, Inc. (doing business as Cardiac Pacemakers, Inc., a Boston Scientific Company) used in connection with the goods or services indicated:

Q-TRAK	SQ-RX	VIRTUS
S-ICD	UNIPASS	

## OTHER

GORE and associated logos are trademarks of W.L. Gore & Associates.

# Country Offices Contact Information

## Argentina

Boston Scientific Argentina S.A.  
Av. Cabildo 2677 — Piso 6  
C1428AAI Buenos Aires  
Argentina  
Tel: 54 11 4896 8500  
Fax: 54 11 4896 8550

## Australia & New Zealand

Boston Scientific ANZ  
Level 5, 247 Coward Street  
MASCOT NSW 202 0  
Australia  
Tel: 61-2-8063 8100  
Fax: 61-2-9330 1404

## Austria

Boston Scientific  
Vienna Twin Tower  
Turm A/ 190G  
Wienerbergstrasse 11  
1100 Vienna  
Austria  
Tel: +43 1 60 810

## Belgium

Boston Scientific Company  
Green Square, Lambroekstraat 5D  
1831 Diegem  
Belgium  
Tel: +32-2-4167011

## Canada

Boston Scientific  
5060 Spectrum Way. Suite 500A  
Mississauga, Ontario L4W 5N5  
Canada

## China

BSC International Medical Trading  
(Shanghai) Co. Ltd.  
#68, Rijing Road  
Waigaoqiao Free Trade Zone  
Shanghai, 200131, China  
Telephone: 021-61415959  
Fax: 021-614159000

## Czech Republic

Guidant CR s.r.o.  
Karla Englise 3219/4  
150 00 Praha 5  
Tel: +420 296 331 900  
Fax: +420 296 331 902

## Denmark

Boston Scientific A/S  
Standvejen 70  
DK-2900 Hellerup  
Denmark

## France

Boston Scientific Headquarters  
Le Capitole  
55, avenue des Champs Pierreux  
92729 Nanterre  
France

## Germany

Boston Scientific Medizintechnik  
GmbH  
Daniel-Goldbach Strasse 17–27  
40880 Ratingen  
Germany  
Tel: 49 (0) 2102 489 3  
Fax: +49 (0) 2102 489 439

## Hong Kong

Boston Scientific Hong Kong Ltd.  
22/F., Bank of East Harbour View  
Centre  
56 Gloucester Road  
Wanchai, Hong Kong 135 984  
Tel: 852-2960 7100  
Fax: 852-2563 5276  
Ireland  
Boston Scientific — Clonmel  
Cashel Rd.  
Clonmel  
Co Tipperary  
Ireland  
Tel: 353.52.6181000

## Italy

Boston Scientific SpA  
Viale Enrico Forlanini, 23  
20134 Milan  
Italy  
Tel: 39.02.269831

## Japan

Boston Scientific Japan KK  
Headquarters  
Nikko Bldg.  
1-14-11 Nishi Shinjuku  
Shinjuku-ku, Tokyo 160-0023  
Japan

## Netherlands

Boston Scientific B.V.  
Kuifmees 56  
3435 RG Nieuwegein  
Netherlands  
Tel: + 31 – 30 - 6025555

## Singapore

Boston Scientific Asia Pacific Pte Ltd  
No. 1 Pickering Street  
#07–02, Great Eastern Centre  
Singapore 048659  
Tel: 65-6418 8888  
Fax: 65-6418 8899

## Portugal

Boston Scientific  
Amoreiras Plaza  
R. Prof. Carlos Alberto Mota Pinto  
N° 9 — ° 4 C  
1070–374 Lisboa  
Portugal

## Spain

Boston Scientific Iberica S.A.  
Parque Empresarial Puerta de las  
Naciones  
C/ ribera del Loira, 38. Edificio 4–1°  
planta  
28042 Madrid  
Spain  
Tel: 34-91 657 25 45

## Sweden

Boston Scientific Sweden AB  
Berga Alle 1  
254 52 Helsingborg  
Sweden  
Tel: 46 42 25 69 00

## Switzerland

Boston Scientific AG  
Dornacherplatz 7  
CH-4500 Solothurn  
Switzerland  
Tel: 0041 32 626 57 00

## United Kingdom

Boston Scientific  
Breakspear Park  
Breakspear Way  
Hemel Hempstead  
Hertz, HP2 4TZ  
United Kingdom  
Tel: 01442 411600  
Fax: 01442 411601

## United States

Boston Scientific Corporation  
4100 Hamline Avenue North  
St. Paul, MN 55112-5798  
USA  
Tel: 651.582.4000  
Fax: 651.582.4166  
Medical Professionals:  
1.800.CARDIAC (227.3422)  
Patients and Families:  
1.866.484.3268  
171



**Cardiac Rhythm Management**

One Boston Scientific Place  
Natick, MA 01760-1537 USA  
[www.bostonscientific.com](http://www.bostonscientific.com)

*Medical Professionals:*  
1.800.CARDIAC (227.3422)

*Patients and Families:*  
1.866.484.3268

© 2013 Boston Scientific Corporation  
or its affiliates. All rights reserved.

CRM-133901-AA MAR2013

Data as of January 17, 2014