

February 2022

Dear Patient,

We recently provided your heart doctor with new information about a software update for Boston Scientific's EMBLEM™ subcutaneous implantable cardioverter defibrillator (S-ICD) device. The information we provided to your doctor is available on our website and can be located at www.BostonScientific.com/advisory.

This new software improves a device's ability to monitor battery status. Your doctor will be able to tell you more about this software, how your system may be affected, and can work with you to make the best plan for you and your device. Your doctor may recommend that you come in for an office visit at the next scheduled follow-up so that your device can receive the new software.

To confirm if your S-ICD is affected, please visit this website: www.BostonScientific.com/lookup. You will need to enter your language, as well as the model and serial numbers of your device, which can be found on your patient ID card.

What should you do?

- Keep and attend all your scheduled medical and device check appointments.
- Talk to your doctor about the best steps for managing your device system.

At Boston Scientific, patient safety is our priority and our constant focus. We are committed to working closely with you and your doctor. We recommend that you discuss this letter with your doctor, who knows best how this new information may affect you.



Alexandra Naughton
Vice President, Quality Assurance



Kenneth Stein, MD, FACC, FHRS
Senior Vice President and Chief Medical Officer, RM