

# IMAGEREADY™

MR-Conditional Systems

[www.bostonscientific.com/imageready](http://www.bostonscientific.com/imageready)

ImageReady™ MR Conditional System Components for 1.5T and 3T<sup>1</sup>

Component	Model Number(s)	3T	1.5T
<b>Pulse Generators</b>			
<b>Pacemakers</b>			
ESSENTIO™ MRI, PROPONENT™ MRI, ACCOLADE™ MRI	L110, L111, L131, L210, L211, L231, L310, L311, L331	✓	✓
<b>Defibrillators</b>			
EMBLEM™ S-ICD System	A209		✓
EMBLEM™ MRI S-ICD System	A219		✓
AUTOGEN™, DYNAGEN™, INOGEN™, ORIGEN™	D044, D046, D174, D176, D020, D022, D150, D152, D010, D012, D140, D142, D000, D002, D050, D052		✓
RESONATE™, VIGILANT™, PERCIVA™	D232, D233, D412, D413, D432, D433		✓
<b>CRT-Ds</b>			
AUTOGEN™, DYNAGEN™, INOGEN™	G148, G158, G179		✓
RESONATE™, VIGILANT™	G247, G447		✓
<b>Leads</b>			
INGEVITY™ MRI Pacing Leads*	7731, 7732, 7735, 7736, 7740, 7741, 7742, 7631, 7632, 7635, 7636, 7640, 7641, 7642	✓	✓
FINELINE™ II Pacing Leads	4456, 4457, 4458, 4459, 4469, 4470, 4471, 4472, 4473, 4474, 4479, 4480		✓
ENDOTAK RELIANCE 4-SITE™ ICD Leads	0262, 0263, 0265, 0266, 0272, 0273, 0275, 0276, 0282, 0283, 0285, 0286, 0292, 0293, 0295, 0296,		✓
RELIANCE 4-FRONT™ ICD Leads	0636, 0654, 0655, 0657, 0658, 0675, 0676, 0682, 0683, 0685, 0686, 0692, 0693, 0695, 0696, 0658, 0650, 0651, 0652, 0653, 0662, 0663, 0672, 0673		✓
EMBLEM™ S-ICD System Electrode	3401, 3501		✓
Q-Trak Subcutaneous Electrode	3400		✓
ACUITY™ X4 CRT Leads	4671, 4672, 4674, 4675, 4677, 4678		✓

\*Note: There are no design differences between the 77XX and 76XX Ingevity™ leads. Therefore, 76XX models of Ingevity™ leads are also deemed MR Conditional in Canada.

Component	Model Number(s)	3T	1.5T
<b>Accessories</b>			
Suture Sleeve for INGEVITY™ MRI Pacing Leads	6402	✓	✓
Suture Sleeve for RELIANCE 4-FRONT™ ICD Leads	4603		✓
Suture Sleeves for FINELINE™ II Pacing Leads	6220, 6221		✓
IS-1 Lead Port Plug	7145	✓	✓
IS-4/DF-4 Lead Port Plug	7148		✓

1. For conditions of use and scan conditions, refer to the appropriate ImageReady™ MR Conditional System manual found at: [www.bostonscientific.com/manuals](http://www.bostonscientific.com/manuals)

© 2017 Boston Scientific Corporation or its affiliates. All rights reserved. All cited trademarks are the property of their respective owners. CAUTION: Law restricts these devices to the sale by or on the order of a physician. Indications, contraindications, warnings and instructions for use can be found in the product labelling supplied with each device. CANCRM-007-AA