

Endoscopic Bite Block with Elastic Cloth Strap

Refer to the device directions for use for complete instructions on device use.

Caution: Federal Law (USA) restricts this device to sale by or on the order of a physician.

Warning

Contents supplied NON-STERILE. Do not use if damaged. If damage is found, call your Boston Scientific representative.

For single use only. Do not reuse, reprocess or sterilize. Reuse, reprocessing or sterilization may compromise the structural integrity of the device and/or lead to device failure which, in turn, may result in patient injury, illness or death. Reuse or reprocessing may also create a risk of contamination of the device and/or cause patient infection or cross-infection, including, but not limited to, the transmission of infectious disease(s) from one patient to another. Contamination of the device may lead to injury, illness or death of the patient.

After use, dispose of product and packaging in accordance with hospital, administrative and/or local government policy.

Intended Use/ Indications For Use

The Blox Endoscopic Bite Block with Elastic Cloth Strap is intended to provide safe passage through the patient's mouth during upper endoscopic procedures thus protecting the patient's teeth and gums as well as protecting endoscopic equipment.

Contraindications

There are no known contraindications.

Warnings

Do not use if there is any apparent shipping or handling damage.

Precaution

Care should be exercised to avoid placing fingers between the bite block and the patient's teeth.

Adverse Events

None known.

All trademarks are the property of their respective owners. CAUTION: U.S. Federal law restricts this device to sale by or on the order of a physician.

Indications, contraindications, warnings and instructions for use can be found in the product labeling supplied with each device.

2023 Copyright © Boston Scientific Corporation or its affiliates. All rights reserved.