Stone Retrieval Basket

Prescriptive Information

Refer to the device directions for use for complete instructions on device use.

Intended Use/Indications for Use

The Stone Retrieval Basket is intended to be used during urological procedures to endoscopically grasp, manipulate and remove calculi and other foreign objects.

Contraindications

None known.

Warnings

- For single use only. Do not reuse, reprocess or resterilize.
- Some objects may be too large to be removed endoscopically with a retrieval basket. To avoid object impaction, fluoroscopy or x-ray should be used to determine the size of the object; do not use the Stone Retrieval Basket if the object is too large to be removed endoscopically or held in basket.
- This device must not come in contact with any electrified instrument.
- This device should not be directly fired upon by any lithotrite. To do so may damage the device and could result in patient injury.
- Use of device without introducer may cause damage to the basket.
- Warnings can be found in the product labeling supplied with each device.

Potential Adverse Events

Potential complications generally recognized with the use of stone retrieval basket devices include, but are not limited to:

- Perforation
- Evulsion
- Edema
- Basket Inversion
- Entrapment
- Hemorrage
- Inability to disengage from irretrievable object

Precautions

- Kinks in the sheath will hinder the mechanical operation of the basket.
- It is recommended that the ureter be predilated before endoscopy and object retrieval to prevent object impaction when retracting the object.
- Precautions can be found in the product labeling supplied with each device.
- Caution: Federal (USA) law restricts this device to sale by or on the order of a physician.

URO-238206-AA APR 2014
©2014 Boston Scientific Corporation or its affiliates. All rights reserved.