

CONTAK RENEWAL[®] TR

Models H120 , H125

Details About Your Boston Scientific CONTAK RENEWAL[®] TR Cardiac Resynchronization Therapy Pacemaker (CRT-P)

What is a CRT-P?

CRT-P stands for *cardiac resynchronization therapy pacemaker*. A CRT-P is a small, battery-powered device that holds a tiny computer. It is implanted in the body to provide special pacing therapy to treat heart failure. Your doctor programs your CRT-P to respond appropriately based on your individual heart rhythm needs.

Treating dyssynchrony

In heart failure, the chambers of the heart do not contract quite as they should. They are *dyssynchronous*, or *out of synch*. When that happens, your heart cannot provide your body with the blood flow it needs to function properly. Cardiac resynchronization therapy (CRT) is designed to help your heart pump better. It delivers specially timed electrical impulses to coordinate, or *resynchronize*, the timing of your heart's contractions. This can help restore your heart's pumping function or ability to fill with blood and work more efficiently.

Treating slow heart rhythms

If an abnormally slow heart rhythm occurs, your CONTAK RENEWAL TR CRT-P can deliver small electrical signals that help your heart beat more normally. Your CRT-P can also help increase your heart rate to meet your body's needs when you are active.

How small is my CRT-P?

Height	Width	Thickness	Weight	Volume
2.13 in. 5.40 cm	1.77 in. 4.50 cm	0.33 in. 0.85 cm	0.92 oz. 26 g	14 cc

How long will my CRT-P last?

Like anything that operates on a battery, the life of your CRT-P will depend on how much the battery is used. The battery is used every time your device delivers therapy. So the amount of therapy you receive, and the amount of energy your doctor programs into those therapies, affects the life of the battery. A CONTAK RENEWAL TR CRT-P typically lasts 5–7 years.



ACTUAL SIZE
(when printed at 100%)

**Boston
Scientific**

What happens after my implant procedure?

Care and recovery at home

Your doctor or nurse will give you specific instructions to follow when returning home. Full recovery from surgery can take from several days to a few months, depending on your condition. It is important that you are actively involved in your own recovery. Follow your doctor's instructions about caring for the implant site and letting your other doctors, dentists, and emergency personnel know that you have an implanted CRT-P system.

Follow-up visits

Once you have a CRT-P, it is important to follow your doctor's guidelines for follow-up visits. Your doctor establishes a periodic follow-up schedule based on his/her preference. Your doctor uses the follow-up visits to check your device and monitor your heart condition, which can change over time. Your doctor may also adjust your device settings.

Important Safety Information

Cardiac Resynchronization Therapy Devices

Cardiac resynchronization therapy pacemakers (CRT-P) and defibrillators (CRT-D) are designed to treat heart failure patients who have symptoms despite the best available drug therapy. They are also designed to help your heart pump more effectively and meet your body's need for blood flow. CRT-P and CRT-D systems are not for everyone, including people with separate implantable cardioverter-defibrillators (CRT-P only) or certain steroid allergies. Procedure risks include infection, tissue damage, and kidney failure. Patients who should not receive this device include: patients who have additional medical conditions that may not allow the pacemaker to function appropriately (CRT-P only) and patients whose ventricular rhythm disturbances or heart failure have a reversible or temporary cause. In some cases, the device may be unable to detect or appropriately respond to your heart rhythm (CRT-P and CRT-D) or may deliver inappropriate shocks (CRT-D only). In rare cases severe complications or device failures can occur. Electrical or magnetic fields can affect the device. Only your doctor knows what is right for you. These devices are available by prescription only. Individual results may vary. Your physician should discuss all potential benefits and risks with you. For further safety information, refer to the Physicians Instructions for Use or call 1.866.484.3268.

Device Quality and Reliability

It is Boston Scientific's intent to provide implantable devices of high quality and reliability. However, these devices may exhibit malfunctions that may result in lost or compromised ability to deliver therapy. Refer to Boston Scientific's CRM product performance report on www.bostonscientific.com for more information about device performance, including the types and rates of malfunctions that these devices have experienced historically. While historical data may not be predictive of future device performance, such data can provide important context for understanding the overall reliability of these types of products. Also, it is important that you talk with your doctor about the risks and benefits associated with the implantation of a device.

(Rev. E)

**Boston
Scientific**

Delivering what's next.™

Cardiac Rhythm Management

Boston Scientific Corporation
4100 Hamline Avenue North
St. Paul, MN 55112-5798 USA
Tel: 651.582.4000 Fax: 651.582.4166
Medical Professionals:
1.800.CARDIAC (227.3422)
Patients and Families:
1.866.484.3268
www.bostonscientific.com

Copyright © 2010 by
Boston Scientific Corporation
or its affiliates. All rights reserved.

CRM9-4755-0610