



## **Important Software Update Information COGNIS 100-D and TELIGEN 100**

Boston Scientific has released Model 2868 software application version 1.05. With this software update, the Respiratory Sensor will be programmed to Off and you will no longer have the option to program it to On. Any Respiratory Rate Trend data you have previously collected will be stored and viewable to you on the PRM. However, no new data will be collected.

If you have any questions regarding this update, please contact Boston Scientific Technical Services at the phone number shown below.

Boston Scientific  
4100 Hamline Avenue North  
St. Paul, MN 55112-5798 USA

1.800.CARDIAC (227.3422)  
+1.651.582.4000  
[www.bostonscientific.com](http://www.bostonscientific.com)

© 2009 Boston Scientific or its affiliates.  
All rights reserved.

**358200-001 EN US 04/09**

