


Blazer II XP™ Temperature Ablation Catheters



8mm & 10mm tips

Product Information

Tip: 8F/8mm Straight **Electrode Configuration:** Quadripolar **Electrode Spacing:** 2.5mm

Catheter Model No.	Shaft Size	Curve Style	Distal Shaft Length	Cable Model No.*
M004 4500TH 0	7F	Standard	Standard	M004 651 0
M004 4500THM 0	7F	Standard	Medium	M004 651 0
M004 4500THK2 0	7F	Large	Standard	M004 651 0
M004 4500THMK2 0	7F	Large	Medium	M004 651 0
M004 4500THN4 0	7F	Asymmetric 4	Standard	M004 651 0

Tip: 8F/10mm Straight **Electrode Configuration:** Quadripolar **Electrode Spacing:** 2.5mm

Catheter Model No.	Shaft Size	Curve Style	Distal Shaft Length	Cable Model No.*
M004 4790TH 0	7F	Standard	Standard	M004 651 0
M004 4790THM 0	7F	Standard	Medium	M004 651 0
M004 4790THK2 0	7F	Large	Standard	M004 651 0
M004 4790THMK2 0	7F	Large	Medium	M004 651 0
M004 4790THN4 0	7F	Asymmetric 4	Standard	M004 651 0

Tip: 8F/8mm Contour **Electrode Configuration:** Quadripolar **Electrode Spacing:** 2.5mm

Catheter Model No.	Shaft Size	Curve Style	Distal Shaft Length	Cable Model No.*
M004 4770TH 0	7F	Standard	Standard	M004 651 0
M004 4770THM 0	7F	Standard	Medium	M004 651 0
M004 4770THK2 0	7F	Large	Standard	M004 651 0
M004 4770THMK2 0	7F	Large	Medium	M004 651 0
M004 4770THN4 0	7F	Asymmetric 4	Standard	M004 651 0

Note: For use with the EPT-1000XP™ Controllers and Maestro 3000™ Controller.

All Blazer II XP Catheters are 110cm in length.

*If an extension cable is required, use Model Number M004 613 0 (10 ft. length).

Blazer II XP Catheters require the use of 2 Valleylab® Ground Pads (Model M004 354 0).

Indications, contraindications, warnings and instructions for use can be found in the instructions for use supplied with each device. CAUTION: Federal (USA) law restricts these devices to sale by or on the order of a physician.




©2007 Boston Scientific Corporation or its affiliates. All rights reserved. EPT-0015 _ 04/07

**Boston
Scientific**

For additional product information call Customer Service at (888) 272-1001

www.bostonscientific.com/electrophysiology

Blazer II™ Catheter Bidirectional Curve Options



Indications, contraindications, warnings and instructions for use can be found in the instructions for use supplied with each device. CAUTION: Federal (USA) law restricts these devices to sale by or on the order of a physician.

©2007 Boston Scientific Corporation or its affiliates. All rights reserved. EPT-0015 _ 04/07

**Boston
Scientific**

For additional product information call Customer Service at (888) 272-1001

www.bostonscientific.com/electrophysiology