

CRE™ Fixed Wire Balloon Dilators

CRE SINGLE-USE FIXED WIRE BALLOON DILATORS

Order Number	Balloon Inflated O.D. (mm)	Balloon O.D. (Fr)	Balloon Length (cm)	Inflation Pressures (ATM)	Minimum Working Channel (mm)	Catheter Size (Fr)	Working Length (cm)	Unit
M00558330	6-7-8	18-21-24	8	3-6-10	2.8	6	180	Each
M00558340	8-9-10	24-27-30	8	3-5-9	2.8	6	180	Each
M00558350	10-11-12	30-33-36	8	3-5-8	2.8	6	180	Each
M00558360	12-13.5-15	36-40.5-45	8	3-4.5-8	2.8	6	180	Each
M00558370	15-16.5-18	45-49.5-54	8	3-4.5-7	2.8	6	180	Each
M00558380	18-19-20	54-57-60	8	3-4.5-6	2.8	6	180	Each
M00558361	12-13.5-15	36-40.5-45	8	3-4.5-8	2.8	6	180	Box 5
M00558371	15-16.5-18	45-49.5-54	8	3-4.5-7	2.8	6	180	Box 5
M00558381	18-19-20	54-57-60	8	3-4.5-6	2.8	6	180	Box 5

Recommended Inflation Device: Alliance™ II Integrated Inflation System Re-usable Handle.
This product contains no detectable latex.

INSTRUCTIONS FOR USE

Refer to the operator's manual for complete instructions for use.

INDICATIONS

Indicated for use in adult and adolescent populations to endoscopically dilate strictures of the esophagus.

CONTRAINDICATIONS

None known.

WARNINGS

Do not exceed the inflation pressure given.

STERILE. For single use only. Do not reuse, reprocess or resterilize. Reuse, reprocessing or resterilization may compromise the structural integrity of the device and/or lead to device failure which, in turn, may result in patient injury, illness or death. Reuse, reprocessing or resterilization may also create a risk of contamination of the device and/or cause patient infection or cross-infection, including, but not limited to, the transmission of infectious disease(s) from one patient to another. Contamination of the device may lead to injury, illness or death of the patient.

After use, dispose of product and packaging in accordance with hospital, administrative and/or local government policy.

Check for proper positioning of the balloon catheter using endoscopic visualization. Balloon inflation in an improper location may lead to patient injury.

POTENTIAL ADVERSE EVENTS

Potential adverse events may result from an esophageal balloon dilation procedure and include, but are not limited to perforation; hemorrhage; hematoma; septicemia/infection; and an allergic reaction to contrast medium.

Please be aware that potential adverse effects may arise even with the proper use of medical devices. Accordingly, this device should be used by persons qualified in the procedures for which it is indicated.

CAUTIONS

Cautions can be found in the product labeling supplied with each device. CAUTION: Federal (USA) law restricts this device to sale by or on the order of a physician.

Boston Scientific
Tel 508.650.8000

www.bostonscientific.com

Ordering Information
1.800.225.3226

©2008 Boston Scientific Corporation
or its affiliates. All Rights Reserved.