

Lynx™ Suprapubic Mid-Urethral Sling System

Mid-Urethral Sling System for the Treatment of Stress Urinary Incontinence

LYNX™ SUPRAPUBIC MID-URETHRAL SLING SYSTEM

Order Number	Description
M0068503000	Lynx Mid-Urethral Sling System (1 Pack = 2 Delivery Devices, 1 Mesh Assembly)
M0068503001	Lynx Mid-Urethral Sling System (5 Pack)

INSTRUCTIONS FOR USE

Refer to the operator's manual for complete instructions for use.

INDICATIONS

The Advantage™ Mesh implant is intended as a suburethral sling for the treatment of stress urinary incontinence resulting from hypermobility and/or intrinsic sphincter deficiency.

CONTRAINDICATIONS

The Advantage Mesh Assembly is contraindicated in pregnant patients. Additionally, this procedure should not be performed in patients with potential for future growth or patients with plans for future pregnancy. Use of the surgical mesh is contraindicated for use in any patient in whom soft tissue implants are contraindicated.

WARNINGS

For single patient use only. Do not reuse, reprocess or resterilize.

POTENTIAL ADVERSE EFFECTS

Potential complications that may occur in a suburethral sling procedure are allergic reaction, fistula, abscess, irritative voiding symptoms including urgency and urge incontinence, detrusor instability, infection, pelvic, vaginal pain, urinary retention, dysparenia, vaginal bleeding, vaginal discharge, erosion of the vaginal or urethral mucosa or bladder wall, dehiscence of vaginal incision, edema at the wound site, erythema at the wound site.

CAUTIONS

Federal (USA) law and governing law outside the USA restrict these devices to sale by or on the order of a physician.



Boston Scientific
Tel 508.650.8000
www.bostonscientific.com

Ordering Information
1.888.272.1001

©2004 Boston Scientific Corporation
or its affiliates. All rights reserved.