

Obtryx™ Transobturator Mid-Urethral Sling System

Mid-Urethral Sling System for the Treatment of Stress Urinary Incontinence

OBTRYX™ TRANSOBTURATOR MID-URETHRAL SLING SYSTEM

Order Number	Description
M0068504000	Obtryx Sling System – Curved (Single Unit)
M0068504001	Obtryx Sling System – Curved (5 Pack)
M0068505000	Obtryx Sling System – Halo (Single Unit)
M0068505001	Obtryx Sling System – Halo (5 Pack)

INSTRUCTIONS FOR USE

Refer to the operator's manual for complete instructions for use.

INDICATIONS

The Advantage™ Mesh implant is intended as a suburethral sling for the treatment of stress urinary incontinence resulting from hypermobility and/or intrinsic sphincter deficiency.

CONTRAINDICATIONS

The Advantage Mesh Assembly is contraindicated in the following patients: Pregnant patients, patients with potential for future growth, patients that are considering future pregnancies, any patient with soft tissue pathology into which the implant is to be placed, patients with any pathology which would compromise implant placement, patients with any pathology such as blood supply limitation or infection that would compromise healing.

WARNINGS

For single patient use only. Do not reuse, reprocess or resterilize.

POTENTIAL ADVERSE EFFECTS

Potential complications that may occur in a suburethral sling procedure are allergic reaction, fistula, abscess, irritative voiding symptoms including urgency and urge incontinence, detrusor instability, infection, pelvic, vaginal pain, urinary retention, dysparenia, vaginal bleeding, vaginal discharge, erosion of the vaginal or urethral mucosa or bladder wall, dehiscence of vaginal incision, edema at the wound site, erythema at the wound site.

CAUTIONS

Federal (USA) law and governing law outside the USA restrict these devices to sale by or on the order of a physician.



Boston Scientific
Tel 508.650.8000
www.bostonscientific.com

Ordering Information
1.888.272.1001

©2004 Boston Scientific Corporation
or its affiliates. All rights reserved.