



ImageReady™ MR-Conditional Model List

Therapy	Device Name	Model Number(s)	MR Status	1.5T	3T
Pacemakers	ESSENTIO™ MRI	L110, L111, L131	MR Conditional	✓	✓
	PROponent™ MRI	L210, L211, L231	MR Conditional	✓	✓
	ACCOLADE™ MRI	L310, L311, L331	MR Conditional	✓	✓
	ADVANTIO™ MRI	J065, J066, J067	MR Conditional	✓	
	INGENIO™ MRI	J175, J176, J177	MR Conditional	✓	
	VITALIO™ MRI	J275, J276, J277	MR Conditional	✓	
	FORMIO™ MRI	J279	MR Conditional	✓	
S-ICD	EMBLEM™ MRI	A219	MR Conditional	✓	
	EMBLEM™ S-ICD	A209	MR Conditional	✓	
CRT-P	VALITUDE™ X4	U128	MR Conditional	✓	✓
	VISIONIST™ X4	U228	MR Conditional	✓	✓
ICDs	AUTOGEN™ ICD	D174, D175, D176, D177	MR Conditional	✓	
	AUTOGEN™ MINI	D044, D045, D046, D047	MR Conditional	✓	
	CHARISMA™ DF1 ICD	D321, D320	MR Conditional	✓	
	CHARISMA™ ICD	D332, D333	MR Conditional	✓	✓
	DYNAGEN™ ICD	D020, D021, D022, D023, D150, D151, D152, D153	MR Conditional	✓	
	INOGEN™ ICD	D010, D011, D012, D013, D140, D141, D142, D143	MR Conditional	✓	
	MOMENTUM™ ICD	D120, D121	MR Conditional	✓	
	ORIGEN™ ICD	D000, D001, D002, D003, D050, D051, D052, D053	MR Conditional	✓	
	PERCIVA DF1 ICD	D400, D401	MR Conditional	✓	
	PERCIVA™ ICD	D412, D413	MR Conditional	✓	✓
	RESONATE™ DF4 ICD	D432, D433	MR Conditional	✓	✓
	VIGILANT™ ICD	D232, D233	MR Conditional	✓	✓
	CRT-Ds	AUTOGEN™ CRT-D	G172, G173, G177, G179	MR Conditional	✓
CHARISMA™ CRT-D		G348, G324, G325	MR Conditional	✓	
CHARISMA™ X4 CRT-D		G347	MR Conditional	✓	✓
DYNAGEN™ CRT-D		G150, G151, G156, G158	MR Conditional	✓	
INOGEN™ CRT-D		G140, G141, G146, G148	MR Conditional	✓	
MOMENTUM™ CRT-D		G124, G125, G138	MR Conditional	✓	
ORIGEN™ CRT-D		G050, G051, G056, G058	MR Conditional	✓	
RESONATE™ IS-1 CRT-D		G424	MR Conditional	✓	
RESONATE™ X4 CRT-D		G447	MR Conditional	✓	✓
VIGILANT™ IS-1 CRT-D		G224	MR Conditional	✓	
VIGILANT™ X4 CRT-D		G247	MR Conditional	✓	✓
ICM	LUX-DX™	M301	MR Conditional	✓	✓
Leads and Accessories labeled as MR-Conditional	ACUITY™ Spiral Pacing lead	4591, 4592, 4593	MR Conditional	✓	
	ACUITY™ X4 Pacing leads	4671, 4672, 4674, 4675, 4677, 4678	MR Conditional	✓	✓
	EASYTRAK™ 2 Pacing lead	4542, 4543, 4544	MR Conditional	✓	
	ENDOTAK RELIANCE™ (DF1) Defibrillation leads	0127, 0128, 0129, 0137, 0138, 0139, 0143, 0147, 0148, 0149, 0153, 0157, 0158, 0159, 0170, 0171, 0172, 0173, 0174, 0175, 0176, 0177, 0180, 0181, 0182, 0183, 0184, 0185, 0186, 0187	MR Conditional	✓	
	ENDOTAK RELIANCE™ (DF4) Defibrillation leads	0262, 0263, 0265, 0266, 0272, 0273, 0275, 0276, 0282, 0283, 0285, 0286, 0292, 0293, 0295, 0296	MR Conditional	✓	✓
	FINELINE™ II Sterox EZ Pacing Leads	4469, 4470, 4471, 4472, 4473, 4474	MR Conditional	✓	✓
	FINELINE™ II Sterox Pacing Leads	4456, 4457, 4458, 4459, 4479, 4480	MR Conditional	✓	✓
	INGEVITY™ MRI Pacing Leads	7731, 7732, 7735, 7736, 7740, 7741, 7742	MR Conditional	✓	✓
	INGEVITY™+ Pacing Leads	7840, 7841, 7842	MR Conditional	✓	✓
	IS-1 Lead Port Plug	7145	MR Conditional	✓	✓
	IS4/DF4 Lead Port Plug	7148	MR Conditional	✓	✓
	DF1 Lead Port Plug	6996	MR Conditional	✓	
	RELIANCE 4-FRONT™ Defibrillation leads	0636, 0650, 0651, 0652, 0653, 0654, 0655, 0657, 0658, 0662, 0663, 0665, 0672, 0673, 0675, 0676, 0682, 0683, 0685, 0686, 0692, 0693, 0695, 0696	MR Conditional	✓	✓
	S-ICD electrodes	3010, 3400, 3401, 3501	MR Conditional	✓	
	Suture Sleeve for ACUITY™ X4 leads	4603	MR Conditional	✓	✓
	Suture Sleeve for FINELINE™ II pacing leads	6220, 6221	MR Conditional	✓	✓
	Suture Sleeve for INGEVITY™ MRI pacing lead	6402	MR Conditional	✓	✓
Suture Sleeve for RELIANCE 4-FRONT™ lead	6403	MR Conditional	✓	✓	

***Warning:** The Programmer/Recorder/Monitor (PRM) is MR Unsafe and must remain outside the MRI site Zone III (and higher) as defined by the American College of Radiology Guidance Document for Safe MR Practices. Under no circumstances should the PRM be brought into the MRI scanner room, the control room, or the MRI site Zone III or IV areas.

MRI TECHNICAL GUIDE IMAGEREADY™ MR CONDITIONAL DEFIBRILLATION SYSTEM 360205-037 EN EUROPE 2019-12
MRI TECHNICAL GUIDE IMAGEREADY™ MR CONDITIONAL S-ICD SYSTEM 359474-001 EN US 2015-11
MRI TECHNICAL GUIDE IMAGEREADY™ MR CONDITIONAL PACING SYSTEM 359259-037 EN EUROPE 2020-01

All cited trademarks are the property of their respective owners. All trademarks are property of their respective owner. CAUTION: The law restricts these devices to sale by or on the order of a physician. Indications, contraindications, warnings, and instructions for use can be found in the product labelling supplied with each device or at www.IFU-BSCI.com. Products shown for INFORMATION purposes only and may not be approved or for sale in certain countries. This material not intended for use in France.

CRM-842401-AD Printed in Germany by medicalvision. C € 2797

**Boston
Scientific**
Advancing science for life™

www.bostonscientific.eu

© 2024 Boston Scientific Corporation
or its affiliates. All rights reserved.