



Your ImageReady™ MR Conditional Pacing System and MRI Scans

Overview

Magnetic Resonance Imaging (MRI) is a diagnostic test that uses a powerful magnet and radio waves to produce images of the human body. Most pacing systems are not appropriate for use during an MRI scan or near MRI scanners. However, some pacemaker patients need to have MRI scans to diagnose their health conditions and determine proper treatment. Boston Scientific has developed a pacing system that can be scanned when certain conditions are met. This includes programming the pacemaker to special settings for MRI.

You MAY be eligible to have an MRI scan if you are implanted with the Boston Scientific ImageReady MR Conditional pacing system (pacemaker and lead wires designed to work in the MRI environment). Your heart doctor or healthcare provider will work with you to determine if you and your pacing system can undergo an MRI scan.

Make sure you and your physician(s) are comfortable with the risks and benefits of having an MRI scan.

Common Questions

To be eligible for an MRI scan, you must have the right pacing system implanted. This includes the pacemaker and the lead wires. The clinic performing the scan must use an MRI machine that meets the technical requirements of the ImageReady pacing system.

Common questions include:

1. Am I able to have an MRI scan?
 - If a healthcare provider recommends an MRI scan, talk to your heart doctor before scheduling your MRI scan.
 - Your heart doctor will check you and your ImageReady pacing system to determine if you are eligible for an MRI scan.
 - Even if your pacing system is eligible, you may have other implanted devices or metal in your body that may prevent you from having an MRI scan.

- You must be physically capable of having an MRI scan. This means that you can lay flat during the scan, and tolerate how the pacemaker functions during the scan.
- Your implanted pacemaker and lead wires are listed on your current medical device ID card. You should keep this card with you at all times, and take it with you to the MRI facility.
- Always check with your heart doctor or a healthcare provider if you have questions before or after the MRI scan.

2. What can I expect during an MRI scan?

- Before the scan, your ImageReady pacing system will be programmed to the MRI settings. These settings are necessary for you to receive an MRI scan. While your pacing system is programmed to these settings, your heart rate may be different from what you are accustomed to.
- Your heart function will be monitored during the scan.
- Your ImageReady pacing system may stay in the new MRI settings for a limited time. If so, you or your caregiver may be notified of the duration of time your device will remain in the MRI settings. The MRI scan needs to be completed before that time expires.
- You may see signs at the MRI facility that warn you not to enter if you have a pacemaker. These signs apply to pacing systems that are not eligible, programmed, and cleared to have an MRI scan. Always check with a healthcare provider if you have questions.
- During the MRI scan, you may experience the following:
 - Loud noises are part of a normal MRI scan; the noises are not coming from your pacemaker.
 - You might feel slight movement or vibration of the pacemaker, or a warm sensation from the pacemaker.

3. What happens after the scan?

- After the scan, follow the directions provided by your heart doctor or healthcare provider.
- Your implanted ImageReady pacing system may be checked to ensure it is working normally. The pacing system may be programmed out of the MRI settings, or it may return to your normal pacing settings automatically after a specific period of time set by your doctor.
- If you experience any new symptoms after the scan, contact your heart doctor or healthcare provider.
- If you need additional MRI scans, you and your pacing system must be checked for eligibility for a scan each time. If your implanted pacing system has changed, for example a new pacemaker or lead wires, or if aspects of your health have changed, it is possible that you are no longer eligible for an MRI scan.

ImageReady is a trademark of Boston Scientific or its affiliates.

Boston Scientific

Boston Scientific
4100 Hamline Avenue North
St. Paul, MN 55112-5798 USA
Tel: 651.582.4000

Medical Professionals:
1.800.CARDIAC (227.3422)

Patients and Families:
1.866.484.3268

www.bostonscientific.com

© 2014 Boston Scientific Corporation or its affiliates.
All rights reserved.

359260-001 EN US 2014-07

