

Metal Composition

Promus PREMIER™ Everolimus-Eluting Platinum Chromium Coronary Stent System

The Promus PREMIER Stent is made from Platinum Chromium alloy and nominally consists of the following elements (with additional trace elements):

Promus PREMIER Stent	Elemental Composition by Weight
Iron	37%
Platinum	33%
Chromium	18%
Nickel	9%
Molybdenum	3%
Manganese	Trace

Directions for Use (DFU):

Section 4.0 Contraindications

Use of the Promus PREMIER Stent System is contraindicated in patients with known hypersensitivity to:

- 316L stainless steel or platinum
- everolimus or structurally related compounds
- the polymer or its individual components (see details in Section 2.2.2, Primer polymer and drug matrix copolymer carrier).

Coronary artery stenting is contraindicated for use in the following:

- Patients judged to have a lesion that prevents complete inflation of an angioplasty balloon or proper placement of the stent or delivery device.
- Patients with uncorrected bleeding disorders or patients who cannot receive anticoagulation or anti-platelet aggregation therapy (see section 6.2, Pre- and Post-Antiplatelet Regimen).

For more information, contact your local sales representative or call the Boston Scientific Customer Service Center at 888-272-1001.

[View the full DFU>>](#)

Promus PREMIER is an unregistered or registered trademark of Boston Scientific or its affiliates. All other trademarks are property of their respective owners. © 2013 Boston Scientific Corporation or its affiliates. All rights reserved.