

Metal Composition

REBEL™ Platinum Chromium Coronary Stent System

The REBEL Stent is made from Platinum Chromium alloy and nominally consists of the following elements (with additional trace elements):

Promus PREMIER Stent	Elemental Composition by Weight
Iron	37%
Platinum	33%
Chromium	18%
Nickel	9%
Molybdenum	3%
Manganese	Trace

Directions for Use (DFU):

Section 4 Contraindications

Use of the REBEL Stent System is contraindicated in patients with the following:

- Known hypersensitivity to platinum, the platinum chromium alloy, or similar alloy types such as stainless steel.
- Known severe reaction to contrast agents that cannot be adequately premedicated prior to the REBEL Stent placement procedure.

Coronary artery stenting is contraindicated for use in the following:

- Patients who cannot receive recommended anti-platelet and/or anticoagulant therapy.
- Patients judged to have a lesion that prevents complete inflation of an angioplasty balloon or proper placement of the stent or delivery device.

For more information, contact your local sales representative or call the Boston Scientific Customer Service Center at 888-272-1001.

REBEL is an unregistered or registered trademark of Boston Scientific or its affiliates. © 2014 Boston Scientific Corporation or its affiliates. All rights reserved.