


# JETSTREAM™ Atherectomy System SELECTION GUIDE



## Jetstream XC Atherectomy Catheters

**2.4/3.4**

Ordering information: **112266-003**

Catheter Length	Min. Introducer Size	Max. Guidewire Diameter	Tip Diameter
120 cm	7 F (2.33 mm)	0.014" (0.36 mm)	2.4 mm 3.4 mm

**2.1/3.0**

Ordering information: **112264-003**

Catheter Length	Min. Introducer Size	Max. Guidewire Diameter	Tip Diameter
135 cm	7 F (2.33 mm)	0.014" (0.36 mm)	2.1 mm 3.0 mm

## Jetstream SC Atherectomy Catheters

**1.85**

Ordering information: **112262-003**

Catheter Length	Min. Introducer Size	Max. Guidewire Diameter	Tip Diameter
145 cm	7 F (2.33 mm)	0.014" (0.36 mm)	1.85

**1.6**

Ordering information: **112260-003**

Catheter Length	Min. Introducer Size	Max. Guidewire Diameter	Tip Diameter
145 cm	7 F (2.33 mm)	0.014" (0.36 mm)	1.6 mm

### Caution

Use only listed compatible guidewires and introducers with the Jetstream System. The use of any supplies not listed as compatible may damage or compromise the performance of the Jetstream System.

### Wire Compatibility

Compatible Guidewires:

- Boston Scientific Thruway™ 0.014 in (0.36 mm) 300 cm
- Abbott Vascular Hi-Torque Spartacore™ 14
- Abbott Vascular Hi-Torque Iron Man™ 0.014 in (0.36 mm)
- Abbott Vascular Extra S'port® 0.014 in (0.36 mm) 300 cm
- Abbott Vascular ASAHI™ MIRACLEbros™ 3 300 cm
- Abbott Vascular ASAHI™ MIRACLEbros™ 6 300 cm
- Abbott Vascular ASAHI™ Grand Slam™ Coronary Guide Wire

### Sheath Compatibility

7 F (2.33 mm) or larger with minimum inner diameter of 0.098 in or 2.5 mm. Do not use Tuohy Borst valve type introducers.

Indications, operating specifications and availability may vary by country. Check with local product representation and country-specific Information For Use for your country. This material is not approved for use or distribution in France.

All cited trademarks are the property of their respective owners. CAUTION: The law restricts these devices to sale by or on the order of a physician. Indications, contraindications, warnings and instructions for use can be found in the product labeling supplied with each device. Information for the use only in countries with applicable health authority product registrations.