

ESP™ Radiation Reduction Examination Gloves Prescriptive Information

Caution: Federal Law (USA) restricts this device to sale by or on the order of a physician.

CAUTION: The law restricts these devices to sale by or on the order of a physician. Indications, contraindications, warnings and instructions for use can be found in the product labelling supplied with each device. Information for use only in countries with applicable health authority registrations. This material not intended for use in France. Products shown for INFORMATION purposes only and may not be approved or for sale in certain countries. Please check availability with your local sales representative or customer service.

Warning

Contents supplied STERILE using an ethylene oxide (EO) process. Do not use if sterile barrier is damaged. If damage is found, call your Boston Scientific representative.

For single use only. Do not reuse, reprocess or resterilize. Reuse, reprocessing or resterilization may compromise the structural integrity of the device and/ or lead to device failure which, in turn, may result in patient injury, illness or death. Reuse, reprocessing or resterilization may also create a risk of contamination of the device and/or cause patient infection or cross-infection, including, but not limited to, the transmission of infectious disease(s) from one patient to another. Contamination of the device may lead to injury, illness or death of the patient.

After use, dispose of product and packaging in accordance with hospital, administrative and/or local government policy.

Intended Use/Indications for Use

The ESP Radiation Reduction Examination Glove is a specialty glove intended for specific examination purposes. The intended use of this glove is to reduce the amount of scattered radiation exposure to the hands from the primary x-ray beam during fluoroscopic examinations or other procedures performed under fluoroscopy.

Contraindications

None known.

Warnings

None known.

Precautions

This product contains natural rubber latex which may cause allergic reactions in some individuals.

The ESP Radiation Reduction Examination Glove is not intended for use in lieu of a Radiographic Protective Glove, which is used in radiography for those studies requiring the physician's hand or forearm to be into the primary x-ray beam but not requiring sterility or the degree of manual dexterity afforded by the ESP Radiation Reduction Examination Glove.